

Peregrine simultaneous capacity clarification

3 messages

Rina Ortega <rina.ortega@peregrine.example>

Tue, 25 Aug 2026 at 10:30

To: Mira Tan <mira.tan@example.invalid>

Following the discussion of laboratory throughput, the CE-24 can process two successive 24-specimen runs within a planned test day where profiles and sample preparation allow. Our application team can help organise the run sequence and preserve separate file identifiers for each batch. This may suit a programme that can compare cohorts across cycles and values a compact installation.

EMAIL-04#PRG-M01

Mira Tan <mira.tan@example.invalid>

Wed, 26 Aug 2026 at 15:22

To: Rina Ortega <rina.ortega@peregrine.example>

Thank you for explaining the daily workflow. The issued M3 requirement concerns simultaneous specimens within one controlled run, rather than the number processed over a day. Please confirm the number of positions in the submitted CE-24 and whether the quoted configuration includes any additional hardware that permits 48 specimens in one run. We will use that clarification to read the technical schedule consistently.

EMAIL-04#PRG-M02

Rina Ortega <rina.ortega@peregrine.example>

Thu, 27 Aug 2026 at 16:01

To: Mira Tan <mira.tan@example.invalid>

The quoted CE-24 has 24 simultaneous positions as stated in PRG-02. Two successive runs can provide 48 results over time, but they remain separate runs. A 48-position single-run configuration would require an expansion chassis. We have not included or priced that chassis in this submission. Our implementation and service schedules relate to the compact instrument as quoted.

EMAIL-04#PRG-M03

Fictional training case; all people, organisations, products, rules and events are invented.