

Helix warranty, maintenance and support schedule



HLX-05 · Version 1.0 · 2026-08-27 · Helix Test Infrastructure · signed

HLX-05 §1.1 — Coverage begins with the committed Site Acceptance Test on 30 November 2026 and runs through 31 March 2030 inclusive. Warranty, preventive maintenance and technical response are included in the submitted price.

HLX-05 §1.2 — Helix includes three preventive-maintenance visits each year and technical response within one business day. These exceed the RFQ minimum of two visits and two-business-day response.

HLX-05 §1.3 — The service package in this schedule is priced at S\$42,000 excluding GST for the full covered term and is included in the submitted price.

HLX-05 §2.1 — A later service change requires signed written agreement. The tendered term does not depend on any unpriced option or external approval.

HLX-05 §4 Service operating procedure

HLX-05 §4.1 — Our support model allocates an incident owner when a ticket is opened and records triage, action and closure in one service history. Scheduled preventive visits include a status review with the laboratory. The additional annual visit and faster response undertaking are included in the fixed bid and are described in the signed schedule, rather than offered as separately purchased upgrades.

HLX-05 §4.2 — A service request should include instrument identification, observed symptoms, the affected run where relevant, and any error code or log extract. The support desk will acknowledge the case and document diagnostic steps and the next action. Technical response is measured under the schedule's stated business-day commitment; a response is not the same event as complete repair.

HLX-05 §5 Maintenance and records

HLX-05 §5.1 — Preventive visits are planned with the laboratory to avoid critical runs. The visit record should list checks performed, measured condition, consumables or parts replaced, recommended actions and the next due visit. Findings that may affect a research run should be communicated to the project technical lead so the data record can be annotated when needed.

HLX-05 §5.2 — The service desk and scheduled visits operate within the fixed coverage term stated in §1.1. Any different service scope would require a documented variation; the quoted support commitment does not depend on a future option or an additional hosted service order.

HLX-05 T2

Record	Contents	Custodian
Incident ticket	Instrument, timestamp, symptom, action	Supplier support desk
Maintenance report	Checks, findings, next visit	Supplier service engineer
Laboratory log	Run impact and follow-up	Institute technical lead

HLX-05 §6 Incident and continuity guidance

HLX-05 §6.1 — When a fault affects an active experiment, the laboratory should preserve the affected run identifier, controller event log and last usable export before making configuration changes. The support desk may request diagnostic information, but the Institute retains its local research files. The incident record should identify whether data were recovered, whether a run was repeated and when routine operation resumed. This creates a traceable link between instrument service and the quality record.

HLX-05 §6.2 — Routine service planning should account for the research calendar. The supplier will propose maintenance windows and document completed checks; the laboratory will identify critical runs that cannot be interrupted. The signed service commitment controls the number of visits, response time and covered dates. Service records can describe performance and actions, but they do not alter the contractual period or add an unpriced obligation.

Fictional training case; all people, organisations, products, rules and events are invented.