



NST-001 §1.1 — Northstar offers the Encabulator NX-48 for S\$348,000 excluding GST. Our modular chamber is designed for reliable 48-position, synchronised stress testing and straightforward laboratory handover.

NST-001 §2.1 — The fixed submission includes instrument, data-export controls, installation, Site Acceptance Test and on-site training for eight staff. Technical, commercial, implementation and service terms are incorporated in NST-002 through NST-005.

NST-001 §3.1 — The implementation schedule targets Site Acceptance Test on 15 January 2027 after dispatch on 1 October 2026. Warranty and service coverage are specified in the signed NST-005 schedule; proposed future enhancements would require a separately agreed revision.

NST-001 §4.1 — Our price excludes GST, consumable specimens and any future service change not specified in the signed appendices. No continuing cloud subscription is required for raw-data export.

NST-001 §5 Covering offer and design intent

NST-001 §5.1 — Northstar submits the Encabulator NX-48 as a fixed-price response to RFQ-ARER-26-014. The proposal is based on modular configuration and a compact installation footprint. The submitted commercial total is S\$348,000 excluding GST. It covers the configuration and services identified in the signed schedules, subject to the stated exclusions. This covering letter, technical schedule, commercial schedule, implementation plan and service schedule are intended to be read as one offer.

NST-001 §5.2 — Northstar's offer emphasises a direct route from installation to routine laboratory use. The proposed modular configuration is priced as a complete submitted package, with research consumables treated separately. Our schedules describe the delivered hardware, acceptance sequence, quoted service package and exclusions.

NST-001 §6 Technical approach

NST-001 §6.1 — The NX-48 uses a common chamber controller for all installed positions.

Northstar proposes a practical operator interface with pre-set profiles, locally recorded run histories and a simple export path. The design is intended to let a small laboratory team repeat qualification runs without re-entering the full control sequence each time.

NST-001 §6.2 — The submitted instrument provides 48 simultaneous specimen positions per run. The programme may use several runs over time, but the specified position count describes one controlled cycle. The configured range, instrumentation and export method are recorded in NST-002. Operators can identify an individual specimen position and review its readings against the associated stress profile and event record.

NST-001 §6.3 — The local export procedure is part of ordinary instrument operation. The laboratory may use the supplier interface for review and diagnostics, while retaining a portable

copy of raw readings and event logs for its research archive. Any optional analytical service is independent of the basic export capability. The proposal does not require a continuing hosted subscription for mandatory data access.

NST-001 §7 Delivery, acceptance and training

NST-001 §7.1 — Our sequence begins with pre-shipment inspection and a dispatch notice, then moves through delivery, utility checks, controller calibration and operator familiarisation. The Institute should make the prepared installation location and designated trainees available for the agreed visits. We will record punch-list items during installation and review them before the acceptance demonstration.

NST-001 §7.2 — The plan in NST-004 commits to Site Acceptance Test on 15 January 2027. The acceptance demonstration will identify the installed configuration, representative controlled run, data export and event log, with a signed record of results. Eight staff are included in on-site training; attendance and a handover list will be captured. Scheduling calls and practical site dependencies will be tracked before the visit.

NST-001 §8 Commercial and service particulars

NST-001 §8.1 — The evaluated total is the sum of the submitted line items in NST-003. GST, where applicable, is separate. The quote identifies instrument controls, required accessories, installation, SAT, training and the service line. Optional research accessories or future changes, where mentioned, are not part of the evaluated configuration and cannot be assumed to fulfil a mandatory requirement.

NST-001 §8.2 — Incidents are logged against the instrument serial number, symptom and most recent run identifier. Remote triage is the first response; a field visit or replacement component is arranged when the fault cannot be resolved remotely. Planned maintenance includes inspection of chamber seals, controller checks and calibration review. The applicable dates and any change route are those in the signed service schedule.

NST-001 §8.3 — The signed NST-005 schedule defines the service commencement, covered period, preventive-maintenance cadence and technical-response undertaking. Requests for any different equipment or service scope would require documented agreement through the applicable procurement process. The price and dates in this response are those submitted by tender close.

NST-001 T2

Submission component	Document	Purpose
Main response	NST-001	Offer overview and inclusions
Technical compliance	NST-002	Delivered capability and export
Commercial schedule	NST-003	Quoted evaluated amount
Implementation plan	NST-004	Delivery, training and SAT
Service schedule	NST-005	Warranty, maintenance and response

NST-001 §9 Configuration control at handover

NST-001 §9.1 — The NX-48 is supplied as a modular assembly of chamber, position controls and local data interface. Northstar will identify the modules installed at the Institute and reconcile them with the technical and commercial schedules before acceptance. The handover pack will distinguish the as-installed configuration from optional accessories or later expansions. That record gives the laboratory a stable basis for operator instructions and future service calls without requiring it to reconstruct the equipment list from purchase correspondence.

NST-001 §9.2 — Our commissioning engineer will review the channel map with the project team, load a representative stress profile and inspect the readings and event sequence. The team can use the same profile to train operators in routine startup, recovery after an interrupted run and local export. A sample file will be retained with the acceptance pack so field names and units can be checked when the first research cohort begins. Northstar expects the laboratory to nominate a technical contact for installation and service coordination.

NST-001 §10 Operating assumptions and dependencies

NST-001 §10.1 — The proposed schedule assumes timely access to the prepared laboratory area, suitable utilities and a coordinated training session for the nominated staff. Northstar will provide the normal installation checklist ahead of the site visit. If a site item is incomplete, the engineer will document its effect on the work plan and discuss the next action with the Institute contact. An internal planning discussion does not itself change the dates in the submitted implementation schedule.

NST-001 §10.2 — Routine operation uses locally retained run records. Operators can label specimens and runs according to the Institute's research convention, export CSV or JSON files and pass those files to the programme data store. Northstar's interface may also display summary traces, but the research team can work from the raw export and event history. We will provide an explanation of channel identifiers and event codes in the handover material.

NST-001 §11 Commercial and document administration

NST-001 §11.1 — The tendered amount follows the line-item allocation in NST-003, with GST treated separately. It covers the equipment and work expressly included in the signed response set. Research consumables and future changes are identified outside the quoted package. Northstar's service organisation will receive the as-installed equipment list and acceptance record so incidents can be logged against the correct configuration and the signed service terms.

NST-001 §11.2 — If the Institute requests a different configuration, service arrangement or scope after submission, Northstar would document the proposal and route it through the applicable approval process. The responsible commercial and technical contacts would reconcile the changed items before any revised agreement is made. Until such a document is agreed, the signed schedules remain the reference for price, delivery and service. This approach lets the laboratory plan against a stable offered baseline while still allowing future requests to be evaluated on their own merits.