

Northstar technical compliance schedule



NST-002 · Version 1.0 · 2026-08-25 · Northstar Metrology · submitted

NST-002 T1

Row	Requirement	NX-48 response
T-01	Per-run positions	48
T-02	Temperature range, °C	−40 to 125
T-03	Load range, N	0 to 500
T-04	Electrical bias, V	0 to 24
T-05	Raw export	CSV and JSON, local without hosted subscription
T-06	Training	Eight staff on site

NST-002 §3 Configuration and measurement record

NST-002 §3.1 — The NX-48 uses a common chamber controller for all installed positions.

Northstar proposes a practical operator interface with pre-set profiles, locally recorded run histories and a simple export path. The design is intended to let a small laboratory team repeat qualification runs without re-entering the full control sequence each time.

NST-002 §3.2 — The position count in row T-01 is a simultaneous capacity for the delivered Encabulator NX-48; it does not aggregate separate runs. Each active position is associated with a channel identifier in the exported record. The controller associates setpoints, readings and event timestamps with a run so the laboratory can compare specimens under the same programmed conditions.

NST-002 §4 Export and verification

NST-002 §4.1 — CSV and JSON exports can be saved to the Institute's local storage without an ongoing hosted account. The supplier will provide a field guide covering timestamp format, units, channel identifiers, event codes and file naming. During SAT, a sample file and an event log will be produced from the installed configuration and reconciled with selected values displayed by the controller.

NST-002 §5 Test record conventions

NST-002 §5.1 — For the delivered NX-48, each specimen position is given a channel identifier before a controlled run starts. The operator records the run identifier, specimen mapping and selected profile, then reviews any warnings before release. Northstar's local controller records actual readings separately from programmed setpoints so the research team can distinguish a prescribed condition from a measured response. The 48-position count is the installed simultaneous configuration described in row T-01.

NST-002 §5.2 — At acceptance, a representative run should exercise the installed modular chamber, export a raw file and verify that the event record can be linked back to a position and timestamp. The operator guide should show which fields identify units, profile state, alarms and missing observations. A derivative chart or summary report may help interpretation, but the retained CSV or JSON source remains the laboratory's reproducible research record.

Fictional training case; all people, organisations, products, rules and events are invented. This is not organisational policy or Singapore procurement law.