

Peregrine warranty, maintenance and support schedule



PRG-005 · Version 1.0 · 2026-08-24 · Peregrine Instrumentation · signed

PRG-005 §1.1 — Warranty, maintenance and technical support begin at the committed SAT on 10 January 2027 and run through 31 March 2030 inclusive. The fixed tendered price includes this full period.

PRG-005 §1.2 — Two preventive-maintenance visits per year and technical response within two business days are included. There is no service-period option or qualification.

PRG-005 §4 Service operating procedure

PRG-005 §4.1 — Our application team provides first-line guidance on profiles, run setup and interpretation of logged warnings. Hardware incidents are recorded through the support desk and handled under the signed warranty terms. Preventive visits are coordinated around the laboratory's planned test cycles. The service duration and response target in the signed schedule apply to the quoted compact system.

PRG-005 §4.2 — A service request should include instrument identification, observed symptoms, the affected run where relevant, and any error code or log extract. The support desk will acknowledge the case and document diagnostic steps and the next action. Technical response is measured under the schedule's stated business-day commitment; a response is not the same event as complete repair.

PRG-005 §5 Maintenance and records

PRG-005 §5.1 — Preventive visits are planned with the laboratory to avoid critical runs. The visit record should list checks performed, measured condition, consumables or parts replaced, recommended actions and the next due visit. Findings that may affect a research run should be communicated to the project technical lead so the data record can be annotated when needed.

PRG-005 §5.2 — The service desk and scheduled visits operate within the fixed coverage term stated in §1.1. Any different service scope would require a documented variation; the quoted support commitment does not depend on a future option or an additional hosted service order.

PRG-005 T2

Record	Contents	Custodian
Incident ticket	Instrument, timestamp, symptom, action	Supplier support desk
Maintenance report	Checks, findings, next visit	Supplier service engineer
Laboratory log	Run impact and follow-up	Institute technical lead

PRG-005 §6 Incident and continuity guidance

PRG-005 §6.1 — When a fault affects an active experiment, the laboratory should preserve the affected run identifier, controller event log and last usable export before making configuration changes. The support desk may request diagnostic information, but the Institute retains its local research files. The incident record should identify whether data were recovered, whether a run was repeated and when routine operation resumed. This creates a traceable link between instrument service and the quality record.

PRG-005 §6.2 — Routine service planning should account for the research calendar. The supplier will propose maintenance windows and document completed checks; the laboratory will identify critical runs that cannot be interrupted. The signed service commitment controls the number of visits, response time and covered dates. Service records can describe performance and actions, but they do not alter the contractual period or add an unpriced obligation.

Fictional training case; all people, organisations, products, rules and events are invented. This is not organisational policy or Singapore procurement law.