

Document ID
Title
Version
Date
Issuer
Status

Case notice

NST-002

Northstar technical compliance schedule

1.0

2026-08-25

Northstar Metrology

submitted

Fictional training case; all people, organisations, products, rules and events are invented. This is not A
Singapore procurement law.



Row	Requirement	NX-48 response	
T-01	Per-run positions		48
T-02	Temperature range, °C	–40 to 125	
T-03	Load range, N	0 to 500	
T-04	Electrical bias, V	0 to 24	
T-05	Raw export	CSV and JSON, local	
T-06	Training places		8

Block locator	Type
NST-002 T1	table
NST-002 §3	heading
NST-002 §3.1	paragraph
NST-002 §3.2	paragraph
NST-002 §4	heading
NST-002 §4.1	paragraph
NST-002 §5	heading
NST-002 §5.1	paragraph
NST-002 §5.2	paragraph

Content

Row | Requirement | NX-48 response

T-01 | Per run positions | 48
Configuration and measurement record

The NX-48 uses a common chamber controller for all installed positions. Northstar proposes a practical set of profiles, locally recorded run histories and a simple export path. The design is intended to let a small position locally record runs a simultaneous capacity for the delivered NX-48; it does not export and verification. Each active position is associated with a channel identifier in the exported record. The controller associated CSV and JSON exports can be saved to the Institute's local storage without an ongoing hosted account. The field guide covering timestamp format, units, channel identifiers, event codes and file naming. During the acceptance, a representative runs and selected profile, then reviews any warnings before release. A record can be linked back to a position and timestamp. The operator guide should show which fields i

Source / fact references