



Helix technical compliance schedule

HLX-002 · Version 1.0 · 2026-08-27 · Helix Test Infrastructure · submitted

HLX-002 T1

Row	Requirement	V60 response
T-01	Per-run positions	60
T-02	Temperature range, °C	−40 to 125
T-03	Load range, N	0 to 500
T-04	Electrical bias, V	0 to 24
T-05	Raw export	CSV and JSON, local without hosted subscription
T-06	Training	Eight staff on site

HLX-002 §3 Configuration and measurement record

HLX-002 §3.1 — The V60 groups positions into monitored banks while retaining a single run identifier and synchronised event clock. Its larger delivered chamber is intended to absorb repeat runs and mixed cohort scheduling without a hardware change. The controller exposes raw channel histories and event logs for local export in the required formats.

HLX-002 §3.2 — The position count in row T-01 is a simultaneous capacity for the delivered Vector Encabulator V60; it does not aggregate separate runs. Each active position is associated with a channel identifier in the exported record. The controller associates setpoints, readings and event timestamps with a run so the laboratory can compare specimens under the same programmed conditions.

HLX-002 §4 Export and verification

HLX-002 §4.1 — CSV and JSON exports can be saved to the Institute's local storage without an ongoing hosted account. The supplier will provide a field guide covering timestamp format, units, channel identifiers, event codes and file naming. During SAT, a sample file and an event log will be produced from the installed configuration and reconciled with selected values displayed by the controller.

HLX-002 §5 Test record conventions

HLX-002 §5.1 — For the delivered V60, each specimen position is given a channel identifier before a controlled run starts. The operator records the run identifier, specimen mapping and selected profile, then reviews any warnings before release. Helix's banked-channel controller records actual readings separately from programmed setpoints so the research team can distinguish a prescribed condition from a measured response. The 60-position count is the installed simultaneous configuration described in row T-01.

HLX-002 §5.2 — At acceptance, a representative run should exercise the installed higher-capacity chamber, export a raw file and verify that the event record can be linked back to a position and timestamp. The operator guide should show which fields identify units, profile state, alarms and missing observations. A derivative chart or summary report may help interpretation, but the retained CSV or JSON source remains the laboratory's reproducible research record.

Fictional training case; all people, organisations, products, rules and events are invented. This is not A*STAR policy or Singapore procurement law.