

GAS DETECTOR CALIBRATION TEST CERTIFICATE

Description	Manufacturer	Serial Number
SGC-C Single Gas Clip CO	GCT	Activate before:
Cartridge Sensors CO	Range 0 - 300 ppm	

This product has been assembled, inspected, tested and calibrated with trained service professionals using NIST traceable calibration gases

Lot Number of Certified Gases used for this calibration test :

CO 100ppm

Gas Production:

Calibrated on :

Calibration is valid through instrument life. Detector life is 24 months from Activation. Remaining time to end of life and calibration expiration is shown on detector display following double click on the power/menu button.

SGC-C

sn :

Detector information

Activation

*To activate the detector, press and briefly hold down the power / menu button
Once activated, the detector will begin to display Gas Readings immediately and
will run continuously for 24 months*

Activation date : d d . m m . y y y y

Detector Life Countdown

*Once activated , the detector will diplay 24 months of life remaining
Detector will continue to countdown and display months remaining until 30 days remain
Detector will then show days remaining until end of life*

Activation date + 730 days = Calibration expiration d d . m m . y y y y

Battery

*The detector is designed so that the battery will not need charging or replacing
during the life of the detector*