

GAS DETECTOR CALIBRATION TEST CERTIFICATE

Description	Manufacturer	Serial Number
-------------	--------------	---------------

MGC-S+

GCT

MGC-SIMPLEPLUS

Activate before:

Cartridge Sensors	Range
<i>H2S</i>	<i>0 - 100 ppm</i>
<i>CO</i>	<i>0 - 500 ppm</i>
<i>O2</i>	<i>0 - 30 %</i>
<i>LEL</i>	<i>0 - 100 %</i>

This product has been assembled, inspected, tested and calibrated with trained service professionals using NIST traceable calibration gases

Lot Number of Certified Gases used for this calibration test :

H2S 25 ppm

CO 100 ppm

O2 18 Vol.%

50% LEL

Gas Production:

Calibrated on :

Calibration is valid through instrument life. Detector life i 36 months from Activation. Remaining time to end of life and calibration expiration is shown on detector display following double click on the power/menu button.

MGC-S+

sn :

Detector information

Activation

*To activate the detector, press and briefly hold down the power / menu button
Once activated, the detector will begin to display Gas Readings immediately and
will run continuously fo 36 months*

Activation date : d d . m m . y y y y

Detector Life Countdown

*Once activated , the detector will diplay 36 months of life remaining
Detector will continue to countdown and display months remaining until 30 days remain
Detector will then show days remaining until end of life*

Activation date + 1095 days = Calibration expiration d d . m m . y y y y

Battery

*The detector is designed so that the battery will not need charging or replacing
during the life of the detector*