

CER-AI SINGLE-PAGE CONCLUSION REPORT

Corneal Ectasia Risk Assessment Intelligence

The final surgical decision and all associated responsibility and liability rest with the surgeon. This application is a clinical decision-support aid only.

Patient	Masked for owner	Age	19
Assessment date	2026-10-05	Reviewer	Ümit Yapıcı
Eyes	OD; OS	Overall disposition	PASS WITH CAUTION

Overall disposition: PASS WITH CAUTION — PASS WITH CAUTION — surgeon review required.

OD — PASS WITH CAUTION

Assessment	Canonical result	Conclusion detail
Procedure	LASIK	
Randleman / ERSS	PASS	Total: 2; Classification: NORMAL_SYMMETRIC
NICE	CAUTION	Total: 5; Classification: CAUTION
Final BAD-D	PASS	Value: 0.93; Classification: NORMAL
PS3	PASS	Moderate / High: 0 / 0; Procedure disposition: Raw PS3 LASIK disposition: ALLOWED.
Procedural safety	PASS	LASIK_RSB_um: 458.5; LASIK_PTA_percent: 21.084; estimated_final_Kmean_D: 43; final_Kmean_model: MYOPIC_0.8_D_PER_D_ESTIMATE; hyperopic_treatment_component_D: 0
Selected procedure plan	OD — PASS WITH CAUTION — LASIK	Plan A — OZ 6.5 mm — TZ 9 mm — Flap 100 µm — Ablation 22.5 µm
ML7 planning	Temporal	Vacuum ring: 9 mm; Vacuum pressure: 550 mmHg
Decision basis	PASS WITH CAUTION	CAUTION: NICE total: 5
Review notes / warnings		Corneal Thickness Map morphology: not evaluated; surgeon review required; not counted in automated PS3.; Relative Thickness Map: not evaluated; surgeon review required; not counted in automated PS3.; PTI/CTSP thickness-profile morphology: not evaluated; surgeon review required; not counted in automated PS3.; Ablation/transition zone is not 0.4-0.5 mm smaller than the selected ring (active ML7 reference optimum).

clinical_policy: CER-AI-2026-10-05-SIGNED-IS-NEGATIVE-BOUNDARY-V13 · deployment_sha: f4b2950cb7492664458047cd528684a1a5e2d597 · software: 2.0.1 · source_registry: 2026-09-10-ML7-REUSE-CANONICAL-K · srax_algorithm: FRONT_MAP_SRAX_GT20-2026-09-07

OS — PASS WITH CAUTION

Assessment	Canonical result	Conclusion detail
Procedure	LASIK	
Randleman / ERSS	PASS	Total: 2; Classification: NORMAL_SYMMETRIC
NICE	CAUTION	Total: 5; Classification: CAUTION
Final BAD-D	PASS	Value: 0.54; Classification: NORMAL
PS3	PASS	Moderate / High: 0 / 0; Procedure disposition: Raw PS3 LASIK disposition: ALLOWED.
Procedural safety	PASS	LASIK_RSB_um: 448.5; LASIK_PTA_percent: 21.454; estimated_final_Kmean_D: 42.9; final_Kmean_model: MYOPIC_0.8_D_PER_D_ESTIMATE; hyperopic_treatment_component_D: 0
Selected procedure plan	OS — PASS WITH CAUTION — LASIK	Plan A — OZ 6.5 mm — TZ 9 mm — Flap 100 µm — Ablation 22.5 µm
ML7 planning	Superior	Vacuum ring: 9 mm; Vacuum pressure: 550 mmHg
Decision basis	PASS WITH CAUTION	CAUTION: NICE total: 5
Review notes / warnings		Corneal Thickness Map morphology: not evaluated; surgeon review required; not counted in automated PS3.; Relative Thickness Map: not evaluated; surgeon review required; not counted in automated PS3.; PTI/CTSP thickness-profile morphology: not evaluated; surgeon review required; not counted in automated PS3.; Ablation/transition zone is not 0.4-0.5 mm smaller than the selected ring (active ML7 reference optimum).

clinical_policy: CER-AI-2026-10-05-SIGNED-IS-NEGATIVE-BOUNDARY-V13 · deployment_sha: f4b2950cb7492664458047cd528684a1a5e2d597 · software: 2.0.1 · source_registry: 2026-09-10-ML7-REUSE-CANONICAL-K · srax_algorithm: FRONT_MAP_SRAX_GT20-2026-09-07

This report is generated under the CER-AI Preoperative Ectasia Risk Assessment Protocol for corneal refractive surgery. CAUTION requires explicit surgeon review but does not automatically defer surgery. STOP-DEFER means surgery must not proceed unless the stated stop/defer condition is resolved. DATA INSUFFICIENT / NOT ASSESSED does not permit PASS. This clinical decision-support report does not replace independent surgeon review.