

Automatic prostate cancer detection through DCE-MRI images: all you need is a good normalization

Guillaume Lemaître^{a,b,*}, Robert Martí^b, Fabrice Meriaudeau^{a,c}

^a*LE2I UMR6306, CNRS, Arts et Métiers, Univ. Bourgogne Franche-Comté, 12 rue de la
Fonderie, 71200 Le Creusot, France*

^b*ViCOROB, Universitat de Girona, Campus Montilivi, Edifici P4, 17071 Girona, Spain*

^c*CISIR, Electrical & Electronic Engineering Department, Universiti Teknologi Petronas, 32610
Seri Iskandar, Perak, Malaysia*

Abstract

This template helps you to create a properly formatted L^AT_EX manuscript.

Keywords: DCE-MRI, prostate cancer, normalization, classification,
quantification

1. Introduction

Prostate Cancer (PCa) is the second most frequently diagnosed men cancer,
accounting for 899,000 cases leading to 258,100 deaths (Ferlay et al., 2010). As
highlighted by the PI-RADS Steering Committee, the two main challenges to be
5 addressed are (Weinreb et al., 2016): (i) the improvement of detecting clinically
significant PCa and (ii) an increase of the confidence in benign or dormant cases,
avoiding unnecessary invasive medical exams. In this regard, multiparametric
Magnetic Resonance Imaging (MRI) (mpMRI) is frequently used to build robust
Computer-Aided Detection and Diagnosis (CAD) systems to detect, localize,
10 and grade PCa. In general, CAD systems are based on mpMRI which combines
several of the following modalities (Lemaître et al., 2015): T₂ Weighted (T₂-W)-
MRI, Dynamic Contrast-Enhanced (DCE)-MRI, Apparent Diffusion Coefficient
(ADC) maps, and Magnetic Resonance Spectroscopy Imaging (MRSI).

*Corresponding author.

Email address: g.lemaitre58@gmail.com (Guillaume Lemaître)

In DCE-MRI, a contrast media is injected intravenously and a set of images
15 is acquired over time. Consequently, each voxel in the image is a dynamic
signal which is related to the vascular properties of the tissue. In fact, these
properties are automatically extracted using quantitative or semi-quantitative
approaches (Lemaître et al., 2015).

The former group of approaches uses pharmacokinetic modelling based on a
20 bicompartiment model, namely Brix (Brix et al., 1991) and Tofts (Tofts et al.,
1995) models. The parameters of the Brix model are found assuming a linear
relationship between the media concentration and MRI signal intensity. This as-
sumption has shown, however, to lead to inaccurate parameter calculation (Heil-
mann et al., 2006). In the contrary, Tofts model only requires a conversion from
25 MRI signal intensity to concentration, which can become a non-linear relation-
ship using specific equation of MRI sequences (e.g., FLASH sequence). Tofts
modelling suffers, however, from an higher complexity (Gliozzi et al., 2011). The
conversion using the non-linear approach requires to acquire a T_1 map which
is not always possible during clinical examination. Furthermore, the parameter
30 calculation require the Arterial Input Function (AIF) which is challenging to
measure and can also lead to inaccurate estimation of the parameters.

The latter group of approaches are rather mathematical than pharmacoki-
netic modelling (Huisman et al., 2001; Gliozzi et al., 2011). These methods
offer the advantages to not require any knowledge about the MRI sequence nor
35 any conversion from signal intensity to concentration. However, the heuristic
approach propose by Huisman et al. requires an estimate regarding the noise
standard deviation of the signal as well as manual tuning.

Nevertheless, all presented methods suffer from two major drawbacks: (i)
the inter-patient variability of the data lead to a variation of the parameters
40 estimated and to poor classification performance while designing CAD systems,
and (ii) only few parameters are used to characterize the dynamic signal imply-
ing that some information are discarded.

In this work, we propose a fully automatic normalization method for DCE-
MRI that reduce the inter-patient variability of the data. Furthermore, we show

45 that using the full normalized signal lead to the best classification performance.

The paper is organized as follows: Section 2 outlines our normalization strategy (Section 2.1) as well as specificity regarding the state-of-the-art methods used for comparison (Section 2.2). The dataset, experiments, and results are reported in Section 3 while discussed in Section 4 followed by a concluding
50 section.

2. Methods

2.1. Normalization of DCE-MRI images

2.2. Quantification of DCE-MRI

2.2.1. Brix and Hoffmann models

55 In the Brix model (Brix et al., 1991), the MRI signal intensity is assumed to be proportional to the media concentration. Therefore, the model is expressed as in Eq. (1):

$$s_n(t) = 1 + A \left[\frac{\exp(k_{el}t') - 1}{k_{ep}(k_{ep} - k_{el})} \exp(-k_{el}t) - \frac{\exp(k_{ep}t') - 1}{k_{el}(k_{ep} - k_{el})} \exp(-k_{ep}t) \right], \quad (1)$$

with

$$s_n(t) = \frac{s(t)}{S_0}, \quad (2)$$

where $s(t)$ and S_0 are the MRI signal intensity at time t and the average pre-contrast MRI signal intensity, respectively; A , k_{el} , and k_{ep} are a constant proportional to the transfer constant, the diffusion rate constant, and the rate constant, respectively. Additionally, during the injection time $0 \leq t \leq \tau$, $t' = t$ and afterwards while $t > \tau$, $t' = \tau$.

Following this model, Hoffmann et al. propose the following similar model
65 as expressed in Eq. (3):

$$s_n(t) = 1 + \frac{A}{\tau} \left[\frac{k_{ep}(\exp(k_{el}t') - 1)}{k_{el}(k_{ep} - k_{el})} \exp(-k_{el}t) - \frac{\exp(k_{ep}t') - 1}{(k_{ep} - k_{el})} \exp(-k_{ep}t) \right]. \quad (3)$$

The parameters are estimated by fitting the model using non-linear least-squares optimization solved with Levenberg-Marcquardt.

2.2.2. Tofts model

The extended Tofts model is formulated as in Eq. (4):

$$C_t(t) = K_{trans}C_p(t) * \exp(-k_{ep}t) + v_pC_p(t), \quad (4)$$

70 where $*$ is the convolution operator; $C_t(t)$ and $C_p(t)$ is the concentration of contrast agent in the tissue and in the plasma, respectively; K_{trans} , k_{ep} , and v_p are the volume transfer constant, the diffusion rate constant, and the plasma volume fraction, respectively.

Therefore, Tofts model requires to: (i) detect candidate voxels from the femoral or iliac arteries and estimate a patient-based AIF signal, (ii) convert
75 the MRI signal intensity (i.e., AIF and dynamic signal) to a concentration, and (iii) in the case of a population-based AIF, estimate an AIF signal.

Segmentation of artery voxels and patient-based AIF estimation

The AIF signal from DCE-MRI can be manually estimated by selecting the most-enhanced voxels from the femoral or iliac arteries (Meng et al., 2010).
80 Few methods have been proposed to address the automated extraction of AIF signal. Chen et al. filter successively the possible candidates (Chen et al., 2008): (i) dynamic signals with small peak are rejecting by thresholding, (ii) voxels with a small wash-in are rejected by thresholding, (iii)
85 a blob detector is used and large enough regions are kept, and (iv) circular and cylindricity are used to reject the last false positive. Zhu et al. propose an iterative method selecting voxels which best fit a gamma variate function (Zhu et al., 2011). However, it requires to compute first and second derivatives as well as maximum curvature points. Shanbhag et al.
90 propose a 4-steps algorithm (Shanbhag et al., 2012; Fennessy et al., 2015): (i) remove slices with artefacts and find the best slices based on intrinsic anatomic landmarks and enhancement characteristics, (ii) find the voxel

95 candidates using the maximum enhanced voxels and a multi-label maximum entropy based thresholding algorithm, (iii) excluding region next to the endorectal coil, and (iv) selecting the best 5 candidates which meet enhancement characteristics and that are correlated.

All the above methods are rather complex and thus we propose a method which is based on the following simple assumptions: (i) all possible AIF signal candidates should have a similar shape, (ii) an high enhancement, 100 and (iii) the arteries should be almost round and within a size range. Therefore, each slice is clustered into regions using K-means clustering with $k = 6$. The cluster with the highest enhancement—i.e. corresponding to the 90th percentile of the maximum of each dynamic signal—contain the arteries and is selected. Finally, regions with an eccentricity smaller than 0.5 and an area in the range of $[100, 400]$ voxels are kept. Additionally, 105 to remove voxels contaminated by partial volume effect, only the 10% most enhanced voxels of the possible candidates are kept as proposed by (Schabel and Parker, 2008) and the average signal is computed. A summary of the different segmentation steps is presented in Fig. ??.

Conversion of MRI signal intensity to concentration To estimate the free parameters of the Tofts model (see Eq. (4)), the concentration $C_t(t)$ and $C_p(t)$ need to be computed from the MRI signal intensity and the AIF signal, respectively. This conversion is based on the equation of the FLASH sequence—see Appendix A for details—and is formulated as in Eq. (5):

$$c(t) = \frac{1}{TR \cdot r_1} \ln \left(\frac{1 - \cos \alpha \cdot S^* \frac{s(t)}{S_0}}{1 - S^* \frac{s(t)}{S_0}} \right) - \frac{R_{10}}{r_1}, \quad (5)$$

with,

$$S^* = \frac{1 - \exp(-TR \cdot R_{10})}{1 - \cos \alpha \cdot \exp(-TR \cdot R_{10})}, \quad (6)$$

110 where $s(t)$ is the MRI signal, S_0 is the MRI signal prior to the injection of the contrast media, α is the flip angle, TR is the Repetition Time (TR), R_{10} is the pre-contrast tissue relaxation time also equal to $\frac{1}{T_{10}}$, r_1 is the relaxativity coefficient of the contrast agent.

115 T_{10} can be estimated from the acquisition of a T_1 map. However, this modality was not part of the clinical trial in this research and the value of T_{10} was fixed to 1600 ms for both blood and prostate as stated in the literature (Fennessy et al., 2015; De Bazelaire et al., 2004; Carr and Carroll, 2011).

Estimation of population-based AIF While estimating the pharmacokinetic parameters from Tofts model, the AIF concentration $C_p(t)$ can be computed either from the patient or a population. We presented in the two previous sections the algorithms which allows to estimate the patient-based AIF concentration. To compare with the previous approach, we also computed a population-based AIF which will be also later used to compare the performance of both approaches. In that regard, the population-based AIF was estimated as in (Meng et al., 2010) by fitting the average patient-based AIFs to the model of Parker et al. (2006) which is formulated as in Eq. (7):

$$C_p(t) = \sum_{n=1}^2 \frac{A_n}{\sigma_n \sqrt{2\pi}} \exp\left(\frac{-(t - T_n)^2}{2\sigma_n^2}\right) + \frac{\alpha \exp(-\beta t)}{1 + \exp(-s(t - \tau))}, \quad (7)$$

120 where A_n , T_n , and σ_n are the scaling constants, centers, and widths of the n^{th} Gaussian, α and β are the amplitude and decay constant of the exponential; and s and τ are the width and center of the sigmoid function, respectively.

The parameters are estimated by fitting the model using non-linear least-squares optimization solved with Levenberg-Marcquardt.

125 2.2.3. PUN model

Gliozzi et al. show that Phenomenological Universalities (PUN) approach can be used for DCE-MRI analysis (Gliozzi et al., 2011). The model has been successfully used in a CAD system proposed by Giannini et al. (2015). This

model can be expressed as in Eq. (8):

$$s_n(t) = \exp \left[rt + \frac{1}{\beta} (a_0 - r) (\exp(\beta t) - 1) \right], \quad (8)$$

130 with

$$s_n(t) = \frac{s(t) - S_0}{S_0}, \quad (9)$$

where $s(t)$ and S_0 are the MRI signal intensity at time t and the average pre-contrast MRI signal intensity, respectively; r , a_0 , and β are the free parameters of the model.

The parameters are estimated by fitting the model using non-linear least-squares optimization solved with Levenberg-Marcquardt.

2.2.4. Semi-quantitative analysis

The semi-quantitative analysis of the DCE-MRI is equivalent to extract curve characteristics directly from the signal without a strict theoretical pharmacokinetic meaning. In this work, we use the model presented by Huisman et al. (2001) which formulate the MRI signal as in Eq. (10):

$$s(t) = \begin{cases} S_0 & 0 \leq t \leq t_0 \\ S_M - (S_M - S_0) \exp\left(\frac{-(t-t_0)}{\tau}\right) & t_0 < t \leq t_0 + 2\tau \\ S_M - (S_M - S_0) \exp\left(\frac{-(t-t_0)}{\tau}\right) + w(t - t_0 + 2\tau) & t > t_0 + 2\tau \end{cases} \quad (10)$$

where $s(t)$ is the MRI signal intensity, S_0 is the pre-contrast signal intensity, t_0 is the time corresponding to the start of enhancement, S_M and τ is the maximum of the signal and the exponential time constant, and w is the slope of the linear part.

145 Huisman et al. argue that curve fitting via least-squares minimization using Nelder-Mead algorithm leads to inaccurate estimation of the free parameters: mainly the issue come from an incorrect estimation of the start of enhancement t_0 leading to incorrect estimation of the other parameters. Therefore, they

propose to: (i) estimate robustly t_0 , (ii) estimate S_0 by averaging the samples
150 between 0 and t_0 (ii) estimate w depending if the slope is significant or not, (iii)
estimate S_M which should be the point at the intersection of the most probable
slope line and the plateau.

Instead of these successive estimations, we propose a unified optimization in
which t_0 is fixed since that this is a key parameter. Therefore, t_0 is robustly
155 estimated from the AIF signal since that this is the most enhanced signal in
which the start of enhancement is easily identifiable. The AIF signal is computed
as in Section 2.2.2. t_0 is estimated by finding the maximum in the beginning
of the first derivative of the MRI signal. Then, the function in Eq.(10) is
fitted using non-linear least squares with Trust Region Reflective algorithm.
160 Furthermore, the parameters τ and S_M are bounded during the optimization to
ensure robust estimations.

From Eq.(10), the following features are extracted: (i) the wash-in corre-
sponding to the slope between t_0 and $t_0 + 2\tau$, (ii) the wash-out corresponding
to the parameter w , (iii) the area under the curve between t_0 and the end of the
165 signal, (iv) the exponential time constant τ , and (v) the relative enhancement
 $S_M - S_0$.

3. Experiments and results

4. Discussions

5. Conclusions and future works

170 Appendix A. Conversion from FLASH signal to media concentration

In this appendix, we show the demonstration used to extract the agent con-
centration from the MRI signal.

The signal equation in FLASH sequence (Haase et al., 1986) is defined as:

$$s(t) = S_{eq} \sin \alpha \cdot \frac{1 - \exp(-TR(R_{10} + r_1 c(t)))}{1 - \cos \alpha \cdot \exp(-TR(R_{10} + r_1 c(t)))}, \quad (\text{A.1})$$

Table 1: AUC of the individual features for each method.

Features	Un-normalized data		Normalized data	
	RF	NB	RF	NB
Brix model				
A	0.54	0.62	0.58	0.67
k_{el}	0.55	0.52	0.54	0.61
k_{ep}	0.51	0.52	0.51	0.58
Hoffmann model				
A	0.52	0.50	0.51	0.56
k_{el}	0.55	0.53	0.54	0.64
k_{ep}	0.55	0.50	0.53	0.66
Tofts model with population AIF				
K_{trans}	0.56	0.62	0.56	0.65
v_e	0.51	0.50	0.50	0.52
v_p	0.53	0.63	0.55	0.53
Tofts model with patient AIF				
K_{trans}	0.57	0.66	0.56	0.65
v_e	0.49	0.50	0.51	0.52
v_p	0.53	0.37	0.57	0.65
PUN model				
a_0	0.52	0.53	0.53	0.51
r	0.53	0.59	0.55	0.55
β	0.55	0.56	0.53	0.44
Semi-quantitative analysis				
wash-in	0.59	0.64	0.55	0.51
wash-out	0.52	0.50	0.56	0.66
IAUC	0.51	0.61	0.52	0.64
τ	0.57	0.57	0.56	0.61
$S_M - S_0$	0.56	0.63	0.53	0.64




Figure 1: ROC analysis using a RF classifier using the DCE signal with and without normalization.




Figure 2: ROC analysis using a NB classifier using the DCE signal with and without normalization.




Figure 3: ROC analysis using a RF classifier for the different quantification methods without data normalization.




Figure 4: ROC analysis using a NB classifier for the different quantification methods without data normalization.




Figure 5: ROC analysis using a RF classifier for the different quantification methods with data normalization.




Figure 6: ROC analysis using a NB classifier for the different quantification methods with data normalization.

where $s(t)$ is the MRI signal, S_{eq} is the maximum signal amplitude of the spoiled
175 gradient at the Echo Time (TE) which is proportional to the Proton Density
(PD), α is the flip angle, TR is the Repetition Time (TR), R_{10} is the pre-
contrast tissue relaxation time also equal to $\frac{1}{T_{10}}$, r_1 is the relaxivity coefficient
of the contrast agent, and $c(t)$ is the media concentration.

Therefore, the pre-contrast signal prior to bolus injection of the media is
180 defined as:

$$S_0 = S_{eq} \sin \alpha \cdot \frac{1 - \exp(-TR \cdot R_{10})}{1 - \cos \alpha \cdot \exp(-TR \cdot R_{10})}. \quad (\text{A.2})$$

To simplify the demonstration, let us define:

$$A = \exp(-TR \cdot R_{10}), \quad (\text{A.3})$$

$$B = \exp(-TR \cdot r_1 c(t)). \quad (\text{A.4})$$

Let us define:

$$S^* = \frac{S_0}{S_{eq} \sin \alpha}, \quad (\text{A.5})$$

$$= \frac{1 - A}{1 - A \cos \alpha}. \quad (\text{A.6})$$

Thus,

$$S^* \frac{s(t)}{S_0} = \frac{S_0}{S_{eq} \sin \alpha} \frac{s(t)}{S_0}, \quad (\text{A.7})$$

$$= \frac{1 - AB}{1 - AB \cos \alpha}. \quad (\text{A.8})$$

Now, let us define:

$$\frac{1 - \cos \alpha \cdot S^* \frac{s(t)}{S_0}}{1 - S^* \frac{s(t)}{S_0}} = \frac{1 - \cos \alpha \left(\frac{1-AB}{1-AB \cos \alpha} \right)}{1 - \frac{1-AB}{1-AB \cos \alpha}}, \quad (\text{A.9})$$

$$= \frac{1 - AB \cos \alpha - \cos \alpha (1 - AB)}{1 - AB \cos \alpha - (1 - AB)}, \quad (\text{A.10})$$

$$= \frac{1 - AB \cos \alpha - \cos \alpha + AB \cos \alpha}{1 - AB \cos \alpha - 1 + AB}, \quad (\text{A.11})$$

$$= \frac{1 - \cos \alpha}{AB(1 - \cos \alpha)}, \quad (\text{A.12})$$

$$= \frac{1}{AB}. \quad (\text{A.13})$$

185 Thus,

$$-TR \cdot R_{10} - TR \cdot r_1 c(t) = \ln \left(\frac{1 - \cos \alpha \cdot S^* \frac{s(t)}{S_0}}{1 - S^* \frac{s(t)}{S_0}} \right). \quad (\text{A.14})$$

Therefore,

$$c(t) = \frac{1}{TR \cdot r_1} \ln \left(\frac{1 - \cos \alpha \cdot S^* \frac{s(t)}{S_0}}{1 - S^* \frac{s(t)}{S_0}} \right) - \frac{R_{10}}{r_1}. \quad (\text{A.15})$$

References

- Brix, G., Semmler, W., Port, R., Schad, L.R., Layer, G., Lorenz, W.J., 1991. Pharmacokinetic parameters in cns gd-dtpa enhanced mr imaging. Journal of computer assisted tomography 15, 621–628.
- 190 Carr, J.C., Carroll, T.J., 2011. Magnetic resonance angiography: principles and applications. Springer Science & Business Media.
- Chen, J., Yao, J., Thomasson, D., 2008. Automatic determination of arterial input function for dynamic contrast enhanced mri in tumor assessment, in: International Conference on Medical Image Computing and Computer-Assisted Intervention, Springer. pp. 594–601. doi:10.1007/978-3-540-85988-8_71.
- 195 De Bazelaire, C.M., Duhamel, G.D., Rofsky, N.M., Alsop, D.C., 2004. Mr imaging relaxation times of abdominal and pelvic tissues measured in vivo at

- 3.0 t: preliminary results 1. *Radiology* 230, 652–659. doi:10.1148/radiol.
200 2303021331.
- Fennessy, F.M., Fedorov, A., Penzkofer, T., Kim, K.W., Hirsch, M.S., Vangel, M.G., Masry, P., Flood, T.A., Chang, M.C., Tempany, C.M., et al., 2015. Quantitative pharmacokinetic analysis of prostate cancer dce-mri at 3t: comparison of two arterial input functions on cancer detection with digitized whole
205 mount histopathological validation. *Magnetic resonance imaging* 33, 886–894. doi:10.1016/j.mri.2015.02.008.
- Ferlay, J., Shin, H.R., Bray, F., Forman, D., Mathers, C., Parkin, D.M., 2010. Estimates of worldwide burden of cancer in 2008: Globocan 2008. *International journal of cancer* 127, 2893–2917. doi:10.1002/ijc.25516.
- 210 Giannini, V., Mazzetti, S., Vignati, A., Russo, F., Bollito, E., Porpiglia, F., Stasi, M., Regge, D., 2015. A fully automatic computer aided diagnosis system for peripheral zone prostate cancer detection using multi-parametric magnetic resonance imaging. *Computerized Medical Imaging and Graphics* 46, 219–226. doi:10.1016/j.compmedimag.2015.09.001.
- 215 Gliozzi, A., Mazzetti, S., Delsanto, P.P., Regge, D., Stasi, M., 2011. Phenomenological universalities: a novel tool for the analysis of dynamic contrast enhancement in magnetic resonance imaging. *Physics in medicine and biology* 56, 573.
- Haase, A., Frahm, J., Matthaei, D., Hanicke, W., Merboldt, K.D., 1986. Flash
220 imaging. rapid nmr imaging using low flip-angle pulses. *Journal of Magnetic Resonance (1969)* 67, 258–266. doi:10.1016/0022-2364(86)90433-6.
- Heilmann, M., Kiessling, F., Enderlin, M., Schad, L.R., 2006. Determination of pharmacokinetic parameters in dce mri: consequence of nonlinearity between contrast agent concentration and signal intensity. *Investigative radiology* 41,
225 536–543. doi:10.1097/01.rli.0000209607.99200.53.

- Hoffmann, U., Brix, G., Knopp, M.V., Heß, T., Lorenz, W.J., 1995. Pharmacokinetic mapping of the breast: a new method for dynamic mr mammography. *Magnetic resonance in medicine* 33, 506–514. doi:10.1002/mrm.1910330408.
- 230 Huisman, H.J., Engelbrecht, M.R., Barentsz, J.O., 2001. Accurate estimation of pharmacokinetic contrast-enhanced dynamic mri parameters of the prostate. *Journal of Magnetic Resonance Imaging* 13, 607–614. doi:10.1002/jmri.1085.
- Lemaître, G., Martí, R., Freixenet, J., Vilanova, J.C., Walker, P.M., Meriaudeau, F., 2015. Computer-aided detection and diagnosis for prostate cancer based on mono and multi-parametric mri: A review. *Computers in biology and medicine* 60, 8–31. doi:10.1016/j.compbiomed.2015.02.009.
- 235 Meng, R., Chang, S.D., Jones, E.C., Goldenberg, S.L., Kozlowski, P., 2010. Comparison between population average and experimentally measured arterial input function in predicting biopsy results in prostate cancer. *Academic radiology* 17, 520–525. doi:10.1016/j.acra.2009.11.006.
- 240 Parker, G.J., Roberts, C., Macdonald, A., Buonaccorsi, G.A., Cheung, S., Buckley, D.L., Jackson, A., Watson, Y., Davies, K., Jayson, G.C., 2006. Experimentally-derived functional form for a population-averaged high-temporal-resolution arterial input function for dynamic contrast-enhanced mri. *Magnetic resonance in medicine* 56, 993–1000. doi:10.1002/mrm.21066.
- 245 Schabel, M.C., Parker, D.L., 2008. Uncertainty and bias in contrast concentration measurements using spoiled gradient echo pulse sequences. *Physics in medicine and biology* 53, 2345. doi:10.1088/0031-9155/53/9/010.
- 250 Shanbhag, D., Gupta, S.N., Rajamani, K., Zhu, Y., Mullick, R., 2012. A generalized methodology for detection of vascular input function with dynamic contrast enhanced perfusion data, in: ISMRM, p. 10.

- Tofts, P.S., Berkowitz, B., Schnall, M.D., 1995. Quantitative analysis of dynamic gd-dtpa enhancement in breast tumors using a permeability model. *Magnetic Resonance in Medicine* 33, 564–568. doi:10.1002/mrm.1910330416.
- Weinreb, J.C., Barentsz, J.O., Choyke, P.L., Cornud, F., Haider, M.A., Macura, K.J., Margolis, D., Schnall, M.D., Shtern, F., Tempany, C.M., et al., 2016. Pi-rads prostate imaging-reporting and data system: 2015, version 2. *European urology* 69, 16–40.
- Zhu, Y., Chang, M.C., Gupta, S., 2011. Automated determination of arterial input function for dce-mri of the prostate, in: *SPIE Medical Imaging*, International Society for Optics and Photonics. pp. 79630W–79630W. doi:10.1117/12.878213.