

Computer-Aided Detection for Prostate Cancer Detection based on Multi-Parametric Magnetic Resonance Imaging

Guillaume Lemaître*, Robert Martí†, Mojdeh Rastgoo‡, Fabrice Mériaudeau§¶

*Parietal team, Inria, CEA, Université Paris-Saclay, 1 Rue Honoré d'Estienne d'Orves, 91120 Palaiseau

†ViCOROB, Universitat de Girona, Campus Montilivi, Edifici P4, 17071 Girona

‡LE2I UMR6306, CNRS, Arts et Métiers, Univ. Bourgogne Franche-Comté, 12 rue de la Fonderie, 71200 Le Creusot

§CISIR, Electrical & Electronic Engineering Department, Universiti Teknologi Petronas, 32610 Seri Iskandar, Perak

¶Corresponding author: fabrice.meriaudeau@utp.edu.my

Abstract—Prostate cancer (CaP) is the second most diagnosed cancer in men all over the world. In the last decades, new imaging techniques based on magnetic resonance imaging (MRI) have been developed improving diagnosis. In practice, diagnosis is affected by multiple factors such as observer variability and visibility and complexity of the lesions. In this regard, computer-aided detection and diagnosis (CAD) systems are being designed to help radiologists in their clinical practice. We propose a CAD system taking advantage of all MRI modalities (i.e., T₂-W-MRI, DCE-MRI, diffusion weighted (DW)-MRI, MRSI). This system has been extensively tested on a dataset which has been made publicly available.

Index Terms—Computer-Aided Diagnosis, Prostate Cancer, Normalization, Multi-Parametric MRI

I. INTRODUCTION

Current prostate cancer (CaP) screening consists of 3 different stages. First, prostate-specific antigen (PSA) control is performed to distinguish between low- and high-risk CaP. To assert such diagnosis, samples are taken during prostate biopsy and analyzed to make an accurate prognosis of the CaP.

Although PSA screening has been shown to improve early detection of CaP [1], its lack of reliability motivates further investigations using magnetic resonance imaging (MRI)-based computer-aided detection and diagnosis (CAD). Consequently, current research is focused on identifying new biological markers to replace PSA-based screening [2]. Until such research comes to fruition, these needs can be met through active-surveillance strategy using multiparametric MRI (mp-MRI) techniques [3].

Lemaître *et al.* recently reviewed more than 50

research works that focused on CAD system for CaP [4]. These studies are based on CAD systems that consists of the following steps: (i) pre-processing, (ii) segmentation, (iii) registration, (iv) feature detection, (v) feature selection-extraction, and (vi) finally classification.

The reviewed mp-MRI-based CAD used 2 to 3 MRI modalities among T₂ Weighted (T₂-W)-MRI, dynamic contrast-enhanced (DCE)-MRI, and diffusion weighted (DW)-MRI, discarding the potential discriminative power of magnetic resonance spectroscopy imaging (MRSI). Furthermore, only half of these studies tackled the challenging detection of CaP in the central gland (CG). Additionally, none of the works investigated the issue related to feature balancing when developing their CAD systems. Finally, none of the datasets nor source codes used have been released, making impossible the possibilities to compare the methods.

In this work, we propose a CAD system to detect CaP in peripheral zone (PZ) and CG, using the 4 aforementioned MRI modalities. In addition, our framework include a step of feature balancing. The dataset used and the source code developed are released for future comparisons and reproducibility.

II. METHODOLOGY

A. Materials

The mp-MRI data are acquired from a cohort of patients with higher-than-normal level of PSA. Acquisition is achieved with a 3T whole body MRI scanner (Siemens Magnetom Trio TIM, Erlangen, Germany) using sequences to obtain T₂-W-MRI,

DCE-MRI, DW-MRI, and MRSI. In addition of the MRI examination, these patients also have undergone a transrectal ultrasound (TRUS) guided-biopsy. The dataset is composed of a total of 19 patients, 17 of which have biopsies that were positive for CaP and 2 patients are considered “healthy” because they have negative biopsies. In all 12 patients have a CaP in the PZ, 3 patients have CaP in the CG, 2 patients have invasive CaP in both the PZ and the CG, and 2 patients are considered “healthy”. An experienced radiologist segmented the prostate organ — on T₂-W-MRI, DCE-MRI, and apparent diffusion coefficient (ADC) — as well as the prostate zones — i.e., PZ and CG —, and CaP on the T₂-W-MRI. The full description and the data set are available at I2Cvb website¹ [5].

B. CAD pipeline for CaP

Our mp-MRI CAD system consists of 7 different steps: pre-processing, segmentation, registration, feature detection, feature balancing, feature selection/extraction, and finally classification.

1) *Pre-processing*: Normalization is, a crucial step to reduce the inter-patient variations which allows to improve the learning during the classification stage. However, the MRI modalities provide specific type of data — static *vs.* dynamic information, images *vs.* signals — that required a dedicated pre-processing. Therefore, we pre-process differently the data: T₂-W-MRI is normalized using a Rician a-priori that has been shown to be better than the traditional *z*-score [6]. In contrast to T₂-W-MRI, in ADC map the probability density function (PDF) within the prostate does not follow a known distribution and thus one cannot use a parametric model to normalize these images and a non-parametric piecewise-linear normalization [7] is the best option for this case. DCE-MRI is a dynamic sequence and the data are normalized based on a mean kinetic expression registration as proposed in [5]. Finally, the MRSI modality has been pre-processed to correct the phase, suppress the baseline, and align the frequencies [8].

¹<http://i2cvb.github.io/>

2) *Segmentation and registration*: For this work, our radiologist has manually segmented the prostate organs on the different modalities. However, the segmented prostate needs to be registered before to extract features. The T₂-W-MRI is used as reference and each segmented prostate in other modalities are registered to this reference. Indeed, three registrations are used to correct: (i) the patient motion during the DCE-MRI acquisition, (ii) the patient motion between the T₂-W-MRI and the DCE-MRI acquisitions, and (iii) the patient motion between the T₂-W-MRI and the ADC map acquisition. Additionally, volumes from all modalities have been interpolated to the resolution of T₂-W-MRI.

3) *Feature detection*: Similarly to the pre-processing, specific features are extracted depending of the specificity of each MRI modality.

T₂-W-MRI and ADC map features Additionally to the normalized intensity, edge- and texture-based features are commonly extracted from T₂-W-MRI and ADC map. The following set of filters characterizing edges have been used: (i) Kirsch, (ii) Laplacian, (iii) Prewitt, (iv) Scharr, (v) Sobel, and (vi) Gabor. Except for the Kirsch filter, the other filters are applied in 3D, taking advantage of the volume information instead of slice information, as it is usually done. Additionally, features based on phase congruency are computed [9]. To characterize the local texture, both second-order gray-level co-occurrence matrix (GLCM)-based features [10] and rotation invariant and uniform local binary pattern (LBP) [11] are extracted. To encode 3D information, the 13 first Haralick features are computed for the 13 possible directions. For the same reason, the LBP codes are computed for the three-orthogonal-planes of each MRI volume. All these features are extracted at each voxel of the volume.

DCE-MRI features In brief, the entire enhanced signal, semi-quantitative [12], and quantitative-based models [13]–[16] are computed.

MRSI features Three different techniques are used to extract discriminative features: (i) relative quantification based on metabolite quantification, (ii) relative quantification based on bounds integration, and (iii) spectra extraction from 2 ppm to 4 ppm [5].

Anatomical features Four different metrics are

computed based on the relative distance to the prostate boundary as well as the prostate center, and the relative position in the Euclidean and cylindrical coordinate systems [17], [18].

4) *Feature balancing*: Imbalanced dataset is a common problem in medical imaging. The number of cancerous voxels is much lower than the number of “healthy” voxels for a patient. However, the problem of imbalanced dataset compromises the learning process. solving the problem of imbalanced is equivalent to under- or over-sampling part of the dataset to obtain equal number of samples in both classes. We used several methods and selected the most efficient for our dataset [19].

5) *Feature selection and extraction*: Feature selection and extraction are used in our experiment. MRSI and DCE-MRI are decomposed using three feature extraction methods: principal components analysis (PCA), sparse-PCA, and independent components analysis (ICA) are used to decompose signal-based data. Additionally to feature extraction, two methods of feature selection are used: (i) the one-way analysis of variance (ANOVA) and (ii) the Gini importance obtained while learning the random forest (RF) classifiers.

6) *Classification*: RF has been chosen as our base classifier — allowing for feature selection as well — to perform classification of individual modality as well as the combination of modalities. Additionally, we use stacking to create ensemble of base learners using a meta-classifier [20], namely AdaBoost (AdB) and Gradient Boosting (GB).

III. RESULTS AND EVALUATION

Various experiments were run in order to optimize the balancing and the feature selection strategies [5]. We found that once all features are concatenated together, nearmiss-3 (NM-3) [21] is the method providing the best enhancement of the classification performance with an area under the curve (AUC) of 0.824 ± 0.076 . Therefore, with this optimal balancing, we here report the final step consisting of three strategies: (i) the selected features from each modality (i.e., 331 features) are concatenated together and used in a RF classifier, (ii) the selected features from each modality (i.e.,




Fig. 1. Analysis of feature combination approaches after fine tuning through balancing and feature selection/extraction.




Fig. 2. Illustration the resulting detection of our mp-MRI CAD for CaP detection. The blue contours corresponds to the CaP while the jet overlay represents the probability.

331 features) are used to train a stacking classifier with a GB as meta-classifier, and (iii) the selected features from the concatenated set of feature (i.e., 267 features) are used to train a single RF classifier.

The experiments were performed in a leave-one-patient-out cross-validation (LOPO CV) fashion and a receiver operating characteristic (ROC) analysis is carried out. The comparative results are shown in Fig. 1. In overall, classification using the fine-tuned features improve the classification performance. The third classification configuration is, however, the one which outperforms others with an AUC of 0.836 ± 0.083 . The improvement in terms of AUC is of 0.028 and 0.050 compared with the 1st and 2nd configurations, respectively.

In clinical setting, the AUC score is categorized in 3 levels: (i) “acceptable” discrimination for an AUC ranging from 0.7 to 0.8, (ii) “excellent” discrimination for an AUC ranging from 0.8 to 0.9, and

“outstanding” discrimination when the AUC is over 0.9 [22]. Therefore, the combination of all MRI modalities in conjunction with fine-tuning allow to upgrade our CAD system from an “acceptable” to an “excellent” discrimination level.

To illustrate qualitatively the results of our mp-MRI CAD system, 2 diverse examples are presented in Fig. 2 by overlapping the probability map of having a CaP with the original T₂-W-MRI slice.

IV. CONCLUSION

In this paper, we presented one of the the first CAD system using all the mp-MRI modalities for prostate cancer detection. Indeed, MRSI has nearly never been used together with the other modalities. With an extensive validation approach to select the best features, the best balancing strategy as well as the best classifier, we obtained results on a rather complicated dataset of 17 patients with an average AUC of 0.836 ± 0.083 which put our system in the state-of-the-art, even so different CADs were tested on different datasets.

REFERENCES

- [1] R. Chou, J. M. Croswell, T. Dana, C. Bougatsos, I. Blazina, R. Fu, K. Gleitsmann, H. C. Koenig, C. Lam, A. Maltz, J. B. Ruge, and K. Lin, “Screening for prostate cancer: a review of the evidence for the U.S. Preventive Services Task Force,” *Ann. Intern. Med.*, vol. 155, no. 11, pp. 762–771, Dec 2011.
- [2] J. Brenner, A. Chinnaiyan, and S. Tomlins, “ETS fusion genes in prostate cancer,” in *Prostate Cancer*, ser. Protein Reviews, D. J. Tindall, Ed. Springer New York, 2013, vol. 16, pp. 139–183.
- [3] C. M. Moore, A. Ridout, and M. Emberton, “The role of MRI in active surveillance of prostate cancer,” *Curr Opin Urol*, vol. 23, no. 3, pp. 261–267, May 2013.
- [4] G. Lemaitre, R. Marti, J. Freixenet, J. C. Vilanova, P. M. Walker, and F. Meriaudeau, “Computer-aided detection and diagnosis for prostate cancer based on mono and multi-parametric mri: A review,” *Computers in Biology and Medicine*, vol. 60, pp. 8–31, 2015.
- [5] G. Lemaitre, “Computer-Aided Diagnosis for Prostate Cancer using Multi-Parametric Magnetic Resonance Imaging,” Ph.D. dissertation, Universitat de Girona and Université de Bourgogne, 2016.
- [6] G. Lemaitre, M. R. Dastjerdi, J. Massich, J. C. Vilanova, P. M. Walker, J. Freixenet, A. Meyer-Baese, F. Mériaudeau, and R. Marti, “Normalization of t2w-mri prostate images using rician a priori,” in *SPIE Medical Imaging*. International Society for Optics and Photonics, 2016, pp. 978 529–978 529.
- [7] L. G. Nyul, J. K. Udupa, and X. Zhang, “New variants of a method of MRI scale standardization,” *IEEE Trans Med Imaging*, vol. 19, no. 2, pp. 143–150, Feb 2000.
- [8] S. Parfait, P. Walker, G. Crhang, X. Tizon, and J. Mitran, “Classification of prostate magnetic resonance spectra using Support Vector Machine,” *Biomedical Signal Processing and Control*, vol. 7, no. 5, pp. 499–508, 2012.
- [9] P. Kovsi, “Image features from phase congruency,” *Videre: Journal of computer vision research*, vol. 1, no. 3, pp. 1–26, 1999.
- [10] R. Haralick, K. Shanmugam, and I. Dinstein, “Textural features for image classification,” *Systems, Man and Cybernetics, IEEE Transactions on*, vol. SMC-3, no. 6, pp. 610–621, 1973.
- [11] T. Ojala, M. Pietikainen, and T. Maenpaa, “Multiresolution gray-scale and rotation invariant texture classification with local binary patterns,” *IEEE Transactions on pattern analysis and machine intelligence*, vol. 24, no. 7, pp. 971–987, 2002.
- [12] H. J. Huisman, M. R. Engelbrecht, and J. O. Barentsz, “Accurate estimation of pharmacokinetic contrast-enhanced dynamic MRI parameters of the prostate,” *J Magn Reson Imaging*, vol. 13, no. 4, pp. 607–614, Apr 2001.
- [13] G. Brix, W. Semmler, R. Port, L. R. Schad, G. Layer, and W. J. Lorenz, “Pharmacokinetic parameters in cns gd-dtpa enhanced mr imaging,” *Journal of computer assisted tomography*, vol. 15, no. 4, pp. 621–628, 1991.
- [14] U. Hoffmann, G. Brix, M. V. Knopp, T. Heß, and W. J. Lorenz, “Pharmacokinetic mapping of the breast: a new method for dynamic mr mammography,” *Magnetic resonance in medicine*, vol. 33, no. 4, pp. 506–514, 1995.
- [15] P. S. Tofts, B. Berkowitz, and M. D. Schnall, “Quantitative analysis of dynamic gd-dtpa enhancement in breast tumors using a permeability model,” *Magnetic Resonance in Medicine*, vol. 33, no. 4, pp. 564–568, 1995.
- [16] V. Giannini, S. Mazzetti, A. Vignati, F. Russo, E. Bollito, F. Porpiglia, M. Stasi, and D. Regge, “A fully automatic computer aided diagnosis system for peripheral zone prostate cancer detection using multi-parametric magnetic resonance imaging,” *Computerized Medical Imaging and Graphics*, vol. 46, pp. 219–226, 2015.
- [17] L. Chen, Z. Weng, L. Goh, and M. Garland, “An efficient algorithm for automatic phase correction of {NMR} spectra based on entropy minimization,” *Journal of Magnetic Resonance*, vol. 158, no. 12, pp. 164–168, 2002.
- [18] G. Litjens, O. Debats, J. Barentsz, N. Karssemeijer, and H. Huisman, “Computer-aided detection of prostate cancer in MRI,” *Medical Imaging, IEEE Transactions on*, vol. 33, no. 5, pp. 1083–1092, May 2014.
- [19] G. Lemaitre, F. Nogueira, and C. K. Aridas, “Imbalanced-learn: A python toolbox to tackle the curse of imbalanced datasets in machine learning,” *CoRR*, vol. abs/1609.06570, 2016. [Online]. Available: <http://arxiv.org/abs/1609.06570>
- [20] D. H. Wolpert, “Stacked generalization,” *Neural networks*, vol. 5, no. 2, pp. 241–259, 1992.
- [21] I. Mani and I. Zhang, “knn approach to unbalanced data distributions: a case study involving information extraction,” in *Proceedings of Workshop on Learning from Imbalanced Datasets*, 2003.
- [22] D. W. Hosmer Jr and S. Lemeshow, *Applied logistic regression*. John Wiley & Sons, 2004.