

# **Advances in Quantitative MRI: Acquisition, Estimation, and Applications**

by

Gopal Nataraj

A dissertation submitted in partial fulfillment  
of the requirements for the degree of  
Doctor of Philosophy  
(Electrical Engineering and Computer Science)  
in the University of Michigan  
2016

## Doctoral Committee:

Professor Jeffrey A. Fessler, Co-Chair  
Assistant Research Scientist Jon-Fredrik Nielsen, Co-Chair  
Professor Thomas Chenevert (?)  
Professor Alfred O. Hero III (?)  
Professor Douglas C. Noll  
Associate Professor Clayton Scott (?)  
Associate Research Scientist Scott Swanson

©Gopal Nataraj

---

2016

# TABLE OF CONTENTS

<b>List of Figures</b> . . . . .	<b>iii</b>
<b>List of Appendices</b> . . . . .	<b>iv</b>
<b>List of Abbreviations</b> . . . . .	<b>v</b>
<b>Abstract</b> . . . . .	<b>vi</b>
<b>Chapter</b>	
<b>1 Introduction</b> . . . . .	<b>1</b>
1.1 Thesis Overview . . . . .	2
1.2 Thesis Organization . . . . .	2
<b>2 Background</b> . . . . .	<b>4</b>
2.1 Relevant MR Physics . . . . .	4
2.1.1 Bloch Equations . . . . .	4
2.1.2 Steady-State Sequences . . . . .	8
2.2 Optimization in QMRI . . . . .	15
2.2.1 Iterative Local Optimization with Constraints . . . . .	15
2.2.2 Partially Linear Models and the Variable Projection Method . . . . .	16
<b>3 MRI Parameter Estimation from Likelihood Models</b> . . . . .	<b>18</b>
3.1 Introduction . . . . .	18
3.2 Likelihood-Based Estimation in QMRI . . . . .	18
3.2.1 The QMRI Scan Profile . . . . .	18
3.2.2 Latent Object Parameter Estimation . . . . .	20
3.3 Experimentation . . . . .	23
3.4 Summary . . . . .	23
<b>4 Optimizing MR Scan Design for Model-Based Relaxometry</b> . . . . .	<b>24</b>
4.1 Introduction . . . . .	24
<b>5 MRI Parameter Estimation via Kernel Regression</b> . . . . .	<b>27</b>

<b>6 Myelin Water Fraction Estimation from Steady-State Sequences . . . . .</b>	<b>28</b>
<b>7 Steady-State RF Pulse Design . . . . .</b>	<b>29</b>
<b>8 Future Work . . . . .</b>	<b>30</b>
<b>Appendices . . . . .</b>	<b>31</b>
<b>Bibliography . . . . .</b>	<b>33</b>

## **LIST OF FIGURES**

## LIST OF APPENDICES

<b>A Coil Data Combination from Multiple Datasets . . . . .</b>	<b>31</b>
<b>B DESS in the Presence of Diffusion . . . . .</b>	<b>32</b>

## **LIST OF ABBREVIATIONS**

## **ABSTRACT**

### **Advances in Quantitative MRI: Acquisition, Estimation, and Applications**

**by**

**Gopal Nataraj**

**Co-Chairs: Jeffrey A. Fessler and Jon-Fredrik Nielsen**

We show that it is possible to get approximate solutions to analytically intractable equations using iterative methods. Thus we show that the author could pass an undergraduate class in numerical analysis. In addition, a unique extension to Brent's method is proposed that results in slight improvements in convergence.



# CHAPTER 1

## Introduction

{c,intro}

Magnetic resonance imaging (MRI) is a non-invasive tool that has earned widespread clinical adoption due (among other factors) to its potential for excellent soft tissue contrast, its avoidance of ionizing radiation, and its flexibility to characterize many physical phenomena. Despite its numerous advantages, MRI requires highly specialized hardware, ongoing liquid-helium cooling of its superconducting main magnet, and comparably long scan times. For these reasons, MRI is (somewhat inherently) expensive relative to other medical imaging modalities. Accordingly, one broad initiative recently advocated by the MR community is to increase the *value* of MRI examinations.

Two reasonable measures of an acquisition’s value are its sensitivity to a given disorder and its specificity in distinguishing it from others. The field of *pathology* seeks to ascribe physical processes to disorders of interest with high sensitivity and specificity. The field of *quantitative MRI* (QMRI) seeks to build MRI biomarkers that measurably describe such physical processes and thereby provide indirect information about the onset and progression of underlying conditions.

QMRI poses several challenges beyond those of commonplace anatomical MRI and thus remains yet to be widely adopted clinically. For example, latent parameter “maps” that describe relevant physical processes are often related to the received MR signal through complicated, highly nonlinear relationships. Furthermore, practical MR pulse sequences produce signals that are usually functions of not only desired but also nuisance parameters. Scan repetition is often necessary for accurate estimation of multiple desired and nuisance parameters, which can increase scan times. Mitigating these challenges (and likely others) is essential to furthering widespread clinical adoption of QMRI techniques.

## 1.1 Thesis Overview

In this thesis, we seek to build a systematic framework towards QMRI. We borrow tools from optimization, statistics, and machine learning to construct time-efficient workflows for quantifiably characterizing physical processes of interest. We apply this framework to challenging QMRI problems that are motivated by pathological studies. Our goal is to introduce tools that aid in identifying clinical tasks for which QMRI should (or should not) be part of a targeted, high-value MRI examination.

We consider two distinct subproblems in our framework. Questions in *acquisition design* (Chapters 4,6) ask how to assemble fast collections of scans that yield data rich in information about physical processes of interest. Questions in *parameter estimation* (Chapters 3,5) ask how to quickly and reliably quantify parameters associated with these relevant physical processes. The overall framework seeks to first design fast and informative scans based on the application, and to then accurately and precisely estimate application-specific parameters of interest.

## 1.2 Thesis Organization

The main body of this thesis is organized as follows:

- Chapter 2 reviews relevant background material on MRI and optimization.
- Chapter 3 discusses methods for MRI parameter estimation from likelihood models and applies these methods for model-based MR relaxometry, (*i.e.*, estimation of relaxation parameters  $T_1, T_2$ ), of interest for many neurological applications. It derives some content (especially regarding applications) from conference papers [1, 2].
- Chapter 4 introduces a minimax optimization approach to aid robust and application-specific MR scan selection and optimization for precise latent parameter estimation. It optimizes several practical acquisitions and uses the likelihood-based estimation techniques introduced in Chapter 3 to assess the utility of scan optimization through simulations, phantom studies, and *in vivo* experiments. It derives content mainly from journal paper [3] and conference paper [4].
- Chapter 5 describes MRI parameter estimation using kernel ridge regression. It derives content from conference paper [5].
- Chapter 6 introduces a multi-compartmental model for relevant MR pulse sequences and proposes a new acquisition useful for myelin water fraction estimation, of in-

terest in white matter disorders. It applies kernel-based MR parameter estimation to estimate myelin water fraction, in simulations and *in vivo* experiments. It derives some content from conference paper [6].

- Chapter 7 presents some relatively immature ideas on steady-state radiofrequency (RF) pulse design as well as associated challenges. This work is presently unpublished and may offer avenues for further research.
- Chapter 8 summarizes several items of possible future work (on both short- and long-term timescales) and presents a timeline for completion of this thesis.

The appendices are organized as follows:

- Appendix A proposes an algorithm for combining multiple MRI datasets (as is necessary for many parameter estimation problems), when each dataset is acquired using multiple receiver coils.
- Appendix B presents an analysis of model mismatch due to the presence of diffusion, shows that neglecting diffusive effects during  $T_2$  estimation can cause significant bias, and suggests acquisition modifications for mitigating this bias.

## CHAPTER 2

# Background

{c,bkgrd}

This chapter focuses only on background information pertinent to multiple subsequent chapters. We present further topic-specific information at the beginnings of corresponding chapters. Section 2.1 places emphasis on reviewing necessary MR fundamentals, and Section 2.2 proceeds to a shorter discussion regarding optimization as it pertains to QMRI.

## 2.1 Relevant MR Physics

{s,bkgrd,mri}

This section begins with the fundamental Bloch equations and derives the signal models associated with two MR pulse sequences used extensively in this thesis. Our coverage of MRI is far from comprehensive, and omits fundamental but tangential topics such as signal localization. We refer the interested reader to books such as [7, 8, 9].

### 2.1.1 Bloch Equations

{ss,bkgrd,mri,bloch}

The Bloch equations [10] describe the magnetization dynamics of *spins*, or (loosely) atomic nuclei with nonzero angular momentum and thus nonzero magnetic moment, *e.g.*  $^1\text{H}$ . If the dominant source of magnetic flux arises (as is typical in MRI) from a main magnetic field that is oriented along the  $z$ -axis, the equations read

{eq:bloch-mxy}

$$\frac{\partial}{\partial t} m_{xy}(\mathbf{r}, t) = i\gamma(m_z(\mathbf{r}, t)b_{xy}(\mathbf{r}, t) - m_{xy}(\mathbf{r}, t)b_z(\mathbf{r}, t)) - \frac{m_{xy}(\mathbf{r}, t)}{T_2(\mathbf{r})}; \quad (2.1)$$

{eq:bloch-mz}

$$\frac{\partial}{\partial t} m_z(\mathbf{r}, t) = \gamma(m_x(\mathbf{r}, t)b_y(\mathbf{r}, t) - m_y(\mathbf{r}, t)b_x(\mathbf{r}, t)) - \frac{m_z(\mathbf{r}, t) - m_0(\mathbf{r})}{T_1(\mathbf{r})}. \quad (2.2)$$

Here,  $m_{xy}(\mathbf{r}, t) := m_x(\mathbf{r}, t) + im_y(\mathbf{r}, t) \in \mathbb{C}$  and  $m_z(\mathbf{r}, t) \in \mathbb{R}$  are the transverse and longitudinal components of the magnetization vector at position  $\mathbf{r} := [x, y, z]^T \in \mathbb{R}^3$  and time  $t \geq 0$ ;  $b_{xy}(\mathbf{r}, t) := b_x(\mathbf{r}, t) + ib_y(\mathbf{r}, t) \in \mathbb{C}$  and  $b_z(\mathbf{r}, t) \in \mathbb{R}$  are the transverse and longitudinal components (in an inertial reference frame) of the applied magnetic field;  $T_1(\mathbf{r})$

and  $T_2(\mathbf{r})$  are spin-lattice and spin-spin relaxation time constants;  $m_0(\mathbf{r})$  is the equilibrium magnetization and is proportional to the density of spins per unit volume as well as the main field strength;  $\gamma$  is the gyromagnetic ratio; and  $i := \sqrt{-1}$ . As written, equations (2.1)-(2.2) only model dominant temporal dynamics; later chapters consider second-order effects such as multiple magnetization compartments (Chapter 6) and diffusion (Appendix B).

It is often convenient to study Bloch dynamics in a non-inertial reference frame rotating clockwise about the  $z$ -axis at Larmor frequency  $\omega_0 := \gamma B_0$ , where  $B_0 \hat{k}$  is the (nearly uniform) main magnetic field. In these coordinates, the apparent transverse magnetic field  $b'_{xy}(\mathbf{r}, t) = b'_x(\mathbf{r}, t) + ib'_y(\mathbf{r}, t) := b_{xy}(\mathbf{r}, t)e^{i\omega_0 t}$  transforms only in phase, but the apparent longitudinal magnetic field  $b'_z(\mathbf{r}, t) := b_z(\mathbf{r}, t) - B_0$  is greatly reduced in magnitude. The magnetization components transform more simply as  $m'_{xy}(\mathbf{r}, t) = m'_x(\mathbf{r}, t) + im'_y(\mathbf{r}, t) := m_{xy}(\mathbf{r}, t)e^{i\omega_0 t}$  and  $m'_z(\mathbf{r}, t) := m_z(\mathbf{r}, t)$ . Remarkably, inserting these coordinate transformations into (2.1)-(2.2) does not change the form of the dynamical equations:

$$\text{\{eq:bloch-mxyp\}} \quad \frac{\partial}{\partial t} m'_{xy}(\mathbf{r}, t) = i\gamma(m'_z(\mathbf{r}, t)b'_{xy}(\mathbf{r}, t) - m'_{xy}(\mathbf{r}, t)b'_z(\mathbf{r}, t)) - \frac{m'_{xy}(\mathbf{r}, t)}{T_2(\mathbf{r})}; \quad (2.3)$$

$$\text{\{eq:bloch-mzp\}} \quad \frac{\partial}{\partial t} m'_z(\mathbf{r}, t) = \gamma(m'_x(\mathbf{r}, t)b'_y(\mathbf{r}, t) - m'_y(\mathbf{r}, t)b'_x(\mathbf{r}, t)) - \frac{m'_z(\mathbf{r}, t) - m_0(\mathbf{r})}{T_1(\mathbf{r})}. \quad (2.4)$$

It thus suffices to consider how perturbations  $\mathbf{b}'(\mathbf{r}, t)$  to main field  $B_0 \hat{k}$  influence rotating-frame magnetization  $\mathbf{m}'(\mathbf{r}, t)$  via Eqs. (2.3)-(2.4). The inertial-frame magnetization  $\mathbf{m}(\mathbf{r}, t)$  is then easily constructed via  $m_{xy}(\mathbf{r}, t) = m'_{xy}(\mathbf{r}, t)e^{-i\omega_0 t}$  and  $m_z(\mathbf{r}, t) = m'_z(\mathbf{r}, t)$ .

It is challenging to explicitly solve Eqs. (2.3)-(2.4) for arbitrary field perturbations  $\mathbf{b}'(\mathbf{r}, t)$ . We discuss relevant special cases in the following.

### 2.1.1.1 Non-Selective Excitation

Here, we derive solutions to Eqs. (2.3)-(2.4) in the case of short, spatially non-selective excitations. We take the following common assumptions:

- We assume negligible spatial variation in the longitudinal magnetic field, so  $b'_z(\mathbf{r}, t) \approx 0$ . This lack of spatial variation is reason for non-selective excitation.
- We assume the transverse field separates in position and time; oscillates at the Larmor frequency (commonly in the radiofrequency (RF) range); and aligns at initial time  $t \leftarrow t_0$  with the  $x$ -axis. Together, these assumptions restrict the so-called RF excitation to take form  $b'_{xy}(\mathbf{r}, t) \approx s^t(\mathbf{r})b'_{1,x}(t)\hat{i} + 0\hat{j}$ , where  $s^t(\mathbf{r}) \in \mathbb{R}$  is the RF transmit coil spatial variation and  $b'_{1,x}(t) \in \mathbb{R}$  is the RF excitation envelope.

- We assume that the duration  $T_P$  of RF excitation (often  $T_P \sim 1\text{ms}$ ) is much shorter than relaxation time constants (typically  $T_1 \sim 1000\text{ms}$  and  $T_2 \sim 50\text{ms}$  in brain tissue) and thus neglect relaxation effects during excitation.

Under these assumptions, Eqs. (2.3)-(2.4) reduce to the linear system

$$\frac{\partial}{\partial t} \begin{bmatrix} m'_x(\mathbf{r}, t) \\ m'_y(\mathbf{r}, t) \\ m'_z(\mathbf{r}, t) \end{bmatrix} = \begin{bmatrix} 0 & 0 & 0 \\ 0 & 0 & \gamma s^t(\mathbf{r}) b'_{1,x}(t) \\ 0 & -\gamma s^t(\mathbf{r}) b'_{1,x}(t) & 0 \end{bmatrix} \begin{bmatrix} m'_x(\mathbf{r}, t) \\ m'_y(\mathbf{r}, t) \\ m'_z(\mathbf{r}, t) \end{bmatrix}. \quad (2.5)$$

Eq. (2.5) admits the simple solution (for  $t \geq t_0$ )

$$\begin{bmatrix} m'_x(\mathbf{r}, t) \\ m'_y(\mathbf{r}, t) \\ m'_z(\mathbf{r}, t) \end{bmatrix} = \begin{bmatrix} 1 & 0 & 0 \\ 0 & \cos(\alpha(\mathbf{r}, t; t_0)) & \sin(\alpha(\mathbf{r}, t; t_0)) \\ 0 & -\sin(\alpha(\mathbf{r}, t; t_0)) & \cos(\alpha(\mathbf{r}, t; t_0)) \end{bmatrix} \begin{bmatrix} m'_x(\mathbf{r}, t_0) \\ m'_y(\mathbf{r}, t_0) \\ m'_z(\mathbf{r}, t_0) \end{bmatrix}, \quad (2.6)$$

where  $\mathbf{m}'(\mathbf{r}, t_0) := [m'_x(\mathbf{r}, t_0), m'_y(\mathbf{r}, t_0), m'_z(\mathbf{r}, t_0)]^T$  is the initial magnetization and

$$\alpha(\mathbf{r}, t; t_0) := \gamma s^t(\mathbf{r}) \int_{t_0}^t b'_{1,x}(\tau) d\tau \quad (2.7)$$

is the nutation (or “flip”) angle at time  $t$ . Eq. (2.6) reveals that on-resonance RF excitation causes the magnetization vector to rotate clockwise about an axis parallel to the direction of excitation. The nutation angle accumulated over an RF pulse of duration  $T_P$  is often decomposed as  $\alpha(\mathbf{r}, t_0 + T_P; t_0) =: \alpha_0 s^t(\mathbf{r})$ , where  $\alpha_0$  is a prescribed nominal flip angle.

For deriving signal models in later sections, it is convenient and intuitive to define matrix operators that summarize relevant dynamics. Here, we rewrite Eq. (2.6) as

$$\mathbf{m}'(\mathbf{r}, t) = \mathbf{R}_{x'}(\alpha(\mathbf{r}, t; t_0)) \mathbf{m}'(\mathbf{r}, t_0), \quad (2.8)$$

where  $\mathbf{R}_{x'}(\alpha(\mathbf{r}, t; t_0))$  denotes a clockwise rotation of angle  $\alpha(\mathbf{r}, t; t_0)$  about the  $x'$ -axis.

### 2.1.1.2 Free Precession and Relaxation

Next, we derive solutions to the rotating-frame Bloch equations when no RF excitation is present, *i.e.*  $b'_{xy}(\mathbf{r}, t) \approx 0$ . In this case, Eqs. (2.3)-(2.4) decouple, yielding separate

dynamical equations for the transverse and longitudinal magnetization components:

$$\frac{\partial}{\partial t} m'_{xy}(\mathbf{r}, t) = -i\gamma m'_{xy}(\mathbf{r}, t) b'_z(\mathbf{r}, t) - \frac{m'_{xy}(\mathbf{r}, t)}{T_2(\mathbf{r})}; \quad (2.9)$$

$$\frac{\partial}{\partial t} m'_z(\mathbf{r}, t) = -\frac{m'_z(\mathbf{r}, t) - m_0(\mathbf{r})}{T_1(\mathbf{r})}. \quad (2.10)$$

Eqs. (2.9)-(2.10) admit simple solutions with no further assumptions:

$$m'_{xy}(\mathbf{r}, t) = m'_{xy}(\mathbf{r}, t_0) e^{-(t-t_0)/T_2(\mathbf{r})} e^{-i\phi'(\mathbf{r}, t; t_0)}; \quad (2.11)$$

$$m'_z(\mathbf{r}, t) = m'_z(\mathbf{r}, t_0) e^{-(t-t_0)/T_1(\mathbf{r})} + m_0(\mathbf{r}) (1 - e^{-(t-t_0)/T_1(\mathbf{r})}), \quad (2.12)$$

where  $m'_{xy}(\mathbf{r}, t_0)$  and  $m'_z(\mathbf{r}, t_0)$  are the initial magnetization components and

$$\phi'(\mathbf{r}, t; t_0) := \gamma \int_{t_0}^t b'_z(\mathbf{r}, \tau) d\tau \quad (2.13)$$

denotes the phase accumulation due to main field inhomogeneity (often called off-resonance effects). Eq. (2.11) reveals that without RF excitations, the transverse magnetization  $m'_{xy}(\mathbf{r}, t)$  relaxes to zero exponentially fast with time constant  $T_2(\mathbf{r})$ , while accruing phase due to off-resonance effects. Eq. (2.12) similarly reveals that without RF excitations, longitudinal magnetization  $m'_z(\mathbf{r}, t)$  recovers to  $m_0(\mathbf{r})$  exponentially fast with time constant  $T_1(\mathbf{r})$ .

As in Section 2.1.1.2, we rewrite Eqs. (2.11)-(2.12) for  $t \geq t_0$  using matrix operators:

$$\mathbf{m}'(\mathbf{r}, t) = \mathbf{R}_{z'}(\phi'(\mathbf{r}, t; t_0)) \mathbf{E}(\mathbf{r}, t; t_0) \mathbf{m}'(\mathbf{r}, t_0) + \mathbf{m}_0(\mathbf{r}, t; t_0) \quad (2.14)$$

where  $\mathbf{m}_0(\mathbf{r}, t; t_0) := m_0(\mathbf{r}) (1 - e^{-(t-t_0)/T_1(\mathbf{r})}) \hat{k}$ ;

$$\mathbf{R}_{z'}(\phi'(\mathbf{r}, t; t_0)) := \begin{bmatrix} \cos(\phi'(\mathbf{r}, t; t_0)) & \sin(\phi'(\mathbf{r}, t; t_0)) & 0 \\ -\sin(\phi'(\mathbf{r}, t; t_0)) & \cos(\phi'(\mathbf{r}, t; t_0)) & 0 \\ 0 & 0 & 1 \end{bmatrix} \quad (2.15)$$

denotes a clockwise rotation of angle  $\phi'(\mathbf{r}, t; t_0)$  about the  $z'$ -axis; and

$$\mathbf{E}(\mathbf{r}, t; t_0) := \begin{bmatrix} e^{-(t-t_0)/T_2(\mathbf{r})} & 0 & 0 \\ 0 & e^{-(t-t_0)/T_2(\mathbf{r})} & 0 \\ 0 & 0 & e^{-(t-t_0)/T_1(\mathbf{r})} \end{bmatrix} \quad (2.16)$$

is an exponential relaxation operator. Section 2.1.2 (and later chapters) use matrix dynamical representations (2.8) and (2.14) to succinctly describe pulse sequence signal models.

{ss,bkgrd,mri,ss}

## 2.1.2 Steady-State Sequences

MRI experiments typically involve repeated cycles of (pulsed) RF excitation; signal localization (not discussed here); and transverse  $T_2$  relaxation and free precession, alongside (relatively slow) longitudinal  $T_1$  recovery. We can build models of the received MR signal by considering the magnetization dynamics induced by specific pulse sequences.

Classical pulse sequences use relatively long cycle repetition times  $T_R$  to ensure near-complete  $T_1$  recovery of the magnetization vector back to equilibrium state  $m_0(\mathbf{r})\hat{k}$  prior to the start of each RF cycle. For such long- $T_R$  sequences, it suffices to approximate the magnetization as fully recovered (*i.e.*,  $\mathbf{m}'(\mathbf{r}, t_0 + rT_R) \approx m_0(\mathbf{r})\hat{k}, \forall r \in \{0, 1, 2, \dots\}$ ) just prior to each RF excitation. This approximation yields a sequence of initial conditions and allows computation of the magnetization at corresponding times of data acquisition via direct application of Bloch dynamics (2.8) and (2.14). Resulting signal models are typically simple expressions of relaxation parameters  $T_1(\mathbf{r})$  and  $T_2(\mathbf{r})$ ; however, model accuracy often depends strongly on the long- $T_R$  assumption, which requires long acquisitions.

Steady-state (SS) sequences [11] utilize short  $T_R$ , and can thus achieve much faster scan times. Due to short repetition times, SS sequences achieve only partial  $T_1$  recovery in between RF excitations; thus, their magnetization responses do not obey the simple classical initial conditions (for the second RF cycle onwards). Although their transient magnetization dynamics can be complicated, SS sequences produce (under certain assumptions [12]) long-time magnetization responses that eventually<sup>1</sup> achieve a steady-state condition:

{eq:ss-cond}

$$\lim_{t_0 \rightarrow \infty} \mathbf{m}'(\mathbf{r}, t_0 + rT_R) = \mathbf{m}'(\mathbf{r}, t_0), \quad (2.17)$$

where repetition count  $r \in \{1, 2, \dots\}$  for fixed RF excitations and off-resonance induced phase increments (as is assumed in the following). Subsections 2.1.2.1 and 2.1.2.2 use SS condition (2.17) and Bloch equation matrix operators introduced in (2.8) and (2.14) to derive long-time signal models for Spoiled Gradient-Recalled Echo (SPGR) and Dual-Echo Steady-State (DESS), two SS pulse sequences useful for quantitative MRI.

### 2.1.2.1 Spoiled Gradient-Recalled Echo (SPGR) Sequence

SPGR [14] is a fast pulse sequence that repeats cycles of fixed RF excitation (such that  $b'_{1,x}(t + rT_R) = b'_{1,x}(t), \forall t \in [t_0, t_0 + T_P], r \in \{1, 2, \dots\}$ ); data acquisition; relaxation and recovery; and residual transverse magnetization “spoiling” (discussed later). Here we

<sup>1</sup>The progression to steady state takes on the order of  $5T_2/T_R$  RF cycles [12], typically a small but not insignificant period during which data acquisition is often foregone. This transition can (in some cases) be accelerated by prepending SS sequences with tailored “magnetization-catalyzing” modules [13].

{sss,bkgrd,mri,ss,spgr}



develop a simple and popular steady-state SPGR signal model.

Let  $\mathbf{m}'(\mathbf{r}, t_0)$  denote the magnetization at an initial time  $t_0$  selected well into the steady-state and just prior to excitation. The SPGR sequence first applies an RF excitation, which rotates the initial magnetization as per (2.8):

$$\{\text{eq:spgr-ex}\} \quad \mathbf{m}'(\mathbf{r}, t_0 + T_P) = \mathbf{R}_{x'}(\alpha(\mathbf{r}, t_0 + T_P; t_0))\mathbf{m}'(\mathbf{r}, t_0). \quad (2.18)$$

The excited magnetization then precesses and relaxes as per (2.14) until data acquisition, defined to occur at “echo time”  $T_E \in [T_P, T_R]$  after the (midpoint of) RF excitation:

$$\{\text{eq:spgr-da}\} \quad \begin{aligned} \mathbf{m}'\left(\mathbf{r}, t_0 + \frac{T_P}{2} + T_E\right) &= \mathbf{R}_{z'}\left(\phi'\left(\mathbf{r}, \frac{T_P}{2} + T_E; T_P\right)\right)\mathbf{E}\left(\mathbf{r}, \frac{T_P}{2} + T_E; T_P\right)\mathbf{m}'(\mathbf{r}, t_0 + T_P) \\ &+ \mathbf{m}_0\left(\mathbf{r}, \frac{T_P}{2} + T_E; T_P\right). \end{aligned} \quad (2.19)$$

Following signal reception, the remaining transverse magnetization is spoiled<sup>2</sup> while the longitudinal component is unaffected. We model an ideal spoiling operation as

$$\{\text{eq:spgr-spoil}\} \quad \mathbf{S}\mathbf{m}'\left(\mathbf{r}, \frac{T_P}{2} + T_E\right), \text{ where } \mathbf{S} := \begin{bmatrix} 0 & 0 & 0 \\ 0 & 0 & 0 \\ 0 & 0 & 1 \end{bmatrix}. \quad (2.20)$$

After spoiling, the longitudinal magnetization (partially) recovers until  $t \leftarrow t_0 + T_R$ :

$$\{\text{eq:spgr-pr}\} \quad \begin{aligned} \mathbf{m}'(\mathbf{r}, t_0 + T_R) &= \mathbf{R}_{z'}\left(\phi'\left(\mathbf{r}, T_R; \frac{T_P}{2} + T_E\right)\right)\mathbf{E}\left(\mathbf{r}, T_R; \frac{T_P}{2} + T_E\right)\mathbf{S}\mathbf{m}'\left(\mathbf{r}, t_0 + \frac{T_P}{2} + T_E\right) \\ &+ \mathbf{m}_0\left(\mathbf{r}, T_R; \frac{T_P}{2} + T_E\right). \end{aligned} \quad (2.21)$$

In steady-state, one cycle of excitation, acquisition, spoiling, and recovery returns the magnetization back to its initial state. We enforce this through the steady-state condition

$$\{\text{eq:spgr-ss}\} \quad \mathbf{m}'(\mathbf{r}, t_0 + T_P) = \mathbf{R}_{x'}(\alpha(\mathbf{r}, t_0 + T_P; t_0))\mathbf{m}'(\mathbf{r}, t_0 + T_R) \quad (2.22)$$

---

<sup>2</sup>Transverse signal spoiling is often (nearly) achieved in practice using strong induced field inhomogeneities (which cause rapid transverse signal dephasing) in tandem with RF excitations that additionally impart nonlinear (often quadratically increasing) transverse magnetization phase [14]. Though the nonlinear RF phase used in so-called “RF-spoiling” prevents any one spin from reaching a true steady-state, the signal integrated over a typically-sized voxel achieves SS-like behavior [15].

which yields an algebraic system of equations. When it exists, the solution is

$$\mathbf{m}'(\mathbf{r}, t_0 + T_P) = \frac{1}{1 - e^{-(T_R - T_P)/T_1(\mathbf{r})} \cos(\alpha(\mathbf{r}))} \begin{bmatrix} 0 \\ m_0(\mathbf{r}) \sin(\alpha(\mathbf{r})) (1 - e^{-(T_R - T_P)/T_1(\mathbf{r})}) \\ m_0(\mathbf{r}) \cos(\alpha(\mathbf{r})) (1 - e^{-(T_R - T_P)/T_1(\mathbf{r})}) \end{bmatrix}, \quad (2.23)$$

{eq:spgr-bmmp-t0}

where  $\alpha(\mathbf{r}) := \alpha(\mathbf{r}, t_0 + T_P; t_0)$  is a slight abuse of notation. Remarkably, the SPGR steady-state magnetization immediately after excitation is approximately independent of both off-resonance effects and  $T_2(\mathbf{r})$ . Researchers more often cite the expression

$$\begin{aligned} m'_{xy}(\mathbf{r}, t_0 + T_P) &= m'_x(\mathbf{r}, t_0 + T_P) + im'_y(\mathbf{r}, t_0 + T_P) \\ &= \frac{im_0(\mathbf{r}) \sin(\alpha(\mathbf{r})) (1 - e^{-T_R/T_1(\mathbf{r})})}{1 - e^{-T_R/T_1(\mathbf{r})} \cos(\alpha(\mathbf{r}))} \end{aligned} \quad (2.24)$$

{eq:spgr-mxyp-t0}

for the complex transverse magnetization as it modifies (2.23) to include a simple first-order correction for unaccounted  $T_1$  recovery during the RF pulse. Substituting (2.24) into (2.19) yields an expression for the transverse magnetization at the echo time:

$$\begin{aligned} m'_{xy}\left(\mathbf{r}, t_0 + \frac{T_P}{2} + T_E\right) &= m'_{xy}(\mathbf{r}, t_0 + T_P) e^{-(T_E - T_P/2)/T_2(\mathbf{r})} e^{-i\phi'(\mathbf{r}, t_0 + \frac{T_P}{2} + T_E; t_0 + T_P)} \\ &\approx m'_{xy}(\mathbf{r}, t_0 + T_P) e^{-T_E/T_2(\mathbf{r})} e^{-i\phi'(\mathbf{r}, t_0 + \frac{T_P}{2} + T_E; t_0 + \frac{T_P}{2})}, \end{aligned} \quad (2.25)$$

{eq:spgr-mxyp-te}

where the approximation again keeps in line with literature expressions.

The received signal is approximately proportional to the integrated transverse magnetization over a volume  $\mathbb{V}$ . To derive expressions, we take a few more usual assumptions:

- We assume that the signal is localized to a scale over which there is off-resonance phase variation, but minimal variation of  $m_0(\mathbf{r})$ ,  $T_1(\mathbf{r})$ ,  $T_2(\mathbf{r})$ , and  $\alpha(\mathbf{r})$ . This assumption is reasonable<sup>3</sup> when describing the signal arising from a typical voxel.
- We assume that (free-precession) off-resonance phase grows linearly with time, *i.e.*  $\phi'(\mathbf{r}, t_0 + \frac{T_P}{2} + T_E; t_0 + \frac{T_P}{2}) \approx \omega'(\mathbf{r})T_E$ . We further assume that off-resonance frequency  $\omega'(\mathbf{r})$  is distributed over the localized voxel as  $p_{\omega'} \leftarrow \text{Cauchy}(\bar{\omega}', R'_2)$ , where  $\bar{\omega}'(\mathbf{r})$  is the median off-resonance frequency and  $R'_2(\mathbf{r})$  is the broadening bandwidth.

With these additional assumptions, the received steady-state SPGR (noiseless) signal model

---

<sup>3</sup>Model mismatch due to within-voxel spatial variation of relaxation parameters can be significant, especially for large voxels. Chapter 6 studies so-called partial volume effects and uses them for QMRI.

for a typically sized voxel centered at position  $\mathbf{r}$  is (to within constants):

$$\{eq:spgr-int\} \quad s_S\left(\mathbf{r}, t_0 + \frac{T_P}{2} + T_E\right) \propto \int_{\mathbb{V}(\mathbf{r})} m'_{xy}\left(\mathbf{r}, t_0 + \frac{T_P}{2} + T_E\right) d^3 \mathbf{r} \quad (2.26)$$

$$\begin{aligned} &\approx m_{xy}(\mathbf{r}, t_0 + T_P) e^{-T_E/T_2(\mathbf{r})} \int_{\mathbb{R}} e^{-i\omega' T_E} p_{\omega'}(\omega') d\omega' \\ &= m_{xy}(\mathbf{r}, t_0 + T_P) e^{-T_E/T_2(\mathbf{r})} e^{-R'_2(\mathbf{r}) T_E - i\bar{\omega}'(\mathbf{r}) T_E} \\ \{eq:spgr-model\} &= \frac{im_0(\mathbf{r}) \sin(\alpha(\mathbf{r})) (1 - e^{-T_R/T_1(\mathbf{r})})}{1 - e^{-T_R/T_1(\mathbf{r})} \cos(\alpha(\mathbf{r}))} e^{-T_E/T_2^*(\mathbf{r})} e^{-i\bar{\omega}'(\mathbf{r}) T_E}, \end{aligned} \quad (2.27)$$

where  $T_2^*(\mathbf{r}) := \left(\frac{1}{T_2} + R'_2\right)^{-1}$  is a modified spin-spin relaxation time that accounts for additional transverse magnetization decay due to off-resonance effects.

### 2.1.2.2 Dual-Echo Steady-State (DESS) Sequence

DESS [16, 17] is a fast pulse sequence that interlaces fixed, constant-phase RF excitations with fixed dephasing “gradients” (*i.e.*, induced main field inhomogeneities that vary nearly linearly with space) to produce two distinct signals per RF excitation. Here we develop simple steady-state DESS signal models.

As in Subsection 2.1.2.1, let  $\mathbf{m}'(\mathbf{r}, t_0)$  denote the magnetization at an initial time  $t_0$  selected well into the steady-state and just prior to excitation. The DESS sequence first applies a fixed RF rotation  $\alpha(\mathbf{r}) := \alpha(\mathbf{r}, t_0 + rT_R + T_P; t_0 + rT_R), \forall r \in \{0, 1, 2, \dots\}$ :

$$\{eq:dess-ex\} \quad \mathbf{m}'(\mathbf{r}, t_0 + T_P) = \mathbf{R}_{x'}(\alpha(\mathbf{r})) \mathbf{m}'(\mathbf{r}, t_0). \quad (2.28)$$

The excited transverse magnetization contributes to a first acquired signal; dephases (but does not spoil completely) due to gradient dephasing<sup>4</sup> and contributes again to a second (smaller, but nonzero) acquired signal. Since (with proper selection) dephasing gradients mainly contribute to off-resonance phase accrual, the net effect after data acquisition and gradient spoiling is well described simply by precession and relaxation:

$$\{eq:dess-pr\} \quad \mathbf{m}'(\mathbf{r}, t_0 + T_R) = \mathbf{R}_{z'}(\phi'(\mathbf{r})) \mathbf{E}(\mathbf{r}, T_R; T_P) \mathbf{m}'(\mathbf{r}, t_0 + T_P) + \mathbf{m}_0(\mathbf{r}, T_R; T_P), \quad (2.29)$$

where the abbreviation  $\phi'(\mathbf{r}) := \phi'(\mathbf{r}, t_0 + (r+1)T_R; t_0 + rT_R + T_P), \forall r \in \{0, 1, 2, \dots\}$  implies fixed phase accrual (due to gradient dephasing, field inhomogeneity, and other

<sup>4</sup>It is worth distinguishing gradient dephasing (commonly but somewhat misleadingly referred to as gradient spoiling) from RF spoiling. Gradient dephasing (used in DESS) primarily affects magnetization phase and is modeled simply as precession. RF spoiling (used in SPGR) combines gradient dephasing with nonlinear RF phase cycling and suppresses magnetization magnitude in steady-state.

unaccounted effects) over each repetition cycle.

In steady-state, one cycle of excitation, first acquisition, gradient spoiling, second acquisition, and (partial) recovery returns the magnetization back to its initial state. We enforce this through the steady-state condition

$$\{\text{eq:dess-ss}\} \quad \mathbf{m}'(\mathbf{r}, t_0) = \mathbf{m}'(\mathbf{r}, t_0 + T_R) \quad (2.30)$$

which yields an algebraic system of equations. When it exists, the solution gives the steady-state magnetization just prior to RF excitation:

$$\{\text{eq:dess-bmmp-t0}\} \quad \mathbf{m}'(\mathbf{r}, t_0) = \begin{bmatrix} E_2(\mathbf{r}, T_F) \sin \alpha(\mathbf{r}) \sin \phi'(\mathbf{r}) \\ -E_2(\mathbf{r}, T_F) \sin \alpha(\mathbf{r}) (E_2(\mathbf{r}, T_F) - \cos \phi'(\mathbf{r})) \\ 1 - E_2(\mathbf{r}, T_F) \cos \phi'(\mathbf{r}) + E_2(\mathbf{r}, T_F) \cos \alpha(\mathbf{r}) (E_2(\mathbf{r}, T_F) - \cos \phi'(\mathbf{r})) \end{bmatrix} q(\mathbf{r}, T_F), \quad (2.31)$$

where  $T_F := T_R - T_P$  is the free precession interval;  $E_1(\mathbf{r}, t) := e^{-t/T_1(\mathbf{r})}$  and  $E_2(\mathbf{r}, t) := e^{-t/T_2(\mathbf{r})}$  are relaxation operators; and  $q(\mathbf{r}, t) :=$

$$\frac{m_0(\mathbf{r})(1 - E_1(\mathbf{r}, t))}{(1 - E_1(\mathbf{r}, t) \cos \alpha(\mathbf{r}))(1 - E_2(\mathbf{r}, t) \cos \phi'(\mathbf{r})) - E_2(\mathbf{r}, t)(E_1(\mathbf{r}, t) - \cos \alpha(\mathbf{r}))(E_2(\mathbf{r}, t) - \cos \phi'(\mathbf{r}))}.$$

Substituting (2.31) into (2.28) produces a similar expression for the steady-state magnetization immediately following RF excitation:

$$\{\text{eq:dess-bmmp-tp}\} \quad \mathbf{m}'(\mathbf{r}, t_0 + T_P) = \begin{bmatrix} E_2(\mathbf{r}, T_F) \sin \alpha(\mathbf{r}) \sin \phi'(\mathbf{r}) \\ \sin \alpha(\mathbf{r})(1 - E_2(\mathbf{r}, T_F) \cos \phi'(\mathbf{r})) \\ \cos \alpha(\mathbf{r})(1 - E_2(\mathbf{r}, T_F) \cos \phi'(\mathbf{r})) + E_2(\mathbf{r}, T_F)(E_2(\mathbf{r}, T_F) - \cos \phi'(\mathbf{r})) \end{bmatrix} q(\mathbf{r}, T_F). \quad (2.32)$$

The transverse magnetizations before and after RF excitation are then

$$\{\text{eq:dess-mxyp-t0}\} \quad m'_{xy}(\mathbf{r}, t_0) = -i \sin \alpha(\mathbf{r}) E_2(\mathbf{r}, T_R) \left( E_2(\mathbf{r}, T_R) - e^{-i\phi'(\mathbf{r})} \right) q(\mathbf{r}, T_R); \quad (2.33)$$

$$\{\text{eq:dess-mxyp-tp}\} \quad m'_{xy}(\mathbf{r}, t_0 + T_P) = +i \sin \alpha(\mathbf{r}) \left( 1 - E_2(\mathbf{r}, T_R) e^{i\phi'(\mathbf{r})} \right) q(\mathbf{r}, T_R), \quad (2.34)$$

where (2.33)-(2.34) include simple first-order corrections for yet-unaccounted relaxation and recovery during excitation. Frequently, the DESS signals are acquired at symmetric echo times  $T_E$  before and after the center of each RF pulse. Substituting (2.34) into (2.9)

gives the magnetization at the data acquisition time after RF excitation:

$$m'_{xy}\left(\mathbf{r}, t_0 + \frac{T_P}{2} + T_E\right) = m'_{xy}(\mathbf{r}, t_0 + T_P) e^{-(T_E - T_P/2)/T_2(\mathbf{r})} e^{-i\phi'(\mathbf{r}, t_0 + \frac{T_P}{2} + T_E; t_0 + T_P)}$$

$$\approx m'_{xy}(\mathbf{r}, t_0 + T_P) e^{-T_E/T_2(\mathbf{r})} e^{-i\phi'(\mathbf{r}, t_0 + \frac{T_P}{2} + T_E; t_0 + \frac{T_P}{2})} \quad (2.35)$$

{eq:dess-mxyp-te1-ph}

{eq:dess-mxyp-te1}

$$\approx m'_{xy}(\mathbf{r}, t_0 + T_P) e^{-T_E/T_2(\mathbf{r})} e^{-i\omega'(\mathbf{r})T_E}, \quad (2.36)$$

where in (2.35) we again approximately correct for relaxation during excitation and in (2.36) we assume linear off-resonance phase accrual during free precession. To compute the magnetization at the acquisition time before excitation, we consider the free precession and relaxation that occurs between<sup>5</sup> signal reception and excitation:

$$m'_{xy}(\mathbf{r}, t_0) = m'_{xy}\left(\mathbf{r}, t_0 - \left(T_E - \frac{T_P}{2}\right)\right) e^{-(T_E - T_P/2)/T_2(\mathbf{r})} e^{-i\phi'(\mathbf{r}, t_0; t_0 - (T_E - \frac{T_P}{2}))}. \quad (2.37)$$

Rearranging (2.37) and applying approximations similar to those of (2.35)-(2.36),

$$m'_{xy}\left(\mathbf{r}, t_0 + \frac{T_P}{2} - T_E\right) = m'_{xy}(\mathbf{r}, t_0) e^{+(T_E - T_P/2)/T_2(\mathbf{r})} e^{+i\phi'(\mathbf{r}, t_0; t_0 - (T_E - \frac{T_P}{2}))}$$

$$\approx m'_{xy}(\mathbf{r}, t_0) e^{+T_E/T_2(\mathbf{r})} e^{+i\phi'(\mathbf{r}, t_0 + \frac{T_P}{2}; t_0 + \frac{T_P}{2} - T_E)} \quad (2.38)$$

{eq:dess-mxyp-te2-ph}

{eq:dess-mxyp-te2}

$$\approx m'_{xy}(\mathbf{r}, t_0) e^{+T_E/T_2(\mathbf{r})} e^{+i\omega'(\mathbf{r})T_E}. \quad (2.39)$$

The received signal is approximately proportional to the integrated transverse magnetization over a volume  $\mathbb{V}$ . To derive expressions, we retake assumptions used in Subsection 2.1.2.1 and append an additional assumption on the full-repetition phase accrual  $\phi'(\mathbf{r})$ :

- We assume that the signal is localized to a scale over which there is off-resonance phase variation, but minimal variation of  $m_0(\mathbf{r})$ ,  $T_1(\mathbf{r})$ ,  $T_2(\mathbf{r})$ , and  $\alpha(\mathbf{r})$ . This assumption is reasonable<sup>6</sup> when describing the signal arising from a typical voxel.
- We assume that free precession off resonance frequency  $\omega'(\mathbf{r})$  is distributed over the localized voxel as  $p_{\omega'} \leftarrow \text{Cauchy}(\bar{\omega}', R'_2)$ , where  $\bar{\omega}'(\mathbf{r})$  is the median off-resonance frequency and  $R'_2(\mathbf{r})$  is the broadening bandwidth.
- We assume that the dephasing gradient imparts an integral number  $n_{\text{cyc}}$  of across-

<sup>5</sup>Observe that we do not attempt to express the magnetization prior to (the next) RF excitation by simply operating on the magnetization after (the current) RF excitation with further precession and relaxation. The reason is due to the intermediate dephasing gradient, which causes phase accrual in excess of off-resonance effects and thus forbids an approximation akin to (2.36).

<sup>6</sup>Model mismatch due to within-voxel spatial variation of relaxation parameters can be significant, especially for large voxels. Chapter 6 studies so-called partial volume effects and uses them for QMRI.

voxel phase cycles<sup>7</sup> such that full-repetition phase accrual  $\phi'(\mathbf{r})$  is distributed essentially uniformly as  $p_{\phi'} \leftarrow \text{uniform}(0, 2\pi n_{\text{cyc}}), n_{\text{cyc}} \in \{1, 2, 3, \dots\}$ .

With these assumptions, the received steady-state DESS (noiseless) signal models for a typically sized voxel centered at position  $\mathbf{r}$  are (to within constants):

$$\begin{aligned}
 \text{\{eq:dess-def-int\}} \quad s_{\text{D}}\left(\mathbf{r}, t_0 + \frac{T_{\text{P}}}{2} + T_{\text{E}}\right) &\propto \int_{\mathbb{V}(\mathbf{r})} m'_{xy}\left(\mathbf{r}, t_0 + \frac{T_{\text{P}}}{2} + T_{\text{E}}\right) d^3 \mathbf{r} \\
 &\approx \int_{\mathbb{R}} \int_{\mathbb{R}} m'_{xy}\left(\mathbf{r}, t_0 + \frac{T_{\text{P}}}{2} + T_{\text{E}}\right) p_{\phi'}(\phi') p_{\omega'}(\omega') d\phi' d\omega' \\
 &\approx e^{-T_{\text{E}}/T_2(\mathbf{r})} \int_{\mathbb{R}} m'_{xy}(\mathbf{r}, t_0 + T_{\text{P}}) p_{\phi'}(\phi') d\phi' \int_{\mathbb{R}} e^{-i\omega' T_{\text{E}}} p_{\omega'}(\omega') d\omega' \\
 &= +im_0(\mathbf{r}) E_2(\mathbf{r}, T_{\text{E}}) e^{-(R'_2(\mathbf{r}) - i\bar{\omega}'(\mathbf{r}))T_{\text{E}}} \tan \frac{\alpha(\mathbf{r})}{2} \left[1 - \frac{\eta(\mathbf{r}, T_{\text{R}})}{\xi(\mathbf{r}, T_{\text{R}})}\right]; \\
 \text{\{eq:dess-def-model\}} & \tag{2.41}
 \end{aligned}$$

$$\begin{aligned}
 \text{\{eq:dess-ref-int\}} \quad s_{\text{D}}\left(\mathbf{r}, t_0 + \frac{T_{\text{P}}}{2} - T_{\text{E}}\right) &\propto \int_{\mathbb{V}(\mathbf{r})} m'_{xy}\left(\mathbf{r}, t_0 + \frac{T_{\text{P}}}{2} - T_{\text{E}}\right) d^3 \mathbf{r} \\
 &\approx \int_{\mathbb{R}} \int_{\mathbb{R}} m'_{xy}\left(\mathbf{r}, t_0 + \frac{T_{\text{P}}}{2} - T_{\text{E}}\right) p_{\phi'}(\phi') p_{\omega'}(\omega') d\phi' d\omega' \\
 &\approx e^{+T_{\text{E}}/T_2(\mathbf{r})} \int_{\mathbb{R}} m'_{xy}(\mathbf{r}, t_0) p_{\phi'}(\phi') d\phi' \int_{\mathbb{R}} e^{+i\omega' T_{\text{E}}} p_{\omega'}(\omega') d\omega' \\
 &= -im_0(\mathbf{r}) E_2^{-1}(\mathbf{r}, T_{\text{E}}) e^{-(R'_2(\mathbf{r}) + i\bar{\omega}'(\mathbf{r}))T_{\text{E}}} \tan \frac{\alpha(\mathbf{r})}{2} [1 - \eta(\mathbf{r}, T_{\text{R}})], \\
 \text{\{eq:dess-ref-model\}} & \tag{2.43}
 \end{aligned}$$

where (2.41) and (2.43) introduce intermediate variables

$$\begin{aligned}
 \eta(\mathbf{r}, t) &:= \sqrt{\frac{1 - E_2^2(\mathbf{r}, t)}{1 - E_2^2(\mathbf{r}, t)/\xi^2(\mathbf{r}, t)}}; \\
 \xi(\mathbf{r}, t) &:= \frac{1 - E_1(\mathbf{r}, t) \cos \alpha(\mathbf{r})}{E_1(\mathbf{r}, t) - \cos \alpha(\mathbf{r})}.
 \end{aligned}$$

In steady-state, the DESS signal is typically greatest immediately following excitation and defocuses with rate  $\frac{1}{T_2} + R'_2$  until what we hereafter denote the *defocusing* echo time.

---

<sup>7</sup>In theory, it suffices to design dephasing gradients to impart as few as one complete cycle of net phase variation across a voxel. In practice, field inhomogeneities will induce spurious through-voxel field gradients that modify the effective dephasing gradient moment and thereby create partial phase cycles that distort the nominally uniform phase distribution. To reduce model mismatch due to such “partial spoiling” effects, dephasing gradients are usually designed to nominally impart multiple complete cycles of across-voxel phase variation. However, larger dephasing gradients cause greater DESS model mismatch due to diffusive signal loss. Appendix B studies diffusion in DESS and discusses regimes of dephasing gradient moments which balance partial-spoiling versus diffusive sources of model mismatch.

After a low-signal period between RF pulses, the DESS signal then refocuses with rate  $\frac{1}{T_2} - R'_2$  from what we hereafter denote the *refocusing* echo time until just prior the next excitation. Fortuitously, the defocusing (2.41) and refocusing (2.43) DESS signal models have significantly different dependence on relaxation parameters (especially  $T_2$ ) and thus together are quite useful for relaxation parameter estimation.

## 2.2 Optimization in QMRI

{s,bkgrd,opt}

This section overviews how optimization methods are leveraged in a substantial portion of this thesis to solve practical QMRI problems. For such problems, the central idea is to construct a suitable scalar cost function  $\Psi$  of some design variables  $\mathbf{x}$ , whose output  $\Psi(\mathbf{x}) \in \mathbb{R}$  is designed to provide a measure of the undesirability of  $\mathbf{x}$ . We then employ tailored optimization algorithms to find an  $\mathbf{x}$  that minimizes  $\Psi$  over a set  $\mathbb{X}$ , written as

{eq:opt-global}

$$\mathbf{x}^* \in \left\{ \arg \min_{\mathbf{x} \in \mathbb{X}} \Psi(\mathbf{x}) \right\}. \quad (2.44)$$

In either optimization-based parameter estimation (Chapter 3) or acquisition design (Chapter 4), we have reason to design  $\Psi$  to depend on corresponding design variables  $\mathbf{x}$  through MR signal models. Because these models are often (strongly) nonlinear functions of design variables, corresponding cost functions are usually non-convex in  $\mathbf{x}$  (though the search space  $\mathbb{X}$  is almost always assumed convex in this thesis). Thus, most QMRI problems in the form of (2.44) are non-convex optimization problems.

In general, solving (2.44) is more challenging when  $\Psi$  is non-convex in  $\mathbf{x}$  than otherwise, due in part to the possible presence of local extrema and/or saddle points. In the following, we discuss two strategies used in this thesis to cope with non-convex optimization. Subsection 2.2.1 relaxes (2.44) to instead seek a local minimizer via iterative methods. Subsection 2.2.2 restricts attention to signal models that are linear in a portion of  $\mathbf{x}$  and discusses a specific problem for which (2.44) simplifies for such partially linear structures.

### 2.2.1 Iterative Local Optimization with Constraints

{ss,bkgrd,opt,loc}

This subsection overviews a method for finding a local minimizer  $\hat{\mathbf{x}}$  of possibly non-convex cost function  $\Psi$  over convex constraint set  $\mathbb{X}$ . Such  $\hat{\mathbf{x}} \in \mathbb{X}$  must satisfy for some  $\delta > 0$

{eq:opt-local}

$$\Psi(\hat{\mathbf{x}}) \leq \Psi(\mathbf{x}) \quad \forall \mathbf{x} \in \mathbb{X} : \|\hat{\mathbf{x}} - \mathbf{x}\|_2 < \delta. \quad (2.45)$$

Observe that a global optimizer  $\mathbf{x}^*$  satisfies (2.45) for arbitrarily large  $\delta$ ; thus, any global minimizer is a local minimizer (but the converse is not necessarily true unless  $\Psi$  is convex).

As even locally optimal minimizers are often challenging to compute analytically, many algorithms construct  $\hat{\mathbf{x}}$  by iteratively updating an initial guess  $\mathbf{x}^{(0)}$  until some convergence criterion is satisfied. For a differentiable cost and convex constraints, the gradient projection method [18] is one such iterative algorithm and repeats the following simple update:

$$\{\text{eq:gpm}\} \quad \mathbf{x}^{(i)} \leftarrow P_{\mathbb{X}}(\mathbf{x}^{(i-1)} - \mathbf{\Pi} \nabla_{\mathbf{x}} \Psi(\mathbf{x}^{(i-1)})), \quad (2.46)$$

where  $P_{\mathbb{X}}$  denotes projection onto  $\mathbb{X}$  and  $\mathbf{\Pi}$  is a diagonal preconditioning matrix that permits elements of  $\mathbf{x}$  to take scale-informed step sizes along the negative gradient direction.

If  $\Psi$  is convex and sufficiently smooth, iterates produced via (2.46) converge to a limit point [19] that is a constrained global minimum (for appropriately selected  $\mathbf{\Pi}$ ). If instead  $\Psi$  is non-convex (but  $\mathbb{X}$  is still convex), statements regarding convergence<sup>8</sup> to a particular constrained local minimizer require additional (strong) assumptions regarding initialization and in general are still much weaker than in the convex case.

Since non-convex cost functions can have many local extrema (whose associated costs can vary dramatically), the utility of a locally optimal solution depends strongly on initialization quality. Accordingly, this thesis uses iterative local optimization for non-convex QMRI problems where a reasonable initialization is available and global optimization (to within quantization error) via exhaustive grid search is intractable.

## 2.2.2 Partially Linear Models and the Variable Projection Method

{ss,bkgrd,opt,vpm}

(Constrained, weighted) nonlinear least-squares is a specific non-convex optimization problem that is useful for many parameter estimation problems:

$$\{\text{eq:nonlin-ls}\} \quad \mathbf{x}^* \in \left\{ \arg \min_{\mathbf{x} \in \mathbb{X}} \|\mathbf{y} - \mathbf{f}(\mathbf{x})\|_{\mathbf{W}}^2 \right\}, \quad (2.47)$$

where  $\mathbf{f} : \mathbb{X} \mapsto \mathbb{C}^D$  is a nonlinear forward model that (barring noise) relates parameters  $\mathbf{x} \in \mathbb{X} \subseteq \mathbb{C}^L$  to data  $\mathbf{y} \in \mathbb{C}^D$ ; weighted 2-norm  $\|\boldsymbol{\iota}\|_{\mathbf{W}} := \sqrt{\boldsymbol{\iota}^H \mathbf{W} \boldsymbol{\iota}}$  for a symmetric, positive-semidefinite weighting matrix  $\mathbf{W} \in \mathbb{R}^{D \times D}$  and arbitrary vector  $\boldsymbol{\iota} \in \mathbb{C}^D$ ; and  $(\cdot)^H$  denotes conjugate transpose. The variable projection method [20] reduces the complexity

---

<sup>8</sup>For example, it suffices to assume that  $\mathbf{x}^{(0)}$  lies in the *attraction basin*  $\mathbb{B}_{\tilde{\mathbf{x}}}$  of a given unconstrained local minimum  $\tilde{\mathbf{x}}$ , where attraction basin is defined here as the largest convex set containing  $\tilde{\mathbf{x}}$  over which  $\Psi$  is convex. If  $\mathbb{B}_{\tilde{\mathbf{x}}} \cap \mathbb{X}$  is nonempty and step sizes within  $\mathbf{\Pi}$  are small enough to contain iterates within  $\mathbb{B}_{\tilde{\mathbf{x}}}$ , then iterates converge to the limit point  $P_{\mathbb{X}}(\tilde{\mathbf{x}})$ .



of (2.47) when the forward model takes the partially linear structure  $\mathbf{f}(\mathbf{x}) \equiv \mathbf{A}(\mathbf{x}_N)\mathbf{x}_L$  and the feasible set takes the partially unconstrained form  $\mathbb{X} \equiv \mathbb{C}^{L_L} \times \mathbb{X}_N$ , where  $\mathbf{x}_L \in \mathbb{C}^{L_L}$ ;  $\mathbf{x}_N \in \mathbb{X}_N$ ; and  $\mathbf{A} : \mathbb{X}_N \mapsto \mathbb{C}^{D \times L_L}$  is a matrix function. These restrictions on (2.47) define a so-called separable least-squares problem:

$$\{\text{eq:sep-ls}\} \quad (\mathbf{x}_L^*, \mathbf{x}_N^*) \in \left\{ \arg \min_{\substack{\mathbf{x}_L \in \mathbb{C}^{L_L} \\ \mathbf{x}_N \in \mathbb{X}_N}} \|\mathbf{y} - \mathbf{A}(\mathbf{x}_N)\mathbf{x}_L\|_{\mathbf{W}}^2 \right\}. \quad (2.48)$$

The variable projection method simplifies (2.48) by exploiting the partially linear structure of  $\mathbf{f}$  to explicitly express the optimal  $\mathbf{x}_L^*$  as a function of any fixed  $\mathbf{x}_N \in \mathbb{X}_N$ :

$$\begin{aligned} \mathbf{x}_L^*(\mathbf{x}_N) &= \arg \min_{\mathbf{x}_L \in \mathbb{C}^{L_L}} \|\mathbf{y} - \mathbf{A}(\mathbf{x}_N)\mathbf{x}_L\|_{\mathbf{W}}^2 \\ &= (\mathbf{W}^{1/2} \mathbf{A}(\mathbf{x}_N))^\dagger \mathbf{W}^{1/2} \mathbf{y} \end{aligned} \quad (2.49)$$

$$\begin{aligned} &= (\mathbf{A}^H(\mathbf{x}_N) \mathbf{W} \mathbf{A}(\mathbf{x}_N))^{-1} \mathbf{A}^H(\mathbf{x}_N) \mathbf{W} \mathbf{y}, \end{aligned} \quad (2.50)$$

$\{\text{eq:sep-ls-lin}\}$   
 $\{\text{eq:sep-ls-fullrnk}\}$

where  $(\cdot)^\dagger$  denotes pseudoinverse;  $\mathbf{W}^{1/2}$  denotes principle (matrix) square root; and (2.50) holds if the matrix inversion within exists. Substituting (2.50) into (2.48) yields a new non-convex optimization problem that contains  $L_L$  fewer unknowns than before:

$$\begin{aligned} \mathbf{x}_N^* &\in \left\{ \arg \min_{\mathbf{x}_N \in \mathbb{X}_N} \left\| \mathbf{y} - \mathbf{A}(\mathbf{x}_N) (\mathbf{A}^H(\mathbf{x}_N) \mathbf{W} \mathbf{A}(\mathbf{x}_N))^{-1} \mathbf{A}^H(\mathbf{x}_N) \mathbf{W} \mathbf{y} \right\|_{\mathbf{W}}^2 \right\} \\ &\equiv \left\{ \arg \max_{\mathbf{x}_N \in \mathbb{X}_N} \mathbf{y}^H \mathbf{W} \mathbf{A}(\mathbf{x}_N) (\mathbf{A}^H(\mathbf{x}_N) \mathbf{W} \mathbf{A}(\mathbf{x}_N))^{-1} \mathbf{A}^H(\mathbf{x}_N) \mathbf{W} \mathbf{y} \right\}, \end{aligned} \quad (2.51)$$

$\{\text{eq:sep-ls-nonlin}\}$

where the equivalence leading to (2.51) omits terms independent of  $\mathbf{x}_N$ .

In low-dimensional QMRI applications (*e.g.*, those discussed in Chapter 3), reduced problem (2.51) may be tractable via exhaustive grid search, in which case a global optimum  $(\mathbf{x}_L^*(\mathbf{x}_N^*), \mathbf{x}_N^*)$  is achievable to within quantization error. However, larger estimation problems involving more nonlinear parameters might still be tractable only via iterative optimization (see Subsection 2.2.1) towards a local solution  $(\hat{\mathbf{x}}_L(\hat{\mathbf{x}}_N), \hat{\mathbf{x}}_N)$ . For such higher-dimensional applications, Chapters 5-6 introduce novel methods that tackle problems similar to (2.47), while circumventing initialization-dependent local optimization.

## CHAPTER 3

# MRI Parameter Estimation from Likelihood Models

{c,relax}

### 3.1 Introduction

{s,relax,intro}

This chapter introduces methods for QMRI parameter estimation from statistical likelihood models and applies these methods to simple problems in  $T_1$  and  $T_2$  relaxometry, which are of interest for monitoring the progression of various disorders [21]. Section 3.2 introduces the notion of a QMRI scan profile, describes a signal model for parameter estimation, formulates several likelihood-based estimators using this model, and discuss practical implementation issues. Section 3.3 demonstrates the utility of likelihood-based parameter estimation over conventional methods through simulation, phantom, and *in vivo* experiments. Section 3.4 provides brief concluding remarks and suggests future directions.

### 3.2 Likelihood-Based Estimation in QMRI

{s,relax,method}

#### 3.2.1 The QMRI Scan Profile

{ss,relax,method,profile}

After image reconstruction, many MRI pulse sequences useful for parameter estimation produce at each voxel centered at position  $\mathbf{r}$  a set of noisy voxel values  $\{y_1(\mathbf{r}), \dots, y_D(\mathbf{r})\}$ , each of which can be described with the following general model:

{eq:relax,mod-scalar}

$$y_d(\mathbf{r}) = s_d(\mathbf{x}(\mathbf{r}); \boldsymbol{\nu}(\mathbf{r}), \mathbf{p}_d) + \epsilon_d(\mathbf{r}), \quad (3.1)$$

where  $d \in \{1, \dots, D\}$ . Here,  $\mathbf{x}(\mathbf{r}) \in \mathbb{C}^L$  collects  $L$  *latent* object parameters at  $\mathbf{r}$ ;  $\boldsymbol{\nu}(\mathbf{r}) \in \mathbb{C}^K$  collects  $K$  *known* object parameters at  $\mathbf{r}$ ;  $s_d : \mathbb{C}^L \times \mathbb{C}^K \times \mathbb{R}^A \mapsto \mathbb{C}$  is a (pulse-sequence dependent) function that models the noiseless signal obtained from the  $d$ th dataset

using *acquisition* parameter  $\mathbf{p}_d \in \mathbb{R}^A$ ; and  $\epsilon_d \sim \mathbb{CN}(0, \sigma_d^2)$  is assumed for simplicity<sup>1</sup> to be (circularly-symmetric) complex Gaussian noise [24, 25] with zero mean and variance  $\sigma_d^2$ . Colon positions in signal model (3.1) and similar expressions throughout this thesis distinguish unknown and known parameters. Concrete examples follow shortly.

For accurate, well-conditioned QMRI parameter estimation, it is typically necessary to acquire a collection of datasets, which we refer to hereafter as a *scan profile*. A scan profile consists of  $D$  datasets from up to  $D$  pulse sequences (some sequences yield more than one dataset, *e.g.* DESS). Let  $\mathbf{y}(\mathbf{r}) := [y_1(\mathbf{r}), \dots, y_D(\mathbf{r})]^\top \in \mathbb{C}^D$  collect noisy voxel values centered at  $\mathbf{r}$  from a given scan profile. Then the vector signal model

{eq:relax,mod-vec}

$$\mathbf{y}(\mathbf{r}) = \mathbf{s}(\mathbf{x}(\mathbf{r}); \boldsymbol{\nu}(\mathbf{r}), \mathbf{P}) + \boldsymbol{\epsilon}(\mathbf{r}) \quad (3.2)$$

helps define the noiseless signal  $\mathbf{s} := [s_1, \dots, s_D]^\top : \mathbb{C}^L \times \mathbb{C}^K \times \mathbb{R}^{A \times D} \mapsto \mathbb{C}^D$  and acquisition parameter  $\mathbf{P} := [\mathbf{p}_1, \dots, \mathbf{p}_D] \in \mathbb{R}^{A \times D}$  associated with that scan profile. Here, noise  $\boldsymbol{\epsilon}(\mathbf{r}) := [\epsilon_1(\mathbf{r}), \dots, \epsilon_D(\mathbf{r})]^\top \in \mathbb{C}^D$  typically has diagonal covariance structure  $\boldsymbol{\Sigma} := \text{diag}([\sigma_1, \dots, \sigma_D]^\top)$  due to independence across datasets, where  $\text{diag}(\cdot)$  assigns its argument to the diagonal entries of an otherwise zero (square) matrix.

The following subsections describe two concrete scan profiles whose signals can be modeled via (3.2) and that we study through experiments later in this chapter.

### 3.2.1.1 Example: An SPGR Scan Profile for $T_1$ estimation

We first consider the problem of  $T_1(\mathbf{r})$  estimation at  $\mathbf{r}$  from as few SPGR scans as possible, given a prior estimate of transmit field variation  $s^t(\mathbf{r})$  (see (2.7)). Examining SPGR model (2.27) makes clear that by fixing echo time  $T_E$  across scans, SPGR signal dependence is reduced to just two spatially varying latent parameters: desired parameter  $T_1(\mathbf{r}) \in \mathbb{R}$  and nuisance parameter  $c_1(\mathbf{r}) := im_0(\mathbf{r})e^{-T_E/T_2^*(\mathbf{r})}e^{-i\bar{\omega}'(\mathbf{r})T_E} \in \mathbb{C}$ . We assign  $\mathbf{x} \leftarrow [T_1, c_1]^\top$  and  $\boldsymbol{\nu} \leftarrow s^t$  for  $L \leftarrow 2$  latent and  $K \leftarrow 1$  known parameters, respectively.

With  $T_E$  fixed, prescribed flip angles  $\alpha_0$  and repetition times  $T_R$  are the only remaining  $P \leftarrow 2$  acquisition parameters available to choose that appear explicitly in (2.27). Thus, an SPGR scan profile useful for  $T_1$  estimation must vary  $\mathbf{p}_d \leftarrow [\alpha_0, T_R]^\top \forall d \in \{1, \dots, D\}$  over  $S_{\text{SPGR}}$  scan repetitions to produce  $D \geq L \leftarrow 2$  datasets for well-conditioned estimation.

<sup>1</sup>Though the noise distribution of  $\mathbf{k}$ -space raw data is usually well-modeled as complex white Gaussian, the noise distribution of the  $d$ th reconstructed image  $y_d$  depends both on the acquisition and reconstruction. If single receive channel  $\mathbf{k}$ -space data is fully-sampled on a Cartesian grid, each dataset  $y_d$  is recoverable via separate Fourier transform, and is thus complex Gaussian and independent across datasets. However if  $\mathbf{k}$ -space data is multi-channel, undersampled, and/or Cartesian, it may be preferable that  $y_d$  be estimated by more sophisticated techniques, *e.g.* [22, 23]. In such cases, reconstructed image noise is unlikely to be Gaussian-distributed.

### 3.2.1.2 Example: A DESS Scan Profile for $T_2$ estimation

We next consider the problem of  $T_2(\mathbf{r})$  estimation at  $\mathbf{r}$  from as few DESS scans as possible. Examining DESS models (2.41) and (2.43) makes clear that even with fixed  $T_E$  over possibly several acquisitions, there is signal dependence on five distinct object parameters:  $s^t(\mathbf{r}) \in \mathbb{R}$ ,  $T_1(\mathbf{r}) \in \mathbb{R}$ ,  $\bar{\omega}'(\mathbf{r}) \in \mathbb{R}$ ,  $c_2(\mathbf{r}) := m_0(\mathbf{r})e^{-T_E/T_2^*(\mathbf{r})} \in \mathbb{C}$ , and  $T_2(\mathbf{r}) \in \mathbb{R}$ . In this chapter, we take  $s^t(\mathbf{r}) \in \mathbb{R}$  and  $T_1(\mathbf{r}) \in \mathbb{R}$  as known for simplicity. To avoid (separate or joint)  $\bar{\omega}'(\mathbf{r})$  estimation, we choose to use magnitude DESS data, at the expense of slight model mismatch<sup>2</sup> due to Rician noise. These choices assign  $\boldsymbol{\nu} \leftarrow [s^t, T_1]^T$  as  $K \leftarrow 2$  known parameters and leave  $L \leftarrow 2$  latent parameters  $\mathbf{x} \leftarrow [c_2, T_2]^T$  to be estimated.

With  $T_E$  again fixed,  $\mathbf{p}_d \leftarrow [\alpha_0, T_R]^T \forall d \in \{1, \dots, D\}$  collects the remaining  $P \leftarrow 2$  tunable scan parameters that appear explicitly in (2.41) and (2.43). As in Example 3.2.1.1,  $D \geq L \leftarrow 2$  datasets are necessary for well-conditioned estimation. Unlike before however, a minimum  $D \leftarrow 2$  datasets need not require scan repetition, since  $S_{\text{DESS}}$  DESS scan repetitions produce  $D \leftarrow 2S_{\text{DESS}}$  datasets.

## 3.2.2 Latent Object Parameter Estimation

### 3.2.2.1 Signal Model and Problem Statement

A scan profile's reconstructed images can be modeled to discretize the bulk MR signal into  $V$  localized voxels centered at positions  $\mathbf{r}_1, \dots, \mathbf{r}_V$ :

$$\mathbf{Y} = \mathbf{S}(\mathbf{X}; \mathbf{N}, \mathbf{P}) + \mathbf{E}. \quad (3.3)$$

Here, signal model  $\mathbf{S} : \mathbb{C}^{L \times V} \times \mathbb{C}^{K \times V} \times \mathbb{R}^{A \times D} \mapsto \mathbb{C}^{D \times V}$  is a matrix function that maps latent  $\mathbf{X} := [\mathbf{x}(\mathbf{r}_1), \dots, \mathbf{x}(\mathbf{r}_V)] \in \mathbb{C}^{L \times V}$  and known  $\mathbf{N} := [\boldsymbol{\nu}(\mathbf{r}_1), \dots, \boldsymbol{\nu}(\mathbf{r}_V)] \in \mathbb{C}^{L \times V}$  parameter images (with fixed acquisition parameter  $\mathbf{P}$ ) to reconstructed image data  $\mathbf{Y} := [\mathbf{y}(\mathbf{r}_1), \dots, \mathbf{y}(\mathbf{r}_V)] \in \mathbb{C}^{D \times V}$ , save for noise image  $\mathbf{E} := [\boldsymbol{\epsilon}(\mathbf{r}_1), \dots, \boldsymbol{\epsilon}(\mathbf{r}_V)] \in \mathbb{C}^{D \times V}$ . The goal in QMRI parameter estimation is to estimate latent parameter images  $\mathbf{X}$  from MR image data  $\mathbf{Y}$ , for a fixed scan profile defined by  $\mathbf{S}$  and  $\mathbf{P}$  and given (separately acquired, estimated, and here assumed) known parameter images  $\mathbf{N}$ .

<sup>2</sup>The assumption of complex Gaussian noise in noisy MRI images implies that corresponding magnitude MRI images are Rician-distributed. However, the statistical estimators we will develop in Subsection 3.2.2 are based on Gaussian data. Fortunately, this source of model mismatch is negligible (less than 1%) for signal-to-noise ratio (SNR) in excess of 10 [26], and the acquisitions we examine here are capable of producing SNR in tissue of at minimum 100 and usually more.

### 3.2.2.2 Maximum Likelihood Methods

In maximum likelihood (ML) estimation, one seeks to find model parameters that maximize the likelihood of observing output data. We apply ML estimation to QMRI by first constructing a *likelihood function* that describes the probability of observing image data  $\mathbf{Y}$  given latent parameters  $\mathbf{X}$ . We then formulate ML latent parameter estimate  $\hat{\mathbf{X}}_{\text{ML}}(\mathbf{Y}; \mathbf{N}, \mathbf{P})$  by finding an  $\mathbf{X}$  that maximizes this likelihood function.

We first construct the likelihood function for the  $v$ th voxel's data  $\mathbf{y}(\mathbf{r}_v)$  and latent parameter  $\mathbf{x}(\mathbf{r}_v)$ . For complex Gaussian noise, the likelihood function is

$$\mathcal{L}(\mathbf{x}(\mathbf{r}_v)) \propto \exp \left( -\|\mathbf{y}(\mathbf{r}_v) - \mathbf{s}(\mathbf{x}(\mathbf{r}_v); \boldsymbol{\nu}(\mathbf{r}_v), \mathbf{P})\|_{\boldsymbol{\Sigma}^{-1}}^2 \right), \quad (3.4)$$

where (3.4) omits constants that are independent of  $\mathbf{x}(\mathbf{r}_v)$  and are therefore irrelevant. Assuming noise independence across image voxels, we can next build a simple and practical likelihood function of the full image data as

$$\mathcal{L}(\mathbf{X}) = \prod_{v=1}^V \mathcal{L}(\mathbf{x}(\mathbf{r}_v)). \quad (3.5)$$

We form an ML parameter estimate by finding  $\mathbf{X}$  that maximizes this likelihood function:

$$\begin{aligned} \hat{\mathbf{X}}_{\text{ML}}(\mathbf{Y}; \mathbf{N}, \mathbf{P}) &\in \left\{ \arg \max_{\mathbf{X} \in \mathbb{X}^V} \mathcal{L}(\mathbf{X}) \right\} \\ &\equiv \left\{ \arg \min_{\mathbf{X} \in \mathbb{X}^V} -\log \mathcal{L}(\mathbf{X}) \right\} \end{aligned} \quad (3.6)$$

$$\begin{aligned} &= \left\{ \arg \min_{\mathbf{X} \in \mathbb{X}^V} \sum_{v=1}^V \|\mathbf{y}(\mathbf{r}_v) - \mathbf{s}(\mathbf{x}(\mathbf{r}_v); \boldsymbol{\nu}(\mathbf{r}_v), \mathbf{P})\|_{\boldsymbol{\Sigma}^{-1}}^2 \right\} \\ &= \left\{ \arg \min_{\mathbf{X} \in \mathbb{X}^V} \left\| \boldsymbol{\Sigma}^{-1/2} (\mathbf{Y} - \mathbf{S}(\mathbf{X}; \mathbf{N}, \mathbf{P})) \right\|_{\text{F}}^2 \right\}, \end{aligned} \quad (3.7)$$

where  $\mathbb{X}$  is a (typically convex) latent parameter search space; the set equivalence in (3.6) uses the monotonicity of the log function; and  $\|\cdot\|_{\text{F}}$  denotes the Frobenius matrix norm.

Typically, QMR image model  $\mathbf{S}$  is nonlinear in  $\mathbf{X}$  and so ML estimation problem (3.7) involves non-convex optimization, which is challenging in general (see Section 2.2). Two properties of (3.7) guide our solution strategies. First, (3.7) is separable across voxels, so problem non-convexity is addressable on a voxel-by-voxel basis. Second, MR signal models are usually partially linear, in which case we may employ the variable projection method (described in Section 2.2.2) to further reduce problem complexity. For applications studied in this chapter, these properties allow for (3.7) to be solved via simple grid search.

### 3.2.2.3 Regularized Likelihood Methods

In regularized likelihood (RL) estimation, we modify ML estimation problem (3.6) to include additional information in the form of *regularization*:

$$\widehat{\mathbf{X}}_{\text{RL}}(\mathbf{Y}; \mathbf{N}, \mathbf{P}) \in \left\{ \arg \min_{\mathbf{X} \in \mathbb{X}^V} -\log \mathcal{L}(\mathbf{X}) + \mathbf{R}(\mathbf{X}) \right\}. \quad (3.8)$$

Here, we have freedom to design regularizer  $\mathbf{R} : \mathbb{C}^{L \times V} \mapsto \mathbb{R}$  to encourage desirable structure in estimates of  $\mathbf{X}$ . We observe that it is usually reasonable to assume that each latent object parameter map is *piecewise smooth* as a function of space: that is, each parameter is likely to vary smoothly in space, except for sharp discontinuities at tissue boundaries. To encourage piecewise-smoothness in parameter estimates, we use the regularizer

$$\mathbf{R}(\mathbf{X}) := \sum_{l=1}^L \beta_l \sum_{j=1}^J \phi_l \left( [\mathbf{J}\mathbf{X}^\top]_{jl} \right), \text{ where} \quad (3.9)$$

$$\phi_l(\cdot) := \gamma_l^2 \left( \sqrt{1 + |\cdot/\gamma_l|^2} - 1 \right) \quad (3.10)$$

is a differentiable approximation the absolute value function;  $\mathbf{J} \in \mathbb{R}^{J \times V}$  evaluates  $J$  (multi-dimensional) finite-differencing operations;  $[\cdot]_{jl}$  extracts the  $(j, l)$ th matrix element; and  $\beta_l$  is a regularization parameter that controls the relative importance of smoothing the  $l$ th latent object parameter image. Conceptually, this regularizer penalizes inconsistencies in adjacent latent parameter image voxels, but with a severity that depends on the degree of inconsistency. “Small” voxel-to-voxel differences are likely due to image data noise within a single tissue type and are penalized near-quadratically, while “large” differences are likely due to tissue boundaries and are penalized near-linearly. Useful notions of small versus large differences are governed by shape parameters  $\gamma_l \forall l \in \{1, \dots, L\}$ , and vary for different latent parameter maps based on their units and relative scale.

In general, QMRI image signal model  $\mathbf{S}$  is nonlinear in  $\mathbf{X}$  and so RL estimation problem (3.8) requires non-convex optimization. Unlike in ML estimation, (3.8) is not separable across voxels due to regularization, precluding global optimization (via grid search or other methods). We instead take the corresponding ML estimate as initialization and solve (3.8) via iterative constrained local optimization (detailed in Section 2.2.1).

### **3.3 Experimentation**

{s,relax,exp}

### **3.4 Summary**

{s,relax,summ}

## CHAPTER 4

# Optimizing MR Scan Design for Model-Based Relaxometry

{c,scn-dsgn}

### 4.1 Introduction

{s,scn-dsgn,intro}

Fast, accurate *relaxometry*, or quantification of spin-lattice and spin-spin relaxation parameters  $T_1$  and  $T_2$  has been of longstanding interest in MRI. Many researchers have suggested that  $T_1, T_2$  maps may serve as biomarkers for monitoring the progression of various disorders [21]. Neurological applications include: lesion classification in multiple sclerosis [27]; tumor characterization [28, 29]; and symptom onset prediction in stroke [30, 31]. In addition,  $T_1, T_2$  have shown promise for detecting hip and knee cartilage degeneration [32, 33] and for assessing cardiac dysfunction due to iron overload [34] or edema [35]. Motivated by this broad interest in  $T_1, T_2$  mapping, this chapter describes a systematic method to guide MR scan design.

Classical pulse sequences such as inversion/saturation recovery (IR/SR) or (single) spin echo (SE) yield relatively simple methods for  $T_1$  or  $T_2$  estimation, respectively; however, these methods require several scans, each with long repetition time  $T_R$ , leading to undesirably long acquisitions. Numerous modifications such as the Look-Locker method [36], multi-SE trains [37], or fast k-space trajectories [38, 39, 40] have been proposed to accelerate  $T_1$  [41, 42, 43, 44] and  $T_2$  [45, 46, 47, 48] relaxometry with these classical sequences. These techniques are more sensitive to model non-idealities [49, 50, 51], and are still speed-limited by the long  $T_R$  required for (near)-complete  $T_1$  recovery.

Steady-state (SS) pulse sequences [11, 12] permit short  $T_R$ , and are thus inherently much faster than classical counterparts. SS techniques are well-suited for relaxometry because the signals produced are highly sensitive to  $T_1$  and  $T_2$  variation. However, short  $T_R$  times also cause SS signals to be complex functions of both desired and undesired (*nuisance*) parameters, complicating quantification. Furthermore, some such methods [52, 53] still require scan repetition, though individual scans are now considerably shorter. Despite



these difficulties, the potential for rapid scanning with high  $T_1$ ,  $T_2$  sensitivity has motivated numerous SS relaxometry studies [54, 52, 53, 55, 56, 57, 58, 59, 60, 61].

The dual-echo steady-state (DESS) sequence [17] was recently proposed as a promising SS imaging technique for  $T_2$  estimation [58]. Because it produces two distinct signals per excitation, the DESS sequence can reduce scan repetition requirements by recording twice as much data per scan. As with other SS methods, the resulting signals [62, 63] are complicated functions of  $T_1$ ,  $T_2$ , and other parameters. Prior works have isolated  $T_2$  dependencies using either algebraic manipulations of the first- and second-echo signals [58, 59] or separate scans to first estimate nuisance parameters [2]. Although DESS concurrently encodes rich  $T_1$  and  $T_2$  information, these methods have shied away from using DESS for  $T_1$  estimation, either through bias-inducing approximations, or noise-propagating sequential estimation, respectively.

Whether it be with DESS, other sequences, or even combinations thereof, it is generally unclear how to best assemble a *scan profile* (*i.e.*, a collection of scans) for a fixed amount of scan time. Furthermore, for a given scan profile, it is typically not obvious how to best select acquisition parameters (*e.g.*, flip angles, repetition times, etc.) for relaxometry. In this chapter, the term *scan design* refers to the related problems of scan profile selection and scan parameter optimization.

Historically, scan design for relaxometry has predominantly been explored using figures of merit related to estimator precision. In particular, several studies have used the Cramér-Rao Bound (CRB), a statistical measure that bounds the minimum variance of an unbiased estimator. Earlier works have used the CRB and variations to select inversion times for recovery experiments [64, 65], flip angles for spoiled gradient-recalled echo (SPGR) sequences [66], and echo times for SE experiments [67]. More recent studies have considered additional scan design challenges, including scan time constraints [68], multiple latent parameters [69], multiple scan parameter types [70], and latent parameter spatial variation [71, 72].

The aforementioned studies consider scan parameter optimization for profiles consisting of *only one* pulse sequence. In contrast, this chapter introduces a general framework for robust, application-specific scan design for parameter estimation from *combinations* of pulse sequences. The framework first finds multiple sets of scan parameters that achieve precise estimation within a tight, *application-specific* range of object parameters (*e.g.*,  $T_1$ ,  $T_2$ , etc.). The framework then chooses the one scan parameter set most *robust* to estimator precision degradation over a broader range of object parameters. As a detailed example, we optimize three combinations of DESS and SPGR sequences for  $T_1$ ,  $T_2$  mapping. For a fixed total scan time, we find that well-chosen DESS scans alone can be used to es-

timate both  $T_1$  and  $T_2$  with precision and robustness comparable to combinations of DESS and SPGR. This example illustrates that, with careful scan profile design, well-established pulse sequences can find use in new estimation problems.

This chapter is organized as follows. Section ?? describes a CRB-inspired min-max optimization problem for robust, application-specific scan optimization. Section ?? optimizes three practical DESS/SPGR combinations to show that, even in the presence of radiofrequency (RF) field inhomogeneity, DESS is a promising option for  $T_1, T_2$  relaxometry. Section ?? describes simulation, phantom, and *in vivo* experiments and discusses corresponding results. Section ?? discusses practical challenges and suggests future directions. Section ?? summarizes key contributions.

## CHAPTER 5

# MRI Parameter Estimation via Kernel Regression

{c,krr}

## CHAPTER 6

# Myelin Water Fraction Estimation from Steady-State Sequences

{c,mwf}

## CHAPTER 7

# Steady-State RF Pulse Design

{c,ss-rf}

## CHAPTER 8

# Future Work

{c,future}

## APPENDIX A

### Coil Data Combination from Multiple Datasets

{a,cc-multi}

## **APPENDIX B**

### **DESS in the Presence of Diffusion**

{a, dess-diff}



## BIBLIOGRAPHY

- [1] G. Nataraj, J.-F. Nielsen, and J. A. Fessler, “Regularized, joint estimation of T1 and M0 maps,” in *Proc. Intl. Soc. Mag. Res. Med.*, p. 3128, 2014.
- [2] G. Nataraj, J.-F. Nielsen, and J. A. Fessler, “Model-based estimation of T2 maps with dual-echo steady-state MR imaging,” in *Proc. IEEE Intl. Conf. on Image Processing*, pp. 1877–81, 2014.
- [3] G. Nataraj, J.-F. Nielsen, and J. A. Fessler, “Optimizing MR scan design for model-based T1, T2 estimation from steady-state sequences,” *IEEE Trans. Med. Imag.*, 2016. To appear.
- [4] G. Nataraj, J.-F. Nielsen, and J. A. Fessler, “A min-max CRLB optimization approach to scan selection for relaxometry,” in *Proc. Intl. Soc. Mag. Res. Med.*, p. 1672, 2015.
- [5] G. Nataraj, J.-F. Nielsen, and J. A. Fessler, “Dictionary-free MRI parameter estimation via kernel ridge regression,” in *Proc. IEEE Intl. Symp. Biomed. Imag.*, 2017. Submitted.
- [6] G. Nataraj, J.-F. Nielsen, and J. A. Fessler, “Myelin water fraction estimation from optimized steady-state sequences using kernel ridge regression,” in *Proc. Intl. Soc. Mag. Res. Med.*, 2017. Submitted.
- [7] A. Macovski, *Medical imaging systems*. New Jersey: Prentice-Hall, 1983.
- [8] E. M. Haacke, R. W. Brown, M. R. Thompson, and R. Venkatesan, *Magnetic resonance imaging: Physical principles and sequence design*. New York: Wiley, 1999.
- [9] D. G. Nishimura, “Principles of magnetic resonance imaging,” 1996. Unpublished textbook.
- [10] F. Bloch, “Nuclear induction,” *Phys. Rev.*, vol. 70, pp. 460–74, Oct. 1946.
- [11] W. S. Hinshaw, “Image formation by nuclear magnetic resonance: The sensitive point method,” *J. Appl. Phys.*, vol. 47, p. 3709, Aug. 1976.
- [12] K. Scheffler, “A pictorial description of steady-states in rapid magnetic resonance imaging,” *Concepts in Magnetic Resonance*, vol. 11, no. 5, pp. 291–304, 1999.

- [13] B. A. Hargreaves, S. S. Vasanawala, J. M. Pauly, and D. G. Nishimura, "Characterization and reduction of the transient response in steady-state MR imaging," *Mag. Res. Med.*, vol. 46, pp. 149–58, July 2001.
- [14] Y. Zur, M. L. Wood, and L. J. Neuringer, "Spoiling of transverse magnetization in steady-state sequences," *Mag. Res. Med.*, vol. 21, pp. 251–63, Oct. 1991.
- [15] V. Denolin, C. Azizieh, and T. Metens, "New insights into the mechanisms of signal formation in rf-spoiled gradient echo sequences," *Mag. Res. Med.*, vol. 54, pp. 937–54, Oct. 2005.
- [16] T. Redpath and R. A. Jones, "Fade-a new fast imaging sequence," *Mag. Res. Med.*, vol. 6, pp. 224–34, Feb. 1988.
- [17] H. Bruder, H. Fischer, R. Graumann, and M. Deimling, "A new steady-state imaging sequence for simultaneous acquisition of two MR images with clearly different contrasts," *Mag. Res. Med.*, vol. 7, pp. 35–42, May 1988.
- [18] J. B. Rosen, "The gradient projection method for nonlinear programming, Part I: Linear constraints," *SIAM J. Appl. Math.*, vol. 8, no. 1, pp. 181–217, 1960.
- [19] C. Byrne, "A unified treatment of some iterative algorithms in signal processing and image reconstruction," *Inverse Prob.*, vol. 20, pp. 103–20, Feb. 2004.
- [20] G. Golub and V. Pereyra, "Separable nonlinear least squares: the variable projection method and its applications," *Inverse Prob.*, vol. 19, pp. R1–26, Apr. 2003.
- [21] H.-L. M. Cheng, N. Stikov, N. R. Ghugre, and G. A. Wright, "Practical medical applications of quantitative MR relaxometry," *J. Mag. Res. Im.*, vol. 36, pp. 805–24, Oct. 2012.
- [22] J. A. Fessler and B. P. Sutton, "Nonuniform fast Fourier transforms using min-max interpolation," *IEEE Trans. Sig. Proc.*, vol. 51, pp. 560–74, Feb. 2003.
- [23] M. J. Muckley, D. C. Noll, and J. A. Fessler, "Fast parallel MR image reconstruction via B1-based, adaptive restart, iterative soft thresholding algorithms (BARISTA)," *IEEE Trans. Med. Imag.*, vol. 34, pp. 578–88, Feb. 2015.
- [24] A. Macovski, "Noise in MRI," *Mag. Res. Med.*, vol. 36, pp. 494–7, Sept. 1996.
- [25] T. Lei, "Statistics of MR signals: revisited," in *Proc. SPIE 6510 Medical Imaging 2007: Phys. Med. Im.*, p. 651052, 2007.
- [26] H. Gudbjartsson and S. Patz, "The Rician distribution of noisy MRI data," *Mag. Res. Med.*, vol. 34, pp. 910–4, Dec. 1995.
- [27] H. B. W. Larsson, J. Frederiksen, L. Kjaer, O. Henriksen, and J. Olesen, "In vivo determination of T1 and T2 in the brain of patients with severe but stable multiple sclerosis," *Mag. Res. Med.*, vol. 7, pp. 43–55, May 1988.

- [28] T. Kurki, N. Lundbom, M. Komu, and M. Kormano, "Tissue characterization of inter cranial tumors by magnetization transfer and spin-lattice relaxation parameters in vivo," *J. Mag. Res. Im.*, vol. 6, pp. 573–9, Aug. 1996.
- [29] E. Englund, A. Brun, Z. Gyroffy-Wagner, E. Larsson, and B. Persson, "Relaxation times in relation to grade of malignancy and tissue necrosis in astrocytic gliomas," *Mag. Res. Im.*, vol. 4, no. 5, pp. 425–9, 1986.
- [30] S. Siemonsen, K. Mouridsen, B. Holst, T. Ries, J. Finsterbusch, G. Thomalia, L. Ostergaard, and J. Fiehler, "Quantitative T2 values predict time from symptom onset in acute stroke patients," *Stroke*, vol. 40, pp. 1612–6, May 2009.
- [31] L. D. DeWitt, J. P. Kistler, D. C. Miller, E. P. Richardson, and F. S. Buonanno, "NMR-neuropathologic correlation in stroke," *Stroke*, vol. 18, no. 2, pp. 342–51, 1987.
- [32] S. J. Matzat, J. V. Tiel, G. E. Gold, and E. H. G. Oei, "Quantitative MRI techniques of cartilage composition," *Quant. Imaging Med. Surg.*, vol. 3, pp. 162–74, June 2013.
- [33] T. J. Mosher and B. J. Dardzinski, "Cartilage MRI T2 relaxation time mapping: overview and applications," *Semin. Musculoskelet. Radiol.*, vol. 8, no. 4, pp. 355–68, 2004.
- [34] H. Guo, W.-Y. Au, J. S. Cheung, D. Kim, J. H. Jensen, P.-L. Khong, Q. Chan, K. C. Chan, C. Tosti, H. Tang, T. R. Brown, W. W. M. Lam, S.-Y. Ha, G. M. Brittenham, and E. X. Wu, "Myocardial T2 quantification in patients with iron overload at 3 Tesla," *J. Mag. Res. Im.*, vol. 30, pp. 394–400, Aug. 2009.
- [35] S. Giri, Y. C. Chung, A. Merchant, G. Mihai, S. Rajagopalan, S. V. Raman, and O. P. Simonetti, "T2 quantification for improved detection of myocardial edema," *Cardiovasc. Magn. Reson.*, vol. 11, no. 1, pp. 56–68, 2009.
- [36] D. C. Look and D. R. Locker, "Time saving in measurement of NMR and EPR relaxation times," *Rev Sci Instrum*, vol. 41, pp. 250–1, Feb. 1970.
- [37] H. Y. Carr and E. M. Purcell, "Effects of diffusion on free precession in nuclear magnetic resonance experiments," *Phys. Rev.*, vol. 94, pp. 630–8, May 1954.
- [38] M. K. Stehling, R. Turner, and P. Mansfield, "Echo-planar imaging: magnetic resonance imaging in a fraction of a second," *Science*, vol. 254, pp. 43–50, Oct. 1991.
- [39] C. B. Ahn, J. H. Kim, and Z. H. Cho, "High-speed spiral-scan echo planar NMR imaging - I," *IEEE Trans. Med. Imag.*, vol. 5, pp. 2–7, Mar. 1986.
- [40] C. H. Meyer, B. S. Hu, D. G. Nishimura, and A. Macovski, "Fast spiral coronary artery imaging," *Mag. Res. Med.*, vol. 28, pp. 202–13, Dec. 1992.
- [41] I. Kay and R. M. Henkelman, "Practical Implementation and Optimization of One-shot T1 imaging," *Mag. Res. Med.*, vol. 22, pp. 414–24, Dec. 1991.

- [42] P. A. Gowland and M. O. Leach, "Fast and accurate measurements of T1 using a multi-readout single inversion-recovery sequence," *Mag. Res. Med.*, vol. 26, pp. 79–88, July 1992.
- [43] D. R. Messroghli, A. Radjenovic, S. Kozerke, D. M. Higgins, M. U. Sivananthan, and J. P. Ridgway, "Modified Look-Locker inversion recovery (MOLLI) for high-resolution  $T_1$  mapping of the heart," *Mag. Res. Med.*, vol. 52, pp. 141–6, July 2004.
- [44] M. K. Stehling, R. J. Ordidge, R. Coxon, and P. Mansfield, "Inversion-recovery Echo-planar imaging (IR-EPI) at 0.5T," *Mag. Res. Med.*, vol. 13, pp. 514–7, Mar. 1990.
- [45] J.-M. Bonny, M. Zanca, J.-Y. Boire, and A. Veyre, "T2 maximum likelihood estimation from multiple spin-echo magnitude images," *Mag. Res. Med.*, vol. 36, pp. 287–93, Aug. 1996.
- [46] D. Kumar, T. D. Nguyen, S. A. Gauthier, and A. Raj, "Bayesian algorithm using spatial priors for multiexponential T2 relaxometry from multiecho spin echo MRI," *Mag. Res. Med.*, vol. 68, pp. 1536–43, Nov. 2012.
- [47] N. Ben-Eliezer, D. K. Sodickson, and K. T. Block, "Rapid and accurate T2 mapping from multi-spin-echo data using Bloch-simulation-based reconstruction," *Mag. Res. Med.*, vol. 73, pp. 809–17, Feb. 2015.
- [48] T. D. Nguyen, C. Wisnieff, M. A. Cooper, D. Kumar, A. Raj, P. Spincemaille, Y. Wang, T. Vartanian, and S. A. Gauthier, "T2prep three-dimensional spiral imaging with efficient whole brain coverage for myelin water quantification at 1.5 tesla," *Mag. Res. Med.*, vol. 67, pp. 614–21, Mar. 2012.
- [49] S. Majumdar, S. C. Orphanoudakis, A. Gmitro, M. O'Donnell, and J. C. Gore, "Error in the measurements of T2 using multiple-echo MRI techniques: 1. Effect of radiofrequency pulse imperfections," *Mag. Res. Med.*, vol. 3, pp. 397–417, June 1986.
- [50] S. Majumdar, S. C. Orphanoudakis, A. Gmitro, M. O'Donnell, and J. C. Gore, "Error in the measurements of T2 using multiple-echo MRI techniques: 2. Effects of static field inhomogeneity," *Mag. Res. Med.*, vol. 3, pp. 562–74, Aug. 1986.
- [51] F. Farzaneh, S. J. Riederer, and N. J. Pelc, "Analysis of T2 limitations and off-resonance effects on spatial resolution and artifacts in echo-planar imaging," *Mag. Res. Med.*, vol. 14, pp. 123–39, Apr. 1990.
- [52] S. C. L. Deoni, B. K. Rutt, and T. M. Peters, "Rapid combined T1 and T2 mapping using gradient recalled acquisition in the steady state," *Mag. Res. Med.*, vol. 49, pp. 515–26, Mar. 2003.
- [53] L.-C. Chang, C. G. Koay, P. J. Basser, and C. Pierpaoli, "Linear least-squares method for unbiased estimation of T1 from SPGR signals," *Mag. Res. Med.*, vol. 60, pp. 496–501, Aug. 2008.

- [54] E. K. Fram, R. J. Herfkens, G. A. Johnson, G. H. Glover, J. P. Kaaris, A. Shimakawa, T. G. Perkins, and N. J. Pelc, "Rapid calculation of T1 using variable flip angle gradient refocused imaging," *Mag. Res. Im.*, vol. 5, no. 3, pp. 201–8, 1987.
- [55] H. Wang and Y. Cao, "Spatially regularized T1 estimation from variable flip angles MRI," *Med. Phys.*, vol. 39, pp. 4139–48, July 2012.
- [56] S. C. L. Deoni, H. A. Ward, T. M. Peters, and B. K. Rutt, "Rapid  $T_2$  estimation with phase-cycled variable nutation steady-state free precession," *Mag. Res. Med.*, vol. 52, pp. 435–9, Aug. 2004.
- [57] S. C. L. Deoni, "Transverse relaxation time (T2) mapping in the brain with off-resonance correction using phase-cycled steady-state free precession imaging," *J. Mag. Res. Im.*, vol. 30, pp. 411–7, Aug. 2009.
- [58] G. H. Welsch, K. Scheffler, T. C. Mamisch, T. Hughes, S. Millington, M. Deimling, and S. Trattnig, "Rapid estimation of cartilage T2 based on double echo at steady state (DESS) with 3 Tesla," *Mag. Res. Med.*, vol. 62, pp. 544–9, Aug. 2009.
- [59] R. Heule, C. Ganter, and O. Bieri, "Rapid estimation of cartilage T2 with reduced T1 sensitivity using double echo steady state imaging," *Mag. Res. Med.*, vol. 71, pp. 1137–43, Mar. 2014.
- [60] T. Stöcker, F. Keil, K. Vahedipour, D. Brenner, E. Pracht, and N. J. Shah, "MR parameter quantification with magnetization-prepared double echo steady-state (MP-DESS)," *Mag. Res. Med.*, vol. 72, pp. 103–11, July 2014.
- [61] R. Heule, C. Ganter, and O. Bieri, "Triple echo steady-state (TESS) relaxometry," *Mag. Res. Med.*, vol. 71, pp. 230–7, Jan. 2014.
- [62] M. L. Gyngell, "The steady-state signals in short-repetition-time sequences," *J. Mag. Res.*, vol. 81, pp. 474–83, Feb. 1989.
- [63] W. Hänicke and H. U. Vogel, "An analytical solution for the SSFP signal in MRI," *Mag. Res. Med.*, vol. 49, pp. 771–5, Apr. 2003.
- [64] G. H. Weiss, R. K. Gupta, J. A. Ferretti, and E. D. Becker, "The choice of optimal parameters for measurement of spin-lattice relaxation times. I. Mathematical formulation," *J. Mag. Res.*, vol. 37, pp. 369–79, Feb. 1980.
- [65] Y. Zhang, H. N. Yeung, M. O'Donnell, and P. L. Carson, "Determination of sample time for T1 measurement," *J. Mag. Res. Im.*, vol. 8, pp. 675–81, May 1998.
- [66] H. Z. Wang, S. J. Riederer, and J. N. Lee, "Optimizing the precision in T1 relaxation estimation using limited flip angles," *Mag. Res. Med.*, vol. 5, pp. 399–416, Nov. 1987.
- [67] J. A. Jones, P. Hodgkinson, A. L. Barker, and P. J. Hore, "Optimal sampling strategies for the measurement of spin-spin relaxation times," *J. Mag. Res. B*, vol. 113, pp. 25–34, Oct. 1996.

- [68] J. Imran, François. Langevin, and Hervé. Saint-Jalmes, “Two-point method for T1 estimation with optimized gradient-echo sequence,” *Mag. Res. Im.*, vol. 17, pp. 1347–56, Nov. 1999.
- [69] S. C. L. Deoni, T. M. Peters, and B. K. Rutt, “Determination of optimal angles for variable nutation proton magnetic spin-lattice,  $T_1$ , and spin-spin,  $T_2$ , relaxation times measurement,” *Mag. Res. Med.*, vol. 51, pp. 194–9, Jan. 2004.
- [70] L. Fleysher, R. Fleysher, S. Liu, W. Zaaraoui, and O. Gonen, “Optimizing the precision-per-unit-time of quantitative MR metrics: Examples for  $T_1$ ,  $T_2$ , and DTI,” *Mag. Res. Med.*, vol. 57, pp. 380–7, Feb. 2007.
- [71] M. Akçakaya, S. Weingärtner, Sébastien. Roujol, and R. Nezafat, “On the selection of sampling points for myocardial T1 mapping,” *Mag. Res. Med.*, vol. 73, pp. 1741–53, May 2015.
- [72] C. M. Lewis, S. A. Hurley, M. E. Meyerand, and C. G. Koay, “Data-driven optimized flip angle selection for T1 estimation from spoiled gradient echo acquisitions,” *Mag. Res. Med.*, 2016. To appear.