

Advances in Quantitative MRI: Acquisition, Estimation, and Applications

by

Gopal Nataraj

A dissertation submitted in partial fulfillment
of the requirements for the degree of
Doctor of Philosophy
(Electrical Engineering and Computer Science)
in the University of Michigan
2016

Doctoral Committee:

Professor Jeffrey A. Fessler, Co-Chair
Assistant Research Scientist Jon-Fredrik Nielsen, Co-Chair
Professor Thomas Chenevert (?)
Professor Alfred O. Hero III (?)
Professor Douglas C. Noll
Associate Professor Clayton Scott
Associate Research Scientist Scott Swanson

©Gopal Nataraj

2016

TABLE OF CONTENTS

List of Figures	iii
List of Tables	v
List of Appendices	vi
List of Abbreviations	vii
Abstract	viii
Chapter	
1 Introduction	1
1.1 Thesis Overview	2
1.2 Thesis Organization	2
2 Background	4
2.1 Relevant MR Physics	4
2.1.1 Bloch Equations	4
2.1.2 Steady-State Sequences	8
2.2 Optimization in QMRI	15
2.2.1 Iterative Local Optimization with Constraints	15
2.2.2 Partially Linear Models and the Variable Projection Method	16
3 MRI Parameter Estimation from Likelihood Models	18
3.1 Introduction	18
3.2 Likelihood-Based Estimation in QMRI	18
3.2.1 The QMRI Scan Profile	18
3.2.2 Latent Object Parameter Estimation	20
3.3 Experimentation	23
3.4 Summary	23
4 Optimizing MR Scan Design for Model-Based Relaxometry	24
4.1 Introduction	24
4.2 A CRB-Inspired Scan Selection Method	26
4.2.1 The CRB and its Relevance to QMRI	26
4.2.2 Min-max Optimization Problem for Scan Design	27
4.3 Optimizing SS Sequences for Relaxometry in the Brain	28

4.3.1	Scan Design Details	28
4.3.2	Scan Profile Comparisons	30
4.4	Experimental Validation and Results	32
4.4.1	Numerical Simulations	33
4.4.2	Phantom Experiments	36
4.4.3	<i>In Vivo</i> Experiments	43
5	MRI Parameter Estimation via Kernel Regression	47
6	Myelin Water Fraction Estimation from Steady-State Sequences	48
7	Steady-State RF Pulse Design	49
8	Future Work	50
	Appendices	51
	Bibliography	53

LIST OF FIGURES

4.1	Worst-case standard deviations $\tilde{\sigma}_{T_1}^t$ (top), $\tilde{\sigma}_{T_2}^t$ (middle), and cost $\tilde{\Psi}^t$ (bottom), versus pairs of nominal flip angles, holding other scan parameters fixed at selected profile $\hat{\mathbf{P}}$. Subfigures (a)-(i), (j)-(l), and (m)-(o) correspond to scan profiles containing $(S_{\text{SPGR}}, S_{\text{DESS}}) = (2, 1), (1, 1), \text{ and } (0, 2)$ SPGR and DESS scans, respectively. Selected scan parameters (starred) are within $\delta = 1\%$ of global minimizers and retain as much estimator precision as possible over a wide range of latent object parameters. All axes range from 5 to 90 degrees, in 5-degree increments. Colorbar ranges are $[0, 100]$, $[0, 10]$, and $[0, 20]$ milliseconds for rows of $\tilde{\sigma}_{T_1}^t$, $\tilde{\sigma}_{T_2}^t$, and $\tilde{\Psi}^t$ subfigures, respectively. The optimized $(0, 2)$ profile appears most robust to flip angle variation.	31
4.2	Histograms of T_1 and T_2 estimates from noisy independent measurements of a <i>single</i> nominal WM or GM value. In each plot, two normal distributions are overlaid, each with latent means T_1 and T_2 . In (a)-(b) and (c)-(d), the solid green curve is $\mathcal{N}(T_1, (\tilde{\sigma}_{T_1}^t)^2)$ and $\mathcal{N}(T_2, (\tilde{\sigma}_{T_2}^t)^2)$, respectively. In (a)-(d), the dashed maroon curves have variances computed from the Fisher information at <i>a priori</i> unknown T_1, T_2 values in WM or GM. These plots correspond to an optimized $(0, 2)$ scan profile; analogous plots for other profiles are visually similar. At realistic noise levels, parameter estimates distribute with minimal bias and near-Gaussian shape. Thus, the CRB reliably approximates \hat{T}_1^{ML} and \hat{T}_2^{ML} errors.	35
4.3	Colorized T_1 and T_2 ML and RL estimates from an HPD [®] quantitative phantom. Columns correspond to scan profiles consisting of (2 SPGR, 1 DESS), (1 SPGR, 1 DESS), (0 SPGR, 2 DESS), and (4 IR, 4 SE) acquisitions. Rows distinguish T_1 and T_2 ML and RL estimators. Fig. 4.4 provides identical grayscale images that enumerate vials. Colorbar ranges are in milliseconds.	38
4.4	Grayscale T_1 and T_2 ML and RL estimates from an HPD [®] quantitative phantom. Columns correspond to scan profiles consisting of (2 SPGR, 1 DESS), (1 SPGR, 1 DESS), (0 SPGR, 2 DESS), and (4 IR, 4 SE) acquisitions. Rows distinguish T_1 and T_2 ML and RL estimators. Vials are enumerated and color-coded to correspond with data points in Fig. 4.5. Fig. 4.3 provides identical colorized images. Colorbar ranges are in milliseconds.	39

4.5	Phantom within-ROI sample statistics of T_1 and T_2 ML and RL estimates from optimized SPGR/DESS and reference IR/SE scan profiles, versus NIST NMR measurements [1]. Markers and error bars indicate ROI sample means and ROI sample standard deviations within the 14 labeled and color-coded vials in Fig. 4.4. Tight \mathbb{X}^t and broad \mathbb{X}^b latent parameter ranges are highlighted in orange and yellow, respectively. Table 4.3 replicates sample statistics within Vials 5-8. Our MR measurements are at 293K and NIST NMR measurements are at 293.00K. Within the designed parameter ranges, estimates from different acquisitions are in reasonable agreement with NIST measurements.	40
4.6	Colorized T_1 and T_2 ML and RL estimates from the brain of a healthy volunteer. Columns correspond to profiles consisting of (2 SPGR, 1 DESS), (1 SPGR, 1 DESS), (0 SPGR, 2 DESS), and (4 IR, 4 SE) acquisitions. Rows distinguish T_1 and T_2 ML and RL estimators. Colorbar ranges are in milliseconds.	44

LIST OF TABLES

4.1	Performance summary of different scan profiles, optimized by solving (4.9) subject to scan time constraint $T_{\max} = 41.9\text{ms}$. The first row defines each profile. The next four rows describe $\hat{\mathbf{P}}$. The latter three pairs of rows show how worst-case values degrade from tight to broad ranges. Flip angles are in degrees; all other values are in milliseconds.	32
4.2	Sample means \pm sample standard deviations of T_1 and T_2 ML estimates in WM and GM ROIs of simulated data, compared across different optimized scan profiles. Sample means exhibit insignificant bias, and sample standard deviations are consistent with worst-case standard deviations $\tilde{\sigma}_{T_1}^t$ and $\tilde{\sigma}_{T_2}^t$ reported in Table 4.1. All values are reported in milliseconds.	34
4.3	Phantom within-ROI sample means \pm sample standard deviations of T_1 and T_2 estimates from optimized SPGR/DESS and reference IR/SE scan profiles, versus NIST NMR measurements (<i>cf.</i> slide 22 of e-poster corresponding to [1]). For sake of brevity, sample statistics corresponding only to phantom vials within (or nearly within) tight design range \mathbb{X}^t (color-coded orange in Fig. 4.4) are reported. Fig. 4.5 plots sample statistics for all vials. ‘V#’ abbreviates vial numbers. All values are reported in milliseconds.	41
4.4	Phantom pooled sample standard deviations \pm pooled standard errors of sample standard deviations, from optimized SPGR/DESS scan profiles. Each entry is a measure of uncertainty of a typical voxel’s T_1 or T_2 ML estimate, estimated over 10 repeated acquisitions. For sake of brevity, sample statistics corresponding only to phantom vials within (or nearly within) tight design range \mathbb{X}^t (color-coded orange in Fig. 4.4) are reported. ‘V#’ abbreviates vial numbers. All values are reported in milliseconds.	42
4.5	<i>Left:</i> WM and GM ROIs, overlaid on a representative anatomical (coil-combined, IR) image. Separate WM ROIs are distinguished with anterior-right (AR), anterior-left (AL), posterior-right (PR), and posterior-left (PL) directions. Four small anterior (A) cortical GM polygons are pooled into a single ROI. <i>Right:</i> Within-ROI sample means \pm within-ROI sample standard deviations of T_1 and T_2 ML and RL estimates from the brain of a healthy volunteer. Sample statistics are computed within ROIs indicated in the anatomical image. All values are reported in milliseconds.	45

LIST OF APPENDICES

A Coil Data Combination from Multiple Datasets	51
B DESS in the Presence of Diffusion	52

LIST OF ABBREVIATIONS

ABSTRACT

Advances in Quantitative MRI: Acquisition, Estimation, and Applications

by

Gopal Nataraj

Co-Chairs: Jeffrey A. Fessler and Jon-Fredrik Nielsen

We show that it is possible to get approximate solutions to analytically intractable equations using iterative methods. Thus we show that the author could pass an undergraduate class in numerical analysis. In addition, a unique extension to Brent's method is proposed that results in slight improvements in convergence.

CHAPTER 1

Introduction

{c,intro}

Magnetic resonance imaging (MRI) is a non-invasive tool that has earned widespread clinical adoption due (among other factors) to its potential for excellent soft tissue contrast, its avoidance of ionizing radiation, and its flexibility to characterize many physical phenomena. Despite its numerous advantages, MRI requires highly specialized hardware, ongoing liquid-helium cooling of its superconducting main magnet, and comparably long scan times. For these reasons, MRI is (somewhat inherently) expensive relative to other medical imaging modalities. Accordingly, one broad initiative recently advocated by the MR community is to increase the *value* of MRI examinations.

Two reasonable measures of an acquisition’s value are its sensitivity to a given disorder and its specificity in distinguishing it from others. The field of *pathology* seeks to ascribe physical processes to disorders of interest with high sensitivity and specificity. The field of *quantitative MRI* (QMRI) seeks to build MRI biomarkers that measurably describe such physical processes and thereby provide indirect information about the onset and progression of underlying conditions.

QMRI poses several challenges beyond those of commonplace anatomical MRI and thus remains yet to be widely adopted clinically. For example, latent parameter “maps” that describe relevant physical processes are often related to the received MR signal through complicated, highly nonlinear relationships. Furthermore, practical MR pulse sequences produce signals that are usually functions of not only desired but also nuisance parameters. Scan repetition is often necessary for accurate estimation of multiple desired and nuisance parameters, which can increase scan times. Mitigating these challenges (and likely others) is essential to furthering widespread clinical adoption of QMRI techniques.

1.1 Thesis Overview

{s,intro,over}

In this thesis, we seek to build a systematic framework towards QMRI. We borrow tools from optimization, statistics, and machine learning to construct time-efficient workflows for quantifiably characterizing physical processes of interest. We apply this framework to challenging QMRI problems that are motivated by pathological studies. Our goal is to introduce tools that aid in identifying clinical tasks for which QMRI should (or should not) be part of a targeted, high-value MRI examination.

We consider two distinct subproblems in our framework. Questions in *acquisition design* (Chapters 4,6) ask how to assemble fast collections of scans that yield data rich in information about physical processes of interest. Questions in *parameter estimation* (Chapters 3,5) ask how to quickly and reliably quantify parameters associated with these relevant physical processes. The overall framework seeks to first design fast and informative scans based on the application, and to then accurately and precisely estimate application-specific parameters of interest.

1.2 Thesis Organization

{s,intro,org}

The main body of this thesis is organized as follows:

- Chapter 2 reviews relevant background material on MRI and optimization.
- Chapter 3 discusses methods for MRI parameter estimation from likelihood models and applies these methods for model-based MR relaxometry, (*i.e.*, estimation of relaxation parameters T_1, T_2), of interest for many neurological applications. It derives some content (especially regarding applications) from conference papers [2, 3].
- Chapter 4 introduces a minimax optimization approach to aid robust and application-specific MR scan selection and optimization for precise latent parameter estimation. It optimizes several practical acquisitions and uses the likelihood-based estimation techniques introduced in Chapter 3 to assess the utility of scan optimization through simulations, phantom studies, and *in vivo* experiments. It derives content mainly from journal paper [4] and conference paper [5].
- Chapter 5 describes MRI parameter estimation using kernel ridge regression. It derives content from conference paper [6].
- Chapter 6 introduces a multi-compartmental model for relevant MR pulse sequences and proposes a new acquisition useful for myelin water fraction estimation, of in-

terest in white matter disorders. It applies kernel-based MR parameter estimation to estimate myelin water fraction, in simulations and *in vivo* experiments. It derives some content from conference paper [7].

- Chapter 7 presents some relatively immature ideas on steady-state radiofrequency (RF) pulse design as well as associated challenges. This work is presently unpublished and may offer avenues for further research.
- Chapter 8 summarizes several items of possible future work (on both short- and long-term timescales) and presents a timeline for completion of this thesis.

The appendices are organized as follows:

- Appendix A proposes an algorithm for combining multiple MRI datasets (as is necessary for many parameter estimation problems), when each dataset is acquired using multiple receiver coils.
- Appendix B presents an analysis of model mismatch due to the presence of diffusion, shows that neglecting diffusive effects during T_2 estimation can cause significant bias, and suggests acquisition modifications for mitigating this bias.

CHAPTER 2

Background

{c,bkgrd}

This chapter focuses only on background information pertinent to multiple subsequent chapters. We present further topic-specific information at the beginnings of corresponding chapters. Section 2.1 places emphasis on reviewing necessary MR fundamentals, and Section 2.2 proceeds to a shorter discussion regarding optimization as it pertains to QMRI.

2.1 Relevant MR Physics

{s,bkgrd,mri}

This section begins with the fundamental Bloch equations and derives the signal models associated with two MR pulse sequences used extensively in this thesis. Our coverage of MRI is far from comprehensive, and omits fundamental but tangential topics such as signal localization. We refer the interested reader to books such as [8, 9, 10].

2.1.1 Bloch Equations

{ss,bkgrd,mri,bloch}

The Bloch equations [11] describe the magnetization dynamics of *spins*, or (loosely) atomic nuclei with nonzero angular momentum and thus nonzero magnetic moment, *e.g.* ^1H . If the dominant source of magnetic flux arises (as is typical in MRI) from a main magnetic field that is oriented along the z -axis, the equations read

{eq:bloch-mxy}
$$\frac{\partial}{\partial t} m_{xy}(\mathbf{r}, t) = i\gamma(m_z(\mathbf{r}, t)b_{xy}(\mathbf{r}, t) - m_{xy}(\mathbf{r}, t)b_z(\mathbf{r}, t)) - \frac{m_{xy}(\mathbf{r}, t)}{T_2(\mathbf{r})}; \quad (2.1)$$

{eq:bloch-mz}
$$\frac{\partial}{\partial t} m_z(\mathbf{r}, t) = \gamma(m_x(\mathbf{r}, t)b_y(\mathbf{r}, t) - m_y(\mathbf{r}, t)b_x(\mathbf{r}, t)) - \frac{m_z(\mathbf{r}, t) - m_0(\mathbf{r})}{T_1(\mathbf{r})}. \quad (2.2)$$

Here, $m_{xy}(\mathbf{r}, t) := m_x(\mathbf{r}, t) + im_y(\mathbf{r}, t) \in \mathbb{C}$ and $m_z(\mathbf{r}, t) \in \mathbb{R}$ are the transverse and longitudinal components of the magnetization vector at position $\mathbf{r} := [x, y, z]^T \in \mathbb{R}^3$ and time $t \geq 0$; $b_{xy}(\mathbf{r}, t) := b_x(\mathbf{r}, t) + ib_y(\mathbf{r}, t) \in \mathbb{C}$ and $b_z(\mathbf{r}, t) \in \mathbb{R}$ are the transverse and longitudinal components (in an inertial reference frame) of the applied magnetic field; $T_1(\mathbf{r})$

and $T_2(\mathbf{r})$ are spin-lattice and spin-spin relaxation time constants; $m_0(\mathbf{r})$ is the equilibrium magnetization and is proportional to the density of spins per unit volume as well as the main field strength; γ is the gyromagnetic ratio; and $i := \sqrt{-1}$. As written, equations (2.1)-(2.2) only model dominant temporal dynamics; later chapters consider second-order effects such as multiple magnetization compartments (Chapter 6) and diffusion (Appendix B).

It is often convenient to study Bloch dynamics in a non-inertial reference frame rotating clockwise about the z -axis at Larmor frequency $\omega_0 := \gamma B_0$, where $B_0 \hat{k}$ is the (nearly uniform) main magnetic field. In these coordinates, the apparent transverse magnetic field $b'_{xy}(\mathbf{r}, t) = b'_x(\mathbf{r}, t) + ib'_y(\mathbf{r}, t) := b_{xy}(\mathbf{r}, t)e^{i\omega_0 t}$ transforms only in phase, but the apparent longitudinal magnetic field $b'_z(\mathbf{r}, t) := b_z(\mathbf{r}, t) - B_0$ is greatly reduced in magnitude. The magnetization components transform more simply as $m'_{xy}(\mathbf{r}, t) = m'_{xy}(\mathbf{r}, t) + im'_y(\mathbf{r}, t) := m_{xy}(\mathbf{r}, t)e^{i\omega_0 t}$ and $m'_z(\mathbf{r}, t) := m_z(\mathbf{r}, t)$. Remarkably, inserting these coordinate transformations into (2.1)-(2.2) does not change the form of the dynamical equations:

$$\{\text{eq:bloch-mxyp}\} \quad \frac{\partial}{\partial t} m'_{xy}(\mathbf{r}, t) = i\gamma(m'_z(\mathbf{r}, t)b'_{xy}(\mathbf{r}, t) - m'_{xy}(\mathbf{r}, t)b'_z(\mathbf{r}, t)) - \frac{m'_{xy}(\mathbf{r}, t)}{T_2(\mathbf{r})}; \quad (2.3)$$

$$\{\text{eq:bloch-mzp}\} \quad \frac{\partial}{\partial t} m'_z(\mathbf{r}, t) = \gamma(m'_x(\mathbf{r}, t)b'_y(\mathbf{r}, t) - m'_y(\mathbf{r}, t)b'_x(\mathbf{r}, t)) - \frac{m'_z(\mathbf{r}, t) - m_0(\mathbf{r})}{T_1(\mathbf{r})}. \quad (2.4)$$

It thus suffices to consider how perturbations $\mathbf{b}'(\mathbf{r}, t)$ to main field $B_0 \hat{k}$ influence rotating-frame magnetization $\mathbf{m}'(\mathbf{r}, t)$ via Eqs. (2.3)-(2.4). The inertial-frame magnetization $\mathbf{m}(\mathbf{r}, t)$ is then easily constructed via $m_{xy}(\mathbf{r}, t) = m'_{xy}(\mathbf{r}, t)e^{-i\omega_0 t}$ and $m_z(\mathbf{r}, t) = m'_z(\mathbf{r}, t)$.

It is challenging to explicitly solve Eqs. (2.3)-(2.4) for arbitrary field perturbations $\mathbf{b}'(\mathbf{r}, t)$. We discuss relevant special cases in the following.

2.1.1.1 Non-Selective Excitation

Here, we derive solutions to Eqs. (2.3)-(2.4) in the case of short, spatially non-selective excitations. We take the following common assumptions:

- We assume negligible spatial variation in the longitudinal magnetic field, so $b'_z(\mathbf{r}, t) \approx 0$. This lack of spatial variation is reason for non-selective excitation.
- We assume the transverse field separates in position and time; oscillates at the Larmor frequency (commonly in the radiofrequency (RF) range); and aligns at initial time $t \leftarrow t_0$ with the x -axis. Together, these assumptions restrict the so-called RF excitation to take form $b'_{xy}(\mathbf{r}, t) \approx s^t(\mathbf{r})b'_{1,x}(t)\hat{i} + 0\hat{j}$, where $s^t(\mathbf{r}) \in \mathbb{R}$ is the RF transmit coil spatial variation and $b'_{1,x}(t) \in \mathbb{R}$ is the RF excitation envelope.

- We assume that the duration T_P of RF excitation (often $T_P \sim 1\text{ms}$) is much shorter than relaxation time constants (typically $T_1 \sim 1000\text{ms}$ and $T_2 \sim 50\text{ms}$ in brain tissue) and thus neglect relaxation effects during excitation.

Under these assumptions, Eqs. (2.3)-(2.4) reduce to the linear system

$$\frac{\partial}{\partial t} \begin{bmatrix} m'_x(\mathbf{r}, t) \\ m'_y(\mathbf{r}, t) \\ m'_z(\mathbf{r}, t) \end{bmatrix} = \begin{bmatrix} 0 & 0 & 0 \\ 0 & 0 & \gamma s^t(\mathbf{r}) b'_{1,x}(t) \\ 0 & -\gamma s^t(\mathbf{r}) b'_{1,x}(t) & 0 \end{bmatrix} \begin{bmatrix} m'_x(\mathbf{r}, t) \\ m'_y(\mathbf{r}, t) \\ m'_z(\mathbf{r}, t) \end{bmatrix}. \quad (2.5)$$

Eq. (2.5) admits the simple solution (for $t \geq t_0$)

$$\begin{bmatrix} m'_x(\mathbf{r}, t) \\ m'_y(\mathbf{r}, t) \\ m'_z(\mathbf{r}, t) \end{bmatrix} = \begin{bmatrix} 1 & 0 & 0 \\ 0 & \cos(\alpha(\mathbf{r}, t; t_0)) & \sin(\alpha(\mathbf{r}, t; t_0)) \\ 0 & -\sin(\alpha(\mathbf{r}, t; t_0)) & \cos(\alpha(\mathbf{r}, t; t_0)) \end{bmatrix} \begin{bmatrix} m'_x(\mathbf{r}, t_0) \\ m'_y(\mathbf{r}, t_0) \\ m'_z(\mathbf{r}, t_0) \end{bmatrix}, \quad (2.6)$$

where $\mathbf{m}'(\mathbf{r}, t_0) := [m'_x(\mathbf{r}, t_0), m'_y(\mathbf{r}, t_0), m'_z(\mathbf{r}, t_0)]^T$ is the initial magnetization and

$$\alpha(\mathbf{r}, t; t_0) := \gamma s^t(\mathbf{r}) \int_{t_0}^t b'_{1,x}(\tau) d\tau \quad (2.7)$$

is the nutation (or “flip”) angle at time t . Eq. (2.6) reveals that on-resonance RF excitation causes the magnetization vector to rotate clockwise about an axis parallel to the direction of excitation. The nutation angle accumulated over an RF pulse of duration T_P is often decomposed as $\alpha(\mathbf{r}, t_0 + T_P; t_0) =: \alpha_0 s^t(\mathbf{r})$, where α_0 is a prescribed nominal flip angle.

For deriving signal models in later sections, it is convenient and intuitive to define matrix operators that summarize relevant dynamics. Here, we rewrite Eq. (2.6) as

$$\mathbf{m}'(\mathbf{r}, t) = \mathbf{R}_{x'}(\alpha(\mathbf{r}, t; t_0)) \mathbf{m}'(\mathbf{r}, t_0), \quad (2.8)$$

where $\mathbf{R}_{x'}(\alpha(\mathbf{r}, t; t_0))$ denotes a clockwise rotation of angle $\alpha(\mathbf{r}, t; t_0)$ about the x' -axis.

2.1.1.2 Free Precession and Relaxation

Next, we derive solutions to the rotating-frame Bloch equations when no RF excitation is present, *i.e.* $b'_{xy}(\mathbf{r}, t) \approx 0$. In this case, Eqs. (2.3)-(2.4) decouple, yielding separate

dynamical equations for the transverse and longitudinal magnetization components:

$$\{\text{eq:bloch-free-mxyp}\} \quad \frac{\partial}{\partial t} m'_{xy}(\mathbf{r}, t) = -i\gamma m'_{xy}(\mathbf{r}, t) b'_z(\mathbf{r}, t) - \frac{m'_{xy}(\mathbf{r}, t)}{T_2(\mathbf{r})}; \quad (2.9)$$

$$\{\text{eq:bloch-free-mzp}\} \quad \frac{\partial}{\partial t} m'_z(\mathbf{r}, t) = -\frac{m'_z(\mathbf{r}, t) - m_0(\mathbf{r})}{T_1(\mathbf{r})}. \quad (2.10)$$

Eqs. (2.9)-(2.10) admit simple solutions with no further assumptions:

$$\{\text{eq:mxy-fp}\} \quad m'_{xy}(\mathbf{r}, t) = m'_{xy}(\mathbf{r}, t_0) e^{-(t-t_0)/T_2(\mathbf{r})} e^{-i\phi'(\mathbf{r}, t; t_0)}; \quad (2.11)$$

$$\{\text{eq:mz-fp}\} \quad m'_z(\mathbf{r}, t) = m'_z(\mathbf{r}, t_0) e^{-(t-t_0)/T_1(\mathbf{r})} + m_0(\mathbf{r}) (1 - e^{-(t-t_0)/T_1(\mathbf{r})}), \quad (2.12)$$

where $m'_{xy}(\mathbf{r}, t_0)$ and $m'_z(\mathbf{r}, t_0)$ are the initial magnetization components and

$$\{\text{eq:ph-def}\} \quad \phi'(\mathbf{r}, t; t_0) := \gamma \int_{t_0}^t b'_z(\mathbf{r}, \tau) d\tau \quad (2.13)$$

denotes the phase accumulation due to main field inhomogeneity (often called off-resonance effects). Eq. (2.11) reveals that without RF excitations, the transverse magnetization $m'_{xy}(\mathbf{r}, t)$ relaxes to zero exponentially fast with time constant $T_2(\mathbf{r})$, while accruing phase due to off-resonance effects. Eq. (2.12) similarly reveals that without RF excitations, longitudinal magnetization $m'_z(\mathbf{r}, t)$ recovers to $m_0(\mathbf{r})$ exponentially fast with time constant $T_1(\mathbf{r})$.

As in Section 2.1.1.2, we rewrite Eqs. (2.11)-(2.12) for $t \geq t_0$ using matrix operators:

$$\{\text{eq:mtx-pr}\} \quad \mathbf{m}'(\mathbf{r}, t) = \mathbf{R}_{z'}(\phi'(\mathbf{r}, t; t_0)) \mathbf{E}(\mathbf{r}, t; t_0) \mathbf{m}'(\mathbf{r}, t_0) + \mathbf{m}_0(\mathbf{r}, t; t_0) \quad (2.14)$$

where $\mathbf{m}_0(\mathbf{r}, t; t_0) := m_0(\mathbf{r}) (1 - e^{-(t-t_0)/T_1(\mathbf{r})}) \hat{k}$;

$$\{\text{eq:op-rotz}\} \quad \mathbf{R}_{z'}(\phi'(\mathbf{r}, t; t_0)) := \begin{bmatrix} \cos(\phi'(\mathbf{r}, t; t_0)) & \sin(\phi'(\mathbf{r}, t; t_0)) & 0 \\ -\sin(\phi'(\mathbf{r}, t; t_0)) & \cos(\phi'(\mathbf{r}, t; t_0)) & 0 \\ 0 & 0 & 1 \end{bmatrix} \quad (2.15)$$

denotes a clockwise rotation of angle $\phi'(\mathbf{r}, t; t_0)$ about the z' -axis; and

$$\{\text{eq:op-relax}\} \quad \mathbf{E}(\mathbf{r}, t; t_0) := \begin{bmatrix} e^{-(t-t_0)/T_2(\mathbf{r})} & 0 & 0 \\ 0 & e^{-(t-t_0)/T_2(\mathbf{r})} & 0 \\ 0 & 0 & e^{-(t-t_0)/T_1(\mathbf{r})} \end{bmatrix} \quad (2.16)$$

is an exponential relaxation operator. Section 2.1.2 (and later chapters) use matrix dynamical representations (2.8) and (2.14) to succinctly describe pulse sequence signal models.

{ss,bkgrd,mri,ss}

2.1.2 Steady-State Sequences

MRI experiments typically involve repeated cycles of (pulsed) RF excitation; signal localization (not discussed here); and transverse T_2 relaxation and free precession, alongside (relatively slow) longitudinal T_1 recovery. We can build models of the received MR signal by considering the magnetization dynamics induced by specific pulse sequences.

Classical pulse sequences use relatively long cycle repetition times T_R to ensure near-complete T_1 recovery of the magnetization vector back to equilibrium state $m_0(\mathbf{r})\hat{k}$ prior to the start of each RF cycle. For such long- T_R sequences, it suffices to approximate the magnetization as fully recovered (*i.e.*, $\mathbf{m}'(\mathbf{r}, t_0 + rT_R) \approx m_0(\mathbf{r})\hat{k}, \forall r \in \{0, 1, 2, \dots\}$) just prior to each RF excitation. This approximation yields a sequence of initial conditions and allows computation of the magnetization at corresponding times of data acquisition via direct application of Bloch dynamics (2.8) and (2.14). Resulting signal models are typically simple expressions of relaxation parameters $T_1(\mathbf{r})$ and $T_2(\mathbf{r})$; however, model accuracy often depends strongly on the long- T_R assumption, which requires long acquisitions.

Steady-state (SS) sequences [12] utilize short T_R , and can thus achieve much faster scan times. Due to short repetition times, SS sequences achieve only partial T_1 recovery in between RF excitations; thus, their magnetization responses do not obey the simple classical initial conditions (for the second RF cycle onwards). Although their transient magnetization dynamics can be complicated, SS sequences produce (under certain assumptions [13]) long-time magnetization responses that eventually¹ achieve a steady-state condition:

{eq:ss-cond}

$$\lim_{t_0 \rightarrow \infty} \mathbf{m}'(\mathbf{r}, t_0 + rT_R) = \mathbf{m}'(\mathbf{r}, t_0), \quad (2.17)$$

where repetition count $r \in \{1, 2, \dots\}$ for fixed RF excitations and off-resonance induced phase increments (as is assumed in the following). Subsections 2.1.2.1 and 2.1.2.2 use SS condition (2.17) and Bloch equation matrix operators introduced in (2.8) and (2.14) to derive long-time signal models for Spoiled Gradient-Recalled Echo (SPGR) and Dual-Echo Steady-State (DESS), two SS pulse sequences useful for quantitative MRI.

2.1.2.1 Spoiled Gradient-Recalled Echo (SPGR) Sequence

SPGR [15] is a fast pulse sequence that repeats cycles of fixed RF excitation (such that $b'_{1,x}(t + rT_R) = b'_{1,x}(t), \forall t \in [t_0, t_0 + T_P], r \in \{1, 2, \dots\}$); data acquisition; relaxation and recovery; and residual transverse magnetization “spoiling” (discussed later). Here we

¹The progression to steady state takes on the order of $5T_2/T_R$ RF cycles [13], typically a small but not insignificant period during which data acquisition is often foregone. This transition can (in some cases) be accelerated by prepending SS sequences with tailored “magnetization-catalyzing” modules [14].

{sss,bkgrd,mri,ss,spgr}

develop a simple and popular steady-state SPGR signal model.

Let $\mathbf{m}'(\mathbf{r}, t_0)$ denote the magnetization at an initial time t_0 selected well into the steady-state and just prior to excitation. The SPGR sequence first applies an RF excitation, which rotates the initial magnetization as per (2.8):

$$\{\text{eq:spgr-ex}\} \quad \mathbf{m}'(\mathbf{r}, t_0 + T_P) = \mathbf{R}_{x'}(\alpha(\mathbf{r}, t_0 + T_P; t_0))\mathbf{m}'(\mathbf{r}, t_0). \quad (2.18)$$

The excited magnetization then precesses and relaxes as per (2.14) until data acquisition, defined to occur at “echo time” $T_E \in [T_P, T_R]$ after the (midpoint of) RF excitation:

$$\{\text{eq:spgr-daq}\} \quad \begin{aligned} \mathbf{m}'\left(\mathbf{r}, t_0 + \frac{T_P}{2} + T_E\right) &= \mathbf{R}_{z'}\left(\phi'\left(\mathbf{r}, \frac{T_P}{2} + T_E; T_P\right)\right)\mathbf{E}\left(\mathbf{r}, \frac{T_P}{2} + T_E; T_P\right)\mathbf{m}'(\mathbf{r}, t_0 + T_P) \\ &+ \mathbf{m}_0\left(\mathbf{r}, \frac{T_P}{2} + T_E; T_P\right). \end{aligned} \quad (2.19)$$

Following signal reception, the remaining transverse magnetization is spoiled² while the longitudinal component is unaffected. We model an ideal spoiling operation as

$$\{\text{eq:spgr-spoil}\} \quad \mathbf{S}\mathbf{m}'\left(\mathbf{r}, \frac{T_P}{2} + T_E\right), \text{ where } \mathbf{S} := \begin{bmatrix} 0 & 0 & 0 \\ 0 & 0 & 0 \\ 0 & 0 & 1 \end{bmatrix}. \quad (2.20)$$

After spoiling, the longitudinal magnetization (partially) recovers until $t \leftarrow t_0 + T_R$:

$$\{\text{eq:spgr-pr}\} \quad \begin{aligned} \mathbf{m}'(\mathbf{r}, t_0 + T_R) &= \mathbf{R}_{z'}\left(\phi'\left(\mathbf{r}, T_R; \frac{T_P}{2} + T_E\right)\right)\mathbf{E}\left(\mathbf{r}, T_R; \frac{T_P}{2} + T_E\right)\mathbf{S}\mathbf{m}'\left(\mathbf{r}, t_0 + \frac{T_P}{2} + T_E\right) \\ &+ \mathbf{m}_0\left(\mathbf{r}, T_R; \frac{T_P}{2} + T_E\right). \end{aligned} \quad (2.21)$$

In steady-state, one cycle of excitation, acquisition, spoiling, and recovery returns the magnetization back to its initial state. We enforce this through the steady-state condition

$$\{\text{eq:spgr-ss}\} \quad \mathbf{m}'(\mathbf{r}, t_0 + T_P) = \mathbf{R}_{x'}(\alpha(\mathbf{r}, t_0 + T_P; t_0))\mathbf{m}'(\mathbf{r}, t_0 + T_R) \quad (2.22)$$

²Transverse signal spoiling is often (nearly) achieved in practice using strong induced field inhomogeneities (which cause rapid transverse signal dephasing) in tandem with RF excitations that additionally impart nonlinear (often quadratically increasing) transverse magnetization phase [15]. Though the nonlinear RF phase used in so-called “RF-spoiling” prevents any one spin from reaching a true steady-state, the signal integrated over a typically-sized voxel achieves SS-like behavior [16].

which yields an algebraic system of equations. When it exists, the solution is

$$\mathbf{m}'(\mathbf{r}, t_0 + T_P) = \frac{1}{1 - e^{-(T_R - T_P)/T_1(\mathbf{r})} \cos(\alpha(\mathbf{r}))} \begin{bmatrix} 0 \\ m_0(\mathbf{r}) \sin(\alpha(\mathbf{r})) (1 - e^{-(T_R - T_P)/T_1(\mathbf{r})}) \\ m_0(\mathbf{r}) \cos(\alpha(\mathbf{r})) (1 - e^{-(T_R - T_P)/T_1(\mathbf{r})}) \end{bmatrix}, \quad (2.23)$$

{eq:spgr-bmmp-t0}

where $\alpha(\mathbf{r}) := \alpha(\mathbf{r}, t_0 + T_P; t_0)$ is a slight abuse of notation. Remarkably, the SPGR steady-state magnetization immediately after excitation is approximately independent of both off-resonance effects and $T_2(\mathbf{r})$. Researchers more often cite the expression

$$\begin{aligned} m'_{xy}(\mathbf{r}, t_0 + T_P) &= m'_x(\mathbf{r}, t_0 + T_P) + im'_y(\mathbf{r}, t_0 + T_P) \\ &= \frac{im_0(\mathbf{r}) \sin(\alpha(\mathbf{r})) (1 - e^{-T_R/T_1(\mathbf{r})})}{1 - e^{-T_R/T_1(\mathbf{r})} \cos(\alpha(\mathbf{r}))} \end{aligned} \quad (2.24)$$

{eq:spgr-mxyp-t0}

for the complex transverse magnetization as it modifies (2.23) to include a simple first-order correction for unaccounted T_1 recovery during the RF pulse. Substituting (2.24) into (2.19) yields an expression for the transverse magnetization at the echo time:

$$\begin{aligned} m'_{xy}\left(\mathbf{r}, t_0 + \frac{T_P}{2} + T_E\right) &= m'_{xy}(\mathbf{r}, t_0 + T_P) e^{-(T_E - T_P/2)/T_2(\mathbf{r})} e^{-i\phi'(\mathbf{r}, t_0 + \frac{T_P}{2} + T_E; t_0 + T_P)} \\ &\approx m'_{xy}(\mathbf{r}, t_0 + T_P) e^{-T_E/T_2(\mathbf{r})} e^{-i\phi'(\mathbf{r}, t_0 + \frac{T_P}{2} + T_E; t_0 + \frac{T_P}{2})}, \end{aligned} \quad (2.25)$$

{eq:spgr-mxyp-te}

where the approximation again keeps in line with literature expressions.

The received signal is approximately proportional to the integrated transverse magnetization over a volume \mathbb{V} . To derive expressions, we take a few more usual assumptions:

- We assume that the signal is localized to a scale over which there is off-resonance phase variation, but minimal variation of $m_0(\mathbf{r})$, $T_1(\mathbf{r})$, $T_2(\mathbf{r})$, and $\alpha(\mathbf{r})$. This assumption is reasonable³ when describing the signal arising from a typical voxel.
- We assume that (free-precession) off-resonance phase grows linearly with time, *i.e.* $\phi'(\mathbf{r}, t_0 + \frac{T_P}{2} + T_E; t_0 + \frac{T_P}{2}) \approx \omega'(\mathbf{r})T_E$. We further assume that off-resonance frequency $\omega'(\mathbf{r})$ is distributed over the localized voxel as $p_{\omega'} \leftarrow \text{Cauchy}(\bar{\omega}', R'_2)$, where $\bar{\omega}'(\mathbf{r})$ is the median off-resonance frequency and $R'_2(\mathbf{r})$ is the broadening bandwidth.

With these additional assumptions, the received steady-state SPGR (noiseless) signal model

³Model mismatch due to within-voxel spatial variation of relaxation parameters can be significant, especially for large voxels. Chapter 6 studies so-called partial volume effects and uses them for QMRI.

for a typically sized voxel centered at position \mathbf{r} is (to within constants):

$$\{eq:spgr-int\} \quad s_S\left(\mathbf{r}, t_0 + \frac{T_P}{2} + T_E\right) \propto \int_{\mathbb{V}(\mathbf{r})} m'_{xy}\left(\mathbf{r}, t_0 + \frac{T_P}{2} + T_E\right) d^3 \mathbf{r} \quad (2.26)$$

$$\begin{aligned} &\approx m_{xy}(\mathbf{r}, t_0 + T_P) e^{-T_E/T_2(\mathbf{r})} \int_{\mathbb{R}} e^{-i\omega' T_E} p_{\omega'}(\omega') d\omega' \\ &= m_{xy}(\mathbf{r}, t_0 + T_P) e^{-T_E/T_2(\mathbf{r})} e^{-R'_2(\mathbf{r}) T_E - i\bar{\omega}'(\mathbf{r}) T_E} \\ \{eq:spgr-model\} &= \frac{im_0(\mathbf{r}) \sin(\alpha(\mathbf{r})) (1 - e^{-T_R/T_1(\mathbf{r})})}{1 - e^{-T_R/T_1(\mathbf{r})} \cos(\alpha(\mathbf{r}))} e^{-T_E/T_2^*(\mathbf{r})} e^{-i\bar{\omega}'(\mathbf{r}) T_E}, \end{aligned} \quad (2.27)$$

where $T_2^*(\mathbf{r}) := \left(\frac{1}{T_2} + R'_2\right)^{-1}$ is a modified spin-spin relaxation time that accounts for additional transverse magnetization decay due to off-resonance effects.

2.1.2.2 Dual-Echo Steady-State (DESS) Sequence

DESS [17, 18] is a fast pulse sequence that interlaces fixed, constant-phase RF excitations with fixed dephasing “gradients” (*i.e.*, induced main field inhomogeneities that vary nearly linearly with space) to produce two distinct signals per RF excitation. Here we develop simple steady-state DESS signal models.

As in Subsection 2.1.2.1, let $\mathbf{m}'(\mathbf{r}, t_0)$ denote the magnetization at an initial time t_0 selected well into the steady-state and just prior to excitation. The DESS sequence first applies a fixed RF rotation $\alpha(\mathbf{r}) := \alpha(\mathbf{r}, t_0 + rT_R + T_P; t_0 + rT_R), \forall r \in \{0, 1, 2, \dots\}$:

$$\{eq:dess-ex\} \quad \mathbf{m}'(\mathbf{r}, t_0 + T_P) = \mathbf{R}_{x'}(\alpha(\mathbf{r})) \mathbf{m}'(\mathbf{r}, t_0). \quad (2.28)$$

The excited transverse magnetization contributes to a first acquired signal; dephases (but does not spoil completely) due to gradient dephasing⁴ and contributes again to a second (smaller, but nonzero) acquired signal. Since (with proper selection) dephasing gradients mainly contribute to off-resonance phase accrual, the net effect after data acquisition and gradient spoiling is well described simply by precession and relaxation:

$$\{eq:dess-pr\} \quad \mathbf{m}'(\mathbf{r}, t_0 + T_R) = \mathbf{R}_{z'}(\phi'(\mathbf{r})) \mathbf{E}(\mathbf{r}, T_R; T_P) \mathbf{m}'(\mathbf{r}, t_0 + T_P) + \mathbf{m}_0(\mathbf{r}, T_R; T_P), \quad (2.29)$$

where the abbreviation $\phi'(\mathbf{r}) := \phi'(\mathbf{r}, t_0 + (r+1)T_R; t_0 + rT_R + T_P), \forall r \in \{0, 1, 2, \dots\}$ implies fixed phase accrual (due to gradient dephasing, field inhomogeneity, and other

⁴It is worth distinguishing gradient dephasing (commonly but somewhat misleadingly referred to as gradient spoiling) from RF spoiling. Gradient dephasing (used in DESS) primarily affects magnetization phase and is modeled simply as precession. RF spoiling (used in SPGR) combines gradient dephasing with nonlinear RF phase cycling and suppresses magnetization magnitude in steady-state.

unaccounted effects) over each repetition cycle.

In steady-state, one cycle of excitation, first acquisition, gradient spoiling, second acquisition, and (partial) recovery returns the magnetization back to its initial state. We enforce this through the steady-state condition

$$\{\text{eq:dess-ss}\} \quad \mathbf{m}'(\mathbf{r}, t_0) = \mathbf{m}'(\mathbf{r}, t_0 + T_R) \quad (2.30)$$

which yields an algebraic system of equations. When it exists, the solution gives the steady-state magnetization just prior to RF excitation:

$$\{\text{eq:dess-bmmp-t0}\} \quad \mathbf{m}'(\mathbf{r}, t_0) = \begin{bmatrix} E_2(\mathbf{r}, T_F) \sin \alpha(\mathbf{r}) \sin \phi'(\mathbf{r}) \\ -E_2(\mathbf{r}, T_F) \sin \alpha(\mathbf{r}) (E_2(\mathbf{r}, T_F) - \cos \phi'(\mathbf{r})) \\ 1 - E_2(\mathbf{r}, T_F) \cos \phi'(\mathbf{r}) + E_2(\mathbf{r}, T_F) \cos \alpha(\mathbf{r}) (E_2(\mathbf{r}, T_F) - \cos \phi'(\mathbf{r})) \end{bmatrix} q(\mathbf{r}, T_F), \quad (2.31)$$

where $T_F := T_R - T_P$ is the free precession interval; $E_1(\mathbf{r}, t) := e^{-t/T_1(\mathbf{r})}$ and $E_2(\mathbf{r}, t) := e^{-t/T_2(\mathbf{r})}$ are relaxation operators; and $q(\mathbf{r}, t) :=$

$$\frac{m_0(\mathbf{r})(1 - E_1(\mathbf{r}, t))}{(1 - E_1(\mathbf{r}, t) \cos \alpha(\mathbf{r}))(1 - E_2(\mathbf{r}, t) \cos \phi'(\mathbf{r})) - E_2(\mathbf{r}, t)(E_1(\mathbf{r}, t) - \cos \alpha(\mathbf{r}))(E_2(\mathbf{r}, t) - \cos \phi'(\mathbf{r}))}.$$

Substituting (2.31) into (2.28) produces a similar expression for the steady-state magnetization immediately following RF excitation:

$$\{\text{eq:dess-bmmp-tp}\} \quad \mathbf{m}'(\mathbf{r}, t_0 + T_P) = \begin{bmatrix} E_2(\mathbf{r}, T_F) \sin \alpha(\mathbf{r}) \sin \phi'(\mathbf{r}) \\ \sin \alpha(\mathbf{r})(1 - E_2(\mathbf{r}, T_F) \cos \phi'(\mathbf{r})) \\ \cos \alpha(\mathbf{r})(1 - E_2(\mathbf{r}, T_F) \cos \phi'(\mathbf{r})) + E_2(\mathbf{r}, T_F)(E_2(\mathbf{r}, T_F) - \cos \phi'(\mathbf{r})) \end{bmatrix} q(\mathbf{r}, T_F). \quad (2.32)$$

The transverse magnetizations before and after RF excitation are then

$$\{\text{eq:dess-mxyp-t0}\} \quad m'_{xy}(\mathbf{r}, t_0) = -i \sin \alpha(\mathbf{r}) E_2(\mathbf{r}, T_R) (E_2(\mathbf{r}, T_R) - e^{-i\phi'(\mathbf{r})}) q(\mathbf{r}, T_R); \quad (2.33)$$

$$\{\text{eq:dess-mxyp-tp}\} \quad m'_{xy}(\mathbf{r}, t_0 + T_P) = +i \sin \alpha(\mathbf{r}) (1 - E_2(\mathbf{r}, T_R) e^{i\phi'(\mathbf{r})}) q(\mathbf{r}, T_R), \quad (2.34)$$

where (2.33)-(2.34) include simple first-order corrections for yet-unaccounted relaxation and recovery during excitation. Frequently, the DESS signals are acquired at symmetric echo times T_E before and after the center of each RF pulse. Substituting (2.34) into (2.9)

gives the magnetization at the data acquisition time after RF excitation:

$$m'_{xy}\left(\mathbf{r}, t_0 + \frac{T_P}{2} + T_E\right) = m'_{xy}(\mathbf{r}, t_0 + T_P) e^{-(T_E - T_P/2)/T_2(\mathbf{r})} e^{-i\phi'(\mathbf{r}, t_0 + \frac{T_P}{2} + T_E; t_0 + T_P)}$$

$$\approx m'_{xy}(\mathbf{r}, t_0 + T_P) e^{-T_E/T_2(\mathbf{r})} e^{-i\phi'(\mathbf{r}, t_0 + \frac{T_P}{2} + T_E; t_0 + \frac{T_P}{2})} \quad (2.35)$$

{eq:dess-mxyp-te1-ph}
{eq:dess-mxyp-te1}

$$\approx m'_{xy}(\mathbf{r}, t_0 + T_P) e^{-T_E/T_2(\mathbf{r})} e^{-i\omega'(\mathbf{r})T_E}, \quad (2.36)$$

where in (2.35) we again approximately correct for relaxation during excitation and in (2.36) we assume linear off-resonance phase accrual during free precession. To compute the magnetization at the acquisition time before excitation, we consider the free precession and relaxation that occurs between⁵ signal reception and excitation:

$$m'_{xy}(\mathbf{r}, t_0) = m'_{xy}\left(\mathbf{r}, t_0 - \left(T_E - \frac{T_P}{2}\right)\right) e^{-(T_E - T_P/2)/T_2(\mathbf{r})} e^{-i\phi'(\mathbf{r}, t_0; t_0 - (T_E - \frac{T_P}{2}))}. \quad (2.37)$$

Rearranging (2.37) and applying approximations similar to those of (2.35)-(2.36),

$$m'_{xy}\left(\mathbf{r}, t_0 + \frac{T_P}{2} - T_E\right) = m'_{xy}(\mathbf{r}, t_0) e^{+(T_E - T_P/2)/T_2(\mathbf{r})} e^{+i\phi'(\mathbf{r}, t_0; t_0 - (T_E - \frac{T_P}{2}))}$$

$$\approx m'_{xy}(\mathbf{r}, t_0) e^{+T_E/T_2(\mathbf{r})} e^{+i\phi'(\mathbf{r}, t_0 + \frac{T_P}{2}; t_0 + \frac{T_P}{2} - T_E)} \quad (2.38)$$

{eq:dess-mxyp-te2-ph}
{eq:dess-mxyp-te2}

$$\approx m'_{xy}(\mathbf{r}, t_0) e^{+T_E/T_2(\mathbf{r})} e^{+i\omega'(\mathbf{r})T_E}. \quad (2.39)$$

The received signal is approximately proportional to the integrated transverse magnetization over a volume \mathbb{V} . To derive expressions, we retake assumptions used in Subsection 2.1.2.1 and append an additional assumption on the full-repetition phase accrual $\phi'(\mathbf{r})$:

- We assume that the signal is localized to a scale over which there is off-resonance phase variation, but minimal variation of $m_0(\mathbf{r})$, $T_1(\mathbf{r})$, $T_2(\mathbf{r})$, and $\alpha(\mathbf{r})$. This assumption is reasonable⁶ when describing the signal arising from a typical voxel.
- We assume that free precession off resonance frequency $\omega'(\mathbf{r})$ is distributed over the localized voxel as $p_{\omega'} \leftarrow \text{Cauchy}(\bar{\omega}', R'_2)$, where $\bar{\omega}'(\mathbf{r})$ is the median off-resonance frequency and $R'_2(\mathbf{r})$ is the broadening bandwidth.
- We assume that the dephasing gradient imparts an integral number n_{cyc} of across-

⁵Observe that we do not attempt to express the magnetization prior to (the next) RF excitation by simply operating on the magnetization after (the current) RF excitation with further precession and relaxation. The reason is due to the intermediate dephasing gradient, which causes phase accrual in excess of off-resonance effects and thus forbids an approximation akin to (2.36).

⁶Model mismatch due to within-voxel spatial variation of relaxation parameters can be significant, especially for large voxels. Chapter 6 studies so-called partial volume effects and uses them for QMRI.

voxel phase cycles⁷ such that full-repetition phase accrual $\phi'(\mathbf{r})$ is distributed essentially uniformly as $p_{\phi'} \leftarrow \text{uniform}(0, 2\pi n_{\text{cyc}}), n_{\text{cyc}} \in \{1, 2, 3, \dots\}$.

With these assumptions, the received steady-state DESS (noiseless) signal models for a typically sized voxel centered at position \mathbf{r} are (to within constants):

$$\begin{aligned}
 \text{\{eq:dess-def-int\}} \quad s_D\left(\mathbf{r}, t_0 + \frac{T_P}{2} + T_E\right) &\propto \int_{\mathbb{V}(\mathbf{r})} m'_{xy}\left(\mathbf{r}, t_0 + \frac{T_P}{2} + T_E\right) d^3 \mathbf{r} \\
 &\approx \int_{\mathbb{R}} \int_{\mathbb{R}} m'_{xy}\left(\mathbf{r}, t_0 + \frac{T_P}{2} + T_E\right) p_{\phi'}(\phi') p_{\omega'}(\omega') d\phi' d\omega' \\
 &\approx e^{-T_E/T_2(\mathbf{r})} \int_{\mathbb{R}} m'_{xy}(\mathbf{r}, t_0 + T_P) p_{\phi'}(\phi') d\phi' \int_{\mathbb{R}} e^{-i\omega' T_E} p_{\omega'}(\omega') d\omega' \\
 &= +im_0(\mathbf{r}) E_2(\mathbf{r}, T_E) e^{-(R'_2(\mathbf{r}) - i\bar{\omega}'(\mathbf{r})) T_E} \tan \frac{\alpha(\mathbf{r})}{2} \left[1 - \frac{\eta(\mathbf{r}, T_R)}{\xi(\mathbf{r}, T_R)}\right]; \\
 \text{\{eq:dess-def-model\}} & \tag{2.41}
 \end{aligned}$$

$$\begin{aligned}
 \text{\{eq:dess-ref-int\}} \quad s_D\left(\mathbf{r}, t_0 + \frac{T_P}{2} - T_E\right) &\propto \int_{\mathbb{V}(\mathbf{r})} m'_{xy}\left(\mathbf{r}, t_0 + \frac{T_P}{2} - T_E\right) d^3 \mathbf{r} \\
 &\approx \int_{\mathbb{R}} \int_{\mathbb{R}} m'_{xy}\left(\mathbf{r}, t_0 + \frac{T_P}{2} - T_E\right) p_{\phi'}(\phi') p_{\omega'}(\omega') d\phi' d\omega' \\
 &\approx e^{+T_E/T_2(\mathbf{r})} \int_{\mathbb{R}} m'_{xy}(\mathbf{r}, t_0) p_{\phi'}(\phi') d\phi' \int_{\mathbb{R}} e^{+i\omega' T_E} p_{\omega'}(\omega') d\omega' \\
 &= -im_0(\mathbf{r}) E_2^{-1}(\mathbf{r}, T_E) e^{-(R'_2(\mathbf{r}) + i\bar{\omega}'(\mathbf{r})) T_E} \tan \frac{\alpha(\mathbf{r})}{2} [1 - \eta(\mathbf{r}, T_R)], \\
 \text{\{eq:dess-ref-model\}} & \tag{2.43}
 \end{aligned}$$

where (2.41) and (2.43) introduce intermediate variables

$$\begin{aligned}
 \eta(\mathbf{r}, t) &:= \sqrt{\frac{1 - E_2^2(\mathbf{r}, t)}{1 - E_2^2(\mathbf{r}, t)/\xi^2(\mathbf{r}, t)}}; \\
 \xi(\mathbf{r}, t) &:= \frac{1 - E_1(\mathbf{r}, t) \cos \alpha(\mathbf{r})}{E_1(\mathbf{r}, t) - \cos \alpha(\mathbf{r})}.
 \end{aligned}$$

In steady-state, the DESS signal is typically greatest immediately following excitation and defocuses with rate $\frac{1}{T_2} + R'_2$ until what we hereafter denote the *defocusing* echo time.

⁷In theory, it suffices to design dephasing gradients to impart as few as one complete cycle of net phase variation across a voxel. In practice, field inhomogeneities will induce spurious through-voxel field gradients that modify the effective dephasing gradient moment and thereby create partial phase cycles that distort the nominally uniform phase distribution. To reduce model mismatch due to such “partial spoiling” effects, dephasing gradients are usually designed to nominally impart multiple complete cycles of across-voxel phase variation. However, larger dephasing gradients cause greater DESS model mismatch due to diffusive signal loss. Appendix B studies diffusion in DESS and discusses regimes of dephasing gradient moments which balance partial-spoiling versus diffusive sources of model mismatch.

After a low-signal period between RF pulses, the DESS signal then refocuses with rate $\frac{1}{T_2} - R'_2$ from what we hereafter denote the *refocusing* echo time until just prior the next excitation. Fortuitously, the defocusing (2.41) and refocusing (2.43) DESS signal models have significantly different dependence on relaxation parameters (especially T_2) and thus together are quite useful for relaxation parameter estimation.

2.2 Optimization in QMRI

{s,bkgrd,opt}

This section overviews how optimization methods are leveraged in a substantial portion of this thesis to solve practical QMRI problems. For such problems, the central idea is to construct a suitable scalar cost function Ψ of some design variables \mathbf{x} , whose output $\Psi(\mathbf{x}) \in \mathbb{R}$ is designed to provide a measure of the undesirability of \mathbf{x} . We then employ tailored optimization algorithms to find an \mathbf{x} that minimizes Ψ over a set \mathbb{X} , written as

{eq:opt-global}

$$\mathbf{x}^* \in \left\{ \arg \min_{\mathbf{x} \in \mathbb{X}} \Psi(\mathbf{x}) \right\}. \quad (2.44)$$

In either optimization-based parameter estimation (Chapter 3) or acquisition design (Chapter 4), we have reason to design Ψ to depend on corresponding design variables \mathbf{x} through MR signal models. Because these models are often (strongly) nonlinear functions of design variables, corresponding cost functions are usually non-convex in \mathbf{x} (though the search space \mathbb{X} is almost always assumed convex in this thesis). Thus, most QMRI problems in the form of (2.44) are non-convex optimization problems.

In general, solving (2.44) is more challenging when Ψ is non-convex in \mathbf{x} than otherwise, due in part to the possible presence of local extrema and/or saddle points. In the following, we discuss two strategies used in this thesis to cope with non-convex optimization. Subsection 2.2.1 relaxes (2.44) to instead seek a local minimizer via iterative methods. Subsection 2.2.2 restricts attention to signal models that are linear in a portion of \mathbf{x} and discusses a specific problem for which (2.44) simplifies for such partially linear structures.

2.2.1 Iterative Local Optimization with Constraints

{ss,bkgrd,opt,loc}

This subsection overviews a method for finding a local minimizer $\hat{\mathbf{x}}$ of possibly non-convex cost function Ψ over convex constraint set \mathbb{X} . Such $\hat{\mathbf{x}} \in \mathbb{X}$ must satisfy for some $\delta > 0$

{eq:opt-local}

$$\Psi(\hat{\mathbf{x}}) \leq \Psi(\mathbf{x}) \quad \forall \mathbf{x} \in \mathbb{X} : \|\hat{\mathbf{x}} - \mathbf{x}\|_2 < \delta. \quad (2.45)$$

Observe that a global optimizer \mathbf{x}^* satisfies (2.45) for arbitrarily large δ ; thus, any global minimizer is a local minimizer (but the converse is not necessarily true unless Ψ is convex).

As even locally optimal minimizers are often challenging to compute analytically, many algorithms construct $\hat{\mathbf{x}}$ by iteratively updating an initial guess $\mathbf{x}^{(0)}$ until some convergence criterion is satisfied. For a differentiable cost and convex constraints, the gradient projection method [19] is one such iterative algorithm and repeats the following simple update:

$$\{\text{eq:gpm}\} \quad \mathbf{x}^{(i)} \leftarrow \mathbf{P}_{\mathbb{X}}(\mathbf{x}^{(i-1)} - \mathbf{\Pi} \nabla_{\mathbf{x}} \Psi(\mathbf{x}^{(i-1)})), \quad (2.46)$$

where $\mathbf{P}_{\mathbb{X}}$ denotes projection onto \mathbb{X} ; $\nabla_{\mathbf{x}}$ denotes row gradient with respect to \mathbf{x} ; and $\mathbf{\Pi}$ is a diagonal preconditioning matrix that permits elements of \mathbf{x} to take scale-informed step sizes along the negative gradient direction.

If Ψ is convex and sufficiently smooth, iterates produced via (2.46) converge to a limit point [20] that is a constrained global minimum (for appropriately selected $\mathbf{\Pi}$). If instead Ψ is non-convex (but \mathbb{X} is still convex), statements regarding convergence⁸ to a particular constrained local minimizer require additional (strong) assumptions regarding initialization and in general are still much weaker than in the convex case.

Since non-convex cost functions can have many local extrema (whose associated costs can vary dramatically), the utility of a locally optimal solution depends strongly on initialization quality. Accordingly, this thesis uses iterative local optimization for non-convex QMRI problems where a reasonable initialization is available and global optimization (to within quantization error) via exhaustive grid search is intractable.

2.2.2 Partially Linear Models and the Variable Projection Method

(Constrained, weighted) nonlinear least-squares is a specific non-convex optimization problem that is useful for many parameter estimation problems:

$$\{\text{eq:nonlin-ls}\} \quad \mathbf{x}^* \in \left\{ \arg \min_{\mathbf{x} \in \mathbb{X}} \|\mathbf{y} - \mathbf{f}(\mathbf{x})\|_{\mathbf{W}}^2 \right\}, \quad (2.47)$$

where $\mathbf{f} : \mathbb{X} \mapsto \mathbb{C}^D$ is a nonlinear forward model that (barring noise) relates parameters $\mathbf{x} \in \mathbb{X} \subseteq \mathbb{C}^L$ to data $\mathbf{y} \in \mathbb{C}^D$; weighted 2-norm $\|\boldsymbol{\iota}\|_{\mathbf{W}} := \sqrt{\boldsymbol{\iota}^H \mathbf{W} \boldsymbol{\iota}}$ for a symmetric, positive-semidefinite weighting matrix $\mathbf{W} \in \mathbb{R}^{D \times D}$ and arbitrary vector $\boldsymbol{\iota} \in \mathbb{C}^D$; and $(\cdot)^H$

⁸For example, it suffices to assume that $\mathbf{x}^{(0)}$ lies in the *attraction basin* $\mathbb{B}_{\tilde{\mathbf{x}}}$ of a given unconstrained local minimum $\tilde{\mathbf{x}}$, where attraction basin is defined here as the largest convex set containing $\tilde{\mathbf{x}}$ over which Ψ is convex. If $\mathbb{B}_{\tilde{\mathbf{x}}} \cap \mathbb{X}$ is nonempty and step sizes within $\mathbf{\Pi}$ are small enough to contain iterates within $\mathbb{B}_{\tilde{\mathbf{x}}}$, then iterates converge to the limit point $\mathbf{P}_{\mathbb{X}}(\tilde{\mathbf{x}})$.

denotes conjugate transpose. The variable projection method [21] reduces the complexity of (2.47) when the forward model takes the partially linear structure $\mathbf{f}(\mathbf{x}) \equiv \mathbf{A}(\mathbf{x}_N)\mathbf{x}_L$ and the feasible set takes the partially unconstrained form $\mathbb{X} \equiv \mathbb{C}^{L_L} \times \mathbb{X}_N$, where $\mathbf{x}_L \in \mathbb{C}^{L_L}$; $\mathbf{x}_N \in \mathbb{X}_N$; and $\mathbf{A} : \mathbb{X}_N \mapsto \mathbb{C}^{D \times L_L}$ is a matrix function. These restrictions on (2.47) define a so-called separable least-squares problem:

$$\{\text{eq:sep-ls}\} \quad (\mathbf{x}_L^*, \mathbf{x}_N^*) \in \left\{ \arg \min_{\substack{\mathbf{x}_L \in \mathbb{C}^{L_L} \\ \mathbf{x}_N \in \mathbb{X}_N}} \|\mathbf{y} - \mathbf{A}(\mathbf{x}_N)\mathbf{x}_L\|_{\mathbf{W}}^2 \right\}. \quad (2.48)$$

The variable projection method simplifies (2.48) by exploiting the partially linear structure of \mathbf{f} to explicitly express the optimal \mathbf{x}_L^* as a function of any fixed $\mathbf{x}_N \in \mathbb{X}_N$:

$$\begin{aligned} \mathbf{x}_L^*(\mathbf{x}_N) &= \arg \min_{\mathbf{x}_L \in \mathbb{C}^{L_L}} \|\mathbf{y} - \mathbf{A}(\mathbf{x}_N)\mathbf{x}_L\|_{\mathbf{W}}^2 \\ &= (\mathbf{W}^{1/2} \mathbf{A}(\mathbf{x}_N))^\dagger \mathbf{W}^{1/2} \mathbf{y} \end{aligned} \quad \{\text{eq:sep-ls-lin}\} \quad (2.49)$$

$$\{\text{eq:sep-ls-fullrnk}\} \quad = (\mathbf{A}^H(\mathbf{x}_N) \mathbf{W} \mathbf{A}(\mathbf{x}_N))^{-1} \mathbf{A}^H(\mathbf{x}_N) \mathbf{W} \mathbf{y}, \quad (2.50)$$

where $(\cdot)^\dagger$ denotes pseudoinverse; $\mathbf{W}^{1/2}$ denotes principle (matrix) square root; and (2.50) holds if the matrix inversion within exists. Substituting (2.50) into (2.48) yields a new non-convex optimization problem that contains L_L fewer unknowns than before:

$$\begin{aligned} \mathbf{x}_N^* &\in \left\{ \arg \min_{\mathbf{x}_N \in \mathbb{X}_N} \left\| \mathbf{y} - \mathbf{A}(\mathbf{x}_N) (\mathbf{A}^H(\mathbf{x}_N) \mathbf{W} \mathbf{A}(\mathbf{x}_N))^{-1} \mathbf{A}^H(\mathbf{x}_N) \mathbf{W} \mathbf{y} \right\|_{\mathbf{W}}^2 \right\} \\ &\equiv \left\{ \arg \max_{\mathbf{x}_N \in \mathbb{X}_N} \mathbf{y}^H \mathbf{W} \mathbf{A}(\mathbf{x}_N) (\mathbf{A}^H(\mathbf{x}_N) \mathbf{W} \mathbf{A}(\mathbf{x}_N))^{-1} \mathbf{A}^H(\mathbf{x}_N) \mathbf{W} \mathbf{y} \right\}, \end{aligned} \quad \{\text{eq:sep-ls-nonlin}\} \quad (2.51)$$

where the equivalence leading to (2.51) omits terms independent of \mathbf{x}_N .

In low-dimensional QMRI applications (*e.g.*, those discussed in Chapter 3), reduced problem (2.51) may be tractable via exhaustive grid search, in which case a global optimum $(\mathbf{x}_L^*(\mathbf{x}_N^*), \mathbf{x}_N^*)$ is achievable to within quantization error. However, larger estimation problems involving more nonlinear parameters might still be tractable only via iterative optimization (see Subsection 2.2.1) towards a local solution $(\hat{\mathbf{x}}_L(\hat{\mathbf{x}}_N), \hat{\mathbf{x}}_N)$. For such higher-dimensional applications, Chapters 5-6 introduce novel methods that tackle problems similar to (2.47), while circumventing initialization-dependent local optimization.

CHAPTER 3

MRI Parameter Estimation from Likelihood Models

{c,relax}

3.1 Introduction

{s,relax,intro}

This chapter introduces methods for QMRI parameter estimation from statistical likelihood models and applies these methods to simple problems in T_1 and T_2 relaxometry, which are of interest for monitoring the progression of various disorders [22]. Section 3.2 introduces the notion of a QMRI scan profile, describes a signal model for parameter estimation, formulates several likelihood-based estimators using this model, and discuss practical implementation issues. Section 3.3 demonstrates the utility of likelihood-based parameter estimation over conventional methods through simulation, phantom, and *in vivo* experiments. Section 3.4 provides brief concluding remarks and suggests future directions.

3.2 Likelihood-Based Estimation in QMRI

{s,relax,method}

3.2.1 The QMRI Scan Profile

{ss,relax,method,profile}

After image reconstruction, many MRI pulse sequences useful for parameter estimation produce at each voxel centered at position \mathbf{r} a set of noisy voxel values $\{y_1(\mathbf{r}), \dots, y_D(\mathbf{r})\}$, each of which can be described with the following general model:

{eq:relax,mod-scalar}

$$y_d(\mathbf{r}) = s_d(\mathbf{x}(\mathbf{r}); \boldsymbol{\nu}(\mathbf{r}), \mathbf{p}_d) + \epsilon_d(\mathbf{r}), \quad (3.1)$$

where $d \in \{1, \dots, D\}$. Here, $\mathbf{x}(\mathbf{r}) \in \mathbb{C}^L$ collects L *latent* object parameters at \mathbf{r} ; $\boldsymbol{\nu}(\mathbf{r}) \in \mathbb{C}^K$ collects K *known* object parameters at \mathbf{r} ; $s_d : \mathbb{C}^L \times \mathbb{C}^K \times \mathbb{R}^A \mapsto \mathbb{C}$ is a (pulse-sequence dependent) function that models the noiseless signal obtained from the d th dataset

using *acquisition* parameter $\mathbf{p}_d \in \mathbb{R}^A$; and $\epsilon_d \sim \mathbb{CN}(0, \sigma_d^2)$ is assumed for simplicity¹ to be (circularly-symmetric) complex Gaussian noise [25, 26] with zero mean and variance σ_d^2 . Colon positions in signal model (3.1) and similar expressions throughout this thesis distinguish unknown and known parameters. Concrete examples follow shortly.

For accurate, well-conditioned QMRI parameter estimation, it is typically necessary to acquire a collection of datasets, which we refer to hereafter as a *scan profile*. A scan profile consists of D datasets from up to D pulse sequences (some sequences yield more than one dataset, *e.g.* DESS). Let $\mathbf{y}(\mathbf{r}) := [y_1(\mathbf{r}), \dots, y_D(\mathbf{r})]^\top \in \mathbb{C}^D$ collect noisy voxel values centered at \mathbf{r} from a given scan profile. Then the vector signal model

{eq:relax,mod-vec}

$$\mathbf{y}(\mathbf{r}) = \mathbf{s}(\mathbf{x}(\mathbf{r}); \boldsymbol{\nu}(\mathbf{r}), \mathbf{P}) + \boldsymbol{\epsilon}(\mathbf{r}) \quad (3.2)$$

helps define the noiseless signal $\mathbf{s} := [s_1, \dots, s_D]^\top : \mathbb{C}^L \times \mathbb{C}^K \times \mathbb{R}^{A \times D} \mapsto \mathbb{C}^D$ and acquisition parameter $\mathbf{P} := [\mathbf{p}_1, \dots, \mathbf{p}_D] \in \mathbb{R}^{A \times D}$ associated with that scan profile. Here, noise $\boldsymbol{\epsilon}(\mathbf{r}) := [\epsilon_1(\mathbf{r}), \dots, \epsilon_D(\mathbf{r})]^\top \in \mathbb{C}^D$ typically has diagonal covariance structure $\boldsymbol{\Sigma} := \text{diag}([\sigma_1, \dots, \sigma_D]^\top)$ due to independence across datasets, where $\text{diag}(\cdot)$ assigns its argument to the diagonal entries of an otherwise zero (square) matrix.

The following subsections describe two concrete scan profiles whose signals can be modeled via (3.2) and that we study through experiments later in this chapter.

3.2.1.1 Example: An SPGR Scan Profile for T_1 estimation

We first consider the problem of $T_1(\mathbf{r})$ estimation at \mathbf{r} from as few SPGR scans as possible, given a prior estimate of transmit field variation $s^t(\mathbf{r})$ (see (2.7)). Examining SPGR model (2.27) makes clear that by fixing echo time T_E across scans, SPGR signal dependence is reduced to just two spatially varying latent parameters: desired parameter $T_1(\mathbf{r}) \in \mathbb{R}$ and nuisance parameter $c_1(\mathbf{r}) := im_0(\mathbf{r})e^{-T_E/T_2^*(\mathbf{r})}e^{-i\bar{\omega}'(\mathbf{r})T_E} \in \mathbb{C}$. We assign $\mathbf{x} \leftarrow [T_1, c_1]^\top$ and $\boldsymbol{\nu} \leftarrow s^t$ for $L \leftarrow 2$ latent and $K \leftarrow 1$ known parameters, respectively.

With T_E fixed, prescribed flip angles α_0 and repetition times T_R are the only remaining $A \leftarrow 2$ acquisition parameters available to choose that appear explicitly in (2.27). Thus, an SPGR scan profile useful for T_1 estimation must vary $\mathbf{p}_d \leftarrow [\alpha_0, T_R]^\top \forall d \in \{1, \dots, D\}$ over S_{SPGR} scan repetitions to produce $D \geq L \leftarrow 2$ datasets for well-conditioned estimation.

¹Though the noise distribution of \mathbf{k} -space raw data is usually well-modeled as complex white Gaussian, the noise distribution of the d th reconstructed image y_d depends both on the acquisition and reconstruction. If single receive channel \mathbf{k} -space data is fully-sampled on a Cartesian grid, each dataset y_d is recoverable via separate Fourier transform, and is thus complex Gaussian and independent across datasets. However if \mathbf{k} -space data is multi-channel, undersampled, and/or Cartesian, it may be preferable that y_d be estimated by more sophisticated techniques, *e.g.* [23, 24]. In such cases, reconstructed image noise is unlikely to be Gaussian-distributed.

3.2.1.2 Example: A DESS Scan Profile for T_2 estimation

We next consider the problem of $T_2(\mathbf{r})$ estimation at \mathbf{r} from as few DESS scans as possible. Examining DESS models (2.41) and (2.43) makes clear that even with fixed T_E over possibly several acquisitions, there is signal dependence on five distinct object parameters: $s^t(\mathbf{r}) \in \mathbb{R}$, $T_1(\mathbf{r}) \in \mathbb{R}$, $\bar{\omega}'(\mathbf{r}) \in \mathbb{R}$, $c_2(\mathbf{r}) := m_0(\mathbf{r})e^{-T_E/T_2^*(\mathbf{r})} \in \mathbb{C}$, and $T_2(\mathbf{r}) \in \mathbb{R}$. In this chapter, we take $s^t(\mathbf{r}) \in \mathbb{R}$ and $T_1(\mathbf{r}) \in \mathbb{R}$ as known for simplicity. To avoid (separate or joint) $\bar{\omega}'(\mathbf{r})$ estimation, we choose to use magnitude DESS data, at the expense of slight model mismatch² due to Rician noise. These choices assign $\boldsymbol{\nu} \leftarrow [s^t, T_1]^T$ as $K \leftarrow 2$ known parameters and leave $L \leftarrow 2$ latent parameters $\mathbf{x} \leftarrow [c_2, T_2]^T$ to be estimated.

With T_E again fixed, $\mathbf{p}_d \leftarrow [\alpha_0, T_R]^T \forall d \in \{1, \dots, D\}$ collects the remaining $A \leftarrow 2$ tunable scan parameters that appear explicitly in (2.41) and (2.43). As in Example 3.2.1.1, $D \geq L \leftarrow 2$ datasets are necessary for well-conditioned estimation. Unlike before however, a minimum $D \leftarrow 2$ datasets need not require scan repetition, since S_{DESS} DESS scan repetitions produce $D \leftarrow 2S_{\text{DESS}}$ datasets.

3.2.2 Latent Object Parameter Estimation

3.2.2.1 Signal Model and Problem Statement

A scan profile's reconstructed images can be modeled to discretize the bulk MR signal into V localized voxels centered at positions $\mathbf{r}_1, \dots, \mathbf{r}_V$:

$$\mathbf{Y} = \mathbf{S}(\mathbf{X}; \mathbf{N}, \mathbf{P}) + \mathbf{E}. \quad (3.3)$$

Here, signal model $\mathbf{S} : \mathbb{C}^{L \times V} \times \mathbb{C}^{K \times V} \times \mathbb{R}^{A \times D} \mapsto \mathbb{C}^{D \times V}$ is a matrix function that maps latent $\mathbf{X} := [\mathbf{x}(\mathbf{r}_1), \dots, \mathbf{x}(\mathbf{r}_V)] \in \mathbb{C}^{L \times V}$ and known $\mathbf{N} := [\boldsymbol{\nu}(\mathbf{r}_1), \dots, \boldsymbol{\nu}(\mathbf{r}_V)] \in \mathbb{C}^{L \times V}$ parameter images (with fixed acquisition parameter \mathbf{P}) to reconstructed image data $\mathbf{Y} := [\mathbf{y}(\mathbf{r}_1), \dots, \mathbf{y}(\mathbf{r}_V)] \in \mathbb{C}^{D \times V}$, save for noise image $\mathbf{E} := [\boldsymbol{\epsilon}(\mathbf{r}_1), \dots, \boldsymbol{\epsilon}(\mathbf{r}_V)] \in \mathbb{C}^{D \times V}$. The goal in QMRI parameter estimation is to estimate latent parameter images \mathbf{X} from MR image data \mathbf{Y} , for a fixed scan profile defined by \mathbf{S} and \mathbf{P} and given (separately acquired, estimated, and here assumed) known parameter images \mathbf{N} .

²The assumption of complex Gaussian noise in noisy MRI images implies that corresponding magnitude MRI images are Rician-distributed. However, the statistical estimators we will develop in Subsection 3.2.2 are based on Gaussian data. Fortunately, this source of model mismatch is negligible (less than 1%) for signal-to-noise ratio (SNR) in excess of 10 [27], and the acquisitions we examine here are capable of producing SNR in tissue of at minimum 100 and usually more.

3.2.2.2 Maximum Likelihood Methods

In maximum likelihood (ML) estimation, one seeks to find model parameters that maximize the likelihood of observing output data. We apply ML estimation to QMRI by first constructing a *likelihood function* that describes the probability of observing image data \mathbf{Y} given latent parameters \mathbf{X} . We then formulate ML latent parameter estimate $\hat{\mathbf{X}}_{\text{ML}}(\mathbf{Y}; \mathbf{N}, \mathbf{P})$ by finding an \mathbf{X} that maximizes this likelihood function.

We first construct the likelihood function for the v th voxel's data $\mathbf{y}(\mathbf{r}_v)$ and latent parameter $\mathbf{x}(\mathbf{r}_v)$. For complex Gaussian noise, the likelihood function is

$$\mathcal{L}(\mathbf{x}(\mathbf{r}_v)) \propto \exp \left(-\|\mathbf{y}(\mathbf{r}_v) - \mathbf{s}(\mathbf{x}(\mathbf{r}_v); \boldsymbol{\nu}(\mathbf{r}_v), \mathbf{P})\|_{\boldsymbol{\Sigma}^{-1}}^2 \right), \quad (3.4)$$

where (3.4) omits constants that are independent of $\mathbf{x}(\mathbf{r}_v)$ and are therefore irrelevant. Assuming noise independence across image voxels, we can next build a simple and practical likelihood function of the full image data as

$$\mathcal{L}(\mathbf{X}) = \prod_{v=1}^V \mathcal{L}(\mathbf{x}(\mathbf{r}_v)). \quad (3.5)$$

We form an ML parameter estimate by finding \mathbf{X} that maximizes this likelihood function:

$$\begin{aligned} \hat{\mathbf{X}}_{\text{ML}}(\mathbf{Y}; \mathbf{N}, \mathbf{P}) &\in \left\{ \arg \max_{\mathbf{X} \in \mathbb{X}^V} \mathcal{L}(\mathbf{X}) \right\} \\ &\equiv \left\{ \arg \min_{\mathbf{X} \in \mathbb{X}^V} -\log \mathcal{L}(\mathbf{X}) \right\} \end{aligned} \quad (3.6)$$

$$\begin{aligned} &= \left\{ \arg \min_{\mathbf{X} \in \mathbb{X}^V} \sum_{v=1}^V \|\mathbf{y}(\mathbf{r}_v) - \mathbf{s}(\mathbf{x}(\mathbf{r}_v); \boldsymbol{\nu}(\mathbf{r}_v), \mathbf{P})\|_{\boldsymbol{\Sigma}^{-1}}^2 \right\} \\ &= \left\{ \arg \min_{\mathbf{X} \in \mathbb{X}^V} \left\| \boldsymbol{\Sigma}^{-1/2} (\mathbf{Y} - \mathbf{S}(\mathbf{X}; \mathbf{N}, \mathbf{P})) \right\|_{\text{F}}^2 \right\}, \end{aligned} \quad (3.7)$$

where \mathbb{X} is a (typically convex) latent parameter search space; the set equivalence in (3.6) uses the monotonicity of the log function; and $\|\cdot\|_{\text{F}}$ denotes the Frobenius matrix norm.

Typically, QMR image model \mathbf{S} is nonlinear in \mathbf{X} and so ML estimation problem (3.7) involves non-convex optimization, which is challenging in general (see Section 2.2). Two properties of (3.7) guide our solution strategies. First, (3.7) is separable across voxels, so problem non-convexity is addressable on a voxel-by-voxel basis. Second, MR signal models are usually partially linear, in which case we may employ the variable projection method (described in Section 2.2.2) to further reduce problem complexity. For applications studied in this chapter, these properties allow for (3.7) to be solved via simple grid search.

3.2.2.3 Regularized Likelihood Methods

In regularized likelihood (RL) estimation, we modify ML estimation problem (3.6) to include additional information in the form of *regularization*:

$$\widehat{\mathbf{X}}_{\text{RL}}(\mathbf{Y}; \mathbf{N}, \mathbf{P}) \in \left\{ \arg \min_{\mathbf{X} \in \mathbb{X}^V} -\log L(\mathbf{X}) + R(\mathbf{X}) \right\}. \quad (3.8)$$

Here, we have freedom to design regularizer $R : \mathbb{C}^{L \times V} \mapsto \mathbb{R}$ to encourage desirable structure in estimates of \mathbf{X} . We observe that it is usually reasonable to assume that each latent object parameter map is *piecewise smooth* as a function of space: that is, each parameter is likely to vary smoothly in space, except for sharp discontinuities at tissue boundaries. To encourage piecewise-smoothness in parameter estimates, we use the regularizer

$$R(\mathbf{X}) := \sum_{l=1}^L \beta_l \sum_{j=1}^J \phi_l \left([\mathbf{J}\mathbf{X}^\top]_{jl} \right), \text{ where} \quad (3.9)$$

$$\phi_l(\cdot) := \gamma_l^2 \left(\sqrt{1 + |\cdot/\gamma_l|^2} - 1 \right) \quad (3.10)$$

is a differentiable approximation the absolute value function; $\mathbf{J} \in \mathbb{R}^{J \times V}$ evaluates J (multi-dimensional) finite-differencing operations; $[\cdot]_{jl}$ extracts the (j, l) th matrix element; and β_l is a regularization parameter that controls the relative importance of smoothing the l th latent object parameter image. Conceptually, this regularizer penalizes inconsistencies in adjacent latent parameter image voxels, but with a severity that depends on the degree of inconsistency. “Small” voxel-to-voxel differences are likely due to image data noise within a single tissue type and are penalized near-quadratically, while “large” differences are likely due to tissue boundaries and are penalized near-linearly. Useful notions of small versus large differences are governed by shape parameters $\gamma_l \forall l \in \{1, \dots, L\}$, and vary for different latent parameter maps based on their units and relative scale.

In general, QMRI image signal model \mathbf{S} is nonlinear in \mathbf{X} and so RL estimation problem (3.8) requires non-convex optimization. Unlike in ML estimation, (3.8) is not separable across voxels due to regularization, precluding global optimization (via grid search or other methods). We instead take the corresponding ML estimate as initialization and solve (3.8) via iterative constrained local optimization (detailed in Section 2.2.1).

3.3 Experimentation

{s,relax,exp}

3.4 Summary

{s,relax,summ}

CHAPTER 4

Optimizing MR Scan Design for Model-Based Relaxometry

{c,scn-dsgn}

4.1 Introduction

{s,scn-dsgn,intro}

Fast, accurate *relaxometry*, or quantification of spin-lattice and spin-spin relaxation parameters T_1 and T_2 has been of longstanding interest in MRI. Many researchers have suggested that T_1, T_2 “maps” (*i.e.*, estimated parameter images) may serve as biomarkers for monitoring the progression of various disorders [22]. Neurological applications include: lesion classification in multiple sclerosis [28]; tumor characterization [29, 30]; and symptom onset prediction in stroke [31, 32]. In addition, T_1, T_2 have shown promise for detecting hip and knee cartilage degeneration [33, 34] and for assessing cardiac dysfunction due to iron overload [35] or edema [36]. Motivated by this broad interest in T_1, T_2 mapping, this chapter describes a systematic method to guide QMRI scan design.

Classical pulse sequences such as inversion/saturation recovery (IR/SR) or (single) spin echo (SE) yield relatively simple methods for T_1 or T_2 estimation, respectively; however, these methods require several scans, each with long repetition time T_R , leading to undesirably long acquisitions. Numerous modifications such as the Look-Locker method [37], multi-SE trains [38], or fast k-space trajectories [39, 40, 41] have been proposed to accelerate T_1 [42, 43, 44, 45] and T_2 [46, 47, 48, 49] relaxometry with these classical sequences. These techniques are more sensitive to model non-idealities [50, 51, 52], and are still speed-limited by the long T_R required for (near)-complete T_1 recovery.

Steady-state (SS) pulse sequences [12, 13] permit short T_R , and are thus inherently much faster than classical counterparts. SS techniques are well-suited for relaxometry because the signals produced are highly sensitive to T_1 and T_2 variation. However, short T_R times also cause SS signals to be complex functions of both desired and undesired (*nuisance*) parameters, complicating quantification. Furthermore, some such methods [53, 54] still require scan repetition, though individual scans are now considerably shorter. Despite

these difficulties, the potential for rapid scanning with high T_1 , T_2 sensitivity has motivated numerous SS relaxometry studies [55, 53, 54, 56, 57, 58, 59, 60, 61, 62].

The dual-echo steady-state (DESS) sequence [18] was recently proposed as a promising SS imaging technique for T_2 estimation [59]. Because it produces two distinct signals per excitation, the DESS sequence can reduce scan repetition requirements by recording twice as much data per scan. As with other SS methods, the resulting signals [63, 64] are complicated functions of T_1 , T_2 , and other parameters (see Section 2.1.2.2 for derivations). Prior works have isolated T_2 dependencies using either algebraic manipulations of the first- and second-echo signals [59, 60] or separate scans to first estimate nuisance parameters [3]. Although DESS concurrently encodes rich T_1 and T_2 information, these methods have shied away from using DESS for T_1 estimation, either through bias-inducing approximations, or noise-propagating sequential estimation, respectively.

Whether it be with DESS, other sequences, or even combinations thereof, it is generally unclear how to best assemble a *scan profile* (*i.e.*, a collection of scans) for a fixed amount of scan time. Furthermore, for a given scan profile, it is typically not obvious how to best select acquisition parameters (*e.g.*, flip angles, repetition times, etc.) for relaxometry. In this and subsequent chapters, the term *scan design* refers to the related problems of scan profile selection and acquisition parameter optimization.

Historically, scan design for relaxometry has mainly been explored using figures of merit related to estimator precision. In particular, several studies have used the Cramér-Rao Bound (CRB), a statistical tool that bounds the minimum variance of an unbiased estimator. Earlier works have used the CRB and variations to select inversion times for recovery experiments [65, 66], flip angles for spoiled gradient-recalled echo (SPGR) sequences [67], and echo times for SE experiments [68]. More recent studies have considered additional scan design challenges, including scan time constraints [69], multiple latent parameters [70], multiple scan parameter types [71], and latent parameter spatial variation [72, 73].

The aforementioned studies consider scan parameter optimization for profiles consisting of *only one* pulse sequence. In contrast, this chapter introduces a general framework for robust, application-specific scan design for parameter estimation from *combinations* of pulse sequences. The framework first finds multiple sets of scan parameters that achieve precise estimation within a tight, *application-specific* range of object parameters (*e.g.*, T_1 , T_2 , etc.). The framework then chooses the one scan parameter set most *robust* to estimator precision degradation over a broader range of object parameters. As a detailed example, we optimize three combinations of SPGR and DESS sequences for T_1 , T_2 mapping. For a fixed total scan time, we find that well-chosen DESS scans alone can be used to estimate both T_1 and T_2 with precision and robustness comparable to combinations of SPGR

and DESS. This example illustrates that, with careful scan profile design, well-established pulse sequences can find use in new estimation problems.

This chapter is organized as follows. Section 4.2 describes a CRB-inspired min-max optimization problem for robust, application-specific scan optimization. Section 4.3 optimizes three practical DESS/SPGR combinations to show that, even in the presence of radiofrequency (RF) field inhomogeneity, DESS is a promising option for T_1, T_2 relaxometry. Section 4.4 describes simulation, phantom, and *in vivo* experiments and discusses corresponding results. Section ?? discusses practical challenges and suggests future directions. Section ?? summarizes key contributions.

4.2 A CRB-Inspired Scan Selection Method

{s,scn-dsgn,crb}

4.2.1 The CRB and its Relevance to QMRI

Recall from Section 3.2.1 that after image reconstruction, vwe can model the single-voxel MR image domain data associated with a particular scan profile as

$$\mathbf{y} = \mathbf{s}(\mathbf{x}; \boldsymbol{\nu}, \mathbf{P}) + \boldsymbol{\epsilon}, \quad (4.1)$$

where signal model $\mathbf{s} := [s_1, \dots, s_D]^T : \mathbb{C}^L \times \mathbb{C}^K \times \mathbb{R}^{A \times D} \mapsto \mathbb{C}^D$ relates latent $\mathbf{x} \in \mathbb{C}^L$, known $\boldsymbol{\nu} \in \mathbb{C}^K$, and acquisition $\mathbf{P} \in \mathbb{R}^{A \times D}$ parameters to noisy scan profile image data $\mathbf{y} \in \mathbb{C}^D$, barring noise $\boldsymbol{\epsilon} \in \mathbb{C}^D$. Assuming (as in Section 3.2.1) complex Gaussian noise $\boldsymbol{\epsilon} \sim \mathbb{CN}(\mathbf{0}, \boldsymbol{\Sigma})$, the likelihood function (3.4) is (to within constants independent of \mathbf{x})

$$\mathcal{L}(\mathbf{x}|\mathbf{y}) \propto \exp \left(-\|\mathbf{y} - \mathbf{s}(\mathbf{x}; \boldsymbol{\nu}, \mathbf{P})\|_{\boldsymbol{\Sigma}^{-1}}^2 \right). \quad (4.2)$$

Under suitable regularity conditions¹, the Fisher information matrix $\mathbf{F}(\mathbf{x}; \boldsymbol{\nu}, \mathbf{P}) \in \mathbb{C}^{L \times L}$ [74] characterizes the imprecision of unbiased estimates of \mathbf{x} from \mathbf{y} , given $\boldsymbol{\nu}$ and \mathbf{P} :

$$\begin{aligned} \mathbf{F}(\mathbf{x}; \boldsymbol{\nu}, \mathbf{P}) &:= \mathbb{E}_{\mathbf{y}} \left((\nabla_{\mathbf{x}} \log \mathcal{L}(\mathbf{x}|\mathbf{y}))^H \nabla_{\mathbf{x}} \log \mathcal{L}(\mathbf{x}|\mathbf{y}) \right) \\ &= (\nabla_{\mathbf{x}} \mathbf{s}(\mathbf{x}; \boldsymbol{\nu}, \mathbf{P}))^H \boldsymbol{\Sigma}^{-1} \nabla_{\mathbf{x}} \mathbf{s}(\mathbf{x}; \boldsymbol{\nu}, \mathbf{P}), \end{aligned} \quad (4.3)$$

¹In particular, \mathbf{s} must be analytic in complex components of \mathbf{x} .

where $E_y(\cdot)$ denotes element-wise expectation with respect to y . In particular, the matrix CRB [75] ensures that any unbiased estimator $\hat{\mathbf{x}}$ satisfies

$$\text{cov}(\hat{\mathbf{x}}; \boldsymbol{\nu}, \mathbf{P}) \succeq \mathbf{F}^{-1}(\mathbf{x}; \boldsymbol{\nu}, \mathbf{P}), \quad (4.4)$$

where for arbitrary, equally-sized \mathbf{C}_1 and \mathbf{C}_2 , matrix inequality $\mathbf{C}_1 \succeq \mathbf{C}_2$ means $\mathbf{C}_1 - \mathbf{C}_2$ is positive semi-definite. In the following, we design an optimization problem based on the CRB to guide QMRI scan design for relaxometry.

4.2.2 Min-max Optimization Problem for Scan Design

Following [76], we focus on minimizing a weighted average of the variances in each of the L latent object parameter estimates. A reasonable objective function for overall estimator precision is therefore given by

$$\Psi(\mathbf{x}; \boldsymbol{\nu}, \mathbf{P}) = \text{tr}(\mathbf{W}\mathbf{F}^{-1}(\mathbf{x}; \boldsymbol{\nu}, \mathbf{P})\mathbf{W}^T), \quad (4.5)$$

where $\mathbf{W} \in \mathbb{R}^{L \times L}$ is a diagonal, application-specific matrix of weights, preselected to control the relative importance of precisely estimating the L latent object parameters. For scan design, we would like to minimize (4.5) with respect to scan parameters \mathbf{P} .

The CRB depends not only on \mathbf{P} but also on the spatially varying object parameters \mathbf{x} and $\boldsymbol{\nu}$. Thus, one cannot perform scan design by “simply” minimizing Ψ with respect to scan parameters \mathbf{P} . Instead, we pose a *min-max* optimization problem for scan design: we seek candidate scan parameters $\check{\mathbf{P}}$ over a search space \mathbb{P} that *minimize* the worst-case (*i.e.*, *maximum*) cost $\tilde{\Psi}^t$, as viewed over “tight” object parameter ranges \mathbb{X}^t and \mathbb{N}^t :

$$\check{\mathbf{P}} \in \left\{ \arg \min_{\mathbf{P} \in \mathbb{P}} \tilde{\Psi}^t(\mathbf{P}) \right\}, \text{ where} \quad (4.6)$$

$$\tilde{\Psi}^t(\mathbf{P}) := \max_{\substack{\mathbf{x} \in \mathbb{X}^t \\ \boldsymbol{\nu} \in \mathbb{N}^t}} \Psi(\mathbf{x}; \boldsymbol{\nu}, \mathbf{P}). \quad (4.7)$$

Here, we select *latent* parameter set \mathbb{X}^t based on the application and *known* parameter set \mathbb{N}^t based on the spatial variation typically observed in the known parameters $\boldsymbol{\nu}$. Min-max approach (4.9) should ensure good estimation precision over a range of parameter values.

Since Ψ is in general non-convex with respect to \mathbf{P} , it may have multiple global minimizers as well as other scan parameters that are nearly global minimizers. To improve robustness to object parameter variations, we form an expanded set of candidate scan parameters by also including scan parameters that yield costs to within a tolerance $\delta \ll 1$

of the optimum. Mathematically, we define this expanded set of candidate scan parameter combinations (for a given scan profile) as

$$\{\text{eq:scn-dsgn,set}\} \quad \check{\mathbb{P}} := \left\{ \mathbf{P} : \tilde{\Psi}^t(\mathbf{P}) - \tilde{\Psi}^t(\check{\mathbf{P}}) \leq \delta \tilde{\Psi}^t(\check{\mathbf{P}}) \right\}. \quad (4.8)$$

To select amongst these candidate scan parameters, we employ a robustness criterion: we select the single scan parameter $\hat{\mathbf{P}}$ that degrades the least when the worst-case cost is viewed over widened object parameter sets $\mathbb{X}^b \supseteq \mathbb{X}^t$ and $\mathbb{N}^b \supseteq \mathbb{N}^t$:

$$\{\text{eq:scn-dsgn,P-star}\} \quad \hat{\mathbf{P}} = \arg \min_{\mathbf{P} \in \check{\mathbb{P}}} \tilde{\Psi}^b(\mathbf{P}), \text{ where} \quad (4.9)$$

$$\{\text{eq:scn-dsgn,cost-broad}\} \quad \tilde{\Psi}^b(\mathbf{P}) := \max_{\substack{\mathbf{x} \in \mathbb{X}^b \\ \boldsymbol{\nu} \in \mathbb{N}^b}} \Psi(\mathbf{x}; \boldsymbol{\nu}, \mathbf{P}). \quad (4.10)$$

To compare different scan profiles, we select corresponding search spaces \mathbb{P} to satisfy acquisition constraints (*e.g.*, total scan time), but otherwise hold optimization parameters \mathbf{W} , δ , \mathbb{X}^t , \mathbb{X}^b , \mathbb{N}^t , \mathbb{N}^b fixed. Since Ψ is data-independent, we can solve (4.6) and (4.9) offline for each scan profile. The result of each profile’s min-max optimization process (4.9) is a corresponding optimized scan parameter matrix $\hat{\mathbf{P}}$ that is suitable for the range of latent \mathbf{x} and known $\boldsymbol{\nu}$ object parameters specified in \mathbb{X}^t and \mathbb{N}^t , and is robust to variations in those parameters over broader sets \mathbb{X}^b and \mathbb{N}^b , respectively.

4.3 Optimizing SS Sequences for Relaxometry in the Brain

$\{\text{s,scn-dsgn,opt}\}$

This section applies the methods of Section 4.2.2 to the problem of scan design for joint T_1, T_2 estimation from combinations of SS sequences. Section 4.3.1 details how we use optimization problems (4.6) and (4.9) to tailor three SPGR and DESS scan combinations for precise T_1, T_2 estimation in white matter (WM) and grey matter (GM) regions of the brain. Section 4.3.2 compares the predicted performance of the three optimized scan profiles.

4.3.1 Scan Design Details

$\{\text{ss,scn-dsgn,opt,design}\}$

There are numerous candidate scan profiles involving DESS and/or other pulse sequences that may be useful for fast, accurate T_1, T_2 mapping. In this chapter, we consider combinations of magnitude SPGR and DESS scans for estimating the $L \leftarrow 3$ latent parameters T_1, T_2 , and a proportionality constant, given knowledge of transmit field inhomogeneity s^t as $K \leftarrow 1$ known parameter. With proper RF phase cycling and gradient spoiling, the SPGR signal s_S (as expressed in (2.27)) contains no explicit T_2 dependence. SPGR’s

reduced dependence on spatially varying unknowns is reason for its use in T_1 mapping [55, 54, 56] and subsequent T_2 mapping from other sequences [53, 3]. In a similar spirit, we examine scan profiles containing SPGR over other SS sequences because we predict that the SPGR sequence’s T_2 -independence may help estimators disentangle T_2 from other unknown sources of DESS signal contrast.

As respectively discussed in Examples 3.2.1.1-3.2.1.2, each SPGR and DESS scan leaves $\mathbf{p} \leftarrow [\alpha_0, T_R]^\top$ as $A \leftarrow 2$ acquisition parameters available to optimize. A given scan profile consisting of S_{SPGR} SPGR and S_{DESS} DESS scans yields $D \leftarrow S_{\text{SPGR}} + 2S_{\text{DESS}}$ datasets. We optimize such a scan profile by solving (4.9) over a dimension- $AD \leftarrow 2(S_{\text{SPGR}} + 2S_{\text{DESS}})$ space of scan parameters.

We select constraints on search space \mathbb{P} based on hardware limitations and desired scan profile properties. Since each pair of DESS signals must share the same \mathbf{p} , the search space \mathbb{P} is reduced to $\mathbb{A}_{0,\text{SPGR}}^{S_{\text{SPGR}}} \times \mathbb{A}_{0,\text{DESS}}^{S_{\text{DESS}}} \times \mathbb{T}_{\text{R,SPGR}}^{S_{\text{SPGR}}} \times \mathbb{T}_{\text{R,DESS}}^{S_{\text{DESS}}}$ (superscripts denote Cartesian powers). We assign flip angle ranges $\mathbb{A}_{0,\text{SPGR}} \leftarrow [5, 90]^\circ$ and $\mathbb{A}_{0,\text{DESS}} \leftarrow [5, 90]^\circ$ to restrict RF energy deposition. We set feasible T_R solution sets $\mathbb{T}_{\text{R,SPGR}} \leftarrow [12.2, +\infty)\text{ms}$ and $\mathbb{T}_{\text{R,DESS}} \leftarrow [17.5, +\infty)\text{ms}$ based on pulse sequence designs that control for other scan parameters. These control parameters are described in further detail in Section 4.4, and are held fixed in all subsequent SPGR and DESS experiments. To equitably compare optima from different scan profiles, we require

$$\mathbf{T}_R := [T_{R,1}, \dots, T_{R,S_{\text{SPGR}}}, T_{R,S_{\text{SPGR}}+1}, \dots, T_{R,S_{\text{SPGR}}+S_{\text{DESS}}}]^\top$$

to satisfy a total time constraint, $\|\mathbf{T}_R\|_1 \leq T_{\text{max}}$. For a scan profile consisting of S_{SPGR} SPGR and S_{DESS} DESS scans, these constraints collectively reduce the search space dimension from AD to $2(S_{\text{SPGR}} + S_{\text{DESS}}) - 1$.

Prior works have considered T_1 or T_2 estimation from as few as 2 SPGR [67, 53] or 1 DESS [59] scan(s), respectively. We likewise elect to optimize the $(S_{\text{SPGR}}, S_{\text{DESS}}) \leftarrow (2, 1)$ scan profile as a benchmark. We choose $T_{\text{max}} \leftarrow 2(12.2) + 1(17.5) = 41.9\text{ms}$ and select other scan profiles capable of meeting this time constraint. Requiring that candidate profiles contain $S_{\text{DESS}} \geq 1$ DESS scans for T_2 contrast and satisfy $D \geq L(= 3)$ for well-conditioned estimation, we note that $(1, 1)$ and $(0, 2)$ are the only other eligible profiles.

In the ensuing experiments, we focus on precise T_1, T_2 estimation in the brain. Noting that $T_1 \sim 10T_2$, we choose $\mathbf{W} \leftarrow \text{diag}(0.1, 1, 0)$ to place roughly equal importance on precise T_1 vs. T_2 estimation and zero weight on proportionality constant estimation (obviating the need for complex differentiation in (4.3)). Since Ψ then depends on the constant through a scale factor, it suffices to fix the constant as 1 and design the

latent object parameter range as $\mathbb{X}^t \leftarrow \mathbb{T}_1^t \times \mathbb{T}_2^t \times 1$. Here, $\mathbb{T}_1^t \leftarrow [800, 1400]\text{ms}$ and $\mathbb{T}_2^t \leftarrow [50, 120]\text{ms}$ correspond with WM and GM regions of interest (ROIs) at 3T [77, 78]. We take $\mathbb{N}^t \leftarrow [0.9, 1.1]$ to account for 10% transmit field spatial variation. Broadened ranges $\mathbb{X}^b \leftarrow [400, 2000]\text{ms} \times [40, 200]\text{ms} \times 1$ and $\mathbb{N}^b \leftarrow [0.5, 2]$ are constructed to encourage solutions robust to a wide range of object parameters. We assume constant noise variance $\sigma_1^2 = \dots = \sigma_D^2 := \sigma^2$, where $\sigma^2 \leftarrow 1.49 \times 10^{-7}$ is selected to reflect measurements from normalized phantom datasets (*cf.* Sections 4.4.2.1 and ?? for acquisition details). Lastly, we set $\delta \leftarrow 0.01$ to select a robust scan parameter $\hat{\mathbf{P}}$ with associated worst-case cost $\tilde{\Psi}^t(\hat{\mathbf{P}})$ within 1% of global optimum $\tilde{\Psi}^t(\check{\mathbf{P}})$.

4.3.2 Scan Profile Comparisons

We solve (4.6) and (4.9) via grid search to allow illustration of $\tilde{\Psi}^t(\mathbf{P})$ as well as worst-case T_1, T_2 standard deviations $\tilde{\sigma}_{T_1}^t(\mathbf{P})$ and $\tilde{\sigma}_{T_2}^t(\mathbf{P})$, each defined as

$$\tilde{\sigma}_{T_1}^t(\mathbf{P}) := \max_{\substack{\mathbf{x} \in \mathbb{X}^t \\ \boldsymbol{\nu} \in \mathbb{N}^t}} \sigma_{T_1}(\mathbf{x}; \boldsymbol{\nu}, \mathbf{P}); \quad (4.11)$$

$$\tilde{\sigma}_{T_2}^t(\mathbf{P}) := \max_{\substack{\mathbf{x} \in \mathbb{X}^t \\ \boldsymbol{\nu} \in \mathbb{N}^t}} \sigma_{T_2}(\mathbf{x}; \boldsymbol{\nu}, \mathbf{P}), \quad (4.12)$$

where $\sigma_{T_1}(\mathbf{x}; \boldsymbol{\nu}, \mathbf{P})$ and $\sigma_{T_2}(\mathbf{x}; \boldsymbol{\nu}, \mathbf{P})$ are corresponding diagonal elements of inverse Fisher matrix $\mathbf{F}^{-1}(\mathbf{x}; \boldsymbol{\nu}, \mathbf{P})$. Grid searches for the (2, 1), (1, 1), and (0, 2) profiles each took about 4, 43, and 28 minutes, respectively. All experiments described hereafter were carried out using MATLAB[®] R2013a on a 3.5GHz desktop with 32GB RAM.

Table 4.1 compares optimized scan parameters for profiles consisting of (2, 1), (1, 1), and (0, 2) SPGR and DESS scans, respectively. In addition to $\tilde{\sigma}_{T_1}^t(\hat{\mathbf{P}})$ and $\tilde{\sigma}_{T_2}^t(\hat{\mathbf{P}})$, Table 4.1 presents analogous worst-case standard deviations $\tilde{\sigma}_{T_1}^b(\hat{\mathbf{P}})$ and $\tilde{\sigma}_{T_2}^b(\hat{\mathbf{P}})$ over $\mathbb{X}^b \times \mathbb{N}^b$ to show how each estimator degrades over the broadened object parameter range. When viewed over tight range $\mathbb{X}^t \times \mathbb{N}^t$, the (0, 2) profile provides a 11.5% reduction in worst-case cost over the other choices. Extending to broadened range $\mathbb{X}^b \times \mathbb{N}^b$, this reduction grows dramatically to 31.4%. We thus observe that while the different optimized profiles afford similar estimator precision over a narrow range of interest, the (0, 2) profile may be preferable due to its robustness to a wide range of object parameters.

Fig. 4.1 displays heat maps of worst-case latent parameter standard deviations $\tilde{\sigma}_{T_1}^t, \tilde{\sigma}_{T_2}^t$ and worst-case cost $\tilde{\Psi}^t$ as pairs of flip angles are varied away from the optimized scan design $\hat{\mathbf{P}}$. Boxes group subfigures corresponding to the same scan profile. Viewing the bottom row of subfigures, it is evident that $\tilde{\Psi}^t(\hat{\mathbf{P}})$ takes similar values for the different scan




Figure 4.1: Worst-case standard deviations $\tilde{\sigma}_{T_1}^t$ (top), $\tilde{\sigma}_{T_2}^t$ (middle), and cost $\tilde{\Psi}^t$ (bottom), versus pairs of nominal flip angles, holding other scan parameters fixed at selected profile $\hat{\mathbf{P}}$. Subfigures (a)-(i), (j)-(l), and (m)-(o) correspond to scan profiles containing $(S_{\text{SPGR}}, S_{\text{DESS}}) = (2, 1), (1, 1), \text{ and } (0, 2)$ SPGR and DESS scans, respectively. Selected scan parameters (starred) are within $\delta = 1\%$ of global minimizers and retain as much estimator precision as possible over a wide range of latent object parameters. All axes range from 5 to 90 degrees, in 5-degree increments. Colorbar ranges are $[0, 100]$, $[0, 10]$, and $[0, 20]$ milliseconds for rows of $\tilde{\sigma}_{T_1}^t$, $\tilde{\sigma}_{T_2}^t$, and $\tilde{\Psi}^t$ subfigures, respectively. The optimized (0, 2) profile appears most robust to flip angle variation.

Scan	(2, 1)	(1, 1)	(0, 2)
$\hat{\alpha}_0^{\text{SPGR}}$	(15,5) $^\circ$	15 $^\circ$	–
$\hat{\alpha}_0^{\text{DESS}}$	30 $^\circ$	10 $^\circ$	(35,10) $^\circ$
\hat{T}_R^{SPGR}	(12.2, 12.2)	13.9	–
\hat{T}_R^{DESS}	17.5	28.0	(24.4, 17.5)
$\tilde{\sigma}_{T_1}^t(\hat{\mathbf{P}})$	28	27	21
$\tilde{\sigma}_{T_1}^b(\hat{\mathbf{P}})$	154	169	113
$\tilde{\sigma}_{T_2}^t(\hat{\mathbf{P}})$	1.3	2.8	1.5
$\tilde{\sigma}_{T_2}^b(\hat{\mathbf{P}})$	9.1	8.8	6.0
$\tilde{\Psi}^t(\hat{\mathbf{P}})$	4.0	4.9	3.5
$\tilde{\Psi}^b(\hat{\mathbf{P}})$	17.7	17.9	12.2

Table 4.1: Performance summary of different scan profiles, optimized by solving (4.9) subject to scan time constraint $T_{\max} = 41.9\text{ms}$. The first row defines each profile. The next four rows describe $\hat{\mathbf{P}}$. The latter three pairs of rows show how worst-case values degrade from tight to broad ranges. Flip angles are in degrees; all other values are in milliseconds.

{table:profile}

profiles. However, it is apparent that the $(S_{\text{SPGR}}, S_{\text{DESS}}) = (0, 2)$ profile is substantially more robust to flip angle variation than other tested profiles (namely, (2, 1) and (1, 1)). Optimized worst-case cost over broadened latent parameter ranges $\tilde{\Psi}^b(\hat{\mathbf{P}})$ captures this by expanding the range of possible flip angles from $\mathbb{N}^t = [0.9, 1.1]$ to $\mathbb{N}^b = [0.5, 2]$ to account for factor-of-two spatial variation in relative flip angle. As a result, we find that the properties of “broad” search criterion $\tilde{\Psi}^b$ provide a stronger reason to select the (0, 2) scan for joint T_1, T_2 estimation in the brain than the properties of “tight” search criterion $\tilde{\Psi}^t$.

As the DESS sequence has already found success for T_2 mapping from even one scan [59], it is reassuring but unsurprising that our analysis finds two DESS scans to yield the most precise T_2 estimates. More interestingly, our methods suggest that, with a minimum $S_{\text{DESS}} = 2$ scans, DESS can be used to simultaneously estimate T_1 as well. In fact, for certain choices of parameter ranges, a second DESS scan is predicted to afford \hat{T}_1 precision comparable to two SPGR scans.

4.4 Experimental Validation and Results

{s,scn-dsgn,exp}

To test our approach to optimized scan design (described in Section 4.2.2), we estimate \mathbf{T}_1 and \mathbf{T}_2 maps (using maximum likelihood (ML) and regularized likelihood (RL) methods detailed in Section 3.2.2) from datasets collected using the scan profiles optimized in Section 4.3. In Section 4.4.1, we study estimator statistics from simulated data. In Sec-

tions 4.4.2-4.4.3, we progress to phantom and *in vivo* datasets to evaluate scan profile performance under increasingly complex settings. For the latter experiments, we use reference parameter maps from classical (long) pulse sequences, in lieu of ground truth maps.

4.4.1 Numerical Simulations

We select T_1 and T_2 WM and GM values based on previously reported measurements at 3T [77, 78] and extrapolate other nuisance latent object parameters m_0 and T_2^* from measurements at 1.5T [79]. We assign these parameter values to the discrete anatomy of the BrainWeb digital phantom [80, 79] to create ground truth \mathbf{M}_0 , \mathbf{T}_1 , \mathbf{T}_2 , and \mathbf{T}_2^* maps. We then choose acquisition parameters based on Table 4.1 (with fixed $T_E = 4.67\text{ms}$) and apply models (2.27) and (2.41)-(2.43) to the 81st slices of these true maps to compute noiseless 217×181 SPGR and DESS image-domain data, respectively.

For each scan profile, we corrupt the corresponding (complex) noiseless dataset \mathbf{S} with additive complex Gaussian noise, whose variance $\sigma^2 \leftarrow 1.49 \times 10^{-7}$ is set to match CRB calculations. This yields realistically noisy datasets \mathbf{Y} ranging from 105-122 signal-to-noise ratio (SNR), where SNR is defined here as

$$\text{SNR}(\mathbf{S}, \mathbf{Y}) := \frac{\|\mathbf{S}\|_F}{\|\mathbf{Y} - \mathbf{S}\|_F}. \quad (4.13)$$

We use each profile’s noisy magnitude dataset $|\mathbf{Y}|$ to compute estimates $\hat{\mathbf{T}}_1$ and $\hat{\mathbf{T}}_2$. We then evaluate estimator bias and variance from latent ground truth \mathbf{T}_1 and \mathbf{T}_2 maps.

In these simulations, we intentionally neglect to model a number of physically realistic effects because their inclusion would complicate study of estimator statistics. First and foremost, we assume knowledge of a uniform transmit field, to avoid confounding s^t and T_1, T_2 estimation errors. For a similar reason, spatial variation in the sensitivity of a single receive coil is also not considered. We omit modeling partial volume effects to ensure deterministic knowledge of WM and GM ROIs. We will explore the influence of these (and other) nuisance effects on scan design in later subsections and chapters.

To isolate bias due to estimator nonlinearity from regularization bias, we solve ML problem (3.7) only, and do not proceed to solve RL problem (3.8). This permits consideration of T_1, T_2 estimation from each of the 7733 WM or 9384 GM data points as voxel-wise independent realizations of the same estimation problem. To minimize quantization bias, we optimize (3.7) using a finely spaced dictionary of signal vectors from 1000 T_1 and T_2 values logarithmically spaced between $[10^2, 10^{3.5}]$ and $[10^1, 10^{2.5}]$, respectively. Using 10^6 dictionary elements, solving (3.7) took less than 7 minutes for each tested scan design $\hat{\mathbf{P}}$.

Scan	(2, 1)	(1, 1)	(0, 2)	Truth
WM \hat{T}_1^{ML}	830 ± 17	830 ± 15	830 ± 14	832
GM \hat{T}_1^{ML}	$1330 \pm 30.$	1330 ± 24	1330 ± 24	1331
WM \hat{T}_2^{ML}	$80. \pm 1.0$	$80. \pm 2.1$	79.6 ± 0.94	79.6
GM \hat{T}_2^{ML}	$110. \pm 1.4$	$110. \pm 3.0$	$110. \pm 1.6$	110

Table 4.2: Sample means \pm sample standard deviations of \mathbf{T}_1 and \mathbf{T}_2 ML estimates in WM and GM ROIs of simulated data, compared across different optimized scan profiles. Sample means exhibit insignificant bias, and sample standard deviations are consistent with worst-case standard deviations $\tilde{\sigma}_{T_1}^t$ and $\tilde{\sigma}_{T_2}^t$ reported in Table 4.1. All values are reported in milliseconds.

{table:numerical}

Table 4.2 verifies² that, despite model nonlinearity and Rician noise, estimation bias in WM- and GM-like voxels is negligible. Sample standard deviations are consistent with $\tilde{\sigma}_{T_1}^t$ and $\tilde{\sigma}_{T_2}^t$ (cf. Table 4.1). We observe that the (1, 1) and (0, 2) profiles afford high \hat{T}_1^{ML} precision, while the (2, 1) and (0, 2) scans afford high \hat{T}_2^{ML} precision. In agreement with the predictions of $\tilde{\Psi}^t$ and $\tilde{\Psi}^b$, these simulation studies suggest that at these SNR levels, an optimized profile containing 2 DESS scans can permit \mathbf{T}_1 and \mathbf{T}_2 estimation precision in WM and GM comparable to optimized profiles containing SPGR/DESS combinations.

Fig. 4.2 histograms (voxel-wise independent) ML estimates \hat{T}_1^{ML} and \hat{T}_2^{ML} from the (0, 2) scan profile. Each histogram is over a WM or GM ROI, within which all voxels are assigned the same single-component true T_1 and T_2 nominal value, listed in Table 4.2.

Overlaid in dashed maroon are normal distributions with latent means T_1 and T_2 and variances computed from the Fisher matrix at T_1, T_2 values in WM or GM. It is apparent that despite finite SNR and Rician noise, \hat{T}_1^{ML} and \hat{T}_2^{ML} exhibit negligible bias and near-Gaussian shape, suggesting locally linear behavior of the DESS signal model in T_1 and T_2 (\hat{T}_1^{ML} and \hat{T}_2^{ML} distributions from other profiles are similar).

The subfigures of Fig. 4.2 superimpose in solid green a second set of normal distributions, with the same means T_1 and T_2 as before, but worst-case standard deviations $\tilde{\sigma}_{T_1}^t$ and $\tilde{\sigma}_{T_2}^t$. The separations between these distribution pairs visually depict how estimator variances specific to WM or GM T_1 and T_2 values differ from worst-case variances. Using the fixed latent object parameters to optimize scan profiles can tailor scans for precise estimation in *either* WM *or* GM. In contrast, the proposed min-max formulation finds scan parameters that ensure precise estimation in *both* WM *and* GM.

²Each sample statistic presented in this chapter is rounded off to the highest place value of its corresponding uncertainty measure. For simplicity, each uncertainty measure is itself endowed one extra significant figure. Decimal points indicate the significance of trailing zeros.




Figure 4.2: Histograms of T_1 and T_2 estimates from noisy independent measurements of a *single* nominal WM or GM value. In each plot, two normal distributions are overlaid, each with latent means T_1 and T_2 . In (a)-(b) and (c)-(d), the solid green curve is $\mathcal{N}(T_1, (\tilde{\sigma}_{T_1}^t)^2)$ and $\mathcal{N}(T_2, (\tilde{\sigma}_{T_2}^t)^2)$, respectively. In (a)-(d), the dashed maroon curves have variances computed from the Fisher information at *a priori* unknown T_1, T_2 values in WM or GM. These plots correspond to an optimized $(0, 2)$ scan profile; analogous plots for other profiles are visually similar. At realistic noise levels, parameter estimates distribute with minimal bias and near-Gaussian shape. Thus, the CRB reliably approximates \hat{T}_1^{ML} and \hat{T}_2^{ML} errors.

{fig:normality}

4.4.2 Phantom Experiments

This subsection describes two experiments. In the first experiment, we compare SPGR/DESS scan profiles described in Table 4.1 (as well as a reference profile consisting of IR and SE scans) against nuclear magnetic resonance (NMR) measurements from the National Institute for Standards and Technology (NIST) [1]. These measurements provide information about *ROI sample means* and *ROI sample standard deviations* (Fig. 4.5), which we define as first- and second-order statistics computed across voxels within an ROI. In the second experiment, we repeat the SPGR/DESS scan profiles 10 times and compute *sample standard deviation maps* across repetitions. Taking ROI sample means of these maps gives *pooled sample standard deviations* (Table 4.4), which indicate relative scan profile precision.

4.4.2.1 Within-ROI Statistics

We acquire combinations of (2, 1), (1, 1), and (0, 2) SPGR and DESS coronal scans of a High Precision Devices® MR system phantom T_2 array. For each scan profile, we prescribe the optimized flip angles $\hat{\alpha}_0$ and repetition times \hat{T}_R listed in Table 4.1, and hold all other scan parameters fixed. We achieve the desired nominal flip angles by scaling a 20mm slab-selective Shinnar-Le Roux excitation [81], of duration 1.28ms and time-bandwidth product 4. For each DESS (SPGR) scan, we apply 2 (10) spoiling phase cycles over a 5mm slice thickness. We acquire all steady-state phantom and *in vivo* datasets with a $256 \times 256 \times 8$ matrix over a $240 \times 240 \times 30$ mm³ field of view (FOV). Using a 31.25kHz readout bandwidth, we acquire all data at minimum $T_E \leftarrow 4.67$ ms before or after RF excitations. To avoid slice-profile effects, we sample k-space over a 3D Cartesian grid. After Fourier transform of the raw datasets, only one of the excited image slices is used for subsequent parameter mapping. Including time to reach steady-state, each steady-state scan profile requires 1m37s scan time.

To validate a reference scan profile for use in *in vivo* experiments, we also collect 4 IR and 4 SE scans. For (phase-sensitive, SE) IR, we hold $(T_R, T_E) \leftarrow (1400, 14)$ ms fixed and vary (adiabatic) inversion time $T_I \in \{50, 150, 450, 1350\}$ ms across scans. For SE, we similarly hold $T_R \leftarrow 1000$ ms fixed and vary echo time $T_E \in \{10, 30, 60, 150\}$ ms across scans. We prescribe these scan parameters to acquire 256×256 datasets over the same $240 \times 240 \times 5$ mm³ slice processed from the SPGR/DESS datasets. Each IR and SE scan requires 5m58s and 4m16s, for a total 40m58s scan time.

We additionally collect a pair of Bloch-Siegert shifted 3D SPGR scans for separate transmit field estimation [82]. We insert a 9ms Fermi pulse at ± 8 kHz off-resonance into an SPGR sequence immediately following on-resonant excitation. We estimate regularized

transmit field maps [83] from the resulting pair of datasets. We normalize this transmit field map estimate by the 0.075G peak Fermi pulse amplitude to estimate transmit coil spatial variation map s_t . After calibration via separate measurements, we take s_t as known. For consistency, we account for flip angle variation when estimating T_1 and T_2 from both candidate (SPGR/DESS) and reference (IR/SE) scan profiles. With a repetition time of 21.7ms, this transmit field mapping acquisition requires 1m40s total scan time.

We acquire all phantom datasets using a GE DiscoveryTM MR750 3.0T scanner with an 8-channel receive head array. We separately normalize and combine coil data from each scan profile using a natural extension of [84] to the case of multiple datasets. For each optimized SPGR/DESS scan profile \hat{P} , we pre-cluster known parameter maps N into 10 clusters using k -means++ [85] and use each of the 10 cluster means to compute a corresponding dictionary of signal vectors from 300 T_1 and T_2 values logarithmically spaced between $[10^{1.5}, 10^{3.5}]$ and $[10^{0.5}, 10^3]$, respectively. We then iterate over clusters and use each dictionary in conjunction with corresponding coil-combined magnitude image data to produce ML parameter estimates $\hat{X}_{ML}(N, \hat{P})$. We subsequently solve RL problem (3.8) with initialization $\hat{X}_{ML}(N, \hat{P})$ to obtain regularized estimates $\hat{X}_{RL}(N, \hat{P})$ for each \hat{P} . We design regularizer (3.9) to encourage RL parameter estimates from different scan profiles to exhibit similar levels of smoothness. Letting $l \in \{1, 2, 3\}$ enumerate latent object parameters T_1 , T_2 , and the proportionality constant, we choose mild regularization parameters $(\beta_1, \beta_2, \beta_3) := D \times (2^{-21}, 2^{-23}, 2^{-26})$ to scale with the number of datasets and fix shape parameters $(\gamma_1, \gamma_2, \gamma_3) := (2^5 \text{ ms}, 2^2 \text{ ms}, 2^{-2})$ to values on the order of anticipated standard deviations. We iteratively update X until convergence criterion

$$\{\text{eq:conv,crit}\} \quad \|X^{(i)} - X^{(i-1)}\|_F < 10^{-7} \|X^{(i)}\|_F \quad (4.14)$$

is satisfied. For all steady-state profiles tested, ML initializations and RL reconstructions of phantom datasets require less than 3m30s and 9s, respectively.

We next describe sequential³ T_1 , then T_2 estimation from IR and SE reference scans. We first jointly coil-combine all 8-channel IR and SE phantom datasets to produce complex images. We next estimate T_1 along with a (nuisance parameter) inversion efficiency map via (3.7) and (3.8) from the 4 complex coil-combined IR images. By using the same flip angle scaling map s_t as is used for SPGR/DESS profiles, we estimate T_1 using a signal model similar to one proposed in [87], which accounts for imperfect excitation/refocusing

³ We initially attempted to circumvent sequential T_1 , then T_2 estimation by instead jointly estimating M_0 , T_1 , T_2 , and inversion efficiency from the IR and SE datasets together. Even using magnitude data and signal models, this resulted in heavily biased parameter maps, possibly due to the dependence of adiabatic inversion efficiency on relaxation parameters [86].




Figure 4.3: Colorized T_1 and T_2 ML and RL estimates from an HPD[®] quantitative phantom. Columns correspond to scan profiles consisting of (2 SPGR, 1 DESS), (1 SPGR, 1 DESS), (0 SPGR, 2 DESS), and (4 IR, 4 SE) acquisitions. Rows distinguish T_1 and T_2 ML and RL estimators. Fig. 4.4 provides identical grayscale images that enumerate vials. Colorbar ranges are in milliseconds.

{fig:hpj, jet}

and imperfect inversion. We then take both T_1 and s_t as known and estimate T_2 along with nuisance parameter M_0 (accounting for imperfect excitation/refocusing and incomplete recovery) via (3.7) and (3.8) from the 4 complex coil-combined SE images. We hold all other reconstruction details identical to those of SPGR/DESS reconstructions.

Figs. 4.3-4.4 compare in color and grayscale phantom T_1 and T_2 ML and RL estimates from optimized scan profiles. Vials are enumerated in Fig. 4.5 in descending T_1 and T_2 order. Vials corresponding to tight \mathbb{X}^t and broad \mathbb{X}^b parameter ranges are highlighted with




Figure 4.4: Grayscale T_1 and T_2 ML and RL estimates from an HPD[®] quantitative phantom. Columns correspond to scan profiles consisting of (2 SPGR, 1 DESS), (1 SPGR, 1 DESS), (0 SPGR, 2 DESS), and (4 IR, 4 SE) acquisitions. Rows distinguish T_1 and T_2 ML and RL estimators. Vials are enumerated and color-coded to correspond with data points in Fig. 4.5. Fig. 4.3 provides identical colorized images. Colorbar ranges are in milliseconds.

{fig:hpd,gray}

orange and yellow labels, respectively. Within these vials of interest, parameter maps from different scans appear visually similar.

In higher- T_1 vials (and the surrounding water), more bias is apparent in \hat{T}_1 ML and RL estimates from the (0, 2) scan profile than from the (2, 1) and (1, 1) scan profiles. With the signal models used in this study, the images suggest that scan profiles consisting of at least one SPGR scan may offer increased protection against T_1 estimation bias.

Fig. 4.5 plots sample means and sample standard deviations computed within circular




Figure 4.5: Phantom within-ROI sample statistics of T_1 and T_2 ML and RL estimates from optimized SPGR/DESS and reference IR/SE scan profiles, versus NIST NMR measurements [1]. Markers and error bars indicate ROI sample means and ROI sample standard deviations within the 14 labeled and color-coded vials in Fig. 4.4. Tight \mathbb{X}^t and broad \mathbb{X}^b latent parameter ranges are highlighted in orange and yellow, respectively. Table 4.3 replicates sample statistics within Vials 5-8. Our MR measurements are at 293K and NIST NMR measurements are at 293.00K. Within the designed parameter ranges, estimates from different acquisitions are in reasonable agreement with NIST measurements.

{fig:hpdl,ml-rls}

	(2SP,1DE)	(1SP,1DE)	(0SP,2DE)	(4IR,4SE)	NIST NMR
V5 \hat{T}_1^{ML}	1450 \pm 50.	1380 \pm 41	1600 \pm 130	1380 \pm 44	1332 \pm 0.8
V5 \hat{T}_1^{RL}	1450 \pm 26	1370 \pm 16	1540 \pm 98	1380 \pm 37	
V6 \hat{T}_1^{ML}	1100 \pm 30.	1050 \pm 39	1120 \pm 39	1100 \pm 74	1044 \pm 3.2
V6 \hat{T}_1^{RL}	1100 \pm 15	1040 \pm 14	1110 \pm 16	1100 \pm 64	
V7 \hat{T}_1^{ML}	870 \pm 22	830 \pm 29	880 \pm 29	870 \pm 25	801.7 \pm 1.70
V7 \hat{T}_1^{RL}	865 \pm 7.1	820 \pm 11	860 \pm 18	870 \pm 21	
V8 \hat{T}_1^{ML}	680 \pm 12	640 \pm 18	670 \pm 12	658 \pm 8.8	608.6 \pm 1.03
V8 \hat{T}_1^{RL}	674 \pm 7.6	637 \pm 7.4	662 \pm 6.6	658 \pm 7.1	
V5 \hat{T}_2^{ML}	131 \pm 5.5	140 \pm 10.	141 \pm 8.4	143 \pm 4.9	133.27 \pm 0.073
V5 \hat{T}_2^{RL}	131 \pm 5.2	145 \pm 9.1	139 \pm 7.1	142 \pm 4.8	
V6 \hat{T}_2^{ML}	91 \pm 3.5	99 \pm 6.0	95 \pm 4.2	96 \pm 2.7	96.89 \pm 0.049
V6 \hat{T}_2^{RL}	91 \pm 3.4	104 \pm 6.2	93 \pm 3.7	96 \pm 2.6	
V7 \hat{T}_2^{ML}	64 \pm 2.2	69 \pm 3.9	65 \pm 2.1	69 \pm 1.2	64.07 \pm 0.034
V7 \hat{T}_2^{RL}	65 \pm 2.1	71 \pm 4.3	64 \pm 1.9	69 \pm 1.2	
V8 \hat{T}_2^{ML}	46 \pm 1.5	50. \pm 2.3	46 \pm 1.1	47.6 \pm 0.87	46.42 \pm 0.014
V8 \hat{T}_2^{RL}	46 \pm 1.5	50. \pm 2.3	46 \pm 1.0	47.5 \pm 0.85	

Table 4.3: Phantom within-ROI sample means \pm sample standard deviations of T_1 and T_2 estimates from optimized SPGR/DESS and reference IR/SE scan profiles, versus NIST NMR measurements (*cf.* slide 22 of e-poster corresponding to [1]). For sake of brevity, sample statistics corresponding only to phantom vials within (or nearly within) tight design range \mathbb{X}^t (color-coded orange in Fig. 4.4) are reported. Fig. 4.5 plots sample statistics for all vials. ‘V#’ abbreviates vial numbers. All values are reported in milliseconds.

ROIs of phantom T_1 and T_2 ML and RL estimates. The highlighted orange and yellow parameter spaces correspond to design ranges \mathbb{X}^t and \mathbb{X}^b . T_1 estimates from both the candidate (2, 1), (1, 1), and (0, 2) (SPGR, DESS) and reference (4, 4) (IR, SE) profiles are in reasonable agreement with NIST estimates [1] across the vial range. T_2 estimates from all profiles are also in good agreement with NIST for vials within \mathbb{X}^b . SPGR/DESS profiles likely underestimate large T_2 values ($\geq 200\text{ms}$) due to greater influence of diffusion in DESS [88, 89, 90], (studied further in Appendix B). SPGR/DESS profiles possibly overestimate and the IR/SE profile likely underestimates short ($\leq 30\text{ms}$) and very short ($\leq 15\text{ms}$) T_2 values, respectively, due to poorly conditioned estimation. Table 4.3 replicates sample statistics in Fig. 4.5 for vials 5-8. Compared to ML initializations, (weakly) regularized estimates reduce error bars without introducing substantial additional bias.

	(2SP,1DE)	(1SP,1DE)	(0SP,2DE)
V5 $\hat{\sigma}_{T_1^{\text{ML}}}$	50 ± 12	$40 \pm 10.$	39 ± 9.4
V6 $\hat{\sigma}_{T_1^{\text{ML}}}$	70 ± 18	60 ± 15	70 ± 16
V7 $\hat{\sigma}_{T_1^{\text{ML}}}$	60 ± 13	50 ± 13	50 ± 13
V8 $\hat{\sigma}_{T_1^{\text{ML}}}$	23 ± 5.4	$20. \pm 4.7$	18 ± 4.3
V5 $\hat{\sigma}_{T_2^{\text{ML}}}$	2.6 ± 0.63	6 ± 1.4	3.5 ± 0.84
V6 $\hat{\sigma}_{T_2^{\text{ML}}}$	1.9 ± 0.46	5 ± 1.1	2.3 ± 0.54
V7 $\hat{\sigma}_{T_2^{\text{ML}}}$	1.4 ± 0.34	3.4 ± 0.80	1.5 ± 0.35
V8 $\hat{\sigma}_{T_2^{\text{ML}}}$	1.1 ± 0.26	3.5 ± 0.84	1.4 ± 0.33

Table 4.4: Phantom pooled sample standard deviations \pm pooled standard errors of sample standard deviations, from optimized SPGR/DESS scan profiles. Each entry is a measure of uncertainty of a typical voxel’s T_1 or T_2 ML estimate, estimated over 10 repeated acquisitions. For sake of brevity, sample statistics corresponding only to phantom vials within (or nearly within) tight design range \mathbb{X}^t (color-coded orange in Fig. 4.4) are reported. ‘V#’ abbreviates vial numbers. All values are reported in milliseconds.

4.4.2.2 Across-Repetition Statistics

In a second study, we repeat the (2, 1), (1, 1), and (0, 2) scan profiles 10 times each and separately compute T_1 and T_2 ML estimates for each repetition of each scan profile. We then estimate the standard deviation across repetitions on a per-voxel basis, to produce sample standard deviation maps for each profile. Each ROI voxel of the sample standard deviation map is a better estimate of the *population standard deviation* (which the CRB characterizes) than the ROI sample standard deviation from a single repetition, because the latter estimate is contaminated with slight spatial variation of voxel population means (due to imaging non-idealities such as Gibbs ringing due to k-space truncation).

Table 4.4 reports pooled sample standard deviations and pooled standard errors of the sample standard deviations (computed via expressions in [91]) for phantom vials within (or nearly within) tight design range \mathbb{X}^t (marked orange in Fig. 4.4). Due to error propagation from coil combination and s_t estimation, pooled ML sample standard deviations cannot be compared *in magnitude* to worst-case predicted standard deviations (Table 4.1); however, *trends* of empirical and theoretical standard deviations are overall similar. In particular, the optimized (0, 2) DESS-only scan profile affords T_1 ML estimation precision (in vials whose T_1, T_2 is similar to that of WM/GM) comparable to optimized (2, 1) and (1, 1) mixed (SPGR, DESS) profiles. Also in agreement with predictions, the optimized (2, 1) and (0, 2) profiles afford greater T_2 ML estimation precision than the optimized (1, 1) profile.

4.4.3 In Vivo Experiments

In a single long study of a healthy volunteer, we acquire the same optimized scan profiles containing (2, 1), (1, 1), and (0, 2) SPGR/DESS scans (*cf.* Table 4.1), as well as the reference profile containing (4, 4) IR/SE scans. We obtain axial slices from a 32-channel Nova Medical® receive head array. To address bulk motion between acquisitions and to compare within-ROI statistics, we rigidly register⁴ each coil-combined image to an IR image prior to parameter mapping. All acquisition and reconstruction details are otherwise the same as in phantom experiments (*cf.* Section 4.4.2.1). For all SS scan profiles tested, ML and RL reconstructions of brain datasets require less than 3m30s and 9s, respectively.

Fig. 4.6 compares brain T_1 and T_2 ML and RL estimates from optimized scan profiles. Though in-plane motion is largely compensated via registration, through-plane motion and non-bulk motion likely persist, and will influence ROI statistics. Due to motion (and scan duration) considerations, we examine within-ROI statistics from a single repetition as in Section 4.4.2.1, and do not attempt across-repetition statistics as in Section 4.4.2.2.

Visually, \hat{T}_1 maps from steady-state profiles exhibit similar levels of contrast in WM/GM regions well away from cerebrospinal fluid (CSF) as that seen in the reference \hat{T}_1 estimate. Since we did not optimize any scan profiles for estimation in high- T_1 regions, it is expected that greater differences may emerge in voxels containing or nearby CSF. In particular, T_1 is significantly underestimated within and near CSF by the (0, 2) DESS-only profile. This suggests that with the signal models used in this work, including at least one SPGR scan in an optimized profile may offer greater protection against estimation bias in high- T_1 regions.

Table 4.5 summarizes within-ROI sample means and sample standard deviations computed⁵ over four separate WM ROIs containing 96, 69, 224, and 148 voxels and one pooled cortical GM ROI containing 156 voxels. Within-ROI \hat{T}_1 sample standard deviations are comparable across SS profiles. In agreement with Table 4.1, T_2 estimates from the optimized (1, 1) scan profile exhibit higher within-ROI sample variation than corresponding (2, 1) and (0, 2) \hat{T}_2 maps. Compared to ML counterparts, RL estimates generally reduce within-ROI sample variation and do not significantly change within-ROI sample means.

In most cases, \hat{T}_1 within-ROI sample means from optimized SPGR/DESS scan profiles

⁴For each coil-combined dataset, we compute a separate 2D rigid transformation (with respect to the $T_1 = 50\text{ms}$ IR dataset) via the MATLAB® function `imregtform` and then apply the transformation via `imwarp`. We choose to use rigid transformations instead of affine distortions to avoid scaling; however in doing so we sacrifice compensating for small through-plane rotations. We do not find registration to substantially change subsequently estimated relaxation maps; however, this extra step substantially improves alignment of (especially cortical GM) ROIs in T_1 and T_2 estimates from different scan profiles.

⁵We have taken effort to select ROIs that reflect expected anatomy in all coil-combined and registered images, including adjacent slices in images from 3D acquisitions. However, we acknowledge the possibility of some contamination across tissue boundaries, especially WM and/or CSF contamination into cortical GM.





Figure 4.6: Colorized T_1 and T_2 ML and RL estimates from the brain of a healthy volunteer. Columns correspond to profiles consisting of (2 SPGR, 1 DESS), (1 SPGR, 1 DESS), (0 SPGR, 2 DESS), and (4 IR, 4 SE) acquisitions. Rows distinguish T_1 and T_2 ML and RL estimators. Colorbar ranges are in milliseconds.

{fig:brain,jet}

{fig:brain.roi.gray}



	ROI	(2SP,1DE)	(1SP,1DE)	(0SP,2DE)	(4IR,4SE)
\hat{T}_1^{ML}	AR WM	840 ± 32	770 ± 31	840 ± 43	780 ± 22
	AL WM	740 ± 61	660 ± 45	740 ± 55	760 ± 24
	PR WM	890 ± 88	860 ± 72	960 ± 84	810 ± 26
	PL WM	860 ± 70.	850 ± 61	880 ± 79	820 ± 37
	A GM	1200 ± 210	1200 ± 230	1300 ± 230	1300 ± 180
\hat{T}_1^{RL}	AR WM	840 ± 24	770 ± 20.	840 ± 43	780 ± 20.
	AL WM	740 ± 51	670 ± 37	740 ± 54	760 ± 23
	PR WM	890 ± 79	860 ± 61	960 ± 82	810 ± 24
	PL WM	870 ± 62	850 ± 50.	880 ± 78	820 ± 35
	A GM	1200 ± 200	1200 ± 220	1300 ± 230	1300 ± 180
\hat{T}_2^{ML}	AR WM	40. ± 1.3	54 ± 3.8	46 ± 1.5	55 ± 1.9
	AL WM	40. ± 1.7	50. ± 4.5	44 ± 1.7	53 ± 1.8
	PR WM	43 ± 2.7	60. ± 6.9	51 ± 3.6	59 ± 2.1
	PL WM	43 ± 1.8	57 ± 4.9	49 ± 2.5	57 ± 1.8
	A GM	50 ± 12	60 ± 15	60 ± 11	59 ± 6.0
\hat{T}_2^{RL}	AR WM	40. ± 1.3	54 ± 3.4	46 ± 1.5	55 ± 1.9
	AL WM	40. ± 1.7	50. ± 4.4	43 ± 1.7	53 ± 1.8
	PR WM	43 ± 2.8	60. ± 6.7	51 ± 3.7	58 ± 2.3
	PL WM	43 ± 1.7	57 ± 4.7	49 ± 2.5	57 ± 1.8
	A GM	50 ± 12	60 ± 15	60 ± 11	59 ± 6.4

{table:brain}

Table 4.5: *Left*: WM and GM ROIs, overlaid on a representative anatomical (coil-combined, IR) image. Separate WM ROIs are distinguished with anterior-right (AR), anterior-left (AL), posterior-right (PR), and posterior-left (PL) directions. Four small anterior (A) cortical GM polygons are pooled into a single ROI. *Right*: Within-ROI sample means ± within-ROI sample standard deviations of T_1 and T_2 ML and RL estimates from the brain of a healthy volunteer. Sample statistics are computed within ROIs indicated in the anatomical image. All values are reported in milliseconds.

do not deviate substantially from each other or from reference IR/SE measurements. Two notable exceptions are \hat{T}_1^{ML} in anterior left and posterior right WM from (1, 1) and (0, 2) profiles: these estimates are significantly lower and higher than analogous estimates from other profiles, respectively. Results thus suggest that the optimized (2, 1) scan profile yields WM \hat{T}_1^{ML} estimates that are more consistently similar to IR WM \hat{T}_1^{ML} estimates than other optimized SPGR/DESS profiles.

Systematic differences in \hat{T}_2 sample means are evident across scan profiles, particularly within WM ROIs. Curiously, the (1, 1) profile agrees most consistently (in WM/GM \hat{T}_2^{ML} within-ROI sample mean) with reference estimates, though with relatively high sample variation. The (2, 1) and (0, 2) SPGR/DESS profiles produce consistently lower WM \hat{T}_2^{ML} than the reference IR/SE profile, though the (0, 2) profile is in reasonable agreement with other steady-state estimates [92]. These discrepancies may due to differences in sensitivity

to multi-compartmental relaxation [93]. Specifically, different signal models with different scan parameter choices might be more or less sensitive to the model mismatch incurred by neglecting to distinguish the multiple T_2 components within each voxel. Chapter 6 studies multi-compartmental relaxation in much greater detail.

CHAPTER 5

MRI Parameter Estimation via Kernel Regression

{c,krr}

CHAPTER 6

Myelin Water Fraction Estimation from Steady-State Sequences

{c,mwf}

CHAPTER 7

Steady-State RF Pulse Design

{c,ss-rf}

CHAPTER 8

Future Work

{c,future}

APPENDIX A

Coil Data Combination from Multiple Datasets

{a,cc-multi}

APPENDIX B

DESS in the Presence of Diffusion

{a, dess-diff}

BIBLIOGRAPHY

- [1] K. E. Keenan, K. F. Stupic, M. A. Boss, S. E. Russek, T. L. Chenevert, P. V. Prasad, W. E. Reddick, K. M. Cecil, J. Zheng, P. Hu, and E. F. Jackson, “Multi-site, multi-vendor comparison of T1 measurement using ISMRM/NIST system phantom,” in *Proc. Intl. Soc. Mag. Res. Med.*, p. 3290, 2016.
- [2] G. Nataraj, J.-F. Nielsen, and J. A. Fessler, “Regularized, joint estimation of T1 and M0 maps,” in *Proc. Intl. Soc. Mag. Res. Med.*, p. 3128, 2014.
- [3] G. Nataraj, J.-F. Nielsen, and J. A. Fessler, “Model-based estimation of T2 maps with dual-echo steady-state MR imaging,” in *Proc. IEEE Intl. Conf. on Image Processing*, pp. 1877–81, 2014.
- [4] G. Nataraj, J.-F. Nielsen, and J. A. Fessler, “Optimizing MR scan design for model-based T1, T2 estimation from steady-state sequences,” *IEEE Trans. Med. Imag.*, 2016. To appear.
- [5] G. Nataraj, J.-F. Nielsen, and J. A. Fessler, “A min-max CRLB optimization approach to scan selection for relaxometry,” in *Proc. Intl. Soc. Mag. Res. Med.*, p. 1672, 2015.
- [6] G. Nataraj, J.-F. Nielsen, and J. A. Fessler, “Dictionary-free MRI parameter estimation via kernel ridge regression,” in *Proc. IEEE Intl. Symp. Biomed. Imag.*, 2017. Submitted.
- [7] G. Nataraj, J.-F. Nielsen, and J. A. Fessler, “Myelin water fraction estimation from optimized steady-state sequences using kernel ridge regression,” in *Proc. Intl. Soc. Mag. Res. Med.*, 2017. Submitted.
- [8] A. Macovski, *Medical imaging systems*. New Jersey: Prentice-Hall, 1983.
- [9] E. M. Haacke, R. W. Brown, M. R. Thompson, and R. Venkatesan, *Magnetic resonance imaging: Physical principles and sequence design*. New York: Wiley, 1999.
- [10] D. G. Nishimura, “Principles of magnetic resonance imaging,” 1996. Unpublished textbook.
- [11] F. Bloch, “Nuclear induction,” *Phys. Rev.*, vol. 70, pp. 460–74, Oct. 1946.
- [12] W. S. Hinshaw, “Image formation by nuclear magnetic resonance: The sensitive point method,” *J. Appl. Phys.*, vol. 47, p. 3709, Aug. 1976.

- [13] K. Scheffler, "A pictorial description of steady-states in rapid magnetic resonance imaging," *Concepts in Magnetic Resonance*, vol. 11, no. 5, pp. 291–304, 1999.
- [14] B. A. Hargreaves, S. S. Vasanawala, J. M. Pauly, and D. G. Nishimura, "Characterization and reduction of the transient response in steady-state MR imaging," *Mag. Res. Med.*, vol. 46, pp. 149–58, July 2001.
- [15] Y. Zur, M. L. Wood, and L. J. Neuringer, "Spoiling of transverse magnetization in steady-state sequences," *Mag. Res. Med.*, vol. 21, pp. 251–63, Oct. 1991.
- [16] V. Denolin, C. Azizieh, and T. Metens, "New insights into the mechanisms of signal formation in rf-spoiled gradient echo sequences," *Mag. Res. Med.*, vol. 54, pp. 937–54, Oct. 2005.
- [17] T. Redpath and R. A. Jones, "Fade-a new fast imaging sequence," *Mag. Res. Med.*, vol. 6, pp. 224–34, Feb. 1988.
- [18] H. Bruder, H. Fischer, R. Graumann, and M. Deimling, "A new steady-state imaging sequence for simultaneous acquisition of two MR images with clearly different contrasts," *Mag. Res. Med.*, vol. 7, pp. 35–42, May 1988.
- [19] J. B. Rosen, "The gradient projection method for nonlinear programming, Part I: Linear constraints," *SIAM J. Appl. Math.*, vol. 8, no. 1, pp. 181–217, 1960.
- [20] C. Byrne, "A unified treatment of some iterative algorithms in signal processing and image reconstruction," *Inverse Prob.*, vol. 20, pp. 103–20, Feb. 2004.
- [21] G. Golub and V. Pereyra, "Separable nonlinear least squares: the variable projection method and its applications," *Inverse Prob.*, vol. 19, pp. R1–26, Apr. 2003.
- [22] H.-L. M. Cheng, N. Stikov, N. R. Ghugre, and G. A. Wright, "Practical medical applications of quantitative MR relaxometry," *J. Mag. Res. Im.*, vol. 36, pp. 805–24, Oct. 2012.
- [23] J. A. Fessler and B. P. Sutton, "Nonuniform fast Fourier transforms using min-max interpolation," *IEEE Trans. Sig. Proc.*, vol. 51, pp. 560–74, Feb. 2003.
- [24] M. J. Muckley, D. C. Noll, and J. A. Fessler, "Fast parallel MR image reconstruction via B1-based, adaptive restart, iterative soft thresholding algorithms (BARISTA)," *IEEE Trans. Med. Imag.*, vol. 34, pp. 578–88, Feb. 2015.
- [25] A. Macovski, "Noise in MRI," *Mag. Res. Med.*, vol. 36, pp. 494–7, Sept. 1996.
- [26] T. Lei, "Statistics of MR signals: revisited," in *Proc. SPIE 6510 Medical Imaging 2007: Phys. Med. Im.*, p. 651052, 2007.
- [27] H. Gudbjartsson and S. Patz, "The Rician distribution of noisy MRI data," *Mag. Res. Med.*, vol. 34, pp. 910–4, Dec. 1995.

- [28] H. B. W. Larsson, J. Frederiksen, L. Kjaer, O. Henriksen, and J. Olesen, "In vivo determination of T1 and T2 in the brain of patients with severe but stable multiple sclerosis," *Mag. Res. Med.*, vol. 7, pp. 43–55, May 1988.
- [29] T. Kurki, N. Lundbom, M. Komu, and M. Kormano, "Tissue characterization of inter cranial tumors by magnetization transfer and spin-lattice relaxation parameters in vivo," *J. Mag. Res. Im.*, vol. 6, pp. 573–9, Aug. 1996.
- [30] E. Englund, A. Brun, Z. Gyroff-Wagner, E. Larsson, and B. Persson, "Relaxation times in relation to grade of malignancy and tissue necrosis in astrocytic gliomas," *Mag. Res. Im.*, vol. 4, no. 5, pp. 425–9, 1986.
- [31] S. Siemonsen, K. Mouridsen, B. Holst, T. Ries, J. Finsterbusch, G. Thomalia, L. Ostergaard, and J. Fiehler, "Quantitative T2 values predict time from symptom onset in acute stroke patients," *Stroke*, vol. 40, pp. 1612–6, May 2009.
- [32] L. D. DeWitt, J. P. Kistler, D. C. Miller, E. P. Richardson, and F. S. Buonanno, "NMR-neuropathologic correlation in stroke," *Stroke*, vol. 18, no. 2, pp. 342–51, 1987.
- [33] S. J. Matzat, J. V. Tiel, G. E. Gold, and E. H. G. Oei, "Quantitative MRI techniques of cartilage composition," *Quant. Imaging Med. Surg.*, vol. 3, pp. 162–74, June 2013.
- [34] T. J. Mosher and B. J. Dardzinski, "Cartilage MRI T2 relaxation time mapping: overview and applications," *Semin. Musculoskelet. Radiol.*, vol. 8, no. 4, pp. 355–68, 2004.
- [35] H. Guo, W.-Y. Au, J. S. Cheung, D. Kim, J. H. Jensen, P.-L. Khong, Q. Chan, K. C. Chan, C. Tosti, H. Tang, T. R. Brown, W. W. M. Lam, S.-Y. Ha, G. M. Brittenham, and E. X. Wu, "Myocardial T2 quantification in patients with iron overload at 3 Tesla," *J. Mag. Res. Im.*, vol. 30, pp. 394–400, Aug. 2009.
- [36] S. Giri, Y. C. Chung, A. Merchant, G. Mihai, S. Rajagopalan, S. V. Raman, and O. P. Simonetti, "T2 quantification for improved detection of myocardial edema," *Cardiovasc. Magn. Reson.*, vol. 11, no. 1, pp. 56–68, 2009.
- [37] D. C. Look and D. R. Locker, "Time saving in measurement of NMR and EPR relaxation times," *Rev Sci Instrum*, vol. 41, pp. 250–1, Feb. 1970.
- [38] H. Y. Carr and E. M. Purcell, "Effects of diffusion on free precession in nuclear magnetic resonance experiments," *Phys. Rev.*, vol. 94, pp. 630–8, May 1954.
- [39] M. K. Stehling, R. Turner, and P. Mansfield, "Echo-planar imaging: magnetic resonance imaging in a fraction of a second," *Science*, vol. 254, pp. 43–50, Oct. 1991.
- [40] C. B. Ahn, J. H. Kim, and Z. H. Cho, "High-speed spiral-scan echo planar NMR imaging - I," *IEEE Trans. Med. Imag.*, vol. 5, pp. 2–7, Mar. 1986.
- [41] C. H. Meyer, B. S. Hu, D. G. Nishimura, and A. Macovski, "Fast spiral coronary artery imaging," *Mag. Res. Med.*, vol. 28, pp. 202–13, Dec. 1992.

- [42] I. Kay and R. M. Henkelman, "Practical Implementation and Optimization of One-shot T1 imaging," *Mag. Res. Med.*, vol. 22, pp. 414–24, Dec. 1991.
- [43] P. A. Gowland and M. O. Leach, "Fast and accurate measurements of T1 using a multi-readout single inversion-recovery sequence," *Mag. Res. Med.*, vol. 26, pp. 79–88, July 1992.
- [44] D. R. Messroghli, A. Radjenovic, S. Kozerke, D. M. Higgins, M. U. Sivananthan, and J. P. Ridgway, "Modified Look-Locker inversion recovery (MOLLI) for high-resolution T_1 mapping of the heart," *Mag. Res. Med.*, vol. 52, pp. 141–6, July 2004.
- [45] M. K. Stehling, R. J. Ordidge, R. Coxon, and P. Mansfield, "Inversion-recovery Echo-planar imaging (IR-EPI) at 0.5T," *Mag. Res. Med.*, vol. 13, pp. 514–7, Mar. 1990.
- [46] J.-M. Bonny, M. Zanca, J.-Y. Boire, and A. Veyre, "T2 maximum likelihood estimation from multiple spin-echo magnitude images," *Mag. Res. Med.*, vol. 36, pp. 287–93, Aug. 1996.
- [47] D. Kumar, T. D. Nguyen, S. A. Gauthier, and A. Raj, "Bayesian algorithm using spatial priors for multiexponential T2 relaxometry from multiecho spin echo MRI," *Mag. Res. Med.*, vol. 68, pp. 1536–43, Nov. 2012.
- [48] N. Ben-Eliezer, D. K. Sodickson, and K. T. Block, "Rapid and accurate T2 mapping from multi-spin-echo data using Bloch-simulation-based reconstruction," *Mag. Res. Med.*, vol. 73, pp. 809–17, Feb. 2015.
- [49] T. D. Nguyen, C. Wisnieff, M. A. Cooper, D. Kumar, A. Raj, P. Spincemaille, Y. Wang, T. Vartanian, and S. A. Gauthier, "T2prep three-dimensional spiral imaging with efficient whole brain coverage for myelin water quantification at 1.5 tesla," *Mag. Res. Med.*, vol. 67, pp. 614–21, Mar. 2012.
- [50] S. Majumdar, S. C. Orphanoudakis, A. Gmitro, M. O'Donnell, and J. C. Gore, "Error in the measurements of T2 using multiple-echo MRI techniques: 1. Effect of radiofrequency pulse imperfections," *Mag. Res. Med.*, vol. 3, pp. 397–417, June 1986.
- [51] S. Majumdar, S. C. Orphanoudakis, A. Gmitro, M. O'Donnell, and J. C. Gore, "Error in the measurements of T2 using multiple-echo MRI techniques: 2. Effects of static field inhomogeneity," *Mag. Res. Med.*, vol. 3, pp. 562–74, Aug. 1986.
- [52] F. Farzaneh, S. J. Riederer, and N. J. Pelc, "Analysis of T2 limitations and off-resonance effects on spatial resolution and artifacts in echo-planar imaging," *Mag. Res. Med.*, vol. 14, pp. 123–39, Apr. 1990.
- [53] S. C. L. Deoni, B. K. Rutt, and T. M. Peters, "Rapid combined T1 and T2 mapping using gradient recalled acquisition in the steady state," *Mag. Res. Med.*, vol. 49, pp. 515–26, Mar. 2003.

- [54] L.-C. Chang, C. G. Koay, P. J. Basser, and C. Pierpaoli, "Linear least-squares method for unbiased estimation of T1 from SPGR signals," *Mag. Res. Med.*, vol. 60, pp. 496–501, Aug. 2008.
- [55] E. K. Fram, R. J. Herfkens, G. A. Johnson, G. H. Glover, J. P. Kaaris, A. Shimakawa, T. G. Perkins, and N. J. Pelc, "Rapid calculation of T1 using variable flip angle gradient refocused imaging," *Mag. Res. Im.*, vol. 5, no. 3, pp. 201–8, 1987.
- [56] H. Wang and Y. Cao, "Spatially regularized T1 estimation from variable flip angles MRI," *Med. Phys.*, vol. 39, pp. 4139–48, July 2012.
- [57] S. C. L. Deoni, H. A. Ward, T. M. Peters, and B. K. Rutt, "Rapid T_2 estimation with phase-cycled variable nutation steady-state free precession," *Mag. Res. Med.*, vol. 52, pp. 435–9, Aug. 2004.
- [58] S. C. L. Deoni, "Transverse relaxation time (T2) mapping in the brain with off-resonance correction using phase-cycled steady-state free precession imaging," *J. Mag. Res. Im.*, vol. 30, pp. 411–7, Aug. 2009.
- [59] G. H. Welsch, K. Scheffler, T. C. Mamisch, T. Hughes, S. Millington, M. Deimling, and S. Trattnig, "Rapid estimation of cartilage T2 based on double echo at steady state (DESS) with 3 Tesla," *Mag. Res. Med.*, vol. 62, pp. 544–9, Aug. 2009.
- [60] R. Heule, C. Ganter, and O. Bieri, "Rapid estimation of cartilage T2 with reduced T1 sensitivity using double echo steady state imaging," *Mag. Res. Med.*, vol. 71, pp. 1137–43, Mar. 2014.
- [61] T. Stöcker, F. Keil, K. Vahedipour, D. Brenner, E. Pracht, and N. J. Shah, "MR parameter quantification with magnetization-prepared double echo steady-state (MP-DESS)," *Mag. Res. Med.*, vol. 72, pp. 103–11, July 2014.
- [62] R. Heule, C. Ganter, and O. Bieri, "Triple echo steady-state (TESS) relaxometry," *Mag. Res. Med.*, vol. 71, pp. 230–7, Jan. 2014.
- [63] M. L. Gyngell, "The steady-state signals in short-repetition-time sequences," *J. Mag. Res.*, vol. 81, pp. 474–83, Feb. 1989.
- [64] W. Hänicke and H. U. Vogel, "An analytical solution for the SSFP signal in MRI," *Mag. Res. Med.*, vol. 49, pp. 771–5, Apr. 2003.
- [65] G. H. Weiss, R. K. Gupta, J. A. Ferretti, and E. D. Becker, "The choice of optimal parameters for measurement of spin-lattice relaxation times. I. Mathematical formulation," *J. Mag. Res.*, vol. 37, pp. 369–79, Feb. 1980.
- [66] Y. Zhang, H. N. Yeung, M. O'Donnell, and P. L. Carson, "Determination of sample time for T1 measurement," *J. Mag. Res. Im.*, vol. 8, pp. 675–81, May 1998.
- [67] H. Z. Wang, S. J. Riederer, and J. N. Lee, "Optimizing the precision in T1 relaxation estimation using limited flip angles," *Mag. Res. Med.*, vol. 5, pp. 399–416, Nov. 1987.

- [68] J. A. Jones, P. Hodgkinson, A. L. Barker, and P. J. Hore, "Optimal sampling strategies for the measurement of spin-spin relaxation times," *J. Mag. Res. B*, vol. 113, pp. 25–34, Oct. 1996.
- [69] J. Imran, François. Langevin, and Hervé. Saint-Jalmes, "Two-point method for T1 estimation with optimized gradient-echo sequence," *Mag. Res. Im.*, vol. 17, pp. 1347–56, Nov. 1999.
- [70] S. C. L. Deoni, T. M. Peters, and B. K. Rutt, "Determination of optimal angles for variable nutation proton magnetic spin-lattice, T_1 , and spin-spin, T_2 , relaxation times measurement," *Mag. Res. Med.*, vol. 51, pp. 194–9, Jan. 2004.
- [71] L. Fleysher, R. Fleysher, S. Liu, W. Zaaraoui, and O. Gonen, "Optimizing the precision-per-unit-time of quantitative MR metrics: Examples for T_1 , T_2 , and DTI," *Mag. Res. Med.*, vol. 57, pp. 380–7, Feb. 2007.
- [72] M. Akçakaya, S. Weingärtner, Sébastien. Roujol, and R. Nezafat, "On the selection of sampling points for myocardial T1 mapping," *Mag. Res. Med.*, vol. 73, pp. 1741–53, May 2015.
- [73] C. M. Lewis, S. A. Hurley, M. E. Meyerand, and C. G. Koay, "Data-driven optimized flip angle selection for T1 estimation from spoiled gradient echo acquisitions," *Mag. Res. Med.*, 2016. To appear.
- [74] R. A. Fisher, "Theory of statistical estimation," *Proc. Cambridge Philosophical Society*, vol. 22, pp. 700–25, July 1925.
- [75] H. Cramér, *Mathematical methods of statistics*. Princeton: Princeton Univ. Press, 1946.
- [76] H. Chernoff, "Locally optimal designs for estimating parameters," *Ann. Math. Stat.*, vol. 24, pp. 586–602, Dec. 1953.
- [77] J. P. Wansapura, S. K. Holland, R. S. Dunn, and W. S. Ball, "NMR relaxation times in the human brain at 3.0 Tesla," *J. Mag. Res.*, vol. 9, pp. 531–8, Apr. 1999.
- [78] G. J. Stanisz, E. E. Odobina, J. Pun, M. Escaravage, S. J. Graham, M. J. Bronskill, and R. M. Henkelman, " T_1 , T_2 relaxation and magnetization transfer in tissue at 3T," *Mag. Res. Med.*, vol. 54, pp. 507–12, Sept. 2005.
- [79] R. K.-S. Kwan, A. C. Evans, and G. B. Pike, "MRI simulation-based evaluation of image-processing and classification methods," *IEEE Trans. Med. Imag.*, vol. 18, pp. 1085–97, Nov. 1999.
- [80] D. L. Collins, A. P. Zijdenbos, V. Kollokian, J. G. Sled, N. J. Kabani, C. J. Holmes, and A. C. Evans, "Design and construction of a realistic digital brain phantom," *IEEE Trans. Med. Imag.*, vol. 17, pp. 463–8, June 1998.

- [81] J. Pauly, P. Le Roux, D. Nishimura, and A. Macovski, "Parameter relations for the Shinnar-Le Roux selective excitation pulse design algorithm," *IEEE Trans. Med. Imag.*, vol. 10, pp. 53–65, Mar. 1991.
- [82] L. I. Sacolick, F. Wiesinger, I. Hancu, and M. W. Vogel, "B1 mapping by Bloch-Siegert shift," *Mag. Res. Med.*, vol. 63, pp. 1315–22, May 2010.
- [83] H. Sun, W. A. Grissom, and J. A. Fessler, "Regularized estimation of Bloch-Siegert B1+ Maps in MRI," in *Proc. IEEE Intl. Conf. on Image Processing*, pp. 3646–50, 2014.
- [84] L. Ying and J. Sheng, "Joint image reconstruction and sensitivity estimation in SENSE (JSENSE)," *Mag. Res. Med.*, vol. 57, pp. 1196–1202, June 2007.
- [85] D. Arthur and S. Vassilvitskii, "K-means++: The advantages of careful seeding," in *Proc. 18th Annual ACM-SIAM Symp. Disc. Alg. (SODA)*, pp. 1027–35, 2007.
- [86] L. R. Frank, E. C. Wong, and R. B. Buxton, "Slice profile effects in adiabatic inversion: Application to multislice perfusion imaging," *Mag. Res. Med.*, vol. 38, pp. 558–64, Oct. 1997.
- [87] J. K. Barral, E. Gudmundson, N. Stikov, M. Etezadi-Amoli, P. Stoica, and D. G. Nishimura, "A robust methodology for in vivo T1 mapping," *Mag. Res. Med.*, vol. 64, pp. 1057–67, Oct. 2010.
- [88] C. E. Carney, S. T. S. Wong, and S. Patz, "Analytical solution and verification of diffusion effect in SSFP," *Mag. Res. Med.*, vol. 19, pp. 240–6, June 1991.
- [89] E. X. Wu and R. B. Buxton, "Effect of diffusion on the steady-state magnetization with pulsed field gradients," *J. Mag. Res.*, vol. 90, pp. 243–53, Nov. 1990.
- [90] R. Kaiser, E. Bartholdi, and R. R. Ernst, "Diffusion and field-gradient effects in NMR Fourier spectroscopy," *J. Chem. Phys.*, vol. 60, pp. 2966–79, Apr. 1974.
- [91] S. Ahn and J. A. Fessler, "Standard errors of mean, variance, and standard deviation estimators," Tech. Rep. 413, Comm. and Sign. Proc. Lab., Dept. of EECS, Univ. of Michigan, Ann Arbor, MI, 48109-2122, July 2003.
- [92] R. Heule, P. Bär, C. Mirkes, K. Scheffler, S. Trattnig, and O. Bieri, "Triple-echo steady-state T2 relaxometry of the human brain at high to ultra-high fields," *NMR in Biomedicine*, vol. 27, pp. 1037–45, Sept. 2014.
- [93] A. Mackay, K. Whittall, J. Adler, D. Li, D. Paty, and D. Graeb, "In vivo visualization of myelin water in brain by magnetic resonance," *Mag. Res. Med.*, vol. 31, pp. 673–7, June 1994.