

Bid Document

Bid Details	
Bid End Date/Time	16-08-2021 14:00:00
Bid Opening Date/Time	16-08-2021 14:30:00
Bid Life Cycle (From Publish Date)	90 (Days)
Bid Offer Validity (From End Date)	65 (Days)
Ministry/State Name	Delhi
Department Name	Municipal Corporation Of Delhi (mcd)
Organisation Name	N/a
Office Name	Mrs Girdhar Lal Maternity Hospital
Total Quantity	1
Item Category	Steam Sterlizer
Years of Past Experience required	3 Year (s)
MSE Exemption for Years of Experience and Turnover	No
Startup Exemption for Years of Experience and Turnover	No
Document required from seller	Experience Criteria,OEM Authorization Certificate *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer
Bid to RA enabled	Yes
Time allowed for Technical Clarifications during technical evaluation	2 Days
Evaluation Method	Total value wise evaluation

EMD Detail

Required	No
----------	----

ePBG Detail

Required	No
----------	----

Splitting

Bid splitting not applied.

MSE Purchase Preference

MSE Purchase Preference	No
-------------------------	----

MII Purchase Preference

MII Purchase Preference	No
-------------------------	----

1. Experience Criteria: In respect of the filter applied for experience criteria, the Bidder or its OEM {themselves or through reseller(s)} should have regularly, manufactured and supplied same or similar Category Products to any Central / State Govt Organization / PSU / Public Listed Company for number of Financial years as indicated above in the bid document before the bid opening date. Copies of relevant contracts to be submitted along with bid in support of having supplied some quantity during each of the Financial year. In case of bunch bids, the category of primary product having highest value should meet this criterion.

Steam Sterlizer (1 pieces)

Brand Type	Registered Brand
------------	------------------

Technical Specifications

[* As per GeM Category Specification](#)

Specification	Specification Name	Bid Requirement (Allowed Values)
Construction	Type of Steriliser	HORIZONTAL RECTANGULAR/SQUARE- Steam Steriliser Pressure type for Hospital and Pharmaceutical use, conforming to IS: 3829(part 1)/1999, fitted with Silicon door gasket(expansion type). Steriliser Chamber size (L x W x H): 1200 mm x 600 mm x 600 mm, Electric load of the unit shall be 18 KW.
	Steam supply to Steriliser	Built-in electrically heated Steam Generator (made of SS 316 (AISI 316) with low water protection system), it should work with 3 phase 400/440 V 50 Hz AC supply
	Material: Steriliser parts; Chamber, Jacket, Door, Door Ring and shall be made of	SS 316 or SS 316L
	Type of control	Automatic: MICROPROCESSOR CONTROLLED, WITH PRESELECTED STERILISATION PROGRAMS (variable programs customisable as per Buyer's requirement to deal with sterilisation and drying of wrapped/unwrapped loads at 121 & 134 degree C)
CONSTRUCTION	Whether a compatible water softener/reversed osmosis based system to feed water to the steam generator unit is to be supplied with the steriliser:	Yes
	Provision of Lock on Steriliser Chamber door:	Manual and Automatic Locks, both
	Type of Door:	Radial Type

Specification	Specification Name	Bid Requirement (Allowed Values)
	Display facility for Sterilisation Cycle parameters	Chamber Temperature, Cycle number, Alarm indicator, Low water indicator
	Print facility for record generation of dates, time, load, identification number and operating parameters (temperature and holding time etc):	Yes, by an inbuilt printer unit
	Warranty Period	5 Or higher
	No of Doors in Steriliser	2
	Whether Type Test Certificate from a Govt/NABL/ILAC approved testing laboratory is available (If Yes, to be furnished by Seller to Buyer on demand) for the offered product?	Yes
	Type of user friendly Interface (for microprocessor controlled Sterilizer) to specify parameters for running Sterilization cycle:	Keyboard interface
	Data Transfer facility through USB /RS232 port to computer/laptop on real time basis	Yes
Integrated Shredder Facility	Integrated Shredder (for shredding the load in Sterilisation chamber)	Not available
	Material: Shredder parts;	SS 316 or SS 316L
	Power of Motor used for Shredding	NA, no integrated Shredder unit offered
Accessories (made of SS 316) for loading/unloading and handling of Load	Number of Rails:	2 Or higher
	Long Handle:	1
	Sterilisation Carriages	2 Or higher
	Pull out Trays/Tanks	2
	Floor loading carts	2 Or higher
	Transfer carriages:	1, 2 Or higher

Consignees/Reporting Officer and Quantity

S.No.	Consignee/Reporting Officer	Address	Quantity	Delivery Days
1	Megha Tandon	110002,Ajmeri Gate	1	30

Buyer Added Bid Specific Additional Terms and Conditions

- Provisions for workmen compensation, insurance, public liability, property damage, insurance etc. to be ensured by the contractor. Also, safety measures prescribed for working in Extra High Voltage & use of personnel protective equipment must be adhered to.
- Scope of supply (Bid price to include all cost components) : Supply Installation Testing Commissioning of Goods and Training of operators and providing Statutory Clearances required (if any)
- Buyer Added text based ATC clauses
 - Should provide availability of spare and consumables for total 10 years.
 - The unit should be manufactured as per ISI/USFSA/CE (preferably ISI).
 - Electrical safety should be confirmed to standards for electrical safety IEC-60601/IS-12450.
 - should provide CMC rate year wise for next 5 years after warranty period of 5 years.
- Availability of Service Centres: Bidder/OEM must have a Functional Service Centre in the State of each Consignee's Location in case of carry-in warranty. (Not applicable in case of goods having on-site warranty). If service center is not already there at the time of bidding, successful bidder / OEM shall have to establish one within 30 days of award of contract. Payment shall be released only after submission of documentary evidence of having Functional Service Centre.
- Dedicated /toll Free Telephone No. for Service Support : BIDDER/OEM must have Dedicated/toll Free Telephone No. for Service Support.
- Escalation Matrix For Service Support : Bidder/OEM must provide Escalation Matrix of Telephone Numbers for Service Support.
- Bidder / OEM has to give an undertaking that after expiry of warranty period, it will provide Comprehensive Maintenance Service for next 5 years for the offered products at the rate not more than 5 % of contract price per annum. Buyer reserves the right to enter into a CMC agreement with the Successful Bidder / OEM after expiry of the Warranty period at above mentioned rate and the payment for the CMC charges would be made Quarterly after rendering of the CMC Services of the relevant CMC period. Performance Security of the successful bidder shall be forfeited if it fails to accept the CMC contract when called upon by the buyer. CMC would include cost of 2.5 (Upload the undertaking). The original Performance Security of contract will be returned only after submission and verification of AMC Performance Security for 5% of total CMC value valid up to CMC period plus 2 months (if there is no other claim).

Disclaimer

The additional terms and conditions have been incorporated by the Buyer after approval of the Competent Authority in Buyer Organization. Buyer organization is solely responsible for the impact of these clauses on the bidding process, its outcome and consequences thereof including any eccentricity / restriction arising in the bidding process due to these ATCs and due to modification of technical specification and / or terms and conditions governing the bid. Any clause incorporated by the Buyer such as demanding Tender Sample, incorporating any clause against the MSME policy and Preference to make in India Policy, mandating any Brand names or Foreign Certification, changing the default time period for Acceptance of material or payment timeline governed by OM of Department of Expenditure shall be null and void and would not be considered part of bid. Further any reference of conditions published on any external site or reference to external documents / clauses shall also be null and void. If any seller has any objection / grievance against these additional clauses or otherwise on any aspect of this bid, they can raise their representation against the same by using the

Representation window provided in the bid details field in Seller dashboard after logging in as a seller within 4 days of bid publication on GeM. Buyer is duty bound to reply to all such representations and would not be allowed to open bids if he fails to reply to such representations.

[This Bid is also governed by the General Terms and Conditions](#)

In terms of GeM GTC clause 26 regarding Restrictions on procurement from a bidder of a country which shares a land border with India, any bidder from a country which shares a land border with India will be eligible to bid in this tender only if the bidder is registered with the Competent Authority. While participating in bid, Bidder has to undertake compliance of this and any false declaration and non-compliance of this would be a ground for immediate termination of the contract and further legal action in accordance with the laws.

---Thank You---