LaBTECH

FSTxx-RKT-D5731-0-260720

Record Sheet

for on-site verification of the coaxial of platens

Worksheet No.:
Page.:

Equipment ID

Loadcell : _____
Dial Gauge : _____
Tyre Gauge : _____

Check No.	**P**rior to Testing [P] / After **Specimen No.** / **E**nd of Batch [E] / **R**e-verification [R]	**Orthogonal Point Reading, *d* (mm)**				**Offset Readings (mm)**		**Variance, *V* (mm)**	**Compliance Assessment**	**for Re-verification only**	
		Point A *(0°)*	Point B *(90°)*	Point C *(180°)*	Point D *(270°)*	d_{max}	d_{min}	$d_{max} - d_{min}$	*V* < 0.20mm [Y / N]	Pass *Resume testing*	Fail *Lockout the equipment*
1											
2											
3											
4											
5											
6											
7											
8											
9											
10											
11											
12											
13											
14											
15											
16											
17											
18											
19											
20											
21											
22											
23											
24											
25											
26											
27											
28											
29											
30											

Notes:

a) Before every verification,

i) Clean the upper platen, lower platen, and the internal bore of the CNC sleeves thoroughly to remove rock dust or grease. Inspect components for any structural pitting or physical chips.

ii) Check the staightness of the CNC sleeves after fitting to the upper and lower platens.

b) For re-verification, rotate the sleeve 180° and re-perofmr the verification to rule out sleeve error.

Recorded by: _____

Date: _____