Table of contents

\\USER

Khan

NeuroAnalytics

SNSX_TLE_7t(july2021)

localizer mp2rage_sag_700iso_p3 sa2rage_B1map t2_tse_cor_p2_500x500x1400 t2_tse_tra_p2_500x500x1400 mbep2d_diff_b2000_b1000_95dir_AP mbep2d_diff_B0_PA ASPIRE_800iso_7e dark-fluid_tse_cor_p2_.67x.67x2.5 cb_sp2d_diff_C52_Spiral_R3 gre_field_mapping 3D_TOF_3slabs_Flow_FH T2_spc_700iso_p3 cb_sp2d_diff_C26_PA_2iso_UFA cb_sp2d_diff_C26_AP_2iso_UFA_B0 spc_dir_sag_750iso_p2 T2_spc_600iso_Mayo T2_spc_600iso_Mayo_CP spc_dir_sag_750iso_p2

\\USER\Khan\NeuroAnalytics\SNSX_TLE_7t(july2021)\localizer

TA: 0:26 PM: REF Voxel size: 1.2×1.2×5.0 mmPAT: Off Rel. SNR: 1.00 : fl

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	Off
Start measurements	Single measurement

Routine

Routine		
Slice group	1	
Slices	3	
Dist. factor	400 %	
Position	Isocenter	
Orientation	Sagittal	
Phase enc. dir.	A >> P	
Slice group	2	
Slices	3	
Dist. factor	300 %	
Position	Isocenter	
Orientation	Transversal	
Phase enc. dir.	A >> P	
Slice group	3	
Slices	3	
Dist. factor	500 %	
Position	Isocenter	
Orientation	Coronal	
Phase enc. dir.	R >> L	
AutoAlign		
Phase oversampling	0 %	
FoV read	300 mm	
FoV phase	100.0 %	
Slice thickness	5.0 mm	
TR	12.0 ms	
TE	3.33 ms	
Averages	1	
Concatenations	9	
Filter	None	
Coil elements	HC2	

Contrast - Common

TR	12.0 ms
TE	3.33 ms
TD	0 ms
MTC	Off
Magn. preparation	None
Flip angle	15 deg
Fat suppr.	None
Water suppr.	None
SWI	Off

Contrast - Dynamic

Averages	1
Averaging mode	Short term
Reconstruction	Magnitude
Measurements	1

Contrast - Dynamic

Multiple series

Resolution - Common		
FoV read	300 mm	
FoV phase	100.0 %	
Slice thickness	5.0 mm	
Base resolution	256	
Phase resolution	90 %	
Phase partial Fourier	Off	
Interpolation	Off	

Each measurement

Resolution - iPAT

PAT mode	Mana
IPAI mode	None
1 / 11 111000	140110

Resolution - Filter Image

Image Filter	Off	
Distortion Corr.	Off	
Prescan Normalize	Off	
Normalize	Off	
B1 filter	Off	

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off

Geometry - Common

Geometry - AutoAlign

Slice group	1
Position	Isocenter
Orientation	Sagittal
Phase enc. dir.	A >> P
Slice group	2
Position	Isocenter
Orientation	Transversal

Geometry - AutoAlign

Phase enc. dir.	A >> P
Slice group	3
Position	Isocenter
Orientation	Coronal
Phase enc. dir.	R >> L
AutoAlign	
Initial Position	Isocenter
L	0.0 mm
Р	0.0 mm
Н	0.0 mm
Initial Rotation	0.00 deg
Initial Orientation	Sagittal

Geometry - Saturation

Saturation mode	Standard
Fat suppr.	None
Water suppr.	None
Special sat.	None

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table position	Н
Table position	0 mm
Inline Composing	Off

Geometry - Tim CT

Tim CT mode	Off
Slices	3
Slice thickness	5.0 mm
Dist. factor	500 %
FoV read	300 mm
FoV phase	100.0 %
Segments	1

System - Miscellaneous

Positioning mode	REF
Table position	F
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	
Coil Select Mode	Default

System - Adjustments

B0 Shim mode	Tune up
B1 Shim mode	TrueForm
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
A >> P	263 mm
R >> L	350 mm
F >> H	350 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Slice-sel.

System - Tx/Rx

Frequency 1H	296.941213 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	12.0 ms
Concatenations	9
Segments	1

Physio - Cardiac

Tagging	None
Magn. preparation	None
Fat suppr.	None
Dark blood	Off
FoV read	300 mm
FoV phase	100.0 %
Phase resolution	90 %

Physio - PACE

Resp. control	Off
Concatenations	9

Inline - Common

Subtract	Off
Measurements	1
StdDev	Off
Liver registration	Off
Save original images	On

Inline - MIP

		_
MIP-Sag	Off	
MIP-Cor	Off	
MIP-Tra	Off	
MIP-Time	Off	
Save original images	On	

Inline - Soft Tissue

Wash - In	Off
Wash - Out	Off
TTP	Off
PEI	Off
MIP - time	Off
Measurements	1

Inline - Composing

Inline Composing	Off	
Distortion Corr.	Off	

Inline - MapIt

Save original images	On
MapIt	None
Flip angle	15 deg
Measurements	1
Contrasts	1
TR	12.0 ms

SIEMENS MAGNETOM Investigational_Device_7T_Plus

Inline - MapIt

TF	3.33 ms
_	5.55 1118

Sequence - Part 1

Introduction	On
Dimension	2D
Phase stabilisation	Off
Asymmetric echo	Allowed
Contrasts	1
Flow comp.	No
Multi-slice mode	Interleaved
Bandwidth	230 Hz/Px

Sequence - Part 2

Segments	1
Acoustic noise reduction	None
RF pulse type	Normal
Gradient mode	Whisper
Excitation	Slice-sel.
RF spoiling	On

Mode	Off	

\\USER\Khan\NeuroAnalytics\SNSX_TLE_7t(july2021)\mp2rage_sag_700iso_p3

TA: 10:12 PM: FIX Voxel size: 0.7×0.7×0.7 mmPAT: 3 Rel. SNR: 1.00 : tfl

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further	Off
preparation	
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slab group	1
Slabs	1
Dist. factor	50 %
Position	R5.3 A13.3 F28.6 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
AutoAlign	
Phase oversampling	0 %
Slice oversampling	7.1 %
Slices per slab	224
FoV read	240 mm
FoV phase	100.0 %
Slice thickness	0.70 mm
TR	6000.0 ms
TE	2.74 ms
Averages	1
Concatenations	1
Filter	Raw filter
Coil elements	HC2

Contrast - Common

TR	6000.0 ms
TE	2.74 ms
Magn. preparation	Non-sel. IR
TI 1	800 ms
TI 2	2700 ms
Flip angle 1	4.0 deg
Flip angle 2	5.0 deg
Fat suppr.	None
Water suppr.	None

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	1
Multiple series	Each measurement

Resolution - Common

FoV read	240 mm	
FoV phase	100.0 %	
Slice thickness	0.70 mm	
Base resolution	342	
Phase resolution	100 %	
Slice resolution	100 %	
Phase partial Fourier	6/8	

Resolution - Common

Slice partial Fourier	6/8
Interpolation	Off

Resolution - iPAT

PAT mode	GRAPPA
Accel. factor PE	3
Ref. lines PE	24
Accel. factor 3D	1
Reference scan mode	Integrated

Resolution - Filter Image

Image Filter	Off	
Distortion Corr.	Off	
Prescan Normalize	Off	
Normalize	Off	
B1 filter	Off	

Resolution - Filter Rawdata

Raw filter	On	
Elliptical filter	Off	

Geometry - Common

Slab group	1
Slabs	1
Dist. factor	50 %
Position	R5.3 A13.3 F28.6 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
Slice oversampling	7.1 %
Slices per slab	224
FoV read	240 mm
FoV phase	100.0 %
Slice thickness	0.70 mm
TR	6000.0 ms
Multi-slice mode	Single shot
Series	Interleaved
Concatenations	1

Geometry - AutoAlign

Slab group	1
Position	R5.3 A13.3 F28.6 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
AutoAlign	
Initial Position	R5.3 A13.3 F28.6
R	5.3 mm
A	13.3 mm
F	28.6 mm
Initial Rotation	0.00 deg
Initial Orientation	Sagittal

Geometry - Navigator

Geometry - Tim Planning Suite

	_
Set-n-Go Protocol	Off
Table position	Н
Table position	0 mm
Inline Composing	Off

System - Miscellaneous

Positioning mode	FIX
Table position	Н
Table position	0 mm
MSMA	S-C-T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	
Coil Select Mode	Default

System - Adjustments

B0 Shim mode	Brain
B1 Shim mode	Volume-selective
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

! Position	R5.8 A6.9 F22.7 mm
! Orientation	Sagittal
! Rotation	0.00 deg
! A >> P	213 mm
! F >> H	173 mm
! R >> L	171 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	Volume-selective
Excitation	Non-sel.
pTx Volume	1
Vol. Property	B1 Shim Vol.
Position	R5.9 A1.1 F20.4 mm
Orientation	Sagittal
Rotation	32.05 deg
A >> P	172 mm
F >> H	122 mm
R >> L	142 mm
Vol. Visibility	On

System - Tx/Rx

Frequency 1H	296.941213 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	6000.0 ms
Concatenations	1

Physio - Cardiac

Magn. preparation	Non-sel. IR
TI 1	800 ms
TI 2	2700 ms
Fat suppr.	None
Dark blood	Off
FoV read	240 mm
FoV phase	100.0 %

Physio - Cardiac

Phase resolution

Physio - PACE	
Resp. control	Off
Concatenations	1

100 %

Inline - Common

Subtract	Off
Measurements	1
StdDev	Off
Save original images	On

Inline - MIP

MIP-Sag	Off	
MIP-Cor	Off	
MIP-Tra	Off	
MIP-Time	Off	
Save original images	On	

Inline - Composing

Inline Composing	Off	
Distortion Corr.	Off	

Inline - MapIt

Save original images	On
MapIt	T1 map
Flip angle 1	4.0 deg
Flip angle 2	5.0 deg
Measurements	1
TR	6000.0 ms
TE	2.74 ms

Sequence - Part 1

Introduction	Off
Dimension	3D
Elliptical scanning	Off
Reordering	Linear
Asymmetric echo	Allowed
Flow comp.	No
Multi-slice mode	Single shot
Echo spacing	7.9 ms
Bandwidth	140 Hz/Px

Sequence - Part 2

_		
RF pulse type	Fast	
Gradient mode	Fast*	
Excitation	Non-sel.	
RF spoiling	On	
Incr. Gradient spoiling	Off	
Turbo factor	180	

Mode	Off	
------	-----	--

\\USER\Khan\NeuroAnalytics\SNSX_TLE_7t(july2021)\sa2rage_B1map

TA: 1:47 PM: FIX Voxel size: 3.0×3.0×3.0 mmPAT: 2 Rel. SNR: 1.00 : tfl

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further	Off
preparation	
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slab group	1
Slabs	1
Dist. factor	50 %
Position	R5.3 A13.3 F28.6 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
AutoAlign	
Phase oversampling	0 %
Slice oversampling	0.0 %
Slices per slab	64
FoV read	240 mm
FoV phase	100.0 %
Slice thickness	3.00 mm
TR	2400.0 ms
TE	0.63 ms
Averages	1
Concatenations	1
Filter	Raw filter
Coil elements	HC2

Contrast - Common

TR	2400.0 ms
TE	0.63 ms
Magn. preparation	Non-sel. SR
TI 1	45 ms
TI 2	1800 ms
Flip angle 1	4.0 deg
Flip angle 2	11.0 deg
Fat suppr.	None
Water suppr.	None

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magn./Phase
Measurements	1
Multiple series	Each measurement

Resolution - Common

FoV read	240 mm
FoV phase	100.0 %
Slice thickness	3.00 mm
Base resolution	80
Phase resolution	100 %
Slice resolution	100 %
Phase partial Fourier	6/8

Resolution - Common

Slice partial Fourier	6/8	
Interpolation	Off	

Resolution - iPAT

PAT mode	GRAPPA
Accel. factor PE	2
Ref. lines PE	30
Accel. factor 3D	1
Reference scan mode	Integrated

Resolution - Filter Image

Image Filter	Off
Distortion Corr.	Off
Prescan Normalize	Off
Normalize	Off
B1 filter	Off

Resolution - Filter Rawdata

Raw filter	On	
Elliptical filter	Off	

Geometry - Common

Slab group	1
Slabs	1
Dist. factor	50 %
Position	R5.3 A13.3 F28.6 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
Slice oversampling	0.0 %
Slices per slab	64
FoV read	240 mm
FoV phase	100.0 %
Slice thickness	3.00 mm
TR	2400.0 ms
Multi-slice mode	Single shot
Series	Interleaved
Concatenations	1

Geometry - AutoAlign

Slab group	1
Position	R5.3 A13.3 F28.6 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
AutoAlign	
Initial Position	R5.3 A13.3 F28.6
R	5.3 mm
Α	13.3 mm
F	28.6 mm
Initial Rotation	0.00 deg
Initial Orientation	Sagittal

Geometry - Navigator

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table position	Н
Table position	0 mm
Inline Composing	Off

System - Miscellaneous

Positioning mode	FIX
Table position	Н
Table position	0 mm
MSMA	S-C-T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	
Coil Select Mode	Default

System - Adjustments

B0 Shim mode	Brain
B1 Shim mode	Volume-selective
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

R5.8 A6.9 F22.7 mm
Sagittal
0.00 deg
213 mm
173 mm
171 mm
Off

System - pTx Volumes

B1 Shim mode	Volume-selective
Excitation	Non-sel.
pTx Volume	1
Vol. Property	B1 Shim Vol.
Position	R5.9 A1.1 F20.4 mm
Orientation	Sagittal
Rotation	32.05 deg
A >> P	172 mm
F >> H	122 mm
R >> L	142 mm
Vol. Visibility	On

System - Tx/Rx

Frequency 1H	296.941213 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	2400.0 ms
Concatenations	1

Physio - Cardiac

Magn. preparation	Non-sel. SR
TI 1	45 ms
TI 2	1800 ms
Fat suppr.	None
Dark blood	Off
FoV read	240 mm
FoV phase	100.0 %

Physio - Cardiac

Phase resolution

Physio - PACE		
Resp. control Concatenations	Off	
Concatenations	1	

100 %

Inline - Common

Subtract	Off
Measurements	1
StdDev	Off
Save original images	On

Inline - MIP

MIP-Sag	Off	
MIP-Cor	Off	
MIP-Tra	Off	
MIP-Time	Off	
Save original images	On	

Inline - Composing

Inline Composing	Off
Distortion Corr.	Off

Inline - MapIt

Save original images	On
MapIt	None
Flip angle 1	4.0 deg
Flip angle 2	11.0 deg
Measurements	1
TR	2400.0 ms
TE	0.63 ms

Sequence - Part 1

Introduction	On
Dimension	3D
Elliptical scanning	Off
Reordering	Linear
Asymmetric echo	Allowed
Flow comp.	No
Multi-slice mode	Single shot
Echo spacing	2 ms
Bandwidth	1560 Hz/Px

Sequence - Part 2

_	
RF pulse type	Normal
Gradient mode	Fast
Excitation	Non-sel.
RF spoiling	On
Incr. Gradient spoiling	Off
Turbo factor	48

Sequence - Special

PPD Config	0	
Workaround Some Siemens	On	
Ineptitude		

Mode	Off	

\\USER\Khan\NeuroAnalytics\SNSX_TLE_7t(july2021)\t2_tse_cor_p2_500x500x1400

TA: 5:10 PM: REF Voxel size: 0.5×0.5×1.4 mmPAT: 2 Rel. SNR: 1.00 : tse

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further	Off
preparation	
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slice group	1
Slices	34
Dist. factor	10 %
Position	R4.4 A7.4 F36.3 mm
Orientation	C > T-17.9
Phase enc. dir.	R >> L
AutoAlign	
Phase oversampling	0 %
FoV read	192 mm
FoV phase	84.9 %
Slice thickness	1.4 mm
TR	5000.0 ms
TE	102 ms
Averages	3
Concatenations	2
Filter	None
Coil elements	HC2

Contrast - Common

TR	5000.0 ms
TE	102 ms
TD	0.0 ms
MTC	Off
Magn. preparation	None
Flip angle	120 deg
Fat suppr.	None
Water suppr.	None
Restore magn.	Off

Contrast - Dynamic

Averages	3
Averaging mode	Short term
Reconstruction	Magnitude
Measurements	1
Multiple series	Each measurement

Resolution - Common

FoV read	192 mm
FoV phase	84.9 %
Slice thickness	1.4 mm
Base resolution	384
Phase resolution	100 %
Phase partial Fourier	Off
Trajectory	Cartesian
Interpolation	Off

Resolution - iPAT

PAT mode	GRAPPA
Accel. factor PE	2
Ref. lines PE	32
Reference scan mode	Self-calibration

Resolution - Filter Image

Image Filter	Off
Distortion Corr.	Off
Prescan Normalize	Off
Normalize	Off
B1 filter	Off

Resolution - Filter Rawdata

Raw filter	Off	
Elliptical filter	Off	

Geometry - Common

Slice group	1
Slices	34
Dist. factor	10 %
Position	R4.4 A7.4 F36.3 mm
Orientation	C > T-17.9
Phase enc. dir.	R >> L
FoV read	192 mm
FoV phase	84.9 %
Slice thickness	1.4 mm
TR	5000.0 ms
Multi-slice mode	Interleaved
Series	Interleaved
Concatenations	2

Geometry - AutoAlign

Slice group	1
Position	R4.4 A7.4 F36.3 mm
Orientation	C > T-17.9
Phase enc. dir.	R >> L
AutoAlign	
Initial Position	R4.4 A7.4 F36.3
R	4.4 mm
A	7.4 mm
F	36.3 mm
Initial Rotation	0.00 deg
Initial Orientation	C > T
C > T	-17.9
> S	0.0

Geometry - Saturation

Fat suppr.	None
Water suppr.	None
Restore magn.	Off
Special sat.	None

Geometry - Navigator

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table position	Н
Table position	0 mm
Inline Composing	Off

Geometry - Tim CT

Tim CT mode	Off
Slices	34
Slice thickness	1.4 mm
Dist. factor	10 %
FoV read	192 mm
FoV phase	84.9 %

System - Miscellaneous

Positioning mode	REF
Table position	F
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	
Coil Select Mode	Default

System - Adjustments

B0 Shim mode	Brain
B1 Shim mode	Volume-selective
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

! Position	R4.4 A7.4 F36.3 mm
! Orientation	C > T-17.9
! Rotation	0.00 deg
! R >> L	169 mm
! F >> H	192 mm
! A >> P	67 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	Volume-selective
pTx Volume	1
Vol. Property	B1 Shim Vol.
Position	R5.8 A9.3 F30.7 mm
Orientation	C > T-18.5
Rotation	0.00 deg
R >> L	145 mm
F >> H	145 mm
A >> P	62 mm
Vol. Visibility	On

System - Tx/Rx

Frequency 1H	296.941213 MHz
	230.341213 WII IZ
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	5000.0 ms
Concatenations	2

Physio - Cardiac

Magn. preparation	None
Fat suppr.	None
Dark blood	Off
FoV read	192 mm
FoV phase	84.9 %
Phase resolution	100 %
Trajectory	Cartesian

Physio - PACE

Resp. control	Off
Concatenations	2

Inline - Common

Subtract	Off	
Measurements	1	
StdDev	Off	
Save original images	On	

Inline - MIP

MIP-Sag MIP-Cor MIP-Tra MIP-Time	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

Inline - Composing

Inline Composing	Off
Distortion Corr.	Off

Sequence - Part 1

Introduction	Off
Dimension	2D
Compensate T2 decay	Off
Reduce Motion Sens.	On
Contrasts	1
Flow comp.	No
Optimization	In phase
Multi-slice mode	Interleaved
Free echo spacing	Off
Echo spacing	12.7 ms
Bandwidth	217 Hz/Px

Sequence - Part 2

Define	Turbo factor
Echo trains per slice	10
Phase correction	Automatic
Acoustic noise reduction	None
RF pulse type	Low SAR
Gradient mode	Fast
Hyperecho	Off
WARP	Off
Red. EC sensitivity	Off
Turbo factor	18

Mode	Off
Allowed delay	0 s

$\verb|\USER\Khan\NeuroAnalytics\SNSX_TLE_7t(july2021)\t2_tse_tra_p2_500x500x1400||$

TA: 7:14 PM: FIX Voxel size: 0.5×0.5×1.4 mmPAT: 2 Rel. SNR: 1.00 : tse

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slice group	1
Slices	58
Dist. factor	10 %
Position	R6.5 A10.6 F16.5 mm
Orientation	T > C22.6
Phase enc. dir.	R >> L
AutoAlign	
Phase oversampling	0 %
FoV read	192 mm
FoV phase	84.9 %
Slice thickness	1.4 mm
TR	7000.0 ms
TE	102 ms
Averages	3
Concatenations	2
Filter	None
Coil elements	HC2

Contrast - Common

7000.0 ms
102 ms
0.0 ms
Off
None
120 deg
None
None
Off

Contrast - Dynamic

Averages	3
Averaging mode	Short term
Reconstruction	Magnitude
Measurements	1
Multiple series	Each measurement

Resolution - Common

FoV read	192 mm
FoV phase	84.9 %
Slice thickness	1.4 mm
Base resolution	384
Phase resolution	100 %
Phase partial Fourier	Off
Trajectory	Cartesian
Interpolation	Off

Resolution - iPAT

PAT mode	GRAPPA
Accel. factor PE	2
Ref. lines PE	32
Reference scan mode	Self-calibration

Resolution - Filter Image

Image Filter	Off
Distortion Corr.	Off
Prescan Normalize	Off
Normalize	Off
B1 filter	Off

Resolution - Filter Rawdata

Raw filter	Off	
Elliptical filter	Off	

Geometry - Common

Slice group	1
Slices	58
Dist. factor	10 %
Position	R6.5 A10.6 F16.5 mm
Orientation	T > C22.6
Phase enc. dir.	R >> L
FoV read	192 mm
FoV phase	84.9 %
Slice thickness	1.4 mm
TR	7000.0 ms
Multi-slice mode	Interleaved
Series	Interleaved
Concatenations	2

Geometry - AutoAlign

Slice group	1
Position	R6.5 A10.6 F16.5 mm
Orientation	T > C22.6
Phase enc. dir.	R >> L
AutoAlign	
Initial Position	R6.5 A10.6 F16.5
R	6.5 mm
A	10.6 mm
F	16.5 mm
Initial Rotation	90.00 deg
Initial Orientation	T > C
T > C	22.6
> S	0.0

Geometry - Saturation

Fat suppr.	None
Water suppr.	None
Restore magn.	Off
Special sat.	None

Geometry - Navigator

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table position	Н
Table position	0 mm
Inline Composing	Off

Geometry - Tim CT

Tim CT mode	Off
Slices	58
Slice thickness	1.4 mm
Dist. factor	10 %
FoV read	192 mm
FoV phase	84.9 %

System - Miscellaneous

Positioning mode	FIX
Table position	F
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	
Coil Select Mode	Default

System - Adjustments

B0 Shim mode	Brain
B1 Shim mode	Volume-selective
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

! Position	R2.9 A12.3 F16.6 mm
! Orientation	T > C22.8
! Rotation	90.00 deg
! R >> L	169 mm
! A >> P	192 mm
! F >> H	98 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	Volume-selective
pTx Volume	1
Vol. Property	B1 Shim Vol.
Position	R5.1 A0.5 F13.6 mm
Orientation	T > C22.5
Rotation	90.00 deg
R >> L	140 mm
A >> P	132 mm
F >> H	93 mm
Vol. Visibility	On

System - Tx/Rx

Frequency 1H	296.941213 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	7000.0 ms
Concatenations	2

Physio - Cardiac

Magn. preparation	None
Fat suppr.	None
Dark blood	Off
FoV read	192 mm
FoV phase	84.9 %
Phase resolution	100 %
Trajectory	Cartesian

Physio - PACE

Resp. control	Off
Concatenations	2

Inline - Common

Subt	ract	Off
Mea	surements	1
StdE)ev	Off
Save	e original images	On

Inline - MIP

MIP-Sag MIP-Cor MIP-Tra MIP-Time	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

Inline - Composing

Inline Composing	Off
Distortion Corr.	Off

Sequence - Part 1

Introduction	Off
Dimension	2D
Compensate T2 decay	Off
Reduce Motion Sens.	On
Contrasts	1
Flow comp.	No
Optimization	In phase
Multi-slice mode	Interleaved
Free echo spacing	Off
Echo spacing	12.7 ms
Bandwidth	217 Hz/Px

Sequence - Part 2

Define	Turbo factor
Echo trains per slice	10
Phase correction	Automatic
Acoustic noise reduction	None
RF pulse type	Low SAR
Gradient mode	Fast
Hyperecho	Off
WARP	Off
Red. EC sensitivity	Off
Turbo factor	18

Mode	Off
Allowed delay	0 s

\\USER\Khan\NeuroAnalytics\SNSX_TLE_7t(july2021)\mbep2d_diff_b2000_b1000_95dir_AP

TA: 9:14 PM: REF Voxel size: 1.5×1.5×1.5 mmPAT: 3 Rel. SNR: 1.00 : epse

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further	Off
preparation	
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slice group	1
Slices	90
Dist. factor	0 %
Position	R4.4 A6.1 F19.4 mm
Orientation	Transversal
Phase enc. dir.	A >> P
AutoAlign	
Phase oversampling	0 %
FoV read	210 mm
FoV phase	100.0 %
Slice thickness	1.50 mm
TR	5500 ms
TE	58.20 ms
Multi-band accel. factor	2
Filter	None
Coil elements	HC2

Contrast - Common

TR	5500 ms
TE	58.20 ms
MTC	Off
Magn. preparation	None
Flip angle	80 deg
Refocus flip angle	150 deg
Fat suppr.	None
Grad. rev. fat suppr.	Disabled

Contrast - Dynamic

Averaging mode	Long term
Reconstruction	Magnitude
Measurements	1
Delay in TR	0 ms
Multiple series	Off

Resolution - Common

FoV read	210 mm
FoV phase	100.0 %
Slice thickness	1.50 mm
Base resolution	140
Phase resolution	100 %
Phase partial Fourier	6/8
Interpolation	Off

Resolution - iPAT

PAT mode	GRAPPA
Accel. factor PE	3

Resolution - iPAT

Ref. lines PE	54
Reference scan mode	FLEET

Resolution - Filter Image

Distortion Corr.	Off
Prescan Normalize	Off
Dynamic Field Corr.	Off

Resolution - Filter Rawdata

Raw filter	Off	
Elliptical filter	Off	

Geometry - Common

Slice group	1
Slices	90
Dist. factor	0 %
Position	R4.4 A6.1 F19.4 mm
Orientation	Transversal
Phase enc. dir.	A >> P
FoV read	210 mm
FoV phase	100.0 %
Slice thickness	1.50 mm
TR	5500 ms
Multi-slice mode	Interleaved
Series	Interleaved
Multi-band accel. factor	2

Geometry - AutoAlign

, ,	
Slice group	1
Position	R4.4 A6.1 F19.4 mm
Orientation	Transversal
Phase enc. dir.	A >> P
AutoAlign	
Initial Position	R4.4 A6.1 F19.4
R	4.4 mm
A	6.1 mm
F	19.4 mm
Initial Rotation	0.00 deg
Initial Orientation	Transversal

Geometry - Saturation

Fat suppr.	None
Grad. rev. fat suppr.	Disabled
Special sat.	None

Geometry - Navigator

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table position	Н
Table position	0 mm
Inline Composing	Off

System - Miscellaneous

Positioning mode	REF
Table position	Н
Table position	0 mm
MSMA	S-C-T
Sagittal	R >> L

System - Miscellaneous

Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Matrix Optimization	Off
AutoAlign	
Coil Select Mode	Default

System - Adjustments

B0 Shim mode	Brain
B1 Shim mode	Volume-selective
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

! Position	R5.8 A6.9 F22.7 mm
! Orientation	Sagittal
! Rotation	0.00 deg
! A >> P	213 mm
! F >> H	173 mm
! R >> L	171 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	Volume-selective
Excitation	Standard
pTx Volume	1
Vol. Property	B1 Shim Vol.
Position	R5.9 A1.1 F20.4 mm
Orientation	Sagittal
Rotation	32.05 deg
A >> P	172 mm
F >> H	122 mm
R >> L	142 mm
Vol. Visibility	On

System - Tx/Rx

Frequency 1H	296.941213 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	5500 ms
Multi-band accel. factor	2

Physio - PACE

Resp. control	Off
Multi-band accel. factor	2

Diff - Neuro

Diffusion mode	Free
Diff. directions	95
Diffusion Scheme	Monopolar
Diff. weightings	2
b-value 1	0 s/mm²
b-value 2	2000 s/mm ²
b-value 1	1
b-value 2	1

Diff - Neuro

Diff. weighted images	On	
Trace weighted images	On	
ADC maps	On	
FA maps	On	
Mosaic	On	
Tensor	On	
Noise level	80	

Diff - Body

Diffusion mode	Free
Diff. directions	95
Diffusion Scheme	Monopolar
Diff. weightings	2
b-value 1	0 s/mm²
b-value 2	2000 s/mm ²
b-value 1	1
b-value 2	1
Diff. weighted images	On
Trace weighted images	On
ADC maps	On
Exponential ADC Maps	Off
FA maps	On
Invert Gray Scale	Off
Calculated Image	Off
b-Value >=	0 s/mm²
Noise level	80

Diff - Composing

Inline Composing	Off
Distortion Corr.	Off

Sequence - Part 1

Introduction	Off
Multi-slice mode	Interleaved
Free echo spacing	Off
Echo spacing	0.76 ms
Bandwidth	1488 Hz/Px

Sequence - Part 2

EPI factor	140
Gradient mode	Fast
Excitation	Standard
RF spoiling	Off

Sequence - Special

Excite pulse duration	5120 us
Refocus pulse duration	10240 us
FLEET ref. prep. scans	0
FLEET ref. min. TR	0 ms
Single-band images	Off
MB LeakBlock kernel	On
MB dual kernel	Off
MB RF phase scramble	Off
Time-shifted MB RF	Off
SENSE1 coil combine	On
Invert RO/PE polarity	Off
PF omits higher k-space	Off
Disable freq. update	Off
Force equal slice timing	Off
Online multi-band recon.	Online
FFT scale factor	1.00
FLEET iPAT ref. FA	12.0 deg
Physio recording	Off
FLEET iPAT ref. FA	12.0 deg

\\USER\Khan\NeuroAnalytics\SNSX_TLE_7t(july2021)\mbep2d_diff_B0_PA

TA: 0:32 PM: FIX Voxel size: 1.5×1.5×1.5 mmPAT: 3 Rel. SNR: 1.00 : epse

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further	Off
preparation	
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slice group	1
Slices	90
Dist. factor	0 %
Position	R4.4 A6.1 F19.4 mm
Orientation	Transversal
Phase enc. dir.	A >> P
AutoAlign	
Phase oversampling	0 %
FoV read	210 mm
FoV phase	100.0 %
Slice thickness	1.50 mm
TR	5500 ms
TE	58.20 ms
Multi-band accel. factor	2
Filter	None
Coil elements	HC2

Contrast - Common

TR	5500 ms
TE	58.20 ms
MTC	Off
Magn. preparation	None
Flip angle	80 deg
Refocus flip angle	150 deg
Fat suppr.	None
Grad. rev. fat suppr.	Disabled

Contrast - Dynamic

Averaging mode	Long term
Reconstruction	Magnitude
Measurements	1
Delay in TR	0 ms
Multiple series	Off

Resolution - Common

FoV read	210 mm
FoV phase	100.0 %
Slice thickness	1.50 mm
Base resolution	140
Phase resolution	100 %
Phase partial Fourier	6/8
Interpolation	Off

Resolution - iPAT

PAT mode	GRAPPA
Accel. factor PE	3

Resolution - iPAT

Ref. lines PE	54
Reference scan mode	FLEET

Resolution - Filter Image

Distortion Corr.	Off
Prescan Normalize	Off
Dynamic Field Corr.	Off

Resolution - Filter Rawdata

Raw filter	Off	
Elliptical filter	Off	

Geometry - Common

Slice group	1
Slices	90
Dist. factor	0 %
Position	R4.4 A6.1 F19.4 mm
Orientation	Transversal
Phase enc. dir.	A >> P
FoV read	210 mm
FoV phase	100.0 %
Slice thickness	1.50 mm
TR	5500 ms
Multi-slice mode	Interleaved
Series	Interleaved
Multi-band accel. factor	2

Geometry - AutoAlign

, ,	
Slice group	1
Position	R4.4 A6.1 F19.4 mm
Orientation	Transversal
Phase enc. dir.	A >> P
AutoAlign	
Initial Position	R4.4 A6.1 F19.4
R	4.4 mm
A	6.1 mm
F	19.4 mm
Initial Rotation	0.00 deg
Initial Orientation	Transversal

Geometry - Saturation

Fat suppr.	None
Grad. rev. fat suppr.	Disabled
Special sat.	None

Geometry - Navigator

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table position	Н
Table position	0 mm
Inline Composing	Off

System - Miscellaneous

Positioning mode	FIX
Table position	Н
Table position	0 mm
MSMA	S-C-T
Sagittal	R >> L

System - Miscellaneous

Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Matrix Optimization	Off
AutoAlign	
Coil Select Mode	Default

System - Adjustments

B0 Shim mode	Brain
B1 Shim mode	Volume-selective
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

! Position	R5.8 A6.9 F22.7 mm
! Orientation	Sagittal
! Rotation	0.00 deg
! A >> P	213 mm
! F >> H	173 mm
! R >> L	171 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	Volume-selective
Excitation	Standard
pTx Volume	1
Vol. Property	B1 Shim Vol.
Position	R5.9 A1.1 F20.4 mm
Orientation	Sagittal
Rotation	32.05 deg
A >> P	172 mm
F >> H	122 mm
R >> L	142 mm
Vol. Visibility	On

System - Tx/Rx

Frequency 1H	296.941213 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	5500 ms
Multi-band accel. factor	2

Physio - PACE

Resp. control	Off
Multi-band accel. factor	2

Diff - Neuro

Diffusion mode	Free
Diff. directions	95
Diffusion Scheme	Monopolar
Diff. weightings	1
b-value	0 s/mm²
b-value	1
Diff. weighted images	On
Trace weighted images	Off

Diff - Neuro

ADC maps	Off
FA maps	Off
Mosaic	Off
Tensor	Off
Noise level	80

Diff - Body

Diffusion mode	Free
Diff. directions	95
Diffusion Scheme	Monopolar
Diff. weightings	1
b-value	0 s/mm²
b-value	1
Diff. weighted images	On
Trace weighted images	Off
ADC maps	Off
Exponential ADC Maps	Off
FA maps	Off
Invert Gray Scale	Off
Calculated Image	Off
b-Value >=	0 s/mm²
Noise level	80

Diff - Composing

Inline Composing	Off
Distortion Corr.	Off

Sequence - Part 1

Introduction	Off
Multi-slice mode	Interleaved
Free echo spacing	Off
Echo spacing	0.76 ms
Bandwidth	1488 Hz/Px

Sequence - Part 2

EPI factor	140
Gradient mode	Fast
Excitation	Standard
RF spoiling	Off

Sequence - Special

Excite pulse duration	5120 us
Refocus pulse duration	10240 us
FLEET ref. prep. scans	0
FLEET ref. min. TR	0 ms
Single-band images	Off
MB LeakBlock kernel	On
MB dual kernel	Off
MB RF phase scramble	Off
Time-shifted MB RF	Off
SENSE1 coil combine	On
Invert RO/PE polarity	On
PF omits higher k-space	Off
Disable freq. update	Off
Force equal slice timing	Off
Online multi-band recon.	Online
FFT scale factor	1.00
FLEET iPAT ref. FA	12.0 deg
Physio recording	Off

\\USER\Khan\NeuroAnalytics\SNSX_TLE_7t(july2021)\ASPIRE_800iso_7e

TA: 6:06 PM: FIX Voxel size: 0.8×0.8×0.8 mmPAT: 2 Rel. SNR: 1.00 : fl

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further	Off
preparation	
Wait for user to start	On
Start measurements	Single measurement

Routine

Slab group	1
Slabs	1
Dist. factor	20 %
Position	R5.8 A9.3 F13.2 mm
Orientation	Transversal
Phase enc. dir.	R >> L
AutoAlign	
Phase oversampling	0 %
Slice oversampling	10.0 %
Slices per slab	160
FoV read	205 mm
FoV phase	87.5 %
Slice thickness	0.75 mm
TR	30.0 ms
TE 1	1.65 ms
TE 2	3.93 ms
TE 3	6.21 ms
TE 4	8.49 ms
TE 5	10.77 ms
TE 6	13.05 ms
TE 7	15.33 ms
Averages	1
Concatenations	1
Filter	Distortion Corr.(2D)
Coil elements	HC2

Contrast - Common

TR	30.0 ms
TE 1	1.65 ms
TE 2	3.93 ms
TE 3	6.21 ms
TE 4	8.49 ms
TE 5	10.77 ms
TE 6	13.05 ms
TE 7	15.33 ms
MTC	Off
Magn. preparation	None
Flip angle	7 deg
Fat suppr.	None
Water suppr.	None
SWI	Off

Contrast - Dynamic

Averages	1
Averaging mode	Short term
Reconstruction	Magnitude

Contrast - Dynamic

Measurements	1
Multiple series	Each measurement

Resolution - Common

FoV read	205 mm
FoV phase	87.5 %
Slice thickness	0.75 mm
Base resolution	272
Phase resolution	100 %
Slice resolution	100 %
Phase partial Fourier	6/8
Slice partial Fourier	6/8
Interpolation	Off

Resolution - iPAT

PAT mode	GRAPPA
Accel. factor PE	2
Ref. lines PE	24
Accel. factor 3D	1
Reference scan mode	Integrated

Resolution - Filter Image

Image Filter	Off	
Distortion Corr.	On	
Mode	2D	
Unfiltered images	Off	
Prescan Normalize	Off	
Normalize	Off	
B1 filter	Off	

Resolution - Filter Rawdata

Raw filter	Off	
Elliptical filter	Off	

Geometry - Common

_	
Slab group	1
Slabs	1
Dist. factor	20 %
Position	R5.8 A9.3 F13.2 mm
Orientation	Transversal
Phase enc. dir.	R >> L
Slice oversampling	10.0 %
Slices per slab	160
FoV read	205 mm
FoV phase	87.5 %
Slice thickness	0.75 mm
TR	30.0 ms
Multi-slice mode	Interleaved
Series	Interleaved
Concatenations	1

Geometry - AutoAlign

Slab group	1
Position	R5.8 A9.3 F13.2 mm
Orientation	Transversal
Phase enc. dir.	R >> L
AutoAlign	
Initial Position	R5.8 A9.3 F13.2
R	5.8 mm
Α	9.3 mm

Geometry - AutoAlign

F	13.2 mm
Initial Rotation	90.00 deg
Initial Orientation	Transversal

Geometry - Saturation

Saturation mode	Standard
Fat suppr.	None
Water suppr.	None
Special sat.	None

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table position	Н
Table position	0 mm
Inline Composing	Off

Geometry - Tim CT

Tim CT mode	Off
Slabs	1
Slices per slab	160
Slice thickness	0.75 mm
Dist. factor	20 %
FoV read	205 mm
FoV phase	87.5 %
Segments	1

System - Miscellaneous

Positioning mode	FIX
Table position	F
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	
Coil Select Mode	Default

System - Adjustments

B0 Shim mode	Advanced
B1 Shim mode	Volume-selective
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

! Position	R5.8 A6.9 F22.7 mm
! Orientation	Sagittal
! Rotation	0.00 deg
! A >> P	213 mm
! F >> H	173 mm
! R >> L	171 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	Volume-selective
Excitation	Slab-sel.
pTx Volume	1
Vol. Property	B1 Shim Vol.
Position	R5.9 A1.1 F20.4 mm

System - pTx Volumes

Orientation	Sagittal
Rotation	32.05 deg
A >> P	172 mm
F >> H	122 mm
R >> L	142 mm
Vol. Visibility	On

System - Tx/Rx

Frequency 1H	296.941213 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	30.0 ms
Concatenations	1
Segments	1

Physio - Cardiac

Tagging	None
Magn. preparation	None
Fat suppr.	None
Dark blood	Off
FoV read	205 mm
FoV phase	87.5 %
Phase resolution	100 %

Physio - PACE

Resp. control	Off
Concatenations	1

Inline - Common

Subtract	Off
Measurements	1
StdDev	Off
Liver registration	Off
Save original images	On

Inline - MIP

MIP-Sag	Off
549	-
MIP-Cor	Off
MID To	011
MIP-Tra	Off
MIP-Time	Off
	Oli
Save original images	On
eare engina images	U

Inline - Soft Tissue

Wash - In	Off
Wash - Out	Off
TTP	Off
PEI	Off
MIP - time	Off
Measurements	1

Inline - Composing

Inline Composing	Off
Distortion Corr.	On
Mode	2D
Unfiltered images	Off

Inline - MapIt

Save original images	On
MapIt	None
Flip angle	7 deg
Measurements	1
Contrasts	7
TR	30.0 ms
TE 1	1.65 ms
TE 2	3.93 ms
TE 3	6.21 ms
TE 4	8.49 ms
TE 5	10.77 ms
TE 6	13.05 ms
TE 7	15.33 ms

Sequence - Part 1

Introduction	On
Dimension	3D
Elliptical scanning	On
Phase stabilisation	On
Asymmetric echo	Allowed
Contrasts	7
Flow comp. 1	No
Readout mode	Monopolar
Multi-slice mode	Interleaved
Bandwidth 1	610 Hz/Px
Bandwidth 2	610 Hz/Px
Bandwidth 3	610 Hz/Px
Bandwidth 4	610 Hz/Px
Bandwidth 5	610 Hz/Px
Bandwidth 6	610 Hz/Px
Bandwidth 7	610 Hz/Px

Sequence - Part 2

•		
Segments	1	
Acoustic noise reduction	None	
RF pulse type	Fast	
Gradient mode	Fast	
Excitation	Slab-sel.	
RF spoiling	On	

Sequence - Special

•	
Phase Combination	ASPIRE
Advanced Options	On
Additional FFT Scale	1.00
Magnitude Scale	1000
PO Smoothing Sigma	5 mm
Use PO for next measurements	Off
ASPIRE 1. Echo	1
ASPIRE 2. Echo	2
Mag of Complex Sum	On
UMPIRE Unwrapping	Off
T2* / R2*	T2* + R2*

Mode	Off
------	-----

\\USER\Khan\NeuroAnalytics\SNSX_TLE_7t(july2021)\dark-fluid_tse_cor_p2_.67x.67x2.5

TA: 2:44 PM: FIX Voxel size: 0.3×0.3×2.5 mmPAT: 2 Rel. SNR: 1.00 : tir_rr

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further	Off
preparation	
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slice group	1
Slices	26
Dist. factor	10 %
Position	R4.4 A7.4 F36.3 mm
Orientation	C > T-17.9
Phase enc. dir.	R >> L
AutoAlign	
Phase oversampling	0 %
FoV read	192 mm
FoV phase	81.3 %
Slice thickness	2.5 mm
TR	9000.0 ms
TE	84 ms
Averages	1
Concatenations	2
Filter	B1 filter
Coil elements	HC2

Contrast - Common

TR	9000.0 ms	
TE	84 ms	
TD	0.0 ms	
MTC	Off	
Magn. preparation	Slice-sel. IR	
TI	2600 ms	
Flip angle	150 deg	
Fat suppr.	Fat sat.	
Fat sat. mode	Strong	
Water suppr.	None	
Restore magn.	Off	
Freeze suppressed tissue	Off	

Contrast - Dynamic

Averages	1
Averaging mode	Short term
Reconstruction	Magnitude
Measurements	1
Multiple series	Each measurement

Resolution - Common

FoV read	192 mm
FoV phase	81.3 %
Slice thickness	2.5 mm
Base resolution	288
Phase resolution	100 %
Phase partial Four	er Off

Resolution - Common

Trajectory	Cartesian
Interpolation	On

Resolution - iPAT

PAT mode	GRAPPA
Accel. factor PE	2
Ref. lines PE	38
Reference scan mode	Integrated

Resolution - Filter Image

Image Filter	Off
Distortion Corr.	Off
Prescan Normalize	Off
Normalize	Off
B1 filter	On
Unfiltered images	Off

Resolution - Filter Rawdata

Raw filter	Off	
Elliptical filter	Off	

Geometry - Common

Slice group	1
Slices	26
Dist. factor	10 %
Position	R4.4 A7.4 F36.3 mm
Orientation	C > T-17.9
Phase enc. dir.	R >> L
FoV read	192 mm
FoV phase	81.3 %
Slice thickness	2.5 mm
TR	9000.0 ms
Multi-slice mode	Interleaved
Series	Interleaved
Concatenations	2

Geometry - AutoAlign

Slice group	1
Position	R4.4 A7.4 F36.3 mm
Orientation	C > T-17.9
Phase enc. dir.	R >> L
AutoAlign	
Initial Position	R4.4 A7.4 F36.3
R	4.4 mm
A	7.4 mm
F	36.3 mm
Initial Rotation	0.00 deg
Initial Orientation	C > T
C > T	-17.9
> S	0.0

Geometry - Saturation

Fat suppr.	Fat sat.
Fat sat. mode	Strong
Water suppr.	None
Restore magn.	Off
Special sat.	None

Geometry - Navigator

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table position	Н
Table position	0 mm
Inline Composing	Off

Geometry - Tim CT

Tim CT mode	Off
Slices	26
Slice thickness	2.5 mm
Dist. factor	10 %
FoV read	192 mm
FoV phase	81.3 %

System - Miscellaneous

-	
Positioning mode	FIX
Table position	F
Table position	0 mm
MSMA	S-C-T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	
Coil Select Mode	Default

System - Adjustments

-	
B0 Shim mode	Brain
B1 Shim mode	Volume-selective
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

! Position	R4.4 A7.4 F36.3 mm
! Orientation	C > T-17.9
! Rotation	0.00 deg
! R >> L	169 mm
! F >> H	192 mm
! A >> P	67 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	Volume-selective
pTx Volume	1
Vol. Property	B1 Shim Vol.
Position	R2.9 A9.3 F30.7 mm
Orientation	C > T-18.5
Rotation	0.00 deg
R >> L	140 mm
F >> H	150 mm
A >> P	62 mm
Vol. Visibility	On

System - Tx/Rx

Frequency 1H	296.941213 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	9000.0 ms
Concatenations	2

Physio - Cardiac

Magn. preparation	Slice-sel. IR
ТІ	2600 ms
Fat suppr.	Fat sat.
Dark blood	Off
FoV read	192 mm
FoV phase	81.3 %
Phase resolution	100 %
Trajectory	Cartesian

Physio - PACE

Resp. control	Off
Concatenations	2

Inline - Common

Subtract	Off
Measurements	1
StdDev	Off
Save original images	On

Inline - MIP

MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

Inline - Composing

Inline Composing	Off
Distortion Corr	Off

Sequence - Part 1

Introduction	On
Dimension	2D
Compensate T2 decay	Off
Reduce Motion Sens.	On
Contrasts	1
Flow comp.	Read
Optimization	In phase
Multi-slice mode	Interleaved
Free echo spacing	Off
Echo spacing	10.5 ms
Bandwidth	248 Hz/Px

Sequence - Part 2

Define	Turbo factor
Echo trains per slice	8
Phase correction	Automatic
Acoustic noise reduction	None
RF pulse type	Normal
Gradient mode	Fast
Hyperecho	Off
WARP	Off
Red. EC sensitivity	Off
Turbo factor	17

Mode	Min flip angle
Min flip angle	130 deg

SIEMENS MAGNETOM Investigational_Device_7T_Plus

Allowed delay	60 s
---------------	------

\\USER\Khan\NeuroAnalytics\SNSX_TLE_7t(july2021)\cb_sp2d_diff_C52_Spiral_R3

TA: 8:43 PM: REF Voxel size: 1.5×1.5×1.5 mmPAT: 6 Rel. SNR: 1.00 : epse

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	On
Start measurements	Single measurement

Routine

Slice group	1
Slices	88
Dist. factor	0 %
Position	L0.0 P5.4 F4.8 mm
Orientation	Transversal
Phase enc. dir.	A >> P
AutoAlign	
Phase oversampling	0 %
FoV read	200 mm
FoV phase	100.0 %
Slice thickness	1.5 mm
TR	5500 ms
TE	82.0 ms
Averages	1
Concatenations	1
Filter	None
Coil elements	FP

Contrast - Common

TR	5500 ms
TE	82.0 ms
MTC	Off
Magn. preparation	None
Flip angle exc	69 deg
Flip angle fat sat	110 deg
Fat suppr.	Fat sat.
Fat sat. mode	Strong

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	1
Delay in TR	0 ms
Multiple series	Off

Resolution - Common

FoV read	200 mm
FoV phase	100.0 %
Slice thickness	1.5 mm
Base resolution	134
Trajectory	Spiral
Interpolation	Off

Resolution - iPAT

Accel. mode Slice accel.

Resolution - iPAT

Accel. factor PE	3
Ref. lines PE	18
Accel. factor slice	2
Reference scan mode	EPI/separate

Resolution - Filter Image

Distortion Corr.	Off
Prescan Normalize	Off
Dynamic Field Corr.	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off

Geometry - Common

Slice group	1
Slices	88
Dist. factor	0 %
Position	L0.0 P5.4 F4.8 mm
Orientation	Transversal
Phase enc. dir.	A >> P
FoV read	200 mm
FoV phase	100.0 %
Slice thickness	1.5 mm
TR	5500 ms
Multi-slice mode	Interleaved
Series	Interleaved
Concatenations	1

Geometry - AutoAlign

Slice group	1
Position	L0.0 P5.4 F4.8 mm
Orientation	Transversal
Phase enc. dir.	A >> P
AutoAlign	
Initial Position	L0.0 P5.4 F4.8
L	0.0 mm
Р	5.4 mm
F	4.8 mm
Initial Rotation	0.00 deg
Initial Orientation	Transversal

Geometry - Saturation

Fat suppr.	Fat sat.
Fat sat. mode	Strong
Special sat.	None

Geometry - Navigator

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table position	Н
Table position	0 mm
Inline Composing	Off

System - Miscellaneous

Positioning mode	REF
Table position	Н
Table position	0 mm

System - Miscellaneous

MSMA	S-C-T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Matrix Optimization	Off
AutoAlign	
Coil Select Mode	Default

System - Adjustments

B0 Shim mode	Brain
B1 Shim mode	Volume-selective
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	L0.0 P5.4 F4.8 mm
Orientation	Transversal
Rotation	0.00 deg
A >> P	200 mm
R >> L F >> H	200 mm
	132 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	Volume-selective
Excitation	Standard
pTx Volume	1
Vol. Property	B1 Shim Vol.
Position	L0.0 P5.4 F4.8 mm
Orientation	Transversal
Rotation	0.00 deg
A >> P	200 mm
R >> L	200 mm
F >> H	132 mm
Vol. Visibility	On

System - Tx/Rx

Frequency 1H	296.941213 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	5500 ms
Concatenations	1

Physio - PACE

Resp. control	Off
Concatenations	1

Diff - Neuro

J 1104.0	
Diffusion mode	Free
Diff. directions	94
Diffusion Scheme	Monopolar Plus
Diff. weightings	1
b-value	2000 s/mm²
b-value	1

Diff - Neuro

Diff. weighted images	On
Trace weighted images	Off
ADC maps	Off
FA maps	Off
Mosaic	Off
Tensor	Off
Noise level	40

Diff - Body

Diffusion mode	Free
Diff. directions	94
Diffusion Scheme	Monopolar Plus
Diff. weightings	1
b-value	2000 s/mm ²
b-value	1
Diff. weighted images	On
Trace weighted images	Off
ADC maps	Off
Exponential ADC Maps	Off
FA maps	Off
Invert Gray Scale	Off
Calculated Image	Off
b-Value >=	0 s/mm²
Noise level	40

Diff - Composing

Inline Composing	Off	
Distortion Corr.	Off	

Sequence - Part 1

Introduction	Off
Optimization	None
Multi-slice mode	Interleaved
Free echo spacing	Off
Echo spacing	1.08 ms
Bandwidth	2194 Hz/Px

Sequence - Part 2

EPI factor	134
RF pulse type	Normal
Gradient mode	Fast
Excitation	Standard

Sequence - pTX Pulses

Sequence - Special

FFT scale	1.00
OGSE b-Value	300.0 s/mm2
Diff Slew Rate Scale	18.0
Field probe trigger	On
Field probe sync	0 #
Alt OGSE scheme	On
Run spiral PAT Ref scans	Off
ADC-to-Grad delay	0 us
SMS CAIPI	osc
VD Spiral	0.2
Num. PAT Echos	1 #
Slew rate scale	7.0

\\USER\Khan\NeuroAnalytics\SNSX_TLE_7t(july2021)\gre_field_mapping

TA: 1:55 PM: FIX Voxel size: 2.0×2.0×2.0 mmRel. SNR: 1.00 : fm_r

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slice group	1
Slices	74
Dist. factor	0 %
Position	R0.5 P9.6 H5.8 mm
Orientation	Transversal
Phase enc. dir.	A >> P
AutoAlign	
Phase oversampling	0 %
FoV read	210 mm
FoV phase	100.0 %
Slice thickness	2.0 mm
TR	542.0 ms
TE 1	4.08 ms
TE 2	5.1 ms
Averages	1
Concatenations	1
Filter	None
Coil elements	FP

Contrast - Common

TR	542.0 ms
TE 1	4.08 ms
TE 1 TE 2	5.1 ms
MTC	Off
Flip angle	35 deg
Fat suppr.	None

Contrast - Dynamic

Averages	1
Averaging mode	Short term
Reconstruction	Magn./Phase
Measurements	1
Multiple series	Off

Resolution - Common

FoV read	210 mm
FoV phase	100.0 %
Slice thickness	2.0 mm
Base resolution	104
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off

Resolution - Filter Image

Image Filter	Off	
Distortion Corr.	Off	

Resolution - Filter Image

Prescan Normalize	Off
Normalize	Off
B1 filter	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off

Geometry - Common

Slice group	1
Slices	74
Dist. factor	0 %
Position	R0.5 P9.6 H5.8 mm
Orientation	Transversal
Phase enc. dir.	A >> P
FoV read	210 mm
FoV phase	100.0 %
Slice thickness	2.0 mm
TR	542.0 ms
Multi-slice mode	Interleaved
Series	Interleaved
Concatenations	1

Geometry - AutoAlign

<u>-</u>	-
Slice group	1
Position	R0.5 P9.6 H5.8 mm
Orientation	Transversal
Phase enc. dir.	A >> P
AutoAlign	
Initial Position	R0.5 P9.6 H5.8
R	0.5 mm
Р	9.6 mm
Н	5.8 mm
Initial Rotation	0.00 deg
Initial Orientation	Transversal

Geometry - Saturation

Fat suppr.	None
Special sat.	None

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table position	Н
Table position	0 mm
Inline Composing	Off

System - Miscellaneous

Positioning mode	FIX
Table position	Н
Table position	0 mm
MSMA	S-C-T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Adaptive Combine
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	
Coil Select Mode	Default

System - Adjustments

B0 Shim mode	Brain
B1 Shim mode	Volume-selective
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	R0.5 P9.6 H5.8 mm
Orientation	Transversal
Rotation	0.00 deg
A >> P	210 mm
R >> L	210 mm
F >> H	148 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	Volume-selective
pTx Volume	1
Vol. Property	B1 Shim Vol.
Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
A >> P	210 mm
R >> L	210 mm
F >> H	148 mm
Vol. Visibility	On

System - Tx/Rx

Frequency 1H	296.941213 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Sequence - Part 1

Introduction	On
Dimension	2D
Asymmetric echo	Off
Contrasts	2
Flow comp.	Yes
Multi-slice mode	Interleaved
Bandwidth	601 Hz/Px

Sequence - Part 2

RF pulse type	Normal
Gradient mode	Normal
RF spoiling	On

Mode	Off

\\USER\Khan\NeuroAnalytics\SNSX_TLE_7t(july2021)\3D_TOF_3slabs_Flow_FH

TA: 5:21 PM: REF Voxel size: 0.3×0.3×0.4 mmPAT: 3 Rel. SNR: 1.00 : fl_r

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slab group	1
Slabs	3
Dist. factor	-17.50 %
Position	R2.9 A7.1 F44.0 mm
Orientation	T > C19.2
Phase enc. dir.	R >> L
AutoAlign	
Phase oversampling	0 %
Slice oversampling	20.0 %
Slices per slab	40
FoV read	210 mm
FoV phase	81.3 %
Slice thickness	0.40 mm
TR	18.0 ms
TE	3.75 ms
Averages	1
Concatenations	3
Filter	Distortion Corr.(2D)
Coil elements	HC2

Contrast - Common

TR TE	18.0 ms
TE	3.75 ms
TD	0.000 ms
MTC	Off
Flip angle	15 deg
Fat suppr. Water suppr.	None
Water suppr.	None

Contrast - Dynamic

Averages	1
Averaging mode	Short term
Reconstruction	Magnitude
Measurements	1

Resolution - Common

FoV read	210 mm
FoV phase	81.3 %
Slice thickness	0.40 mm
Base resolution	704
Phase resolution	100 %
Slice resolution	96 %
Phase partial Fourier	6/8
Slice partial Fourier	6/8
Interpolation	Off

Resolution - iPAT

PAT mode	GRAPPA	
Accel. factor PE	3	
Ref. lines PE	24	
Accel. factor 3D	1	
Reference scan mode	Integrated	

Resolution - Filter Image

Image Filter	Off
Distortion Corr.	On
Mode	2D
Unfiltered images	On
Prescan Normalize	Off
Normalize	Off
B1 filter	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off
POCS	Off

Geometry - Common

Slab group	1
Slabs	3
Dist. factor	-17.50 %
Position	R2.9 A7.1 F44.0 mm
Orientation	T > C19.2
Phase enc. dir.	R >> L
Slice oversampling	20.0 %
Slices per slab	40
FoV read	210 mm
FoV phase	81.3 %
Slice thickness	0.40 mm
TR	18.0 ms
Multi-slice mode	Sequential
Series	Descending
Concatenations	3

Geometry - AutoAlign

Slab group	1
Position	R2.9 A7.1 F44.0 mm
Orientation	T > C19.2
Phase enc. dir.	R >> L
AutoAlign	
Initial Position	R2.9 A7.1 F44.0
R	2.9 mm
Α	7.1 mm
F	44.0 mm
Initial Rotation	90.00 deg
Initial Orientation	T > C
T > C	19.2
> S	0.0

Geometry - Saturation

Fat suppr.	None
Water suppr.	None
Special sat.	None

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table position	Н

Geometry - Tim Planning Suite

Table position	0 mm
Inline Composing	Off

System - Miscellaneous

Positioning mode	REF
Table position	Н
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	
Coil Select Mode	Default

System - Adjustments

В	0 Shim mode	Brain
В	1 Shim mode	Volume-selective
С	onfirm freq. adjustment	Off
Α	ssume Dominant Fat	Off
Α	ssume Silicone	Off
Α	djustment Tolerance	Auto

System - Adjust Volume

! Position	R2.9 A1.3 F41.6 mm
! Orientation	T > C19.7
! Rotation	90.00 deg
! R >> L	169 mm
! A >> P	192 mm
! F >> H	70 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	Volume-selective
pTx Volume	1
Vol. Property	B1 Shim Vol.
Position	R5.1 P14.0 F36.3 mm
Orientation	T > C19.2
Rotation	90.00 deg
R >> L	140 mm
A >> P	137 mm
F >> H	59 mm
Vol. Visibility	On

System - Tx/Rx

Frequency 1H	296.941213 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	18.0 ms
Concatenations	3

Physio - Cardiac

Fat suppr.	None
Dark blood	Off
FoV read	210 mm

Physio - Cardiac

FoV phase	81.3 %
Phase resolution	100 %

Angio - Common

TONE ramp	70 %
Flow direction	F >> H
Flip angle	15 deg
MTC	Off
Measurements	1
3D centric reordering	On

Angio - Inline

Subtract	Off
Measurements	1
StdDev	Off
Save original images	On

Angio - MIP

MIP-Sag MIP-Cor MIP-Tra MIP-Time	On
MIP-Cor	On
MIP-Tra	On
MIP-Time	Off
Save original images	On

Angio - Composing

Inline Composi	ng Off
Distortion Corr.	On
Mode	2D
Unfiltered imag	es On

Sequence - Part 1

Introduction	On
Dimension	3D
Elliptical scanning	On
Asymmetric echo	Allowed
Contrasts	1
Flow comp.	Yes
Multi-slice mode	Sequential
Bandwidth	203 Hz/Px

Sequence - Part 2

Gradient mode	Fast
RF spoiling	On

Mode	Off	
------	-----	--

\\USER\Khan\NeuroAnalytics\SNSX_TLE_7t(july2021)\T2_spc_700iso_p3

TA: 5:32 PM: FIX Voxel size: 0.7×0.7×0.7 mmPAT: 3 Rel. SNR: 1.00 : spc

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slab group	1
Slabs	1
Position	R3.9 A8.2 F28.9 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
AutoAlign	
Phase oversampling	0 %
Slice oversampling	0.0 %
Slices per slab	224
FoV read	202 mm
FoV phase	100.0 %
Slice thickness	0.70 mm
TR	4000 ms
TE	401 ms
Averages	1.0
Concatenations	1
Filter	Elliptical filter
Coil elements	HC2

Contrast - Common

TR	4000 ms
TE	401 ms
MTC	Off
Magn. preparation	None
Fat suppr.	None
Blood suppr.	Off
Restore magn.	Off

Contrast - Dynamic

Averages	1.0
Reconstruction	Magnitude
Measurements	1
Multiple series	Each measurement

Resolution - Common

FoV read	202 mm
FoV phase	100.0 %
Slice thickness	0.70 mm
Base resolution	288
Phase resolution	100 %
Slice resolution	100 %
Phase partial Fourier	Allowed
Slice partial Fourier	Off
Interpolation	Off

Resolution - iPAT

PAT mode	CAIPIRINHA
Accel. factor PE	3
Ref. lines PE	24
Accel. factor 3D	1
Reordering Shift 3D	0
Reference scan mode	Integrated

Resolution - Filter Image

Image Filter	Off	
Distortion Corr.	Off	
Prescan Normalize	Off	
Normalize	Off	
B1 filter	Off	

Resolution - Filter Rawdata

Raw filter	Off	
Elliptical filter	On	

Geometry - Common

Slab group	1
Slabs	1
Position	R3.9 A8.2 F28.9 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
Slice oversampling	0.0 %
Slices per slab	224
FoV read	202 mm
FoV phase	100.0 %
Slice thickness	0.70 mm
TR	4000 ms
Series	Ascending
Concatenations	1

Geometry - AutoAlign

occinion y riatoring.	
Slab group	1
Position	R3.9 A8.2 F28.9 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
AutoAlign	
Initial Position	R3.9 A8.2 F28.9
R	3.9 mm
A	8.2 mm
F	28.9 mm
Initial Rotation	0.00 deg
Initial Orientation	Sagittal

Geometry - Saturation

Fat suppr.	None
Restore magn.	Off
Special sat.	None

Geometry - Navigator

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table position	Н
Table position	0 mm
Inline Composing	Off

System - Miscellaneous

Positioning mode	FIX
Table position	Н
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Adaptive Combine
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	
Coil Select Mode	Default

System - Adjustments

B0 Shim mode	Brain
B1 Shim mode	Volume-selective
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

! Position	R5.8 A6.9 F22.7 mm
! Orientation	Sagittal
! Rotation	0.00 deg
! A >> P	213 mm
! F >> H	173 mm
! R >> L	171 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	Volume-selective
Excitation	Non-sel.
pTx Volume	1
Vol. Property	B1 Shim Vol.
Position	R5.9 A1.1 F20.4 mm
Orientation	Sagittal
Rotation	32.05 deg
A >> P	172 mm
F >> H	122 mm
R >> L	142 mm
Vol. Visibility	On

System - Tx/Rx

•	
Frequency 1H	296.941213 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
Trigger delay	0 ms
TR	4000 ms
Concatenations	1

Physio - Cardiac

Magn. preparation	None
Fat suppr.	None
Dark blood	Off
FoV read	202 mm
FoV phase	100.0 %
Phase resolution	100 %

Physio - PACE

Resp. control	Off
Concatenations	1

Inline - Common

Subtract	Off
Measurements	1
StdDev	Off
Save original images	On

Inline - MIP

MIP-Sag MIP-Cor MIP-Tra MIP-Time	Off	
MIP-Cor	Off	
MIP-Tra	Off	
MIP-Time	Off	
Save original images	On	

Inline - Composing

Inline Composing	Off
Distortion Corr.	Off

Sequence - Part 1

Introduction	Off
Dimension	3D
Elliptical scanning	On
Reordering	Linear
Flow comp.	No
Echo spacing	3.58 ms
Adiabatic-mode	Off
Bandwidth	789 Hz/Px

Sequence - Part 2

Echo train duration	845 ms
RF pulse type	Low SAR
Gradient mode	Fast
Excitation	Non-sel.
Flip angle mode	T2 var
Turbo factor	250

$\verb|\USER\Khan\NeuroAnalytics\SNSX_TLE_7t(july2021)\cb_sp2d_diff_C26_PA_2iso_UFA||$

TA: 8:58 PM: REF Voxel size: 2.0×2.0×2.0 mmPAT: 4 Rel. SNR: 1.00 : epse

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slice group	1
Slices	74
Dist. factor	0 %
Position	R3.0 A8.4 F23.9 mm
Orientation	Transversal
Phase enc. dir.	P >> A
AutoAlign	
Phase oversampling	0 %
FoV read	220 mm
FoV phase	100.0 %
Slice thickness	2.0 mm
TR	6400 ms
TE	91.0 ms
Averages	1
Concatenations	1
Filter	None
Coil elements	HC2

Contrast - Common

TR	6400 ms	
TE	91.0 ms	
MTC	Off	
Magn. preparation	None	
Flip angle exc	80 deg	
Flip angle fat sat	110 deg	
Fat suppr.	Fat sat.	
Fat sat. mode	Strong	

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	1
Delay in TR	0 ms
Multiple series	Off

Resolution - Common

FoV read	220 mm
FoV phase	100.0 %
Slice thickness	2.0 mm
Base resolution	110
Phase resolution	100 %
Phase partial Fourier	6/8
Trajectory	Cartesian
Interpolation	Off

Resolution - iPAT

Accel. mode	Slice accel.
Accel. factor PE	2
Ref. lines PE	24
Accel. factor slice	2
Reference scan mode	EPI/separate

Resolution - Filter Image

Distortion Corr.	Off
Prescan Normalize	Off
Dynamic Field Corr.	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off

Geometry - Common

Slice group	1
Slices	74
Dist. factor	0 %
Position	R3.0 A8.4 F23.9 mm
Orientation	Transversal
Phase enc. dir.	P >> A
FoV read	220 mm
FoV phase	100.0 %
Slice thickness	2.0 mm
TR	6400 ms
Multi-slice mode	Interleaved
Series	Interleaved
Concatenations	1

Geometry - AutoAlign

Slice group	1
Position	R3.0 A8.4 F23.9 mm
Orientation	Transversal
Phase enc. dir.	P >> A
AutoAlign	
Initial Position	R3.0 A8.4 F23.9
R	3.0 mm
Α	8.4 mm
F	23.9 mm
Initial Rotation	-180.00 deg
Initial Orientation	Transversal

Geometry - Saturation

Fat suppr.	Fat sat.
Fat sat. mode	Strong
Special sat.	None

Geometry - Navigator

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table position	Н
Table position	0 mm
Inline Composing	Off

System - Miscellaneous

Positioning mode	REF
Table position	Н

System - Miscellaneous

Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Matrix Optimization	Off
AutoAlign	
Coil Select Mode	Default

System - Adjustments

B0 Shim mode	Advanced
B1 Shim mode	Volume-selective
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

! Position	R5.8 A6.9 F22.7 mm
! Orientation	Sagittal
! Rotation	0.00 deg
! A >> P	213 mm
! F >> H	173 mm
! R >> L	171 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	Volume-selective
Excitation	Standard
pTx Volume	1
Vol. Property	B1 Shim Vol.
Position	R5.9 A1.1 F20.4 mm
Orientation	Sagittal
Rotation	32.05 deg
A >> P	172 mm
F >> H	122 mm
R >> L	142 mm
Vol. Visibility	On

System - Tx/Rx

Frequency 1H	296.941213 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	6400 ms
Concatenations	1

Physio - PACE

Resp. control	Off
Concatenations	1

Diff - Neuro

Diffusion mode	Free
Diff. directions	79
Diffusion Scheme	Monopolar Plus
Diff. weightings	1
b-value	2000 s/mm ²

Diff - Neuro

b-value	1
Diff. weighted images	On
Trace weighted images	Off
ADC maps	Off
FA maps	Off
Mosaic	Off
Tensor	Off
Noise level	40

Diff - Body

Diffusion mode	Free
Diff. directions	79
Diffusion Scheme	Monopolar Plus
Diff. weightings	1
b-value	2000 s/mm ²
b-value	1
Diff. weighted images	On
Trace weighted images	Off
ADC maps	Off
Exponential ADC Maps	Off
FA maps	Off
Invert Gray Scale	Off
Calculated Image	Off
b-Value >=	0 s/mm²
Noise level	40

Diff - Composing

Inline Composing	Off	
Distortion Corr.	Off	

Sequence - Part 1

Introduction	Off
Optimization	None
Multi-slice mode	Interleaved
Free echo spacing	Off
Echo spacing	0.5 ms
Bandwidth	2272 Hz/Px

Sequence - Part 2

EPI factor	110
RF pulse type	Normal
Gradient mode	Fast
Excitation	Standard

Sequence - pTX Pulses

Sequence - Special

FFT scale	1.00
OGSE b-Value	0.0 s/mm2
Diff Slew Rate Scale	11.0
PE Shift	0.0 x dk
Field probe trigger	On
Field probe sync	0#

TA: 0:39 PM: FIX Voxel size: 2.0×2.0×2.0 mmPAT: 4 Rel. SNR: 1.00 : epse

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slice group	1
Slices	74
Dist. factor	0 %
Position	R3.0 A8.4 F23.9 mm
Orientation	Transversal
Phase enc. dir.	A >> P
AutoAlign	
Phase oversampling	0 %
FoV read	220 mm
FoV phase	100.0 %
Slice thickness	2.0 mm
TR	6400 ms
TE	91.0 ms
Averages	1
Concatenations	1
Filter	None
Coil elements	HC2

Contrast - Common

TR	6400 ms
TE	91.0 ms
MTC	Off
Magn. preparation	None
Flip angle exc	80 deg
Flip angle fat sat	110 deg
Fat suppr.	Fat sat.
Fat sat. mode	Strong

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	1
Delay in TR	0 ms
Multiple series	Off

Resolution - Common

FoV read	220 mm
FoV phase	100.0 %
Slice thickness	2.0 mm
Base resolution	110
Phase resolution	100 %
Phase partial Fourier	6/8
Trajectory	Cartesian
Interpolation	Off

Resolution - iPAT

Accel. mode	Slice accel.
Accel. factor PE	2
Ref. lines PE	24
Accel. factor slice	2
Reference scan mode	EPI/separate

Resolution - Filter Image

Distortion Corr.	Off
Prescan Normalize	Off
Dynamic Field Corr.	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off

Geometry - Common

Slice group	1
Slices	74
Dist. factor	0 %
Position	R3.0 A8.4 F23.9 mm
Orientation	Transversal
Phase enc. dir.	A >> P
FoV read	220 mm
FoV phase	100.0 %
Slice thickness	2.0 mm
TR	6400 ms
Multi-slice mode	Interleaved
Series	Interleaved
Concatenations	1

Geometry - AutoAlign

, ,	
Slice group	1
Position	R3.0 A8.4 F23.9 mm
Orientation	Transversal
Phase enc. dir.	A >> P
AutoAlign	
Initial Position	R3.0 A8.4 F23.9
R	3.0 mm
A	8.4 mm
F	23.9 mm
Initial Rotation	0.00 deg
Initial Orientation	Transversal

Geometry - Saturation

Fat suppr.	Fat sat.
Fat sat. mode	Strong
Special sat.	None

Geometry - Navigator

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table position	Н
Table position	0 mm
Inline Composing	Off

System - Miscellaneous

Positioning mode	FIX
Table position	Н

System - Miscellaneous

Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Matrix Optimization	Off
AutoAlign	
Coil Select Mode	Default

System - Adjustments

B0 Shim mode	Advanced
B1 Shim mode	Volume-selective
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

! Position	R5.8 A6.9 F22.7 mm
! Orientation	Sagittal
! Rotation	0.00 deg
! A >> P	213 mm
! F >> H	173 mm
! R >> L	171 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	Volume-selective
Excitation	Standard
pTx Volume	1
Vol. Property	B1 Shim Vol.
Position	R5.9 A1.1 F20.4 mm
Orientation	Sagittal
Rotation	32.05 deg
A >> P	172 mm
F >> H	122 mm
R >> L	142 mm
Vol. Visibility	On

System - Tx/Rx

Frequency 1H	296.941213 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	6400 ms
Concatenations	1

Physio - PACE

Resp. control	Off
Concatenations	1

Diff - Neuro

Diffusion mode	Free
Diff. directions	79
Diffusion Scheme	Monopolar Plus
Diff. weightings	1
b-value	0 s/mm²

Diff - Neuro

b-value	1
Diff. weighted images	On
Trace weighted images	Off
ADC maps	Off
FA maps	Off
Mosaic	Off
Tensor	Off
Noise level	40

Diff - Body

Diffusion mode	Free
	riee
Diff. directions	79
Diffusion Scheme	Monopolar Plus
Diff. weightings	1
b-value	0 s/mm²
b-value	1
Diff. weighted images	On
Trace weighted images	Off
ADC maps	Off
Exponential ADC Maps	Off
FA maps	Off
Invert Gray Scale	Off
Calculated Image	Off
b-Value >=	0 s/mm²
Noise level	40

Diff - Composing

Inline Composing	Off	
Distortion Corr.	Off	

Sequence - Part 1

Introduction	Off
Optimization	None
Multi-slice mode	Interleaved
Free echo spacing	Off
Echo spacing	0.5 ms
Bandwidth	2272 Hz/Px

Sequence - Part 2

EPI factor	110
RF pulse type	Normal
Gradient mode	Fast
Excitation	Standard

Sequence - pTX Pulses

Sequence - Special

FFT scale	1.00
OGSE b-Value	0.0 s/mm2
Diff Slew Rate Scale	11.0
PE Shift	0.0 x dk
Field probe trigger	On
Field probe sync	0#

\\USER\Khan\NeuroAnalytics\SNSX_TLE_7t(july2021)\spc_dir_sag_750iso_p2

TA: 10:18 PM: FIX Voxel size: 0.8×0.8×0.8 mmPAT: 2 Rel. SNR: 1.00 : spcir

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further	Off
preparation	
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slab group	1
Slabs	1
Position	R2.9 A16.0 F31.2 mm
Orientation	Sagittal
Phase enc. dir.	H >> F
AutoAlign	
Phase oversampling	0 %
Slice oversampling	0.0 %
Slices per slab	208
FoV read	240 mm
FoV phase	84.4 %
Slice thickness	0.75 mm
TR	9500 ms
TE	350 ms
Averages	1.0
Concatenations	1
Filter	Elliptical filter, Image
	Filter
Coil elements	HC2

Contrast - Common

TR	9500 ms
TE	350 ms
MTC	Off
Magn. preparation	Non-sel. DIR
TI 1	3800 ms
TI 2	650 ms
Fat suppr.	None
Blood suppr.	Off
Restore magn.	Off

Contrast - Dynamic

Averages	1.0
Reconstruction	Magnitude
Measurements	1
Multiple series	Each measurement

Resolution - Common

FoV read	240 mm
FoV phase	84.4 %
Slice thickness	0.75 mm
Base resolution	320
Phase resolution	100 %
Slice resolution	100 %
Phase partial Fourier	Allowed
Slice partial Fourier	6/8

Resolution - Common

Interpolation	Off
-	

Resolution - iPAT

PAT mode	GRAPPA
Accel. factor PE	2
Ref. lines PE	24
Accel. factor 3D	1
Reference scan mode	Integrated

Resolution - Filter Image

Image Filter	On
Intensity	Medium
Edge Enhancement	3
Smoothing	3
Unfiltered images	Off
Distortion Corr.	Off
Prescan Normalize	Off
Normalize	Off
B1 filter	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	On

Geometry - Common

Slab group	1
Slabs	1
Position	R2.9 A16.0 F31.2 mm
Orientation	Sagittal
Phase enc. dir.	H >> F
Slice oversampling	0.0 %
Slices per slab	208
FoV read	240 mm
FoV phase	84.4 %
Slice thickness	0.75 mm
TR	9500 ms
Series	Ascending
Concatenations	1

Geometry - AutoAlign

Slab group	1
Position	R2.9 A16.0 F31.2 mm
Orientation	Sagittal
Phase enc. dir.	H >> F
AutoAlign	
Initial Position	R2.9 A16.0 F31.2
R	2.9 mm
A	16.0 mm
F	31.2 mm
Initial Rotation	90.00 deg
Initial Orientation	Sagittal

Geometry - Saturation

Fat suppr.	None
Restore magn.	Off
Special sat.	None

Geometry - Navigator

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table position	Н
Table position	0 mm
Inline Composing	Off

System - Miscellaneous

Positioning mode	FIX
Table position	Н
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	
Coil Select Mode	Default

System - Adjustments

B0 Shim mode	Brain
B1 Shim mode	Volume-selective
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

! Position	R3.6 A8.3 F14.0 mm
! Orientation	Sagittal
! Rotation	0.00 deg
! A >> P	213 mm
! F >> H	166 mm
! R >> L	171 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	Volume-selective
Excitation	Non-sel.
pTx Volume	1
Vol. Property	B1 Shim Vol.
Position	R3.0 A5.4 F16.1 mm
Orientation	Sagittal
Rotation	27.17 deg
A >> P	168 mm
F >> H	120 mm
R >> L	131 mm
Vol. Visibility	On

System - Tx/Rx

Frequency 1H	296.941213 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
Trigger delay	0 ms
TR	9500 ms
Concatenations	1

Physio - Cardiac

Magn. preparation	Non-sel. DIR
TI 1	3800 ms
TI 2	650 ms
Fat suppr.	None
Dark blood	Off
FoV read	240 mm
FoV phase	84.4 %
Phase resolution	100 %

Physio - PACE

Resp. control	Off	
Concatenations	1	

Inline - Common

Subtract	Off
Measurements	1
StdDev	Off
Save original images	On

Inline - MIP

MIP-Sag	Off
MIP-Sag MIP-Cor MIP-Tra	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

Inline - Composing

Inline Composing	Off
Distortion Corr.	Off

Sequence - Part 1

Introduction	Off
Dimension	3D
Elliptical scanning	On
Reordering	Linear
Flow comp.	No
Echo spacing	3.5 ms
Adiabatic-mode	Off
Bandwidth	781 Hz/Px

Sequence - Part 2

Echo train duration	872 ms
RF pulse type	Low SAR
Gradient mode	Fast
Excitation	Non-sel.
Flip angle mode	T2 var
Turbo factor	300

Allowed dela	<i>(</i> 0 s	

$\verb|\USER\Khan\NeuroAnalytics\SNSX_TLE_7t(july2021)\T2_spc_600 iso_Mayo|\\$

TA: 8:14 PM: FIX Voxel size: 0.3×0.3×0.6 mmPAT: 6 Rel. SNR: 1.00 : spc

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slab group	1
Slabs	1
Position	R3.9 A8.2 F28.9 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
AutoAlign	
Phase oversampling	0 %
Slice oversampling	33.3 %
Slices per slab	240
FoV read	225 mm
FoV phase	99.4 %
Slice thickness	0.60 mm
TR	2000 ms
TE	131 ms
Averages	1.0
Concatenations	1
Filter	Distortion Corr.(2D), B1
	filter
Coil elements	HC2

Contrast - Common

TR	2000 ms
TE	131 ms
MTC	Off
Magn. preparation	None
Fat suppr.	None
Blood suppr.	Off
Restore magn.	Off

Contrast - Dynamic

Averages	1.0
Reconstruction	Magnitude
Measurements	1
Multiple series	Each measurement

Resolution - Common

FoV read	225 mm
FoV phase	99.4 %
Slice thickness	0.60 mm
Base resolution	352
Phase resolution	100 %
Slice resolution	100 %
Phase partial Fourier	Allowed
Slice partial Fourier	Off
Interpolation	On

Resolution - iPAT

PAT mode	CAIPIRINHA
Accel. factor PE	3
Ref. lines PE	48
Accel. factor 3D	2
Ref. lines 3D	32
Reordering Shift 3D	0
Reference scan mode	Integrated

Resolution - Filter Image

Image Filter	Off
Distortion Corr.	On
Mode	2D
Unfiltered images	Off
Prescan Normalize	Off
Normalize	Off
B1 filter	On
Unfiltered images	Off

Resolution - Filter Rawdata

Raw filter	Off	
Elliptical filter	Off	

Geometry - Common

_	
Slab group	1
Slabs	1
Position	R3.9 A8.2 F28.9 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
Slice oversampling	33.3 %
Slices per slab	240
FoV read	225 mm
FoV phase	99.4 %
Slice thickness	0.60 mm
TR	2000 ms
Series	Ascending
Concatenations	1

Geometry - AutoAlign

Slab group	1
Position	R3.9 A8.2 F28.9 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
AutoAlign	
Initial Position	R3.9 A8.2 F28.9
R	3.9 mm
Α	8.2 mm
F	28.9 mm
Initial Rotation	0.00 deg
Initial Orientation	Sagittal

Geometry - Saturation

Fat suppr.	None
Restore magn.	Off
Special sat.	None

Geometry - Navigator

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off	
Table position	Н	

Geometry - Tim Planning Suite

Table position	0 mm
Inline Composing	Off

System - Miscellaneous

Positioning mode	FIX
Table position	Н
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Adaptive Combine
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	
Coil Select Mode	Default

System - Adjustments

B0 Shim mode	Brain
B1 Shim mode	TrueForm
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

! Position	R5.8 A6.9 F22.7 mm
! Orientation	Sagittal
! Rotation	0.00 deg
! A >> P ! F >> H	213 mm
! F >> H	173 mm
! R >> L	171 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Non-sel.

System - Tx/Rx

Frequency 1H	296.941213 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
Trigger delay	0 ms
TR	2000 ms
Concatenations	1

Physio - Cardiac

•	
Magn. preparation	None
Fat suppr.	None
Dark blood	Off
FoV read	225 mm
FoV phase	99.4 %
Phase resolution	100 %

Physio - PACE

Resp. control	Off
Concatenations	1

Inline - Common

Subtract	Off
Measurements	1
StdDev	Off
Save original images	On

Inline - MIP

MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

Inline - Composing

Inline Composing	Off
Distortion Corr.	On
Mode	2D
Unfiltered images	Off

Sequence - Part 1

Introduction	On
Dimension	3D
Elliptical scanning	On
Reordering	Linear
Flow comp.	No
Echo spacing	4.96 ms
Adiabatic-mode	Off
Bandwidth	473 Hz/Px

Sequence - Part 2

Echo train duration	283 ms
RF pulse type	Low SAR
Gradient mode	Fast
Excitation	Non-sel.
Flip angle mode	T2 var
Turbo factor	65

Allowed delay	0 s	
---------------	-----	--

\\USER\Khan\NeuroAnalytics\SNSX_TLE_7t(july2021)\T2_spc_600iso_Mayo_CP

TA: 6:44 PM: FIX Voxel size: 0.6×0.6×0.6 mmPAT: 6 Rel. SNR: 1.00 : spc

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slab group	1
Slabs	1
Position	R3.1 A20.6 F21.3 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
AutoAlign	
Phase oversampling	0 %
Slice oversampling	6.7 %
Slices per slab	240
FoV read	225 mm
FoV phase	99.4 %
Slice thickness	0.60 mm
TR	2000 ms
TE	131 ms
Averages	1.0
Concatenations	1
Filter	Distortion Corr.(3D), B1
	filter
Coil elements	HC2

Contrast - Common

TR	2000 ms
TE	131 ms
MTC	Off
Magn. preparation	None
Fat suppr.	None
Fat suppr. Blood suppr.	Off
Restore magn.	Off

Contrast - Dynamic

Averages	1.0
Reconstruction	Magnitude
Measurements	1
Multiple series	Each measurement

Resolution - Common

FoV read	225 mm
FoV phase	99.4 %
Slice thickness	0.60 mm
Base resolution	352
Phase resolution	100 %
Slice resolution	100 %
Phase partial Fourier	Allowed
Slice partial Fourier	Off
Interpolation	Off

Resolution - iPAT

PAT mode	CAIPIRINHA
Accel. factor PE	3
Ref. lines PE	48
Accel. factor 3D	2
Ref. lines 3D	32
Reordering Shift 3D	0
Reference scan mode	Integrated

Resolution - Filter Image

Image Filter	Off
Distortion Corr.	On
Mode	3D
Unfiltered images	On
Prescan Normalize	Off
Normalize	Off
B1 filter	On
Unfiltered images	Off

Resolution - Filter Rawdata

Raw filter	Off	
Elliptical filter	Off	

Geometry - Common

Slab group	1
Slabs	1
Position	R3.1 A20.6 F21.3 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
Slice oversampling	6.7 %
Slices per slab	240
FoV read	225 mm
FoV phase	99.4 %
Slice thickness	0.60 mm
TR	2000 ms
Series	Ascending
Concatenations	1

Geometry - AutoAlign

Slab group	1
Position	R3.1 A20.6 F21.3 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
AutoAlign	
Initial Position	R3.1 A20.6 F21.3
R	3.1 mm
A	20.6 mm
F	21.3 mm
Initial Rotation	0.00 deg
Initial Orientation	Sagittal

Geometry - Saturation

Fat suppr.	None
Restore magn.	Off
Special sat.	None

Geometry - Navigator

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table position	Н

Geometry - Tim Planning Suite

Table position	0 mm
Inline Composing	Off

System - Miscellaneous

Positioning mode	FIX
Table position	Н
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Adaptive Combine
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	
Coil Select Mode	Default

System - Adjustments

B0 Shim mode	Brain
B1 Shim mode	TrueForm
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

! Position	R3.6 A9.1 F0.9 mm
! Orientation	Sagittal
! Rotation	0.00 deg
! A >> P	206 mm
! F >> H	166 mm
! R >> L	168 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Non-sel.

System - Tx/Rx

Frequency 1H	296.941213 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
Trigger delay	0 ms
TR	2000 ms
Concatenations	1

Physio - Cardiac

•	
Magn. preparation	None
Fat suppr.	None
Dark blood	Off
FoV read	225 mm
FoV phase	99.4 %
Phase resolution	100 %

Physio - PACE

Resp. control	Off
Concatenations	1

Inline - Common

Subtract	Off
Measurements	1
StdDev	Off
Save original images	On

Inline - MIP

MIP-Sag	Off
MIP-Cor	Off
	•
MIP-Tra	Off
MIP-Time	Off
Save original images	On

Inline - Composing

Inline Composing	Off
Distortion Corr.	On
Mode	3D
Unfiltered images	On

Sequence - Part 1

Introduction	On
Dimension	3D
Elliptical scanning	On
Reordering	Linear
Flow comp.	No
Echo spacing	4.96 ms
Adiabatic-mode	Off
Bandwidth	473 Hz/Px

Sequence - Part 2

Echo train duration	283 ms
RF pulse type	Low SAR
Gradient mode	Fast
Excitation	Non-sel.
Flip angle mode	T2 var
Turbo factor	65

Allowed delay	0 s	
---------------	-----	--

\\USER\Khan\NeuroAnalytics\SNSX_TLE_7t(july2021)\spc_dir_sag_750iso_p2

TA: 10:18 PM: FIX Voxel size: 0.8×0.8×0.8 mmPAT: 2 Rel. SNR: 1.00 : spcir

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slab group	1
Slabs	1
Position	R2.9 A16.0 F31.2 mm
Orientation	Sagittal
Phase enc. dir.	H >> F
AutoAlign	
Phase oversampling	0 %
Slice oversampling	0.0 %
Slices per slab	208
FoV read	240 mm
FoV phase	84.4 %
Slice thickness	0.75 mm
TR	9500 ms
TE	350 ms
Averages	1.0
Concatenations	1
Filter	Elliptical filter, Image Filter
Coil elements	HC2

Contrast - Common

TR	9500 ms
TE	350 ms
MTC	Off
Magn. preparation	Non-sel. DIR
TI 1	3800 ms
TI 2	650 ms
Fat suppr.	None
Blood suppr.	Off
Restore magn.	Off

Contrast - Dynamic

Averages	1.0
Reconstruction	Magnitude
Measurements	1
Multiple series	Each measurement

Resolution - Common

FoV read	240 mm
FoV phase	84.4 %
Slice thickness	0.75 mm
Base resolution	320
Phase resolution	100 %
Slice resolution	100 %
Phase partial Fourier	Allowed
Slice partial Fourier	6/8

Resolution - Common

Interpolation	Off

Resolution - iPAT

PAT mode	GRAPPA
Accel. factor PE	2
Ref. lines PE	24
Accel. factor 3D	1
Reference scan mode	Integrated

Resolution - Filter Image

Image Filter	On
Intensity	Medium
Edge Enhancement	3
Smoothing	3
Unfiltered images	Off
Distortion Corr.	Off
Prescan Normalize	Off
Normalize	Off
B1 filter	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	On

Geometry - Common

Slab group	1
Slabs	1
Position	R2.9 A16.0 F31.2 mm
Orientation	Sagittal
Phase enc. dir.	H >> F
Slice oversampling	0.0 %
Slices per slab	208
FoV read	240 mm
FoV phase	84.4 %
Slice thickness	0.75 mm
TR	9500 ms
Series	Ascending
Concatenations	1

Geometry - AutoAlign

Slab group	1
Position	R2.9 A16.0 F31.2 mm
Orientation	Sagittal
Phase enc. dir.	H >> F
AutoAlign	
Initial Position	R2.9 A16.0 F31.2
R	2.9 mm
A	16.0 mm
F	31.2 mm
Initial Rotation	90.00 deg
Initial Orientation	Sagittal

Geometry - Saturation

Fat suppr.	None
Restore magn.	Off
Special sat.	None

Geometry - Navigator

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off	
Table position	Н	
Table position	0 mm	
Inline Composing	Off	

System - Miscellaneous

Positioning mode	FIX
Table position	Н
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	
Coil Select Mode	Default

System - Adjustments

B0 Shim mode	Brain
B1 Shim mode	Volume-selective
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

! Position	R3.6 A8.3 F14.0 mm
! Orientation	Sagittal
! Rotation	0.00 deg
! A >> P	213 mm
! F >> H	166 mm
! R >> L	171 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	Volume-selective
Excitation	Non-sel.
pTx Volume	1
Vol. Property	B1 Shim Vol.
Position	R3.0 A5.4 F16.1 mm
Orientation	Sagittal
Rotation	27.17 deg
A >> P	168 mm
F >> H	120 mm
R >> L	131 mm
Vol. Visibility	On

System - Tx/Rx

Frequency 1H	296.941213 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
Trigger delay	0 ms
TR	9500 ms
Concatenations	1

Physio - Cardiac

Magn. preparation	Non-sel. DIR
TI 1	3800 ms
TI 2	650 ms
Fat suppr.	None
Dark blood	Off
FoV read	240 mm
FoV phase	84.4 %
Phase resolution	100 %

Physio - PACE

Resp. control	Off	
Concatenations	1	

Inline - Common

Subtract	Off
Measurements	1
StdDev	Off
Save original images	On

Inline - MIP

MIP-Sag	Off	
MIP-Sag MIP-Cor MIP-Tra	Off	
MIP-Tra	Off	
MIP-Time	Off	
Save original images	On	

Inline - Composing

Inline Composing	Off	
Distortion Corr.	Off	

Sequence - Part 1

Introduction	Off
Dimension	3D
Elliptical scanning	On
Reordering	Linear
Flow comp.	No
Echo spacing	3.5 ms
Adiabatic-mode	Off
Bandwidth	781 Hz/Px

Sequence - Part 2

Echo train duration	872 ms
RF pulse type	Low SAR
Gradient mode	Fast
Excitation	Non-sel.
Flip angle mode	T2 var
Turbo factor	300

Allowed delay	<i>o</i> 0 s	