



# National Institute of Standards & Technology

## Certificate of Analysis

### Standard Reference Material<sup>®</sup> 915b

#### Calcium Carbonate

This Standard Reference Material (SRM) is intended for use as an analytical standard of known purity. It is intended primarily for use in the calibration and standardization of procedures for calcium (Ca) determinations employed in clinical analysis and for routine critical evaluation of the daily working standards used in these procedures. This lot of calcium carbonate ( $\text{CaCO}_3$ ) was prepared to ensure a material of high purity and homogeneity and has been assayed after heating at 200 °C to 210 °C. A unit of SRM 915b consists of a single glass bottle containing 20 g of the material.

**Certified Values:** Table 1 lists the certified values for this SRM, expressed as mass fractions,  $w$ , of  $\text{CaCO}_3$ , Ca, and carbonate ( $\text{CO}_3^{2-}$ ). A NIST certified value is a value for which NIST has the highest confidence in its accuracy in that all known or suspected sources of bias have been investigated or taken into account [1]. The measurands are the mass fractions for Ca,  $\text{CO}_3$ , and  $\text{CaCO}_3$ . Metrological traceability is to the SI units for mass, current, and time in the coulometric assay and to the SI unit for mass in the gravimetric assay.

Table 1. Certified Values<sup>(a)</sup> for SRM 915b Calcium Carbonate

Quantity	Value (%)		
$w_{\text{CaCO}_3}$	99.907	±	0.021
$w_{\text{Ca}}$	40.0104	±	0.0083
$w_{\text{CO}_3}$	59.923	±	0.012

<sup>(a)</sup> Each result is expressed as the certified value  $\pm$  the expanded uncertainty,  $U$ , calculated as  $U = ku_c$ , where  $u_c$  is the combined standard uncertainty calculated according to the ISO/JCGM and NIST Guides [2]. The value of  $u_c$  is intended to represent, at the level of one standard deviation, the combined effect of inherent sources of uncertainty of the assay techniques, and applicable corrections for interfering trace elements. The value of  $k$  is 2, which is the coverage factor corresponding to approximately 95 % confidence based on > 60 overall effective degrees of freedom.

**Expiration of Certification:** The certification of **SRM 915b** is valid, within the measurement uncertainty specified, until **01 March 2026**, provided the SRM is handled and stored in accordance with the instructions given in this certificate (see “Instructions for Use”). The certification is nullified if the SRM is damaged, contaminated, or otherwise modified.

Coulometric and gravimetric analyses were performed by K.W. Pratt and T.W. Vetter, respectively, of the NIST Chemical Sciences Division. Trace element analyses by glow-discharge mass spectrometry and other techniques [3] were performed by commercial laboratories.

Coordination of the technical measurements leading to the certification of SRM 915b was provided by T.W. Vetter.

Statistical consultation was provided by W.F. Guthrie of the NIST Statistical Engineering Division.

Support aspects involved in the issuance of this SRM were coordinated through the NIST Office of Reference Materials.

Carlos A. Gonzalez, Chief  
Chemical Sciences Division

Gaithersburg, MD 20899  
Certificate Issue Date: 13 September 2016  
*Certificate Revision History on Last Page*

Steven J. Choquette, Director  
Office of Reference Materials

**Information Values:** Table 2 lists information values for the mass fractions of trace elements in SRM 915b. Information values are non-certified values that may be of interest and use to the SRM user, but insufficient information is available to provide an uncertainty associated with the value [1]. No other elements were detected at a mass fraction greater than 1 µg/g. Information values cannot be used to establish metrological traceability.

Table 2. Trace Elements in SRM 915b Calcium Carbonate

Element	Mass Fraction (µg/g)
Strontium (Sr)	150
Magnesium (Mg)	40
Sulfur (S)	30
Sodium (Na)	17
Chlorine (Cl)	8
Silicon (Si)	5
Phosphorus (P)	3
Barium (Ba)	2

**Maintenance of SRM Certification:** NIST will monitor this SRM lot over the period of its certification. If substantive technical changes occur that affect the certification before the expiration of this certificate, NIST will notify the purchaser. Registration (see attached sheet or register online) will facilitate notification.

#### NOTICE AND WARNINGS TO USERS

This SRM is intended for research use.

**Stability and Storage:** This SRM should be stored in its original bottle at room temperature. It must be tightly re-capped after use and protected from moisture and acid vapors.

**Homogeneity:** This SRM is homogeneous within the uncertainty limits for the nominal sample mass, 150 mg, used for the coulometric assays. Samples less than 150 mg are not recommended in order to avoid possible heterogeneity with smaller sample sizes.

**Possible Interfering Species:** It is the responsibility of the user to evaluate which species may interfere with the application of this SRM and to apply any necessary corrections that affect the given application. The following information and the values in Table 2 may be useful in this evaluation.

The certified value for  $w_{\text{CaCO}_3}$  is obtained from an equally-weighted combination of the results of independent coulometric analyses, corrected for trace impurities of strontium carbonate ( $\text{SrCO}_3$ ), magnesium carbonate ( $\text{MgCO}_3$ ), sodium carbonate ( $\text{Na}_2\text{CO}_3$ ), barium carbonate ( $\text{BaCO}_3$ ), and calcium hydrogen phosphate ( $\text{CaHPO}_4$ ); and gravimetric analyses, corrected for trace impurities of  $\text{SrCO}_3$ ,  $\text{MgCO}_3$ , and  $\text{Na}_2\text{CO}_3$ .

The certified value for  $w_{\text{Ca}}$  is obtained from an equally-weighted combination of  $w_{\text{Ca}}$  obtained directly from the gravimetric analyses and the indirect  $w_{\text{Ca}}$ , which is calculated from the stoichiometric amount of Ca expected from the coulometric  $w_{\text{CO}_3}$  plus the additional Ca in the trace impurities of calcium sulfate ( $\text{CaSO}_4$ ), calcium chloride ( $\text{CaCl}_2$ ), and  $\text{CaHPO}_4$  minus the excess  $\text{CO}_3^{2-}$  in the trace impurities of  $\text{SrCO}_3$ ,  $\text{MgCO}_3$ ,  $\text{Na}_2\text{CO}_3$ , and  $\text{BaCO}_3$ .

The certified value for  $w_{\text{CO}_3}$  is obtained from an equally-weighted combination of  $w_{\text{CO}_3}$  obtained directly from the coulometric analyses and the indirect  $w_{\text{CO}_3}$ , which is calculated from the stoichiometric amount of  $\text{CO}_3^{2-}$  expected from the gravimetric  $w_{\text{Ca}}$  plus the additional  $\text{CO}_3^{2-}$  in the trace impurities  $\text{SrCO}_3$ ,  $\text{MgCO}_3$ , and  $\text{Na}_2\text{CO}_3$  minus the excess Ca in the trace impurities of  $\text{CaSO}_4$ ,  $\text{CaCl}_2$ , and  $\text{CaHPO}_4$ .

The corrections for trace impurities were obtained from the trace element determinations and the appropriate gravimetric factors [4]. A portion of the Ca is present in SRM 915b as  $\text{CaSO}_4$ ,  $\text{CaCl}_2$ , and  $\text{CaHPO}_4$ , and a portion of the  $\text{CO}_3^{2-}$  is present as  $\text{SrCO}_3$ ,  $\text{MgCO}_3$ ,  $\text{Na}_2\text{CO}_3$ , and  $\text{BaCO}_3$ . Hence, the sum of the certified values for  $w_{\text{Ca}}$  and  $w_{\text{CO}_3}$  does not equal the certified value for  $w_{\text{CaCO}_3}$ .

## INSTRUCTIONS FOR USE

**Drying Instructions:** Dry the material at 200 °C to 210 °C for 4 h. After the SRM has been dried, store it in a desiccator over anhydrous magnesium perchlorate.

**Source of Material:** The  $\text{CaCO}_3$  used for this SRM was obtained from a commercial supplier. The material was examined for compliance with the specification for reagent grade  $\text{CaCO}_3$  as specified by the American Chemical Society [3]. The material was found to meet or exceed the minimum requirements in every respect.

**Assay Techniques:** The coulometric assay value was obtained by automated back-titration [5] using coulometrically-standardized hydrochloric acid (HCl) as the excess added substance, with potentiometric detection of the strong acid endpoint after removal of the product carbon dioxide ( $\text{CO}_2$ ). The gravimetric assay value was obtained by conversion to  $\text{CaSO}_4$  and correcting for trace contaminants in the  $\text{CaSO}_4$  (gravimetric procedure based on reference 6).

## REFERENCES

- [1] May, W.; Parris, R.; Beck II, C.; Fassett, J.; Greenberg, R.; Guenther, F.; Kramer, G.; Wise, S.; Gills, T.; Colbert, J.; Gettings, R.; MacDonald, B.; *Definitions of Terms and Modes Used at NIST for Value Assignment of Reference Materials for Chemical Measurements*; NIST Special Publication 260-136 (2000); available at <http://www.nist.gov/srm/upload/SP260-136.PDF> (accessed Sep 2016).
- [2] JCGM 100:2008; *Evaluation of Measurement Data — Guide to the Expression of Uncertainty in Measurement*; (GUM 1995 with Minor Corrections), Joint Committee for Guides in Metrology (JCGM) (2008); available at [http://www.bipm.org/utis/common/documents/jcgm/JCGM\\_100\\_2008\\_E.pdf](http://www.bipm.org/utis/common/documents/jcgm/JCGM_100_2008_E.pdf) (accessed Sep 2016); see also Taylor, B.N.; Kuyatt, C.E.; *Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Results*; NIST Technical Note 1297; U.S. Government Printing Office: Washington, DC (1994); available at <http://www.nist.gov/pml/pubs/index.cfm> (accessed Sep 2016).
- [3] *Reagent Chemicals*; 9th ed., American Chemical Society: Washington, DC (1999).
- [4] *IUPAC Commission of Atomic Weights and Isotopic Abundances*; *Pure & Appl. Chem.*, Vol. 75 (8), pp. 1107–1122 (2003).
- [5] Pratt, K.W.; *Automated, High-Precision Coulometric Titrimetry Part II. Strong and Weak Acids and Bases*; *Anal. Chim. Acta*, Vol. 289 (2), pp. 135–142 (1994).
- [6] Moody, J.R.; Vetter, T.W.; *Development of the Ion Exchange-Gravimetric Method for Sodium in Serum as a Definitive Method*; *J. Res. Natl. Inst. Stand. Technol.*, Vol. 101, pp. 155–164 (1996); available at <http://nvl.nist.gov/pub/nistpubs/jres/101/2/j2mood.pdf> (accessed Sep 2016).

**Certificate Revision History:** 13 September 2016 (Editorial changes); 09 December 2015 (Change of expiration date; editorial changes); 07 April 2006 (Original certificate date).

*Users of this SRM should ensure that the Certificate of Analysis in their possession is current. This can be accomplished by contacting the SRM Program: telephone (301) 975-2200; fax (301) 948-3730; e-mail [srminfo@nist.gov](mailto:srminfo@nist.gov); or via the Internet at <http://www.nist.gov/srm>.*