DATE: 19 January 2018

## **Product Identifier**

SRM Number: 2216

**SRM Name:** Miniaturized Low-Energy Charpy V-Notch KLST Impact Specimens (Self-Verification)

Under the U.S. Department of Labor, Occupational Safety and Health Administration (OSHA) 29 CFR 1910.1200, this Standard Reference Material (SRM) is NOT classified as a physical hazard or a health hazard, a simple asphyxiant, combustible dust, pyrophoric gas, or hazard not otherwise classified. There are no hazard pictograms, hazard statements or signal word associated with it. Safety Data Sheet information is not required. This document may be used in conjunction with your hazard communication program.

**Exemption:** 1910.1200 (c). This SRM is an Article, as the word is defined by OSHA, where *Article* means a manufactured item other than a fluid or particle: (i) which is formed to a specific shape or design during manufacture; (ii) which has end use function(s) dependent in whole or in part upon its shape or design during end use; and (iii) which under normal conditions of use does not release more than very small quantities, e.g., minute or trace amounts of a hazardous chemical (as determined under paragraph (d) of 1910.1200), and does not pose a physical hazard or health risk to employees.

**Description:** This Standard Reference Material (SRM) is intended for the verification of maximum force and absorbed energy values measured at room temperature using a small-scale Charpy machine in accordance with the current ASTM Standard E2248 or the current ISO Standard 14556. A unit of SRM 2216 consists of a set of three KLST-type miniaturized Charpy specimens needed to perform a single verification. SRM 2216 can be used to verify the absorbed energy scale of a small-scale impact machine at the low-energy level (approximately 1.6 J).

**Disposal:** SRM 2216 should be disposed of in accordance with local, state, and federal regulations.

**Transport Information:** This material is not regulated by the U.S. Department of Transportation (DOT) and/or International Air Transport Association (IATA).

**Disclaimer:** This document was prepared carefully, using current references. Users of this SRM should ensure that this document and the corresponding Certificate in their possession are current. This can be accomplished by contacting the SRM Program: telephone (301) 975-2200; fax (301) 948-3730; e-mail srmmsds@nist.gov; or via the Internet at http://www.nist.gov/srm.

SRM 2216 Page 1 of 1