

White noise sensor calibration processing results

Processed by dbcalbrt:glenn:swarm at 2008302:00:14:55.924

Channel: PS_PS10_BHN		Time: 2008298:23:40:42.000		Sequence: PS_PS10-2008298:23:40:42	
DImodel: q330		DIserial: 0100000A415DB228		Snmodel: -	Snserial: -
Ref Channel: PS_PS10_BHN		Ref Time: 2008298:22:57:48.000		Ref Sequence: PS_PS10-2008298:22:57:48	
Ref DImodel: q330		Ref DIserial: 0100000A415DB228		Ref Snmodel: -	Ref Snserial: -
Cal mode: cmp	Cal Waveform: white	Cal Duration: 1:00 minutes		Samplerate: 50	Cal Amplitude: 0.625 V
Cal processing: ratio	Cal Settle Time: 10.000 seconds	Cal Trailer Time: 20.000 seconds			

Amp Ratio: 0.007542	Norm Freq: 1.000 Hz
------------------------	------------------------

Processing Parameters:	
{bands}[0]{fmax} = 0.02	{bands}[2]{nwindows} = 0
{bands}[0]{fmin} = 0.000001	{bands}[2]{overlap_percent} = 50.0
{bands}[0]{nwindows} = 1	{bands}[2]{taper_percent} = 50.0
{bands}[0]{overlap_percent} = 0.0	{bands}[3]{fmax} = 200.0
{bands}[0]{taper_percent} = 0.0	{bands}[3]{fmin} = 1.00
{bands}[1]{fmax} = 1.0	{bands}[3]{nwindows} = 0
{bands}[1]{fmin} = 0.0025	{bands}[3]{overlap_percent} = 50.0
{bands}[1]{nwindows} = 0	{bands}[3]{taper_percent} = 50.0
{bands}[1]{overlap_percent} = 50.0	{tflag} = 1000.0
{bands}[1]{taper_percent} = 25.0	{tlead} = 30.0
{bands}[2]{fmax} = 10.0	
{bands}[2]{fmin} = 0.05	

