

White noise sensor calibration processing results

Processed by dbcalbrt:glenn.inverse at 2008303:19:38:00.505

Channel: PS_PS04_HNZ		Time: 2008302:12:10:12.000		Sequence: PS_PS04-2008302:12:10:12	
DImodel: q330		DIserial: 0100000A415B7D0F		Snmodel: -	Snsrserial: -
Ref Channel: PS_PS04_HNZ		Ref Time: 2008302:04:10:12.000		Ref Sequence: PS_PS04-2008302:04:10:12	
Ref DImodel: q330		Ref DIserial: 0100000A415B7D0F		Ref Snmodel: -	Ref Snsrserial: -
Cal mode: cmp	Cal Waveform: white	Cal Duration: 1:00 hours		Samplerate: 200	Cal Amplitude: 0.625 V
Cal processing: ratio	Cal Settle Time: 10:00 minutes	Cal Trailer Time: 10:00 minutes			

Amp Ratio: 0.997750	Norm Freq: 1.000 Hz
------------------------	------------------------

Processing Parameters:	
{bands}[0]{fmax} = 0.02	{bands}[2]{nwindows} = 0
{bands}[0]{fmin} = 0.000001	{bands}[2]{overlap_percent} = 50.0
{bands}[0]{nwindows} = 1	{bands}[2]{taper_percent} = 50.0
{bands}[0]{overlap_percent} = 0.0	{bands}[3]{fmax} = 200.0
{bands}[0]{taper_percent} = 0.0	{bands}[3]{fmin} = 1.00
{bands}[1]{fmax} = 1.0	{bands}[3]{nwindows} = 0
{bands}[1]{fmin} = 0.0025	{bands}[3]{overlap_percent} = 50.0
{bands}[1]{nwindows} = 0	{bands}[3]{taper_percent} = 50.0
{bands}[1]{overlap_percent} = 50.0	{tflag} = 1000.0
{bands}[1]{taper_percent} = 25.0	{tlead} = 30.0
{bands}[2]{fmax} = 10.0	
{bands}[2]{fmin} = 0.05	

