

AND

Digital Blood Pressure Monitor

UA-767PC

Preliminary Remarks

The device conforms to the following requirements: European Directive 93/42 EEC for Medical Products Act; Medical Products Act; European Standards for Electrical Medical Equipment EN 60601-1 (General Safety Provisions), EN 60601-1-2 and EN 55011 (Electromagnetic Compatibility); European Standards pertaining to Non Invasive Blood Pressure Instruments EN 1060-1(General Requirements), EN 1060-3 (Supplementary Requirements for Electromechanical Blood Pressure Measuring Systems).

The above is evidenced by the CE mark of conformity accompanied by the reference number of a designated authority.


This device is designed for adults only.

Environment for use




The device is for use in patient's home.

Precaution
















1. Precision components were used in the construction of this device. Extremes in temperature, humidity, direct sunlight, shock or dust should be avoided.
2. Clean the device with a dry, soft cloth. Never use thinner, alcohol, benzene, or wet dusters.
3. Avoid tightly folding the cuff or storing the hose tightly twisted for long periods, as such treatment may shorten the life of the components.
4. The device and cuff are not water resistant. Prevent rain, sweat and water from wetting the device and cuff.
5. Measurements may be impaired if the device is used close to televisions, microwave ovens, cellular telephone, X-ray or other devices with strong electrical fields.



Parts Identification



Symbols

Symbols	Function / Meaning	Treatment
	Turns on or off the device	_____
	Set the clock / Does not store the current data	_____
	Direction guide to install battery	_____
	Direct current	_____
SN	Serial number	_____
1997 	Date of manufacture	_____
	Type BF: Device, cuff and tubing are designed to provide special protection against electric shocks.	_____
	Appears when measurement is in progress and flashes when the pulse is detected.	Measurement is in progress -- remain as still as possible.
	Previous measurement stored in MEMORY	_____
	Appears when the battery voltage is too low for the device to work properly.	Replace all batteries with new ones.
	Appears if the systolic and diastolic measurements are within 10 mmHg of each other.	Verify that the air hose is properly connected and that the exhaust rate is between 2 and 5 mmHg. Check for air leakage.
	Appears if the pressure value is unstable due to movement during the measurement.	Try the measurement again. Remain very still during the measurement.
	Appears if the pressure value did not increase during pump up.	Verify that the air hose is properly connected and that the exhaust rate is between 2 and 5 mmHg. Check for air leakage.
	System error	Reload batteries.
	Appears if pulses are not detected correctly.	Attach the cuff properly and try the measurement again.
	Attention symbol	See instruction for use


Setting the Clock

The UA-767PC blood pressure monitor is equipped with a date and time function. To set the clock on the monitor, follow these steps:

Press the clock button on the front of the monitor.

The blinking digits reflect the year. If they are not currently set on the correct year, press the START button to advance to the correct year. The clock is set to show years 1999 through 2098. If you pass the correct year, continue advancing the clock until you have run through the cycle of years and are back at the correct year.


Year



To set the month and day, press the clock button again.

The month numeral(s) will blink. Press the START button to advance the numbers to the correct month. Press the clock button again to advance to the day setting. When the numbers reflecting the day of the month blink, press the START button to advance the numbers to the correct day.

Month/Day



When you press the clock button again, the display will change and the time will be shown in the same format as a digital clock, with the hour numerals blinking. To set the correct hour, press the START button to advance the numbers to the correct hour. To set the minutes, press the clock button again. The minute setting will blink. Press the START button to advance the minutes until you reach the correct setting. Press the clock again. The hour and minute display appears. This is the standby mode.


Hour/Minute



Preparation for Measurements

1. Installing / Changing the Batteries

Remove the battery cover and insert new batteries into the battery compartment as shown, taking care to observe the proper direction for each battery. Use R6P, LR6 or similar type batteries.



If the symbol \otimes appears on the display panel, this means that the batteries are depleted. Always change all batteries at the same time. If you do not intend to use the device for longer periods, remove the batteries. Only use high quality, leakproof batteries with the specified ratings. Never mix new and old batteries or batteries from different manufactures.


Note: Used equipment, parts, batteries do not fall within the category of household waste, and must be disposed of according to the applicable regulations.

2. Connecting the Air Hose

Insert the air connector plug into the air socket firmly.

3. Attaching the Arm Cuff

Wrap the cuff around the upper arm about 2-3 cm (1 inch) above the elbow as shown. Place the cuff directly against the skin, as clothing may cause a faint heart beat, and result in an error. Constriction of the upper arm caused by rolling up a shirt sleeve may prevent accurate readings.



4. Posture While Measuring

Remain seated in a comfortable position, placing the center of the upper arm at the same height as your heart.

5. Measurement


During the measurement, the cuff will become very tight. Do not be alarmed as this is normal.

6. After the Measurement


After the measurement, remove the cuff and record your data.

Measurements

1. Place the cuff on the arm (preferably the left arm).
2. Set the PRESSURE switch to a value that will be 30 to 40 mmHg above the expected systolic value and press the START button.
 - When the START button is pressed, all of the display symbols will appear for about one second.
 - When the "0" is blinking, the device is calibrating its zero point. If there is still some air in the arm cuff, it will be exhausted at this time.




- The micro pump will pressurize the cuff.
Note: The cuff will become very tight. Do not be alarmed, as this is normal.
 - If the START button or clock button is pressed again during pressurization or during the measurement, the pump will stop and the air will be released from the cuff.
3. When pressurization has completed, the automatic exhaust mechanism will gradually reduce the pressure in the cuff and the Measurement in Progress symbol will appear along with the current pressure reading. This symbol will flash in unison with the pulse of the person being monitored.



Measurements


- Remain still while the device is in operation.
 - When a measurement is made with insufficient pressure, the device will pump again to a higher pressure.
4. When measurement has completed, the buzzer sounds and the air is automatically released from the cuff. The systolic pressure, diastolic pressure readings, and the pulse rate are displayed.



5. **Turning the Power OFF**
Press the START button to turn the power off. The device will be in standby mode.
6. **Subsequent Measurements**
If a subsequent measurement is required, turn off the power and turn it on again. When the "0" is displayed, the device is ready for measurement again.
Note: Please wait for about 10 minutes before repeating the measurement.
7. **Automatic Power OFF Function**
If this device is left on after a measurement, it will turn itself off after about 1 minute. It may be turned off at any time by pressing the START button.

Measurements

- Accurate measurement cannot be made if the exhaust velocity is outside the range of 2 to 5 mmHg. This can be determined by watching the display each time the "Measurement in Progress" symbol blinks. Note the reading as the "Measurement in Progress" symbol blinks.



Notes for Proper Measurements

1. Sit in a comfortable position where you can place the arm to be used for the measurement on a table or other support that will place the center of the upper arm at about the same height as your heart.
2. Relax for about 5 or 10 minutes before starting the measurement. If you are excited or depressed by emotional stress, the measurement will reflect this stress as a higher (or lower) than normal blood pressure reading, the pulse reading will usually be faster than normal.
3. A normal person's blood pressure varies constantly depending on what you are doing and what you have eaten. What you drink can have a very strong and rapid affect on your blood pressure.
4. As this device basses its measurements on the heart beat, if you have a very weak or irregular heart beat, the device may have difficulty determining your proper blood pressure.
5. Should the device detect a condition that is abnormal, it will stop the measurement and display an error. See the section on description of display symbols for details.
6. This blood pressure device is intended for use by adults only. Consult with your physician before using this device on a child. A child should not use this device unattended.

Data in Memory

Storing Data in Memory

This device stores up to 126 sets of measurement data in memory automatically. If an error message appears, the measurement data will not be stored.

Whether the data will be stored or not after the measurement can be selected as follows:

- When the START button is pressed or the automatic power OFF function works after a measurement, the measurement data will be stored in memory.
- When the clock button is pressed after a measurement, the measurement data will not be stored in memory.

Recalling Memory Data

To access measurements stored in memory, wait until the monitor is in standby mode. Then, press and hold the START button until the display shows the most recent measurement.

The display will show all of the stored measurements starting with the most recent and then will return to standby mode.

To interrupt memory display, press the START button to initiate inflation mode. To put the monitor in standby mode, press the START button again.

Erasing Data in Memory

To erase measurements stored in memory, wait until the monitor is in standby mode. Then, press and hold the clock button until the buzzer sounds and the memory access indicator blinks. All data in memory will be erased.

Transmitting Readings to a Computer

The connected computer is not allowed to be in the patient area.

To transmit the readings to a computer, connect the attached communication cable to the data port on the front of the monitor.


When the monitor receives the open command, it will be in the ready-to-send mode.

When the monitor receives the open command again within five seconds, “-” appears one after another in succession and permission will be given to perform various functions.

Measurement data of the UA-767PC can be monitored with “Monitor Pro” Data Analysis Software (UA-767P-13) and Printer terminal (TM-2480 series).

For details, refer to the instruction manual.

* UA-767P-13 and TM-2480 are sold separately.




How to Make Proper Measurements

How is a Proper Measurement Made?

Blood pressure varies depending upon the conditions prevailing at the time of the measurement, consider the following so that the most accurate measurements may be made.

1. In preparation for blood pressure measurement, the subject should urinate and should remain relatively still for 5 to 10 minutes before a measurement.
2. Shirts or other garments which fit tightly on the upper arm should be removed before attaching the arm cuff.
3. Exercise, eating, drinking, and smoking before a measurement can affect the results.
4. Blood pressure varies constantly throughout the day. Measurement should be made regularly at the same time each day.
5. Do not be too impressed by the results of one measurement. Keep a record of blood pressure measurements. Many readings tell a story.
6. When making repeated measurements, the arm becomes heavy with blood, resulting in wide variations in measurement values. For this reason, repeated measurements should be made after a rest period of approximately 5 to 10 minutes or after raising the arm until the engorged condition is relieved.
7. In general, the blood pressure is low in summer (when it is hot) and high in winter (when it is cold).
8. Emotional stresses may cause an increase in blood pressure.



About Blood Pressure

What is Blood Pressure?

The blood pressure is at its highest levels in the large arteries near the heart and drops off towards the peripheral areas of the circulatory system.

The blood pressure varies with the beating of the heart. When the heart contracts, forcing the blood out, the pressure inside the arteries is said to be systolic. When the heart expands, the pressure of the blood inside the arteries is said to be diastolic.

Hypertension

Hypertension, an abnormally high arterial blood pressure, which is most common among older adults, if left unattended, can cause many health problems including stroke and heart attack. It is wise to control the blood pressure to prevent it from becoming high, by reducing salt intake, and by controlling diet and activities. People who were born with high blood pressure can prevent the progress of heart disease by means of adequate control.

Why is it a Good Thing to Measure Blood Pressure at Home

Having your blood pressure measured in a hospital or a clinic may cause you to become nervous, thus artificially raising the blood pressure reading. Also blood pressure varies with a variety of conditions. So an accurate judgment of the heart condition can not be made with a single measurement.

Having the capability to measure blood pressure at home provides a good way to determine your fundamental blood pressure. This is defined as the first blood pressure reading upon waking, before taking any food or drink.


It is possible for you to carry out your own blood pressure control at home. Take blood pressure readings with your own blood pressure device on a regular basis. Record the measurements. If the accumulated results seem to be above normal, it is recommended that a doctor review the record.

About Blood Pressure

WHO Blood Pressure Classifications


Standards for assessment of high blood pressure, without regard to age, have been established by the World Health Organization (WHO), as shown in this chart.

Reference Material:
Journal of Hypertension
1999, Vol 17 No.2



Variations in Blood Pressure

Individual blood pressures vary greatly both on a daily and a seasonal basis. These variations are even more pronounced in the hypertensive patient. Normally the blood pressure rises while at work and is at its lowest during the sleeping period.



Shown is data for measurements taken every 5 minutes. The thick line represents sleep. The rises in blood pressure at 4PM (A in the graph) and 12PM (B in the graph) correspond to an attack of pain and sexual intercourse. (Beven, Honour & Scott: Clin. Sci. 36:329, 1969)

Maintenance

Do not open the case of the device because it uses delicate electrical components and an intricate air unit that could be damaged. If you can not locate and fix the problem, request service from your supplier, or from the A&D service group. The A&D service group will support authorized suppliers about technical information, spare parts and units.

The technical testing procedures – to be carried out at least every 3 years – can be performed either by the manufacturer or by an authorized repair service in accordance with the regulations governing manufactures of medical products.

Specifications

Performance Specifications

Measurement method Oscillometric measurement

Measurement range Pressure :20 - 280 mmHg
Pulse : 40 - 200 beats / minute

Accuracy Pressure : ± 3 mmHg or 2%, whichever is greater
Pulse : $\pm 5\%$

Battery 4 x 1.5V (R6P, LR6, or AA)

Cuff Arm Circumference

Classification Type BF   0366

Device, cuff and tubing are designed to provide special protection against electric shocks.

Clinical test According to ANSI / AAMI SP-10 1987

	Standard deviation	Mean error
Systolic blood pressure	5.43 mmHg	2.24 mmHg
Diastolic blood pressure	5.18 mmHg	1.54 mmHg

Environment Specifications

Operating environment +10 °C - +40 °C, less than 85 %RH

Storage environment -10 °C - +60 °C, less than 85 %RH

Physical Specifications

Dimensions 163.7 [W] x 111 [D] x 66.7 [H] mm

Weight Approx. 320g without batteries



A&D Co., LTD.

1-243 Asahi, Kitomoto-shi, Saitama 364-8585 JAPAN



[81] (48) 593-1111



[FAX] [81] (48) 593-1119

A&D INSTRUMENTS LTD.

<Authorized Representative Established in the European Community>

Abingdon Science Park, Abingdon, Oxford OX14 3YS ENGLAND



[44] (1235) 550420



[FAX] [44] (1235) 550485

A&D MERCURY PTY. LTD.

32 Dew Street, Thebarton, South Australia 5031 AUSTRALIA



[61] (8) 352-3033



[FAX] [61] (8) 352-7409