

TM-2430

**RECORDER FOR AMBULATORY
BLOOD PRESSURE MONITOR**

INSTRUCTION MANUAL

Ambulatory Blood Pressure Monitor



This mark informs you about the operation of the product.

Note This manual and or the TM-2430 may be changed at any time to improve the product without notice.

Windows is the registered trademark of the Microsoft Corporation.

Macintosh is the registered trademark of Apple Computer, Inc.

EXCEL is the registered trademark of the Microsoft Corporation.



Contents

Before use

| | |
|--|---|
| Compliance | 1 |
| Definitions | 1 |
| Precautions for Use | 2 |
| Notes on the Blood Pressure Recorder | 4 |

Welcome

| | |
|-----------------------------|---|
| Welcome and Intention | 4 |
|-----------------------------|---|

Product overview

| | |
|--|---|
| Packing List and Component Names | 5 |
| Display | 7 |
| Symbols | 7 |

Specifications

| | |
|------------------------------------|---|
| Features | 8 |
| Functions and Specifications | 8 |

The complete procedure for use

| | |
|------------------------------|----|
| Step by Step Procedure | 12 |
|------------------------------|----|

Initializing the recorder

| | |
|---|----|
| Replacing Batteries | 14 |
| The State of Turning on the Recorder | 15 |
| Parameters for the Display and Clock | 15 |
| Selection for the Automatic Measurement | 17 |
| Deleting Old Data | 21 |
| Resetting the Recorder | 21 |

Preparing the patient

| | |
|---------------------------------------|----|
| Patient Instructions | 22 |
| Use of the Cuff Cover | 22 |
| Attaching the Cuff and Recorder | 23 |

Operation

| | |
|---|----|
| Automatic Measurement (by Programed Time Intervals) | 25 |
| Manual Measurement | 25 |
| To Stop a Current Measurement | 25 |

Data transfer

| | |
|---|----|
| Data Transmission to a Printer | 26 |
| Data Transmission to a Computer Using Analysis Software | 28 |

Options and accessories

| | |
|--|----|
| Analysis Software and Communication Cables | 29 |
| Cuffs and Other Accessories | 30 |

Maintenance

| | |
|--------------------------------------|----|
| Checking Accuracy | 31 |
| Cleaning the Cuff and Recorder | 32 |
| Periodical inspection | 32 |
| Problem Solving | 32 |
| Error Codes | 33 |
| Index | 35 |



Before use



Compliance

Compliance with FCC Rules

Please note that this equipment generates, uses and can radiate radio frequency energy. This equipment has been tested and has been found to comply with the limits of a Class A computing device pursuant to Subpart J of Part 15 of FCC rules. These rules are designed to provide reasonable protection against interference when this equipment is operated in a commercial environment. If this unit is operated in a residential area it might cause some interference and under these circumstances the user would be required to take, at his own expense, whatever measures are necessary to eliminate the interference. (FCC = Federal Communications Commission in the U.S.A.)

Compliance with European Directive 93/42 EEC for Medical Products

The device conforms to the following requirements: European Directive 93/42 EEC for Medical Products; Medical Products Act; European Standards for Electrical Medical Equipment EN 60601-1 (General Safety Provisions), EN 60601-2-30 (Particular Requirements for the Safety of Automatic Cycling Indirect Blood Pressure Monitoring Equipment), EN 60601-1-2 and EN 55011 (Electromagnetic Compatibility); European Standards pertaining to Non Invasive Blood Pressure Instruments EN 1060-1 (General Requirements), prEN 1060-3 (Supplementary Requirements for Electromechanical Blood Pressure Measuring Systems). The above is evidenced by the CE mark of conformity accompanied by the reference number of a designated authority. This device is designed for adults only.



Definitions

| | |
|----------------------|--|
| SYS | Systolic Blood Pressure |
| DIA | Diastolic Blood Pressure |
| DSD | The Difference between Systolic Blood Pressure and Diastolic Blood Pressure. |
| Exhaust | This means "releasing the cuff air as soon as possible". |
| Constant exhaust | This means "releasing the cuff air in a constant depressurization rate". |
| Exhaust velocity | This means the rate of depressurizing the cuff air. |
| During a measurement | This means "a period between the start of cuff inflation and the end of exhausting the air". |
| Interval | This is called "block". A block consists of a start time and frequency. |
| bpm | beats per minute. |



Precautions for Use

Precautions

Batteries

- ❑ Use alkaline batteries (LR6 type, AA type, Mignon) or suitable Ni-Cd batteries.
- ❑ Do not mix new and used batteries in the recorder.
- ❑ Remove the batteries from the recorder, if it will not be used for a long period of time and unless there is no risk of a SAFETY HAZARD arising.

A malfunctioning recorder

- ❑ If the recorder malfunctions, stop the operation, attach a note "Do not use this recorder" and store in a safe place to avoid mis-use.

Training

- ❑ Instruct the patient on how to stop the operation if there is an abnormal measurement, and how to remove the cuff if there is excessive arm pain.
- ❑ Advise the patient on how to cope with misoperation and contingencies

Repair

- ❑ Do not open the recorder case. Contact the nearest A&D office if you have questions.

Blood pressure measurement

- ❑ Use the recorder for adult.
- ❑ The recorder may not make a measurement when a patient has continuous arrhythmia or the recorder senses noise due to the patients movement.
- ❑ Please check measurement values by other methods, if you suspect an value.
- ❑ Do not use this recorder on a person who is using a heartlung machine and a defibrillator.
- ❑ Do not use this recorder on a person who is critical or is in an intensive care unit.

Saving power

- ❑ Turn off the power switch when not in use.
- ❑ Please transfer the data as soon as possible. All measurement data, clock parameters, measurement parameters and internal system parameters are preserved by backup battery when turning the power switch off. The backup battery life is few days.

Cuff

- ❑ Close the cuff fastener properly when attaching the cuff to a patient or replacing the cuff cloth. If the fastener is closed incorrectly, inflating cuff may damage this cuff.



Notes on the Blood Pressure Recorder

Storage

- ☐ Do not store the recorder in the following places.
 - Where the recorder might be splashed with water or other liquids. If the recorder soaked, it needs a repair. (Do not use the recorder.)
 - Where the temperature and humidity are high, or in direct sunlight.
 - Where the recorder may be influenced by vibration or shock.
 - Where there is dust, salt or sulfur.
 - Where medicines are stored, or medicines are evaporating.

Before use

- ☐ Cover the RS-232C terminal using the rubber cap, to avoid dust.
- ☐ Confirm that the recorder works correctly and measurement values are proper.
- ☐ Confirm that the cuff and air hose are connected properly.
- ☐ Confirm parts of contact with patient directly. Keep clean for these parts.
- ☐ Use clean cuff cover because of contact with patient directly.
- ☐ Clear the old data before starting a new measurement.
- ☐ Avoid strong magnetic field and static electricity.
- ☐ Do not use this recorder with a high frequency surgical equipment.

During use

- ☐ The recorder should be operated by a physician who knows it well.
- ☐ Use the recorder only during the time of diagnosis or medical treatment.
- ☐ Stop the use, if the patient feels pain in his arm or the recorder does not measure properly.
- ☐ It may reduced the measurement cycle due to environment.
- ☐ When an inside part of the recorder has been soaked (wetted), turn off the power switch and request service to your supplier or the A&D service group.

After use

- ☐ Clean the recorder, cuff and accessories for the next use. Do not pull or kink hoses. Do not use organic solvent, antiseptic solution, etc.
- ☐ Turn off the power switch after measurement.
- ☐ Please use the original box for transportation.

Periodic maintenance

- ☐ The recorder is a precision instrument. Please check all functions (every year) periodically. Contact your nearest A&D office for this inspection.

Environmental protection

- ☐ If you disuse the recorder, remove Ni-Cd battery and built-in Li battery from this recorder.
- ☐ Disuse Ni-Cd battery to its exclusive trash can because of recycling it.
- ☐ Dispose of Li coin battery in the recorder as dangerous object properly.





Welcome



Welcome and Intention

Thank you for your Purchase!

The A&D TM-2430 ambulatory blood pressure recorder enables you to accurately take a patient's blood pressure, automatically, at different preset times throughout a 24-hours period.

Recently, in the treatment of patients with hypertension, there has been an increasing need to prescribe medication according to the particular blood pressure fluctuation pattern of the patient. These patterns can be made more evident by using the TM-2430 recorder, and analysis by a physician.

This manual will explain in simple language how this recorder works.

Patient

This blood pressure recorder is designed for an adult patient.

Environment for use

This blood pressure recorder is used in a hospital and / or patient's home.




Product overview




Packing List and Component Names

When you open this box, make sure you have everything as shown here :




| Name | Functions |
|------------------|---|
| Power switch | This is the main power switch. At the OFF state, all data and parameters are preserved by an backup battery. This backup battery life is approximately 10 days with the power off. |
| AUTO ON/OFF key | <ul style="list-style-type: none"> • When you press and hold the AUTO ON/OFF key, the automatic measurement is started or stopped alternately. • When you press the AUTO ON/OFF key at mode II of the automatic measurement, "S" is displayed or turned off alternately. This sign changes the interval for sleep. |
| START STOP key | <ul style="list-style-type: none"> • When you press the START STOP key, a blood pressure measurement is started at once. • When you press and hold the START STOP key for approx. 3 seconds, the recorder proceeds to "Selection for the automatic measurement". • When you press and hold the START STOP key for approx. 6 seconds, the recorder proceeds to "Parameters for the display and clock". • When you press and hold the START STOP key for approx. 9 seconds, the recorder proceeds to "Deleting old data". |
| RS-232C terminal | This terminal is used for data output to a printer or computer. The optional RS-232C cable is necessary to output the data. |
| Reset key | All data and parameters are deleted. |





Display



| Sign | Name | Functions |
|------|-----------------------|---|
| ▶ | Arrow | The arrow points to the kind of current display in the measurement result and function mode. |
| A | Automatic measurement | "A" is displayed when the automatic measurement is selected. When you press and hold the AUTO ON/OFF key, this sign is turned on or off alternately. |
| S | Sleep | When you press the AUTO ON/OFF key while in mode II of the automatic measurement, "S" is displayed or turned off alternately. With the "S" turned off, the time interval is 15 minutes. With the "S" turned on, the time interval is 30 minutes. |
| B | Low battery | When the recorder can not operate all functions due to low battery, this sign is displayed. The clock is still displayed. Please replace the batteries at once. |
| M | Full memory | When data memory is at full capacity, this sign is displayed. In this case, you can not perform another measurement. Transfer the data save in other media and delete. Then the "M" turns off. |



Symbols

- Turning on the recorder.
- Turning off the recorder.
- Direction guide to install batteries.
- Direct current.
- SN Serial number.
- 1997 Date of manufacture.
- Attention symbol. "See instruction for use."
- Recorder, Cuffs and tubings are designed to have special protection against electric shocks.



Specifications



Features

Portability

- ❑ The recorder weighs approx. 215g (including batteries) and is palm top size, because a micro-pump is used.
- ❑ The recorder is powered by LR6 type (Mignon) alkaline batteries. It is possible to replace the LR6 type batteries with Ni-Cd rechargeable batteries.

Operation & management

- ❑ Clock and automatic measurement parameters may be set as needed.
- ❑ If you connect to a computer and use the optional software, clock and automatic measurement parameters can be set easily.
- ❑ There are three modes for automatic measurement.
- ❑ The recorder can transmit data to a printer directly. (An adaptable printer is necessary to print the data. Refer to Section "Data Transmission to a Printer" for specifications of the printer.)
- ❑ The recorder has the built-in chargeable coin Li battery to keep the clock and automatic measurement parameters.

Analysis

- ❑ The time interval may be changed as needed.
- ❑ The patient's blood pressure can be measured immediately at any time.
- ❑ If you use the optional software, you can analyze the data widely.

Smart measurement

- ❑ The measurement time is shortened by proper exhaust velocity control.
- ❑ The exhaust velocity adjustment is unnecessary, because the constant exhaust is properly controlled.
- ❑ In the automatic measurement, these inflation value and stop value at exhaust is managed to reduce the measurement time.



Functions and Specifications

Blood pressure measurement

- ❑ There are two ways of blood pressure measurement.

Automatic measurement --- This automatic measurement works in accordance with internal clock, preset time intervals and preset mode. The measurement data is saved in memory.

Manual measurement ----- Any time you press the **START STOP** key, a blood pressure measurement is performed immediately. The measurement data is saved in the memory.

Automatic measurement

- ❑ This measurement starts or stops using the **AUTO ON/OFF** key. When this measurement is started, the recorder begins to work in accordance with preset time intervals from the preset time of the internal clock. Refer to "Selection for the automatic measurement".
- ❑ In the automatic measurement, an "A" appears in the upper left of the display.
- ❑ The recorder automatically measures the patient's blood pressure at the time that is pointed out by "the frequency" and "the start time" (by the programed time intervals).
- ❑ When a measurement error occurs and there is 8 minutes until the next measurement, the measurement is retried after approx. 30 seconds.
- ❑ If a measurement is retried, only the data from the retry is saved.
- ❑ The recorder automatically adjusts the proper pressure, exhaust velocity and end of measurement.
- ❑ Refer to "Selection for the automatic measurement" and "Automatic measurement by programed time intervals" about operation and entering parameters.

Stopping a measurement

- ❑ If you press the **START STOP** key during a measurement, the recorder exhausts the air and stops the measurement.

Concealing the measurement value

- ❑ This function works only while using automatic measurement.
- ❑ This function does not display the SYS, DIA or pulse rate for the automatic measurement, but the data is saved in memory.
- ❑ This function can select "reveal" or "conceal" at "Parameters for the display and clock". Refer to this section.
- ❑ If you select "conceal", the recorder displays the clock during a measurement.
- ❑ If you reset the recorder, this parameter is set to "reveal".

Pressurization

- ❑ The pressure is automatically selected by the recorder while in the automatic measurement mode.
- ❑ The first pressure is set to approx. 185 mmHg. This value automatically varies to the proper value after first measurement. If the first inflation is not successful, the recorder retries twice.
- ❑ If you reset the recorder, the first inflation value is set to 185 mmHg.

Memory

- ❑ The recorder can store 300 sets of data (the memory capacity is 300).
A data set consists of a SYS, a DIA and a pulse rate.
- ❑ When memory becomes full, the recorder displays an "M". Until you clear the data, you can not measure blood pressure.



When the recorder saves data for more than one patient, data management becomes complication. We encourage that a patient's data is recorded, trans-

ferred and is cleared from memory.



When a "B" is displayed, the backup batteries that is preserving a patient's data are weak. Please transfer the data and save it to other media as soon as possible.

ID number


- ☐ If you reset the recorder, the ID number is set to "1".
- ☐ The ID number can be set using the optional software.

Performance specifications

| | |
|---|---|
| Measurement method | Oscillometric |
| Pressurization | micro-pump Display range 0 ~ 320 mmHg Automatic measurement 85 ~ 300 mmHg(Fitted) Manual measurement 185 mmHg (Fixed) |
| Measurement range | Systolic Blood Pressure 60 ~ 280 mmHg Diastolic Blood Pressure 40 ~ 160 mmHg Pulse rate 30 ~ 200 bpm |
| Accuracy | Pressure ± 3 mmHg Blood pressure Conforming to 1992 AAMI standard Pulse rate ± 5 % |
| Minimum display division | Pressure 1 mmHg Pulse rate 1 bpm |
| Depressurization | Constant exhaust Controlled ceramic valve Exhaust Ceramic valve |
| Measurement | Automatic measurement Manual measurement |
| Number of measurement | Apporx. 200 times (It may reduce this number due to environment and capacity of Ni-Cd batteries) |
| Memory | Up to 300 sets of data |
| Display | Normal Clock During a measurement Pressure value After a measurement SYS, DIA and pulse rate Error code, function of concealing the measurement data |
| Clock | 24-hours (1997~2096year, automatic leap year setting) |
| Batteries | 3 x Alkaline battery (type LR6, type AA , Mignon) or 3 x Ni-Cd battery (type AA , Mignon) |
| Type of protection against electric shock | Internally powered equipment type BF |
| CE Marking | The label of the medical device by the EC directive. |

AAMI : Association for the Advancement of Medical Instrumentation

| | | |
|------------------|--|---|
| Interface | Connected to a computer, you can output the data and enter parameters. Connected to a printer, you can print the data. | |
| | EIA RS-232C, Asynchronous, bi-directional, half duplex | |
| | Baud rate | 9600 bps |
| | Data bits | 8 bits |
| | Stop bits | 2 bits |
| | X parameter | Used (for computer) Not used (for printer) |
| | Parity | None |
| | Code | ASCII |
| | | RS-232C |
| | | 1 -5V ~ -15V |
| | | 0 +5V ~ +15V |




Environment specifications

| | |
|-----------------------|---|
| Operating environment | +10°C ~ +40°C (+50°F ~ +104°F), Less than 85%RH * |
| Transport and Storage | -20°C ~ +55°C (-4°F ~ +131°F), Less than 95%RH * |

* Non Condensing

Physical specifications

| | |
|------------|---|
| Dimensions | 72(W) x 100(D) x 27(H) mm 2.8(W) x 3.9(D) x 1.0(H) in. |
| Weight | Approx. 215 g (0.47lb) excluding cuff |





The complete procedure for use



A Step by Step Procedure

Step 1

Battery replacement.

Replace with new alkaline batteries (note direction).

Refer to page 14, "Replacing Batteries".

Step 2

Turn on the recorder using the power switch.

Step 3

With the recorder turned on, select any operation to set the state.

Case 1 Normal state

The buzzer is sounded once and the clock is displayed.

You can use this recorder at once. The recorder stores parameters for "display and clock" and "automatic measurement". Please proceed to [Step 5](#).

Case 2 Error state

When the recorder displays the $\text{E} \square \square$ error code, it is necessary to set up parameters for "display and clock" and "automatic measurement". Please proceed to the next step.

Step 4

Set up parameters for "display and clock". When you use automatic measurement, set up parameters for automatic measurement. Refer to "Parameters for the Display and Clock" and "Selection for the Automatic Measurement". Please proceed to [Step 6](#).

Step 5


Set up new parameters for "display and clock" and "automatic measurement", if necessary.

Step 6


Delete the old data stored in the recorder. Refer to the section "Deleting Old Data".

Step 7

Explain the "Patient Instructions" and "Notes on the Blood Pressure Recorder" to the patient.



- Step 8** Attach the cuff to the patient. Refer to "Attaching the Cuff and Recorder".
- Step 9** Attach the holder to patient and insert the recorder. Refer to "Attaching the Cuff and Recorder".
- Step 10** Start the automatic measurement. The recorder displays an "A" . Refer to "Automatic Measurement".
- Step 11** The recorder starts the automatic measurement sequence. Check the recorder using the manual measurement with relaxed, but correct posture. Refer to "Manual Blood Pressure Measurement".
- Step 12** When using automatic measurement, consider the following.
- ❑ Consider the sections "Patient instructions" and "Notes on the Blood Pressure Recorder".
 - ❑ When the patient uses mode II of automatic measurement, press the **AUTO ON/OFF** key with each rising and going to bed.
- Step 13** When the automatic measurement sequence is finished, turned off "A" sign . Refer to "Automatic Measurement".
- Step 14** Detach the cuff and recorder from the patient.
- Step 15** Transfer the patient's data and save it to other media. Refer to "Data transfer".
- Step 16** Clean the cuff and recorder and place in storage. Refer to the sections "Before use" and "Maintenance".





Initializing the recorder




Replacing Batteries

Caution

- ❑ When "B" is displayed before a measurement, the recorder can not make a measurement. Please replace the new batteries before using.
- ❑ If "B" is displayed during the measurement, stop the measurement and replace the new batteries at once.
- ❑ Use alkaline batteries or the specified rechargeable batteries for the recorder.
- ❑ Do not use new and used batteries at the same time.

Steps for replacing the batteries

- Step 1** Open the battery cover.
- Step 2** Turn off the power switch.
- Step 3** Replace with new batteries. (note the direction, "+" and "-".)
- Step 4** Turn on the power switch.
- Step 5** Close the battery cover.





The State of Turning on the Recorder

There are three types of state when the recorder is turned on. Select a operation. Refer to the section "The complete procedure for use" about use.


| Action when the re-corder is turned on. | State of recorder | Treatment (Operation) |
|---|---|---|
| The buzzer sounds once and the clock is displayed. (Normal state) | The recorder storage parameters of "display and clock" and "automatic measurement". | You can use the re-corder at once. |
| The buzzer is sounded once and $E \square \square$ is displayed blinking. | All parameters are lost. | Set up parameters of "display and clock" and "Automatic measurement". |
| The buzzer sounds four times and $E \square \square$ is displayed blinking. | The state after reset. All parameters are lost. | |



Parameters for the Display and Clock

This setting selects the display during automatic measurement sequence and adjusts the clock parameters. The sequence number tells you which parameter you are adjusting.

Display & key



Items

| Sequence number | Value & range | Meaning of parameters | |
|-----------------|---------------|---|--|
| 1 | 0 | Displaying clock only in automatic measurement | |
| | 1 | Displaying pressure and result in automatic measurement | |
| 5 | 00 ~ 99 | Years (1997 ~ 2096) | |
| 6 | 1 ~ 12 | Month | |
| 7 | 1 ~ 31 | Day | |
| 8 | 0 ~ 23 | Hour | |
| 9 | 0 ~ 59 | Minute | |

Steps for setting the display and clock

This explanation uses the following examples.

ex. After reset, The measurement value is not displayed.

The clock is adjusted to 1997/ 05/ 27 14:28.

Step 1 Press and hold the **START STOP** key for approx. 6seconds.
The recorder displays **1 1** for adjusting the display and clock.

Step 2 Press the **AUTO ON/OFF** key so as to display **1 0**.
(A selection where a clock is displayed only in automatic measurement)

Step 3 Press the **START STOP** key. The current year is displayed .

Step 4 Press the **START STOP** key. The current month is displayed.

Step 5 Press the **AUTO ON/OFF** key to display 5 (for May).

Step 6 Press the **START STOP** key. The current day is displayed.

Step 7 Press the **AUTO ON/OFF** key to display 27 (27th day).


Step 8 Press the **START STOP** key. The current hour is displayed.

Step 9 Press the **AUTO ON/OFF** key to display 14 (14th hour).

Step 10 Press the **START STOP** key. The current minute is displayed.

Step 11 Press the **AUTO ON/OFF** key to display 28 (28th minute).

Step 12 Press the **START STOP** key to save these parameters. Then the recorder displays the clock.





Selection for the Automatic Measurement

This setting initializes measurement intervals that is based on the internal 24-hour clock.

Mode

mode I 07:00 ~ 21:59 The measurement is performed every quarter hour.
22:00 ~ 06:59 The measurement is performed every half hour.


mode II The **AUTO ON/OFF** key is pressed at rising and going to bed so that the measurement intervals are changed and the time during sleep can be distinguished on the data.

When the "S" is off, the measurement is performed every quarter hour.

When the "S" is displayed, the measurement is performed every half hour.

mode III The measurement interval can change six times within a maximum of 24-hours. (The recorder can store six measurement intervals (blocks) in 24-hours. A block consists of a start time and frequency.)


Display & key



Steps for selecting a mode

ex. Mode II is selected.

- Step 1** Press and hold the **START STOP** key for about 3 seconds. The current mode is displayed.
- Step 2** Press the **AUTO ON/OFF** key so as to display **2** of mode II.
- Step 3** Press the **START STOP** key. The recorder stores the mode and displays the clock.



Mode III Settings


Set up procedure

Before you enter into mode III, read the procedure below. Also, refer to the example on the next page for the setting procedure.

- ❑ Each blocks starting time must match the previous blocks finish time.
- ❑ The end of block 6 automatically equalizes to start time of the block 1.
- ❑ If you enter the block 1 start time in any other block, these parameters are saved and this sequence is finished.
- ❑ When selecting 120 minutes for the current frequency, you must adjust the start time of the next block so that the current block fits a multiple of 120 minutes. If you do not fit to the next start time, an error code is displayed.
- ❑ The recorder displays t_h as 60 minutes and t_h^2 as 120 minutes.
- ❑ When you enter the sequence of mode III settings, the recorder initializes each start time to the start time of block 1 and each frequency to "-" (not used).

To read the current settings, press the **START STOP** key in this sequence.

Display & key



Items


| Sequence number | Parameters (monitor) | Meaning | Initial value |
|-----------------|---------------------------------------|----------------------------|---------------|
| 01 | 0 ~ 23 o'clock | Start time of first block | 01 7 |
| 02 | -, 5, 10, 15, 20, 30, 60, 120 minutes | Frequency of first block | 02 15 |
| 03 | 0 ~ 23 o'clock | Start time of second block | 03 22 |
| 04 | -, 5, 10, 15, 20, 30, 60, 120 minutes | Frequency of second block | 04 30 |
| 05 | 0 ~ 23 o'clock | Start time of third block | 05 7 |
| 06 | -, 5, 10, 15, 20, 30, 60, 120 minutes | Frequency of third block | 06 - |
| 07 | 0 ~ 23 o'clock | Start time of fourth block | 07 - |
| 08 | -, 5, 10, 15, 20, 30, 60, 120 minutes | Frequency of fourth block | 08 - |
| 09 | 0 ~ 23 o'clock | Start time of fifth block | 09 - |
| 10 | -, 5, 10, 15, 20, 30, 60, 120 minutes | Frequency of fifth block | 10 - |
| 11 | 0 ~ 23 o'clock | Start time of sixth block | 11 - |
| 12 | -, 5, 10, 15, 20, 30, 60, 120 minutes | Frequency of sixth block | 12 - |
| 13 | 0 ~ 23 o'clock | End of sixth block | 13 - |

The "-" means "not used".

Steps for automatic measurement

ex. First block 8:00 ~ 21:59 frequency is 30 minutes
 Second block 22:00 ~ 5:59 frequency is 60 minutes
 Third block 6:00 ~ 7:59 frequency is 10 minutes

- Step 1** Press and hold the **START STOP** key for approx. 3 seconds. The current mode is displayed.
- Step 2** Press the **AUTO ON/OFF** key so as to display "3" for mode III.
- Step 3** Press the **START STOP** key. The mode is stored and the current start time of the first block is displayed.
- Step 4** Press the **AUTO ON/OFF** key so as to display "8" for 8:00 hours as the start time of the first block.
- Step 5** Press the **START STOP** key. The current frequency for the first block is displayed.
- Step 6** Press the **AUTO ON/OFF** key so as to display "30" for 30 minutes as the frequency for the first block.
- Step 7** Press the **START STOP** key. The current start time of the second block is displayed.
- Step 8** Press the **AUTO ON/OFF** key so as to display "22" for 22:00 hours as the start time for the second block.
- Step 9** Press the **START STOP** key. The current frequency for the second block is displayed.
- Step 10** Press the **AUTO ON/OFF** key so as to display "1h" for 60 minutes as the frequency of the second block.
- Step 11** Press the **START STOP** key. The current start time of the third block is displayed.
- Step 12** Press the **AUTO ON/OFF** key so as to display "6" for 6:00 hours as the start time of the third block.



- Step 13** Press the **START STOP** key. The current frequency of the third block is displayed.
- Step 14** Press the **AUTO ON/OFF** key so as to display "10" for 10 minutes as the frequency of the third block.
- Step 15** Press the **START STOP** key. The current start time of the fourth block is displayed.
- Step 16** Press the **AUTO ON/OFF** key. The recorder stores these parameters and displays the clock, (because the current start time of the fourth block is the same start time of the first block).

From previous page

START STOP
06 -

AUTO ON/OFF
06 10

START STOP
07 8

START STOP
14:22



Deleting Old Data

- Caution**
- Confirm that the data has already been transferred and saved, when the data is to be deleted. It is not possible to recover data that was deleted.
 - It is not possible to delete data completely, if the **START STOP** key is released while the buzzer sounds at Step 2.

Steps for deleting old data

- Step 1** Press and hold the **START STOP** key for approx. 9 seconds. **LL** is displayed. If you want to cancel this process, press the **AUTO ON/OFF** key.
- Step 2** Press and hold the **START STOP** key once more until the buzzer becomes silent.



Resetting the Recorder


If the recorder does not work correctly, press the reset key. The recorder deletes all data and parameters. The internal system is initialized.

- Caution**
- All data and parameters are deleted and preset initial parameters by the reset.
 - Do not press the reset switch to intensely. Press this key gently so as not to damage the components inside.
 - Keep foreign matter away from the reset switch hole.

Steps for reset

- Step 1** Open the battery cover.
- Step 2** Turn off the power switch.
- Step 3** Remove the batteries from the recorder.
- Step 4** Press the reset key gently.
- Step 5** Place new batteries in the recorder.
- Step 6** Turn on the power switch. The recorder sounds the buzzer four times and **LL** is displayed blinking.
- Step 7** Set the parameters for the display and clock. Also, adjust the parameters for automatic measurement.

Reset key





Preparing the patient



Patient instructions

Advise the patient on how to cope with misoperation and contingencies.

Cautions during automatic measurement

- ☐ Relax and be quiet, when the recorder starts inflating the cuff.
- ☐ Check the patient's blood pressure in the same posture as the measurement.
- ☐ Minimize noise and movement for the measurements.
- ☐ The recorder measures the patient's blood pressure within one minute after the measurement. Also, relax and be quiet during measurement. The maximum measurement time is 90 seconds.
- ☐ There is the case of re-measuring the blood pressure during the last measurement. This is in case the recorder did not acquire usable data and the frequency of the interval is above 8 minutes. The patient should relax and not move during the measurement.
- ☐ Stop the use, if the patient feels pain in his arm.

Stopping or canceling an automatic measurement

- ☐ When the patient is to stop a measurement, press the **START STOP** key. The recorder beeps, releases the cuff air and an error is displayed. The recorder will inflate the cuff for the next time period automatically.
- ☐ When the **AUTO ON/OFF** key is pressed and held for approx. 3 seconds, the recorder stops (re-starts) automatic measurement and the "A" disappears (appears). This key operation can select alternately.

Manual measurement

- ☐ For the patient to start a measurement at once, press the **START STOP** key.
- ☐ For the patient to stop this measurement, press the **START STOP** key.

Attention while attaching the cuff and recorder

- ☐ Do not drop or shock the recorder.
- ☐ The recorder and cuff are not water resistant. Prevent rain, sweat and water from wetting the recorder and cuff.
- ☐ Do not place anything on the recorder.
- ☐ Re-attach the cuff and recorder, when patient moves and they will not slip off.
- ☐ Prevent the air hose from breaking during sleep. Affix the air hose to the patient's body only as shown on page 24.


Replacing the batteries

- ☐ Replace with new batteries quickly, when "B" is displayed.



Use of the cuff cover

- Step 1** Pass the air hose through the slit.
- Step 2** Place the cover on the cuff as shown. Link them using the three velcro strips.



Attaching the cuff and recorder

Caution

- ☐ If the cuff is not attached at the right position, the recorder may not measure the blood pressure correctly and an error may occur.
- ☐ The cuff accessory is for use on the left arm, of about 20cm ~ 31cm. If you need a different cuff, purchase a cuff of the proper size and arm position. Refer to "Option and Accessories".
- ☐ Do not use, if the patient has dermatitis, etc.



Keep the cuff clean. Exchange the cuff cover for each patient. The cuff cover may be used for both right or left.


Steps for attaching the cuff and recorder

- Step 1** Make a circle where the end of the cuff is passed through the ring.
- Step 2** Search for the brachialis artery using palpation.

- Step 3** Attach the cuff directly against the skin so that the yellow mark is directly over the brachialis artery and space it one inch above the inside of the elbow and the lower edge of the cuff .
- Step 4** Wrap the cuff so that the ring is within the slide range, it is flat and does not slip down, but has room to insert two fingers. (If the ring is not within the slide range, you need a proper cuff.)
- Step 5** Position the air hose over the shoulder and affix it on the patient using adhesive tape.
- Step 6** Assemble the belt and holder.
- Step 7** Position the belt so that the holder is on the right (left) side of the patient, when a patient attaches left (right) arm cuff.
- Step 8** Connect the air hose plug to the air socket.
- Step 9** Place the recorder into the holder.

Caution

- Do not disturb the cuff or air hose during the measurement because the recorder measures small pressure variations.





Operation



Automatic Measurement

Caution

- ❑ Automatic measurement uses the internal clock and parameters of automatic measurement. Refer to section "Parameters for the Display and Clock" and "Selection for the automatic measurement" for setting these parameters.
- ❑ Press and hold the **AUTO ON/OFF** key for approx. 3 seconds so as to turn off "A" sign, when the patient stops the automatic measurement or detaches the cuff. If "A" sign is leaved, when the next automatic measurement starts, it damages the cuff.

Starting or re-starting automatic measurement

- Step 1** Confirm the parameters for automatic measurement. Refer to "Selection for the automatic measurement".
- Step 2** Press and hold the **AUTO ON/OFF** key for about 3 seconds. Then the "A" is displayed and the recorder starts an automatic measurement based on the internal clock and the parameters for automatic measurement.

Operation using mode II

- Step 1** Press the **AUTO ON/OFF** key turning off the "S" when the patient wakes up.
- Step 2** Press the **AUTO ON/OFF** key turning on the "S" when the patient goes to bed.

Stopping or canceling automatic measurement

- Step 1** Press and hold the **AUTO ON/OFF** key for about 3 seconds. "A" is not displayed and the recorder stops automatic measurement.



Manual Measurement

- Step 1** Press the **START STOP** key. The recorder starts a measurement at once. The results are displayed and stored in memory.



To stop a Current Measurement

- Step 1** Press the **START STOP** key during measurement. The recorder will stop the measurement at once and releases the air from the cuff.



Data Transfer

- ❑ The recorder transfers data to a printer or computer using the RS-232C terminal.
- ❑ We recommend analysis of the data using the optional analysis software.

Caution

- ❑ Cap the RS-232C terminal to prevent dust and foreign matter from entering when this terminal is not in use.
- ❑ Remove the recorder and cuff from the patient, when the recorder is connected to a printer or computer.



Data Transmission to a Printer

Caution


- ❑ The recorder intensely consumes the battery power while connected to the RS-232C cable. Disconnect the cable when not actually transferring data.
- ❑ Maintain the power-on state while transmitting the data so that the data is not damaged.
- ❑ The optional RS-232C cable is required when connecting to a printer.
- ❑ The printer (to print the data) must have a serial interface and adapt to the RS-232C protocol of the recorder.

Specifications for an adaptable printer

| | | |
|--------------------|---|------------------|
| Transmission | EIA RS-232C | |
| | Asynchronous, bi-directional, half duplex | |
| | Baud rate | 9600 bps |
| | Start bits | 1 bit |
| | Data bits | 8 bits |
| | Parity bit | None |
| | Stop bits | 2 bits |
| | X parameter | Not used |
| | ETX/ACK | Not used |
| | DSR | Not used |
| | Code | ASCII |
| Command | Carriage return | 0Dh |
| | Next line | 0Dh 0Ah |
| | → Next page | 0Ch 0Dh |
| Printer parameters | Next page | Automatic |
| | Characters per line | 72 min. |
| | Buffer size | approx. 32Kbytes |

Steps for data transmission

- Step 1** Enter the parameters into the printer so that the data can be transmitted.
- Step 2** Connect the cable to both the recorder and printer. Then the recorder displays - - - -. Refer to "Analysis Software and Communication Cable" about the cable.
- Step 3** Set the printer to "ON LINE".
- Step 4** Press the **START STOP** key. Then P is displayed and the data is transmitted.
- Step 5** When the transmission is finished, - - - - is displayed.
- Step 6** Remove the cable at once. The clock is displayed.



Print sample

| ABPM DATA TABLE | | | | | | |
|-----------------|-----------|-------|-----------|-----------|----------|-----|
| No. | Date | Time | SYS(mmHg) | DIA(mmHg) | PUL(bpm) | ERR |
| 1 | `97/ 5/17 | 7:43 | 103 | 65 | 55 | - |
| 2 | `97/ 5/17 | 8:00 | 119 | 79 | 65 | - |
| 3 | `97/ 5/17 | 8:30 | 125 | 88 | 132 | - |
| 4 | `97/ 5/17 | 9:00 | 122 | 84 | 116 | - |
| 5 | `97/ 5/17 | 9:30 | 115 | 87 | 63 | - |
| 6 | `97/ 5/17 | 10:00 | 118 | 76 | 61 | - |
| 7 | `97/ 5/17 | 10:30 | - | - | - | 08 |
| 8 | `97/ 5/17 | 10:35 | 116 | 82 | 68 | - |
| 9 | `97/ 5/17 | 11:00 | 114 | 75 | 62 | - |
| 10 | `97/ 5/17 | 11:30 | 122 | 81 | 94 | - |
| 11 | `97/ 5/17 | 12:00 | 123 | | | - |
| 12 | `97/ 5/17 | 12:30 | 112 | | | - |




Data Transmission to a Computer Using Analysis Software

Caution

- ❑ The recorder intensely consumes the battery power while connected to the RS-232C cable. Disconnect the cable when not actually transferring data.
- ❑ Maintain the power-on state while transmitting the data so that the data is not damaged.

Steps for data transmission

- Step 1** Connect the cable to both the recorder and printer. The recorder displays - - - -. Refer to "Analysis Software and Communication Cable" about the cable.
- Step 2** Read the data using the optional analysis software. Refer to the software instruction manual.
- Step 3** Remove the cable at once. The clock is displayed.





Options and accessories



Analysis Software and Communication Cables

The analysis software has functions as follows :

- ☐ Maximum value, minimum value and average is calculated in an arbitrary period of time. (Partial analysis)
- ☐ Correlation graph, trend graphis and histogram is displayed.
- ☐ The patient's data and information is managed.
- ☐ Data can be saved.
- ☐ It is able to delete and copy a data.
- ☐ A saved data file can export to CSV format that EXECL able to open.
- ☐ Data can be output as a report.
- ☐ Data is input from the recorder and parameters are written to the recorder.

| Name | Order code |
|--|------------|
| Windows analysis software for Japanese | TM-2430-15 |
| Windows analysis software for English | TM-2430-13 |
| Communication cable, D-SUB 9pin socket type | AX-KO1502 |
| Communication cable, D-SUB 25pin socket type | AX-KO1503 |
| Communication cable, D-SUB 25pin plug type | AX-KO1504 |



Cuffs and Other Accessories

| Name | | | Order code |
|---|------------|-----------------------------|----------------|
| Large cuff | left arm, | 28 ~ 36 cm (11 ~ 14 inches) | TM-2430-02 |
| Adult cuff | left arm, | 20 ~ 31 cm (8 ~ 12 inches) | TM-2430-06 |
| Small cuff | left arm, | 15 ~ 22 cm (6 ~ 8 inches) | TM-2430-07 |
| Adult cuff | right arm, | 20 ~ 31 cm (8 ~ 12 inches) | TM-2430-09 |
| Cuff cloth & bladder(14x30cm) for large cuff, left arm | | | TM-9111C-1 |
| Cuff cloth & bladder(12x22cm) for adult cuff, left arm | | | TM-9112C-1 |
| Cuff cloth & bladder(9x16cm) for small cuff, left arm | | | TM-9113C-1 |
| Cuff cloth & bladder(12x22cm) for adult cuff, right arm | | | TM-9117C-1 |
| Cuff cloth, 2 sheets | | for large cuff, left arm | AX-133001469-S |
| Cuff cloth, 2 sheets | | for adult cuff, left arm | AX-133001431-S |
| Cuff cloth, 2 sheets | | for small cuff, left arm | AX-133001468-S |
| Cuff cloth, 2 sheets | | for large cuff, right arm | AX-133001514-S |
| Cuff cloth, 2 sheets | | for adult cuff, right arm | AX-133001513-S |
| Cuff cloth, 2 sheets | | for small cuff, right arm | AX-133001515-S |
| Cuff cover, 10 sheets | | for large cuff | AX-133002066-S |
| Cuff cover, 10 sheets | | for adult cuff | AX-133002018-S |
| Cuff cover, 10 sheets | | for small cuff | AX-13A37410-S |

| Name | | Order code |
|-------------------------|-----------|--------------|
| TM-2430 Check connector | | TM-2430-90 |
| Record sheet | 10 sheets | AX-PP155-S |
| Holder | | AX-003001955 |



Maintenance



Checking Accuracy


Required equipment

- ☐ Accurate office mercury sphygmomanometer or aneroid gauge with inflation system.
- ☐ TM-2430 Check connector (TM-2430-90).
- ☐ A rigid cylinder sized to fit the cuff pressured.

Steps for checking accuracy

Step 1 Turn off the TM-2430 and remove the air hose from the unit.

Step 2 Construct the check system as this drawing.



Step 3 Keep the pressure at the air socket to atmospheric pressure.

Step 4 Turn on the power switch, when you press and hold the **START STOP** key. The TM-2430 blinks the "0" of measurement value.

Step 5 Squeeze the inflator bulb until cuff pressure reaches to 50 mmHg. Check the difference between the blinking display of TM-2430 and mercury sphygmomanometer that is within ± 3 mmHg.

Step 6 Squeeze the inflator bulb until cuff pressure reaches to 150 mmHg. Check this difference that is within ± 3 mmHg.

Step 7 Squeeze the inflator bulb until cuff pressure reaches to 250 mmHg. Check this difference that is within ± 3 mmHg.

Step 8 Release the cuff air, turn off the TM-2430 and remove them.

This blood pressure recorder is a precision instrument. Contact your nearest A&D office for this inspection, if you need repair.



Cleaning the cuff and recorder

- ❑ Before cleaning the recorder, remove the battery cover and turn the power switch off. Remove the batteries.
- ❑ The recorder is not water resistant, do not allow liquids to splash on or get into the case while cleaning.
- ❑ After each use, wipe the case of the recorder with a clean lint free cloth, moistened with water and a mild detergent.
- ❑ Do not use antiseptic solutions, Alcohol, etc., to clean the recorder, hose or cuff.
- ❑ Clean the cuff cloth and cuff cover by washing in water with a mild detergent. Do not scrub or wring them by hand. If the cuff cloth and cuff cover become contaminated, replace them with new covers.



Periodical inspection

- ❑ This blood pressure recorder is a precision instrument. Please inspect the functions (every year) periodically. Contact your nearest A&D office for this inspection.



Problem solving

Caution

- ❑ Do not open the case of the recorder because it uses delicate electrical components and intricate air unit that could be damaged.
- ❑ If you can not locate and fix the problem, request service from your supplier, or from the A&D service group.
- ❑ A&D service group will support authorized suppliers about technical information, spare parts and units.

| Problem | Cause | Treatment |
|---|---|--|
| No display at turning on. | Battery power has been consumed. | Replace with new batteries |
| Data lost while replacing batteries.... | Unable to charge the internal sub battery. | Set for clock display for approx. 24 hours. This battery charges during clock display. |
| No pressure. | Air leakage at the connector, hose or cuff. | Confirm the cuff and air hose are not damaged and are connected correctly. |



Error codes

Caution The error code updates without announcement.

| Error code | Meaning | Status | Operation |
|------------|--|--|---|
| E00 | No clock parameter | All parameters are lost. Reset status. | Enter clock parameters. Refer to "Setup of display and clock" |
| E03 | Zero point pressure error | An error code is displayed without cuff inflation. | Exhaust the air from the cuff completely. |
| E04 | Low battery | Measurement is stopped. An error code is displayed. | Replace with new batteries. Auto mode is quit. Restart the auto mode if you use it. |
| E05 | Inflation error | Inflation pressure does not reach the target pressure. | Wrap the cuff and connect to main unit exactly. If you can not clear the error, there may be an air leak and repair is necessary |
| E06 | Above 320mmHg | An error code is displayed. | Do not move and try to relax during the measurement. If you can not clear the error, the product will requier repair. |
| E07 | Controlled stop using STOP key | Air is exhausted. An error code is displayed. | Do not press the STOP key if you do not need to use it. |
| E08 | Pulsation can not be measured | Measurable pulsation is searched to 20mmHg in constant exhaust. An error code is displayed. | Do not move and try to relax during the measurement. The error occurs when measurable pulsations are not received due to thick cloth or quick motion. |
| E10 | Pulsations can not be detected because the patient may have moved. | In the measurement, Quick exhaust is executed. An error code is displayed. | Do not move and try to relax during the measurement. |
| E20 | Pulse rate < 30 200 < Pulse rate | An error code is displayed. DIA : Diastolic Blood Pressure SYS : Systolic Blood Pressure DSD : The Difference between Systolic Blood Pressure and Diastolic Blood Pressure. | Measure the blood pressure by other methods. |
| E21 | DIA < 40 160 < DIA | | |
| E22 | SYS < 60 280 < SYS | | |
| E23 | DSD < 10 150 < DSD | | |

| Error code | Meaning | Status | Operation |
|--------------------------|--|--|---|
| E30 | Greater than 90 seconds after measurement start. | Air is exhausted from the cuff, and an error code is displayed. | Repair is needed because of slow inflation or slow constant exhaust. |
| E31 | Greater than 60 seconds into the constant exhaust. | Air is exhausted from the cuff, and an error code is displayed. | Repair is needed because of slow constant exhaust. |
| E32 | Clock error. | An error code is displayed. | If you not clear this error, the product needs repair. |
| E50 | Pressure offset error for measurable pulsation. | An error code is displayed at restarting the product. | Exhaust the air from the cuff completely, reset the product. If you not clear this error, the product needs repair. |
| E52 | Memory error. | An error code is displayed at restarting the product. | The product needs repair. |
| E53 | Battery contact is defective. | The measurement is stopped, air is exhausted from the cuff and an error code is displayed. | Replace batteries correctly. If you can not clear this error, the product needs repair. |
| E55 E56 E57 | Exhaust error. | An error code is displayed at measurement. | Relax and do not move during the measurement. If this error occurs many times, repair is needed. |
| E60 | Interval setting error. | Start time is not proper, interval of last block not set in the unit of 120 min. | Enter parameters for the interval correctly. |
| E70 E71 E72 E73 | RS-232C error. | The error code is displayed during communications. | Re-connect the communication cable. If you can not clear this error, the product needs repair. |
| E74 | Voltage decrease during communication. | | Replace batteries with new ones and restart communication. |
| E75 | Protocol error due to external equipment. | | Re-connect the communication cable. If you can not clear this error, the product needs repair. |
| Other | | Monitor code is displayed. | Reset. Turn on power switch again. |



Index

Symbols

| | | |
|----|-------|-----------|
| | | 7 |
| | | 7 |
| | | 7 |
| | | 7 |
| | | 7 |
| | | 7 |
| | | 7 |
| | | 7, 15, 17 |
| | | 17 |
| | | 17 |
| | | 18 |
| | | 18 |
| | | 27, 28 |
| SN | | 7 |

A

| | | |
|-----------------------|-------|-------------|
| A | | 7, 9, 25 |
| AAMI | | 11 |
| accuracy | | 10 |
| adult cuff | | 30 |
| air hose | | 6, 23 |
| air plug | | 6, 24 |
| alkaline battery | | 3, 10 |
| analysis | | 8, 29 |
| arrow | | 7 |
| artery | | 23 |
| AUTO ON/OFF key | | 6 |
| automatic measurement | | 6, 7, 9, 25 |

B

| | | |
|---------------|-------|-----------|
| B | | 7, 10, 14 |
| battery | | 3, 14 |
| battery cover | | 6 |
| baud rate | | 11 |
| belt | | 5, 6, 24 |
| bladder | | 31 |

| | | |
|-------|-------|---------------|
| block | | 2, 17, 18, 19 |
| bpm | | 2 |

C

| | | |
|------------------|-------|------------------|
| cable | | 29 |
| Check connector | | 30 |
| | | 21 |
| clock | | 7, 10, 15 |
| code | | 11 |
| Constant exhaust | | 2 |
| CSV | | 29 |
| cuff | | 3, 5, 23, 24, 30 |
| cuff cloth | | 30 |
| cuff cover | | 5, 23, 30 |

D

| | | |
|--------------------------|-------|----|
| D-SUB | | 29 |
| data bits | | 11 |
| Detection method | | 10 |
| Depressurization | | 10 |
| DIA | | 2 |
| diastolic | | 7 |
| Diastolic Blood Pressure | | 2 |
| dimensions | | 11 |
| Display | | 10 |
| DSD | | 2 |

E

| | | |
|------------------|-------|------------|
| | | 15, 21, 33 |
| EEC | | 2 |
| EXECL | | 29 |
| Exhaust | | 2 |
| Exhaust velocity | | 2 |

F

| | | |
|-------------|-------|---------------|
| FCC | | 2 |
| frequency | | 2, 17, 18, 19 |
| full memory | | 7 |

H

hiding the measurement

value 9

holder 5, 6, 24, 30

I

ID number 10

interface 11

interval 2, 17

L

large cuff 30

left arm 30

low battery 7

M

M 7, 9

management 8

manual measurement 8

Measurement 10

memory 9, 10

mode I 17

mode II 17, 25

mode III 17, 18

N

Ni-Cd 4, 10

O

operating environment ... 11

P

P 27

parity 11

patient 22

portability 8

power 11

power switch 6

printer 26

pulse 7

R

record sheet 30

recorder 5, 22

reset key 6, 21

right arm 30

ring 23

RS-232C 6, 11, 26

S

S 7, 17

selecting mode 17

sleep 7

slit 23

small cuff 30

SN 7

START STOP key 6

start time 2, 17, 18, 19

stop bits 11

stop of a measurement 9

Storage 4

storage environment 11

SYS 2

systolic 7

Systolic Blood Pressure ... 2

T

Transmission 26

V

velcro 23

W

weight 11

X

X parameter 11

Y

yellow mark 6, 24



A&D Co., LTD.

1-243 Asahi, Kitamoto-shi, Saitama, 364-8585 JAPAN
Telephone: [81] (485) 93-1111 Fax: [81] (485) 93-1119

A&D ENGINEERING, INC.

1555 McCandless Drive, Milpitas, CA. 95035 U.S.A.
Telephone: [1] (408) 263-5333 Fax: [1] (408) 263-0119

A&D INSTRUMENTS LTD.

<Authorized Representative Established in the European Community>
Abingdon Science Park, Abingdon, Oxford OX14 3YS ENGLAND
Telephone: [44] (1235) 550420 Fax: [44] (1235) 550485

<German Sales Office>

Genfer Strasse 6, Niedereschbach 60437 Frankfurt Main 50 GERMANY
Telephone: [49] (69) 507-1017 Fax: [49] (69) 507-2054

A&D MERCURY PTY. LTD.

32 Dew Street, Thebarton, South Australia 5031 AUSTRALIA
Telephone: [61] (8) 352-3033 Fax: [61] (8) 352-7409

A&D KOREA Limited

8th Floor, Manhattan Bldg. 36-2 Yoido-dong, Youngdeungpo-ku, Seoul, KOREA
Telephone: [82] (2) 780-4101 Fax: [82] (2) 782-4280