

BUREAU OF INDIAN STANDARDS

DRAFT FOR COMMENTS ONLY

(Not to be reproduced without permission of BIS or used as an Indian Standard)

भारतीय मानक मसौदा
मोड़ने योग्य शल्य चिकित्सा के लिए मेज - विशिष्टि
[IS 9132 का पहला पुनरीक्षण]

Draft Indian Standard
Folding Type Operation Table - Specification
[First Revision of IS 9132]

ICS 11.140

Hospital Equipment and Surgical Disposable
Products Sectional Committee, MHD 12

Last date for comments:
12 September 2026

FOREWORD

(Formal clause will be added later)

This standard was originally published in 1979. The first revision of this standard has been brought out to align it with recent developments in the material requirements and to bring the standard in line with the latest style and format of referred standards.

For the purpose of deciding whether a particular requirement of this standard is complied with the final value, observed or calculated, expressing the result of a test or analysis shall be rounded off in accordance with IS 2: 2022 'Rules for rounding off numerical values (*second revision*)'. The number of significant places retained in the rounded off value should be the same as that of the specified value in this standard.

1 SCOPE

This standard specifies the construction and performance requirements of manually controlled folding type operation table.

2 REFERENCES

The standards given below contain provisions which, through reference in this text, constitute provisions of this standard. At the time of publication, the editions indicated were valid. All standards are subject to revision, and parties to agreements based on this standard are encouraged to investigate the possibility of applying the most recent editions of these standards.

<i>IS No.</i>	<i>Title</i>
MHD 12 (33411) WC (Under Development)	General requirements for hospital furniture (<i>first revision</i>)
IS 6911: 2026	Stainless Steel Plate, Sheet and Strip - Specification (<i>third revision</i>)
IS 1068: 2025 ISO 1456: 2009	Metallic and other inorganic coatings - Electrodeposited coatings of nickel, nickel plus chromium, copper plus nickel and of copper plus nickel plus chromium (<i>fourth revision</i>)

3 MATERIAL REQUIREMENTS

3.1 Angles, Flats, Strips and Rods

The components shall be made up of mild steel with cold rolled or hot rolled quality.

3.2 Top

The top shall be made up of aluminium alloy with suitable composition or stainless-steel conforming to designation X07Cr18Ni9S given in **Table 1** of IS 6911 or mild steel sheet with hot rolled or cold rolled quality.

4 DIMENSIONS

4.1 The dimensions of various parts present in operation table shall be as per given below:

<i>Component</i>	<i>Dimensions</i>
Head Section	300 x 500 mm
Trunk Section	1050 x 500 mm
Leg Section	450 x 500 mm
Gap between sections	10 mm
Height from floor to top of table	850 mm

4.2 The tolerance of ± 5 mm shall be allowed on all parts of the operation table except Gap between sections, the tolerance for Gap between sections shall be ± 2 mm.

5 CONSTRUCTION REQUIREMENTS

5.1 Framework

The framework of continuous lengths of mild steel tubes shall be welded to give the desired shape. The floor ends of the tubular framework shall be fitted with metal shoes with rounded edges.

5.2 Table Top

The table top shall have a minimum of three sections that is head, trunk and leg section. The leg section shall be detachable. The entire top shall be capable of satisfying the following requirements:

- a) The head end shall be capable of being raised by 60° up and 45° down.
- b) The trunk section shall support the body and be capable of being raised by 45°.

5.3 The lowering and raising of the two sections shall be arranged with a ratchet and hook arrangement.

5.4 The operation table shall have provision to attach optional accessories like anaesthetic screen, saline drip, arm rest, instrument tray and lithotomy rod at hip level.

6 ACCESSORIES

The following accessories shall be optional:

TABLE 1 LIST OF ACCESSORIES	
<i>Sr No.</i>	<i>Accessories</i>
i)	Detachable instrument tray
ii)	Rubber, mattress in three pieces

7 PERFORMANCE REQUIREMENTS

7.1 Base

The construction of the table shall ensure stability of the base. The metal framework shall have four metal shoes to provide sufficient rigidity to the table.

7.2 Table Top

The load of 750 N shall be applied on the operation table vertically in the middle of the table from height of 100 mm, when load is applied there shall not be wobble or damage to the table.

8 TESTING REQUIREMENTS

8.1 Adhesion Test

The adhesion test required for the operation table shall conform to 6.1.1 of IS 4033.

9 WORKMANSHIP AND FINISH

9.1 The components of the operation table shall be finished smooth to provide easy movements.

9.2 The construction of the operation table shall be rigid.

9.3 The entire table shall be capable of folding to a small unit.

9.4 The edges of the table shall be well finished and rounded off.

9.5 The three sections comprising the table top in their normal position shall be in one plane.

9.6 The welded joints shall have welds fully penetrating.

9.7 The surface of the operation table shall be painted in accordance with 7.1 of IS 4033.

9.8 The stainless steel table top shall be passivated.

9.9 The Aluminium top shall be anodized.

9.10 The resulting finish shall be hard and shall not readily chip or flake. The plating shall conform to **Service Condition 2** of IS 1068.

10 Operation Manual

10.1 The table shall be accompanied by instructions manual giving the following information:

- a) Brief instructions for installation and maintenance for opening and folding the table, and
- b) Any other information considered necessary by the manufacturer.

11 MARKING

11.1 The marking which is used by the manufacturer shall be legibly and indelibly marked with the Identification mark of the manufacturer.

11.2 BIS Certification Marking

The product(s) conforming to the requirements of this standard may be certified as per the conformity assessment schemes under the provisions of the *Bureau of Indian Standards Act*, 2016 and the Rules and Regulations framed thereunder, and the product(s) may be marked with the Standard Mark.

12 PACKAGING

The operation table shall be packed as agreed to between the purchaser and the manufacturer. Each individual package shall contain only one unit and may be marked with the name of the item, and identification of the source of manufacture and batch number and any other marking if so desired.