

BUREAU OF INDIAN STANDARDS

DRAFT FOR COMMENTS ONLY

(Not to be reproduced without permission of BIS or used as an Indian Standard)

भारतीय मानक मसौदा
चिकित्सीय उपयोग के लिए इन्फ्यूजन उपकरण
भाग 16 आयतनमितीय इन्फ्यूजन नियंत्रकों के साथ एकल उपयोग के लिए
इन्फ्यूजन सेट

Draft Indian Standard
Infusion Equipment for Medical Use
Part 16 Infusion Sets for Single Use with Volumetric Infusion Controllers
[ICS 11.040.20]

Hospital Equipment and Surgical Disposable
Products Sectional Committee, MHD 12

Last date for comments:
19 September 2026

NATIONAL FOREWORD

(Adoption clause will be added later)

The text of ISO Standard has been approved as suitable for publication as an Indian Standard without deviations. Certain conventions are however not identical to those used in Indian Standards. Attention is particularly drawn to the following:

- a) Wherever the words ‘International Standard’ appear referring to this standard, they should be read as ‘Indian Standard’; and
- b) Comma (,) has been used as a decimal marker while in Indian Standards, the current practice is to use a point (.) as the decimal marker.

In this adopted standard, reference appears to certain International Standards for which Indian Standards also exist. The corresponding Indian Standards which are to be substituted in their respective places are listed below along with their degree of equivalence for the editions indicated:

<i>International Standard</i>	<i>Corresponding Indian Standard</i>	<i>Degree of Equivalence</i>
ISO 3696 Water for analytical laboratory use — Specification and test methods	IS 1070 : 2023 Reagent grade water — Specification (<i>fourth revision</i>)	Technically Equivalent
ISO 7000 Graphical symbols for use on equipment — Registered symbols	IS 16450 : 2023/ISO 7000 : 2019 Graphical symbols for use on equipment — Registered symbols (<i>first revision</i>)	Identical

ISO 8536-4 Infusion equipment for medical use — Part 4: Infusion sets for single use, gravity feed	IS/ISO 8536-4 : 2019 Infusion equipment for medical use: Part 4 Infusion sets for single use, gravity feed	Identical
ISO 14644-1 Cleanrooms and associated controlled environments — Part 1: Classification of air cleanliness by particle concentration	IS 18637 (Part 1) : 2024 Cleanrooms and associated controlled environments: Part 1 Classification of air cleanliness by particle concentration (ISO 14644-1 : 2015, MOD)	Modified
ISO 15223-1 Medical devices — Symbols to be used with information to be supplied by the manufacturer — Part 1: General requirements	IS 18105 (Part 1) : 2023/ISO 15223-1 : 2021 Medical devices — Symbols to be used with information to be supplied by the manufacturer: Part 1 General requirements (<i>third revision</i>)	Identical

This standard also makes a reference to the BIS Certification Marking of the product. Details of which is given in National Annex A.

For the purpose of deciding whether a particular requirement of this standard is complied with, the final value, observed or calculated expressing the result of a test or analysis, shall be rounded off in accordance with IS 2 : 2022 ‘Rules for rounding off numerical values (*second revision*)’. The number of significant places retained in the rounded off value should be the same as that of the specified value in this standard.

NATIONAL ANNEX A
(National Foreword)

A-1 BIS CERTIFICATION MARKING

The product(s) conforming to the requirements of this standard may be certified as per the conformity assessment schemes under the provisions of the *Bureau of Indian Standards Act, 2016* and the Rules and Regulations framed thereunder, and the product(s) may be marked with the Standard Mark.

Scope

This document specifies the requirements for sterilized, single-use, gravity feed infusion sets, used together with the volumetric infusion controllers of IEC 60601-2-24.

The technical content of the document has not been enclosed as it is identical with the corresponding ISO standard. For details, please refer to ISO 8536-16 : 2025 or kindly contact:

Head
Medical Equipment and Hospital Planning Department
Bureau of Indian Standards
9 Bahadur Shah Zafar Marg
New Delhi-110002
Email: mhd@bis.gov.in
hmhd@bis.gov.in