

**BUREAU OF INDIAN STANDARDS**

**DRAFT FOR COMMENTS ONLY**

*(Not to be reproduced without permission of BIS or used as an Indian Standard)*

भारतीय मानक मसौदा  
**मीटरिंग पंपों के लिए स्नैप-ऑन बोटलें**  
**भाग 3 प्लास्टिक**

*Draft Indian Standard*  
**Snap-on Bottles for Metering Pumps**  
**Part 3 Plastic**

[ICS 11.120.99]

Hospital Equipment and Surgical Disposable  
Products Sectional Committee, MHD 12

Last date for comments:  
**19 September 2026**

**NATIONAL FOREWORD**

*(Adoption clause will be added later)*

The text of ISO Standard has been approved as suitable for publication as an Indian Standard without deviations. Certain conventions are however not identical to those used in Indian Standards. Attention is particularly drawn to the following:

- Wherever the words 'International Standard' appear referring to this standard, they should be read as 'Indian Standard'; and
- Comma (,) has been used as a decimal marker while in Indian Standards, the current practice is to use a point (.) as the decimal marker.

In this adopted standard, reference appears to certain International Standards for which Indian Standards also exist. The corresponding Indian Standards which are to be substituted in their respective places are listed below along with their degree of equivalence for the editions indicated:

| <i>International Standard</i>                                                                                                                       | <i>Corresponding Indian Standard</i>                                                                                                                                                  | <i>Degree of Equivalence</i> |
|-----------------------------------------------------------------------------------------------------------------------------------------------------|---------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|------------------------------|
| ISO 11607-1 Packaging for terminally sterilized medical devices — Part 1: Requirements for materials, sterile barrier systems and packaging systems | IS/ISO 11607-1 : 2019 Packaging for terminally sterilized medical devices: Part 1 Requirements for materials, sterile barrier systems and packaging systems ( <i>first revision</i> ) | Identical                    |

|                                                                                                                                               |                                                                                                                                                                                 |           |
|-----------------------------------------------------------------------------------------------------------------------------------------------|---------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|-----------|
| ISO 11607-2 Packaging for terminally sterilized medical devices — Part 2: Validation requirements for forming, sealing and assembly processes | IS/ISO 11607-2 : 2019 Packaging for terminally sterilized medical devices: Part 2 Validation requirements for forming, sealing and assembly processes ( <i>first revision</i> ) | Identical |
|-----------------------------------------------------------------------------------------------------------------------------------------------|---------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|-----------|

This standard also makes a reference to the BIS Certification Marking of the product. Details of which is given in National Annex A.

For the purpose of deciding whether a particular requirement of this standard is complied with, the final value, observed or calculated expressing the result of a test or analysis, shall be rounded off in accordance with IS 2 : 2022 ‘Rules for rounding off numerical values (*second revision*)’. The number of significant places retained in the rounded off value should be the same as that of the specified value in this standard.

**NATIONAL ANNEX A**  
(National Foreword)

**A-1 BIS CERTIFICATION MARKING**

The product(s) conforming to the requirements of this standard may be certified as per the conformity assessment schemes under the provisions of the *Bureau of Indian Standards Act, 2016* and the Rules and Regulations framed thereunder, and the product(s) may be marked with the Standard Mark.

**Scope**

This document specifies the shape, dimensions, fill capacities and performance requirements of plastic bottles for metering pumps. It also specifies the material for the manufacturing of such containers as well as the secondary packaging.

The document provides requirements for packaging of the plastic bottles and addresses nonsterile, ready to sterilize or sterile as three possible options.

This document is applicable to colourless or coloured containers moulded from plastic and intended to be used in the packaging, storage or transportation of products intended for medicinal use.

---

The technical content of the document has not been enclosed as it is identical with the corresponding ISO standard. For details, please refer to ISO 24166-3 : 2022 or kindly contact:

Head  
Medical Equipment and Hospital Planning Department  
Bureau of Indian Standards  
9 Bahadur Shah Zafar Marg  
New Delhi-110002  
Email: [mhd@bis.gov.in](mailto:mhd@bis.gov.in)  
[hmhd@bis.gov.in](mailto:hmhd@bis.gov.in)