

BUREAU OF INDIAN STANDARDS

DRAFT FOR COMMENTS ONLY

(Not to be reproduced without permission of BIS or used as an Indian Standard)

भारतीय मानक मसौदा

**चिकित्सीय उपकरणों की धुलाई एवं कीटाणुशोधन हेतु यंत्र
भाग 6 गैर-गंभीर चिकित्सा उपकरणों और स्वास्थ्य देखभाल उपकरणों के लिए
तापीय कीटाणुशोधन का उपयोग करने वाले धोनेवाले-कीटाणुशोधकों के लिए
अपेक्षाएँ और परीक्षण**

Draft Indian Standard

Washer-Disinfectors

**Part 6 Requirements and Tests for Washer-Disinfectors Employing
Thermal Disinfection for Non-Critical Medical Devices and Health Care
Equipment**

[ICS 11.080.10]

Hospital Equipment and Surgical Disposable
Products Sectional Committee, MHD 12

Last date for comments:
19 September 2026

NATIONAL FOREWORD

(Adoption clause will be added later)

This standard supersedes IS/ISO 15883-6 : 2011 ‘Washer-disinfectors: Part 6 Requirements and tests for washer-disinfectors employing thermal disinfection for non-invasive, non-critical medical devices and healthcare equipment’.

The text of ISO Standard has been approved as suitable for publication as an Indian Standard without deviations. Certain conventions are however not identical to those used in Indian Standards. Attention is particularly drawn to the following:

- Wherever the words ‘International Standard’ appear referring to this standard, they should be read as ‘Indian Standard’; and
- Comma (,) has been used as a decimal marker while in Indian Standards, the current practice is to use a point (.) as the decimal marker.

In this adopted standard, reference appears to certain International Standards for which Indian Standards also exist. The corresponding Indian Standards which are to be substituted in their respective places are listed below along with their degree of equivalence for the editions indicated:

<i>International Standard</i>	<i>Corresponding Indian Standard</i>	<i>Degree of Equivalence</i>
-------------------------------	--------------------------------------	------------------------------

ISO 15883-1 : 2024 Washer-disinfectors — Part 1: General requirements, terms and definitions and tests	IS 18344 (Part 1) : 2026/ISO 15883-1 : 2024 Washer-disinfectors: Part 1 General requirements, terms and definitions and tests	Identical
ISO 15883-5 Washer-disinfectors — Part 5: Performance requirements and test method criteria for demonstrating cleaning efficacy	IS 18468 (Part 5) : 2024 Washer-disinfectors: Part 5 Performance requirements and test method criteria for demonstrating cleaning efficacy (ISO 15883-5 : 2021, MOD)	Modified

This standard also makes a reference to the BIS Certification Marking of the product. Details of which is given in National Annex A.

For the purpose of deciding whether a particular requirement of this standard is complied with, the final value, observed or calculated expressing the result of a test or analysis, shall be rounded off in accordance with IS 2 : 2022 ‘Rules for rounding off numerical values (*second revision*)’. The number of significant places retained in the rounded off value should be the same as that of the specified value in this standard.

NATIONAL ANNEX A
(National Foreword)

A-1 BIS CERTIFICATION MARKING

The product(s) conforming to the requirements of this standard may be certified as per the conformity assessment schemes under the provisions of the *Bureau of Indian Standards Act, 2016* and the Rules and Regulations framed thereunder, and the product(s) may be marked with the Standard Mark.

Scope

This document specifies particular requirements for washer-disinfectors (WD) intended for use when the level of assurance of disinfection that is necessary can be achieved by cleaning and thermal disinfection (A0 not less than 60) and does not require an independent automated record of critical processes to be kept. It is intended to be used in conjunction with ISO 15883-1, which specifies general requirements for WD.

The range of products on which WD of this particular type can be used is restricted to non-invasive and non-critical devices and equipment (i.e. not penetrating skin or contacting mucosal surfaces).

NOTE Thermal disinfection can be achieved by rinsing the load with hot water, exposure to steam, or combination of the two.

This document does not cover powered devices, lumened devices, and other semi-critical and critical medical devices.

Devices identified within the scopes of ISO 15883-2, ISO 15883-3, ISO 15883-4, and ISO 15883-7 do not fall within the scope of this document.

The technical content of the document has not been enclosed as it is identical with the corresponding ISO standard. For details, please refer to ISO 15883-6 : 2026 or kindly contact:

Head
Medical Equipment and Hospital Planning Department
Bureau of Indian Standards
9 Bahadur Shah Zafar Marg
New Delhi-110002
Email: mhd@bis.gov.in
hmhd@bis.gov.in