UNIVERSITÄT
BERN

UNIVERSITÄT BERN

X-ray microtomography

David Haberthür

December 22, 2023 | 9256-HS2023-0: Advanced Microscopy

Grüessech mitenang!

- David Haberthür
  - Physicist by trade
  - PhD in high resolution imaging of the lung, Institute of Anatomy, University of Bern, Switzerland
  - Post-Doc I: TOMCAT, Swiss Light Source, Paul Scherrer Institute, Switzerland
  - Post-Doc II: µCT group, Institute of Anatomy, University of Bern, Switzerland.

Grüessech from the µCT-group



David.Haberthuer@unibe.ch Ruslan.Hlushchuk@unibe.ch Oleksiy.Khoma@unibe.ch

µCT-group

- microangioCT [1]
    - Angiogenesis: heart, musculature [2] and bones
    - Vasculature: (mouse) brain [3], (human) nerve scaffolds [4], (human) skin flaps [5] and tumors
- Zebrafish musculature and gills [6]
- (Lung) tumor detection and metastasis classification [7]
- Collaborations with museums [8] and scientist at UniBe [9, 10] to scan a wide range of specimens, from human hearing bones to meteorites
- Automate *all* the things! [11, 12]



bruker.com/skyscan1272

Contents

micro-Computed tomography

- Allows for imaging dense and non-transparent samples
- Non-destructive imaging
- Results in three-dimensional images
- Covers a very large range of sample sizes



From [13], Subject *C3L-02465*

micro-Computed tomography

- Allows for imaging dense and non-transparent samples
- Non-destructive imaging, thus compatible with routine sample preparation
- Results in three-dimensional images with µm resolution
- Covers a very large range of sample sizes
- (Small) biological samples
- Enables correlative imaging pipelines, scanning of precious biological samples, as well as museum & collection material



Imaging methods



Stefan Tschanz, with permission

- *Light Sheet Microscopy* by Nadia Mercader Huber
- X-ray imaging
- Electron microscopy
    - *Transmission Electron Microscopy* by Dimitri
    - *Scanning Electron Microscopy* by Sabine Kässmeyer and Ivana Jaric
    - *Cryoelectron Microscopy & Serial Block Face SEM* by Ioan

CT-Scanner

youtu.be/2CWpZKuy-NE

CT History

- 1963: Cormack used a collimated ^{60}Co source and a Geiger counter as a detector [14]
- 1976: Hounsfield worked on first clinical scanner [15]



From [16], Figure 5

CT History

- 1963: Cormack used a collimated ^{60}Co source and a Geiger counter as a detector [14]
- 1976: Hounsfield worked on first clinical scanner [15]
- CT scanner generations
    - First generation



From [17], Figure 1.12

CT History

- 1963: Cormack used a collimated ^{60}Co source and a Geiger counter as a detector [14]
- 1976: Hounsfield worked on first clinical scanner [15]
- CT scanner generations
  - First generation
  - Second generation



From [17], Figure 1.13

UNIVERSITÄT BERN

CT History

- 1963: Cormack used a collimated ^{60}Co source and a Geiger counter as a detector [14]
- 1976: Hounsfield worked on first clinical scanner [15]
- CT scanner generations
  - First generation
  - Second generation
  - Third generation



From [17], Figure 1.14

µCT History I

- X-ray computed tomography began to replace analog focal plane tomography in the early 1970s [18]
- Non-medical use in the late 1970s, for detection of internal defects in fabricated parts and equipment
- Lee Feldkamp [19] developed one of the early laboratory microCT systems by assembling a micro-focus cone beam x-ray source, specimen holder and stagers, and an image intensifier at Ford Motor Company's Scientific Research Laboratory to nondestructively detect damage in ceramic manufactured automobile parts
- Feldkamp met with scientists at Henry Ford Hospital and University of Michigan interested in understanding the relationship between the microstructure and biomechanical function of trabecular bone to study osteoporotic fractures [20]

µCT History II

- µCT was first reported in the 1980s, for scanning gemstones to cut out the largest possible one
- Today: Nondestructive imaging for quantifying the (micros)tructure of organic materials
  - Mineralized bone tissue and the relationships between the mechanical behavior of bone to its structural and compositional properties
  - Teeth and their internal details
  - Soft tissues and vasculature using radio-opaque contrast agents
  - Characterization of anatomical details in high resolution
- $\approx$2500 µCT systems are in use worldwide with over 1000 publications annually

Why µCT?



From [13], Subject *C3L-02465*

Why µCT?



From [13], Subject *C3L-02465*

UNIVERSITÄT BERN

Why µCT?



From [13], Subject *C3L-02465*



Why µCT?



From [13], Subject *C3L-02465*

Why µCT?

From [13], Subject *C3L-02465*

Maximum intensity projection

Projections

Reducing the dimensions of a dataset. For example projecting a volume (3D) to a surface by taking the maximum value across planes for each pixel.





Fundamentals of Digital Image Processing (2023) by Guillaume Witz, Slide 24

Machinery

- Hospital CT
  - Voxel size around 0.5 mm
- Lab/Desktop CT
  - Voxel size around 7 µm (*in vivo*)
  - Voxel size around 0.5 µm (*ex vivo*)
- Synchrotron CT
  - Voxel size down to 160 nm

flic.kr/p/D4rbom

Machinery

- Hospital CT
  - Voxel size around 0.5 mm
- Lab/Desktop CT
  - Voxel size around 7 µm (*in vivo*)
  - Voxel size around 0.5 µm (*ex vivo*)
- Synchrotron CT
  - Voxel size down to 160 nm

flic.kr/p/fpTrGu

Machinery

- Hospital CT
  - Voxel size around 0.5 mm
- Lab/Desktop CT
  - Voxel size around 7 µm (*in vivo*)
  - Voxel size around 0.5 µm (*ex vivo*)
- Synchrotron CT
  - Voxel size down to 160 nm



bruker.com/skyscan1272

Machinery

- Hospital CT
  - Voxel size around 0.5 mm
- Lab/Desktop CT
  - Voxel size around 7 µm (*in vivo*)
  - Voxel size around 0.5 µm (*ex vivo*)
- Synchrotron CT
  - Voxel size down to 160 nm



flic.kr/p/7Xhk2Y

UNIVERSITÄT BERN

What is happening?

No matter what kind of machine, the basic principle is always the same

- an x-ray source
- a sample
- a detector



w.wiki/7g3

Machinery



Examples



Examples



Examples



From [8], *Diancta phoenīx*

UNIVERSITÄT BERN

Examples



Examples

X-ray interaction

- "X-rays interact with tissue in 2 main ways: photoelectric effect and Compton scatter. To a first approximation, the photoelectric effect contributes to contrast while the Compton effect contributes to noise. Both contribute to dose." ([21])
  - Photoelectric absorption (τ) is strongly dependent on the atomic number Z of the absorbing material: $\tau \propto \frac{Z^4}{E^{3.5}}$
  - Compton scattering is one of the principle forms of photon interaction and is directly proportional to the (electron & physical) density of the material. It does *not* depend on the atomic number: $\lambda' - \lambda = \frac{h}{m_e c}(1 - \cos\theta)$
- Lowering x-ray energy increases contrast
- X-ray penetration decreases exponentially with sample thickness [22, i. e. Beer-Lamberts law]: $I(t) = I_0\, e^{-\alpha z}$

Composition of biological tissues

Tissue: content by mass percentage

Element	H	C	N	O	Na	P	S	Cl	K	Ca
Atomic number	1	6	7	8	11	15	16	17	19	20
Fat	11.4	59.8	0.7	27.8	0.1		0.1	0.1		
Water	11.2			88.8						
Blood	10.2	11	3.3	74.5	0.1	0.1	0.2	0.3	0.2	
Liver	10.2	13.9	3	71.6	0.3	0.2	0.3	0.2	0.3	
Brain	10.7	14.5	2.2	71.2	0.2	0.4	0.2	0.3	0.3	
Bone	3.4	15.5	4.2	43.5	0.1	10.3	0.3			22.5

Preparation

- Study design
- Sample preparation

Projections



Contrast, Magnification and Resolution—Laws of Physics for Microscopists (1, 2023) by Martin Frenz, Slide 21

Projection acquisition



Projections



Projections

- A (micro-focus) x-ray source illuminates the object
- The x-rays penetrate the sample and are attenuated
- A scintillator converts the x-rays to visible light
- A (planar) x-ray detector collects (magnified) projection images.
- The projections are recorded on disk

Reconstructions



UNIVERSITÄT BERN

Reconstructions



UNIVERSITÄT BERN

Reconstructions

Reconstructions

- Based on hundreds of angular views acquired while the object rotates, a computer synthesizes a stack of virtual cross section slices through the object.
- Radon Transformation
- Filtered back projection
- Cone beam reconstruction [19]
- Corrections (beam hardening, etc.)
- Writing to stack

Visualization

Visualization

- Based the on reconstructions, a computer synthesizes a three-dimensional view of the scanned sample

What to use?

- ImageJ/Fiji [23]
- Also see *Fundamentals of Digital Image Processing* by Guillaume Witz
- Reproducible research
  - in Jupyter [24]
  - **git**
  - Script all your things!
  - Data repositories; i. e. sharing is caring!

Quantitative data

- Pretty images are nice, but we need quantitative numbers
- Segmentation
- Characterization

Internal morphology of human teeth

Collaboration with zmk bern – Zahnmedizinische Kliniken

- Numbers instead of just pretty images
- Segmentation of teeth and root canal
- (Unbiased) Characterization
- Reproducible and automated image analysis (in Jupyter [24])
- Two publications:
  - [11]: doi.org/gjpw2d
  - [25]: doi.org/g7r8

UNIVERSITÄT BERN

How?

- 104 extracted human permanent mandibular canines
- µCT imaging
- Root canal configuration, according to Briseño-Marroquín et al. [26]
- *Reproducible* analysis [27], e. g. you can click a button to double-check or recalculate the results yourself!

UNIVERSITÄT BERN

How?

- 104 extracted human permanent mandibular canines
- µCT imaging
- Root canal configuration, according to Briseño-Marroquín et al. [26]
- *Reproducible* analysis [27], e. g. you can click a button to double-check or recalculate the results yourself!



bruker.com/skyscan1272

UNIVERSITÄT BERN

How?

- 104 extracted human permanent mandibular canines
- µCT imaging
- Root canal configuration, according to Briseño-Marroquín et al. [26]
- *Reproducible* analysis [27], e. g. you can click a button to double-check or recalculate the results yourself!

```
Scanner=SkyScan1272
Instrument S/N=15G09089-B
Software Version=1.1.19
Filename Prefix=Tooth045~00
Number Of Files=  482
Number Of Rows= 1092
Number Of Columns= 1632
Source Voltage (kV)=  80
Source Current (uA)= 125
Image Pixel Size (um)=9.999986
Exposure (ms)=950
Rotation Step (deg)=0.400
Frame Averaging=ON (3)
Filter=Al 1mm
Study Date and Time=02 Jul 2020
08h:23m:34s
Scan duration=0h:39m:51s
```

UNIVERSITÄT BERN

How?

- 104 extracted human permanent mandibular canines
- µCT imaging
- Root canal configuration, according to Briseño-Marroquín et al. [26]
- *Reproducible* analysis [27], e. g. you can click a button to double-check or recalculate the results yourself!

Sample changer on the SkyScan 1272

In total:

- 13 days of *continuous* µCT scanning
- 819 GB of raw data
  230 648 TIFF projections
- 326 GB data as input for analysis
  282 062 PNG reconstructions

UNIVERSITÄT BERN

How?

- 104 extracted human permanent mandibular canines
- µCT imaging
- Root canal configuration, according to Briseño-Marroquín et al. [26]
- *Reproducible* analysis [27], e. g. you can click a button to double-check or recalculate the results yourself!



From [26], Fig. 2

UNIVERSITÄT BERN

How?

- 104 extracted human permanent mandibular canines
- µCT imaging
- Root canal configuration, according to Briseño-Marroquín et al. [26]
- *Reproducible* analysis [27], e. g. you can click a button to double-check or recalculate the results yourself!

gph.is/2nqkple

How?

- 104 extracted human permanent mandibular canines
- µCT imaging
- Root canal configuration, according to Briseño-Marroquín et al. [26]
- *Reproducible* analysis [27], e. g. you can click a button to double-check or recalculate the results yourself!



UNIVERSITÄT
BERN

µCT imaging



Dataset cropping

- Full datasets: 326 GB
- Cropped datasets: 115 GB







Dataset cropping

- Full datasets: 326 GB
- Cropped datasets: 115 GB



UNIVERSITÄT BERN

Tooth morphology



UNIVERSITÄT BERN

Tooth morphology



From [26], Fig. 2

Tooth morphology



Detection of enamel-dentin border



Detection of enamel-dentin border



Outcome root canal configuration classification

Roots		RCC	#	%
Single (N=98)		1–1–1/1	73	74.5
		1–1–1/2	14	14.3
		1–1–1/3	1	1.0
		1–1–1/4	2	2.1
		1–1–2/1	1	1.0
		1–2–1/1	4	4.1
		1–2–1/2	1	1.0
		1–2–2/2	1	1.0
		2–3–1/1	1	1.0
Double (N=3)	Buccal	1–1–1/1	2	66.6
		1–2–1/1	1	33.3
	Lingual	1–1–1/1	2	66.6
		1–1–1/2	1	33.3

Extraction of root canal space



UNIVERSITÄT BERN

Results of root canal space extraction



Conclusion ZMK

- Efficient use of time, e. g. more teeth does not mean more (human) work
- Reproducible analysis with *free and open-source* software, usable by *anyone*
- Objective analysis, e. g. no operator bias

Thanks!

- Thanks for listening to me!
- What questions do you have for me?

UNIVERSITÄT BERN

Colophon

- This BEAMER presentation was crafted in LaTeX with the (slightly adapted) template from *Corporate Design und Vorlagen* of the University of Bern.
    - Complete source code: git.io/fjpP7
    - The LaTeX code is automatically compiled with a GitHub actionto the (handout) PDF linked on ILIAS (git.io/JeQxO)
- Did you spot an error?
    - File an issue: git.io/fjpPb
    - Submit a pull request: git.io/fjpPN
    - Send me an email: david.haberthuer@unibe.ch

UNIVERSITÄT BERN

References I

[1] Ruslan Hlushchuk et al. “Ex Vivo microangioCT: Advances in Microvascular Imaging”. DOI: 10.1016/j.vph.2018.09.003.

[2] Henry Nording et al. “The C5a/C5a Receptor 1 Axis Controls Tissue Neovascularization through CXCL4 Release from Platelets”. DOI: 10.1038/s41467-021-23499-w.

[3] Ruslan Hlushchuk et al. “Innovative High-Resolution microCT Imaging of Animal Brain Vasculature”. DOI: 10.1007/s00429-020-02158-8.

[4] Tsering Wüthrich et al. “Development of Vascularized Nerve Scaffold Using Perfusion-Decellularization and Recellularization”. DOI: 10.1016/j.msec.2020.111311.

[5] Cédric Zubler et al. “The Anatomical Reliability of the Superficial Circumflex Iliac Artery Perforator (SCIP) Flap”. DOI: 10.1016/j.aanat.2020.151624.

[6] Matthias Messerli et al. “Adaptation Mechanism of the Adult Zebrafish Respiratory Organ to Endurance Training”. DOI: 10.1371/journal.pone.0228333.

[7] Verdiana Trappetti et al. “Synchrotron Microbeam Radiotherapy for the Treatment of Lung Carcinoma: A Pre-Clinical Study”. DOI: 10.1016/j.ijrobp.2021.07.1717.

References II

[8] Estée Bochud et al. “A New Diancta Species of the Family Diplommatinidae (Cyclophoroidea) from Vanua Levu Island, Fiji”. DOI: 10.3897/zookeys.1073.73241.

[9] Sebastian Halm et al. “Micro-CT Imaging of Thiel-embalmed and Iodine-Stained Human Temporal Bone for 3D Modeling”. DOI: 10.1186/s40463-021-00522-0.

[10] Yogita Kadlag et al. “Physical Properties and Average Atomic Numbers of Chondrules Using Computed Tomography”. DOI: 10.1016/j.pss.2023.105799.

[11] David Haberthür et al. “Automated Segmentation and Description of the Internal Morphology of Human Permanent Teeth by Means of Micro-CT”. DOI: gjpw2d.

[12] David Haberthür et al. “Microtomographic Investigation of a Large Corpus of Cichlids”. DOI: 10.1371/journal.pone.0291003.

[13] Kenneth Clark et al. “The Cancer Imaging Archive (TCIA): Maintaining and Operating a Public Information Repository”. DOI: 10.1007/s10278-013-9622-7.

[14] A. M. Cormack. “Representation of a Function by Its Line Integrals, with Some Radiological Applications”. http://is.gd/b0Bnr. DOI: 10.1063/1.1729798.

UNIVERSITÄT BERN

References III

[15] Godfrey Newbold Hounsfield. “Historical Notes on Computerized Axial Tomography.”.

[16] E C Beckmann. “CT Scanning the Early Days.”. DOI: 10.1259/bjr/29444122.

[17] J Hsieh. *Computed Tomography: Principles, Design, Artifacts, and Recent Advances*. Society of Photo Optical.

[18] Angela S.P. Lin et al. “Microcomputed Tomography”. *Springer Handbook of Microscopy*. Springer Handbooks. Cham: Springer International Publishing. ISBN: 978-3-030-00069-1. DOI: 10.1007/978-3-030-00069-1_24.

[19] L. A. Feldkamp et al. “Practical Cone-Beam Algorithm”. DOI: 10.1364/JOSAA.1.000612.

[20] LA Feldkamp et al. “Investigation of 3-Dimensional Structure of Trabecular Bone by Computed-Tomography of Iliac Biopsy Samples”. *Calcified Tissue International*. SPRINGER VERLAG 175 FIFTH AVE, NEW YORK, NY 10010.

[21] Mark Hammer. *X-Ray Physics: X-Ray Interaction with Matter, X-Ray Contrast, and Dose*.

[22] Wikipedia contributors. *Beer–Lambert Law — Wikipedia, The Free Encyclopedia*. [Online; accessed 6-December-2019].

[23] Johannes Schindelin et al. “Fiji: An Open-Source Platform for Biological-Image Analysis”. DOI: 10.1038/nmeth.2019.

References IV

[24] Thomas Kluyver et al. “Jupyter Notebooks – a Publishing Format for Reproducible Computational Workflows”. DOI: 10.3233/978-1-61499-649-1-87.

[25] Thomas Gerhard Wolf et al. “Internal Morphology of 101 Mandibular Canines of a Swiss-German Population by Means of Micro-CT: An Ex Vivo Study”. DOI: 10.1038/s41598-021-00758-w.

[26] Benjamín Briseño-Marroquín et al. “Root Canal Morphology and Configuration of 179 Maxillary First Molars by Means of Micro–Computed Tomography: An Ex Vivo Study”. DOI: 10.1016/j.joen.2015.09.007.

[27] David Haberthür. “Habi/Zmk-Tooth-Cohort: Used for Manuscript about Method”. DOI: 10.5281/ZENODO.3999402.

[28] Thomas Gerhard Wolf et al. “Three-Dimensional Analysis of the Physiological Foramen Geometry of Maxillary and Mandibular Molars by Means of Micro-CT”. DOI: 10.1038/ijos.2017.29.