이 력 카 드

계측기기명	도막 두께 측정기	MODEL	QNIX 9500B-FN 1
MAKER	Automation	이력카드 번호	QGH-039
구 입 일 자	2022-02-21	관리부서/담당자	품질경영팀 QM 1파트 /황동환
구 매 금 액	\ 2,250,000	SER No.	3005343
규 격	0-3mm	용 도	FAN 검사

검교정 /수리내역

교정일	검교정 /수리내역	CERT No.	차기 교정일	검교정 비용	비고
2022-02-21	정기점검	22002233-6	2023-02-21	0	
2023-02-10	정기점검	23B09001-1	2024-02-10	42,000	
2024-02-16	정기점검	24001762-6	2025-02-16	23,628	
2025-02-14	정기점검	25001572-1	2026-02-14	0	
2026-03-13	정기점검	26003582-3	2027-03-13	0	

Calibration Certificate

Korea Instruments Co., Ltd. Busan branch
45, Gwangjang-ro 56beon-gil,
Sasang-gu, Busan, Korea
Tel : 051-322-0880, Fax : 051-322-4066

Certificate No : 26003582-3

1. Client
   Name : HI AIR KOREA Co.,Ltd
   Address : 204, Gomo-ro 324beon-gil, Jillye-myeon, Gimhae-si, Gyeongsangnam-do, Korea
2. Calibration Subject
   Description : Coating thickness tester
   Manufacturer and Model Name : QNIX / 9500
   Serial Number : 3005343
3. Date of Calibration : 2026. 03. 13.
4. Environment
   - Temperature : (20.2 ± 0.2) ℃ - Humidity : (38 ± 2) % R.H.
   - Location ▣ Permanent Calibration Lab ☐ Mobile Lab ☐ On Site Calibration
   (Address : 45, Gwangjang-ro 56beon-gil, Sasang-gu, Busan, Korea)
5. Traceability
   Calibration method and/or brief description

   The above instrument is calibration obtained by using standard equipment which traceability is maintained to SI units by National Calibration Laboratory on the basis of the standard calibration procedure of Coating thickness testers (KIC-102-36).

   List of used standards/specifications

Description	Manufacturer and Model	Serial Number	The due date of next calibration	Calibration Laboratory
Coating thickness specimen	Elcometer / (11.9 ~ 14 919) μm	KIC170621	2027.07.11	KTL

6. Calibration Result : Refer to Calibration Result
7. Measurement uncertainty : Refer to Calibration Result

affirmation	Measurements performed by	Approved by
		Title : Technical Cal. Manager
	Name : Jeong-Ho Kim 김정호	Name : Sang-Hyun Park

The above calibration certificate is the accredited calibration items by Korea Laboratory Accreditation Scheme, which signed the ILAC-MRA.

2026. 03. 13

Korea Instruments Co., Ltd. Busan branch Calibration Laboratory
Accredited by KOLAS, Republic of KOREA

(NOTE) If any significant instability or other adverse factor(overload, temperature, humidity etc.) manifests itself before, during or after calibration, and is likely to affect the validity of the calibration.

Calibration Results

Korea Instruments Co., Ltd. Busan branch
45, Gwangjang-ro 56beon-gil, Sasang-gu, Busan, Korea
Tel : 051-322-0880, Fax : 051-322-4066

Certificate No : 26003582-3

* Description : Coating thickness tester
* Manufacturer : QNIX
* Serial Number : 3005343

1. Instruction accuracy calibration

Nominal value (μm)	Indication value (μm)	Correction value (μm)	Measurement uncertainty(μm) (Confidence level is approximately 95 %, k = 2)
13.0	12.9	0.1	2.8
23.8	23.7	0.1	2.8
51.1	49.4	1.7	2.8
125.3	127	-1	2.9
250.3	248	2	2.9
509.5	499	10	2.9
1 028.5	1 028	-	2.9

◇ The above calibration results are the values calibrated by adjusting the gain to the reference specimen 1028.5 μm for thickness measurement. End.

COPY 복사본