이 력 카 드

계측기기명	버니어 캘리퍼스	MODEL	-
MAKER	MITUTOYO	이력카드 번호	QG J-002
구 입 일 자	2001-03-15	관리부서 /담당자	FAN 생산팀 FAN 2파트 /홍성범
구 매 금 액	\ 90,000	SER No.	1019261
규 격	300mm	용 도	두께 ,내경 ,외경 측정

검 교정 /수리내역

교정일	검 교정 /수리내역	CERT No.	차기 교정일	검교정 비용	비고
2015-07-24	정기점검	15004317-3	2016-07-24	0	
2016-08-26	정기점검	16007569-12	2017-08-26	0	
2017-11-06	정기점검	BCT 173810-8509	2018-11-06	0	
2019-05-13	정기점검	BCT 190494-1688	2020-05-13	0	
2020-07-20	정기점검	BCT 203605-8029	2021-07-20	0	
2021-08-24	정기점검	BCT 213929-9724	2022-08-24	23,000	
2022-09-13	정기점검	22108003-2	2023-09-13	23,000	
2023-12-12	정기점검	23 11003-4	2024-12-12	23,000	
2025-05-26	정기점검	25007055-1	2026-05-26	0	
2026-03-12	정기점검	26003582-1	2027-03-12	0	

Calibration Certificate

Korea Instruments Co., Ltd. Busan branch
45, Gwangjang-ro 56beon-gil,
Sasang-gu, Busan, Korea
Tel : 051-322-0880, Fax : 051-322-4066

Certificate No : 26003582-1

1. Client
   Name : HI AIR KOREA Co.,Ltd
   Address : 204, Gomo-ro 324beon-gil, Jillye-myeon, Gimhae-si, Gyeongsangnam-do, Korea
2. Calibration Subject
   Description : Calipers,inside/outside
   Manufacturer and Model Name : Mitutoyo / (0 ~ 300) mm, 0.05 mm
   Serial Number : 1019261
3. Date of Calibration : 2026. 03. 12.
4. Environment
   - Temperature : (20.2 ± 0.2) ℃ - Humidity : (38 ± 2) % R.H.
   - Location ▣ Permanent Calibration Lab ☐ Mobile Lab ☐ On Site Calibration
   (Address : 45, Gwangjang-ro 56beon-gil, Sasang-gu, Busan, Korea)
5. Traceability
   Calibration method and/or brief description

The above instrument is calibration obtained by using standard equipment which traceability is maintained to SI units by National Calibration Laboratory on the basis of the standard calibration procedure of Inside/Outside calipers (KIC-106-01-1).

List of used standards/specifications

Description	Manufacturer and Model	Serial Number	The due date of next calibration	Calibration Laboratory
Caliper testers	Mitutoyo / 600 mm	2310104	2027.05.02	Korea Instrument Co., Ltd. Branch

6. Calibration Result : Refer to Calibration Result
7. Measurement uncertainty : Refer to Calibration Result

affirmation	Measurements performed by	Approved by
		Title : Technical Cal. Manager
	Name : Seong-Hwan Kim 김성환	Name : Sang-Hyun Park

The above calibration certificate is the accredited calibration items by Korea Laboratory Accreditation Scheme, which signed the ILAC-MRA.

2026. 03. 12

Korea Instruments Co., Ltd. Busan branch Calibration Laboratory
Accredited by KOLAS, Republic of KOREA

(NOTE) If any significant instability or other adverse factor(overload, temperature, humidity etc.) manifests itself before, during or after calibration, and is likely to affect the validity of the calibration.

Calibration Results

Korea Instruments Co., Ltd. Busan branch
45, Gwangjang-ro 56beon-gil, Sasang-gu, Busan, Korea
Tel : 051-322-0880, Fax : 051-322-4066

Certificate No : 26003582-1

* Description : Calipers,inside/outside
* Manufacturer : Mitutoyo
* Serial Number : 1019261

1. Accuration calibration of graduations

Nominal value (mm)	Correction value (mm)		Measurement uncertainty (μm) (Confidence level is approximately 95 %, k = 2)
	Outside	Inside	
0	0.00	-	-
50	0.00	0.00	41
100	0.00	0.00	41
150	0.00	0.00	41
200	0.00	0.00	41
300	0.00	0.00	41

End.

COPY 복사본