



ValProbe RT Qualification Summary Report

Printed on 23-Mar-2026 at 10:52:45 by wusitan

Study Name: 26-Feb-2026 15_47_48	Company: Hisun	ValProbe RT Version: 1.3.0.35
SOP / Protocol #: H3-VD-26684	Run#: 2	Common Reporting Tool Version: 1.4.0.8
Asset Name: Y10sterilizer-2	Asset ID: 152903-109-0008	BS Firmware Version: 0.11.34
Started By: wusitan	Date: 23-Mar-2026 10:20:47 (UTC+08:00)	Stopped By: wusitan
Setup Created By: qinpengyun	Setup Created Date: 26-Feb-2026 15:56:56 (UTC+08:00)	Study File Comments: yt20max
Setup Comments:	Setup Modified Date: 23-Mar-2026 10:20:42	Setup Description:

Comments

Calculation Criteria

Lethality has been selected with the following criteria : Base Temperature : 121.1 °C D Value : 1.00 Z Value : 10.0 TimeScale : Minutes
Lethality Calculated: During Exposure Cycle

Calculations in the report is based on 30 Seconds sampling rate.

Groups included in report

Temp - Include sensor inputs, statistical calculations, Lethality

Base Station Details

Name: VPRT_Basestation	Location: Anywhere	Hardware Version: 2.4.0	IP Address: 192.168.99.50
Serial No: T220806	Model No: VPRT_BS_1	Firmware Version: 0.11.34	

Performed by: _____

Date: _____

Reviewed by: _____

Date: _____

Study Name: 26-Feb-2026 15_47_48

Loggers Included in Study

Serial No	Logger Type	Model No	MFG Cal Date	FW Version	Battery Life (%) at Start
MDV3	5 Ch Temp-Flexible	XC5FVP40-RT	30-Sep-2024	2.3.5	92
NBS5	5 Ch Temp-Flexible	XC5FVP40-RT	17-Dec-2024	2.3.5	90
NBT2	5 Ch Temp-Flexible	XC5FVP40-RT	17-Dec-2024	2.3.5	91

Sensor/Group Definitions

Group Name	Sensor Label	Sensor Type	Sensor SN-Probe ID	Description
Temp	tc1	Temperature	MDV3-A	
	tc2	Temperature	MDV3-B	
	tc3	Temperature	MDV3-C	
	tc4	Temperature	MDV3-D	
	tc5	Temperature	MDV3-E	
	tc6	Temperature	NBS5-A	
	tc7	Temperature	NBS5-B	
	tc8	Temperature	NBS5-C	
	tc9	Temperature	NBS5-D	
	tc10	Temperature	NBS5-E	
	tc11	Temperature	NBT2-A	
	tc12	Temperature	NBT2-B	

Study Name: 26-Feb-2026 15_47_48

Temp															
Asset ID: 152903-109-0008								SOP / Protocol #: H3-VD-26684							
Comments:								Description:							

Temperature Data (°C)																Totals	
		1					Exposure					3					Totals
Sensor Label	S/N	Min	Max	Avg	Max-Min	Cycle ALeth	Min	Max	Avg	Max-Min	Cycle ALeth	Min	Max	Avg	Max-Min	Cycle ALeth	Totals ALeth
tc1	MDV3-A	75.39	122.03	105.72	46.64	0.00	121.67	122.12	122.01	0.45	18.43	120.49	121.09	120.79	0.60	0.00	18.43
tc2	MDV3-B	81.87	121.95	106.40	40.08	0.00	121.73	122.09	122.01	0.36	18.44	120.54	121.12	120.83	0.58	0.00	18.44
tc3	MDV3-C	82.00	122.06	106.45	40.06	0.00	121.71	122.12	122.03	0.41	18.51	120.53	121.12	120.82	0.59	0.00	18.51
tc4	MDV3-D	82.66	122.05	106.50	39.39	0.00	121.70	122.08	122.00	0.38	18.39	120.51	121.11	120.81	0.60	0.00	18.39
tc5	MDV3-E	78.15	122.04	106.22	43.89	0.00	121.67	122.12	122.01	0.45	18.45	120.49	121.07	120.78	0.58	0.00	18.45
tc6	NBS5-A	81.75	122.01	106.28	40.26	0.00	121.68	122.01	121.88	0.33	17.93	120.47	121.07	120.77	0.60	0.00	17.93
tc7	NBS5-B	81.68	122.00	106.29	40.32	0.00	121.68	122.03	121.91	0.35	18.03	120.48	121.08	120.78	0.60	0.00	18.03
tc8	NBS5-C	80.29	121.86	106.16	41.57	0.00	121.60	122.00	121.88	0.40	17.93	120.50	121.09	120.79	0.59	0.00	17.93
tc9	NBS5-D	79.27	122.03	106.05	42.76	0.00	121.73	121.99	121.91	0.26	18.02	120.53	121.12	120.82	0.59	0.00	18.02
tc10	NBS5-E	79.80	121.96	106.16	42.16	0.00	121.69	122.01	121.88	0.32	17.94	120.50	121.08	120.79	0.58	0.00	17.94
tc11	NBT2-A	80.71	121.95	106.17	41.24	0.00	121.71	122.05	121.89	0.34	8.99					0.00	8.99
tc12	NBT2-B	76.41	122.06	105.88	45.65	0.00	121.76	122.07	121.92	0.31	9.06					0.00	9.06

Study Name: 26-Feb-2026 15_47_48

Temp	
Asset ID: 152903-109-0008	SOP / Protocol #: H3-VD-26684
Comments:	Description:

Temperature Summary Data (°C)				Totals
	1	Exposure	3	
Cycle Start	23-Mar-2026 10:20:47	23-Mar-2026 10:32:30	23-Mar-2026 10:47:30	Study Start
Cycle Duration	00:11:43	00:15:00	00:01:00	23-Mar-2026 10:20:47
Min of Min	75.39 Label tc1	121.60 Label tc8	120.47 Label tc6	Study End
Time	23-Mar-2026 10:20:50	23-Mar-2026 10:32:30	23-Mar-2026 10:48:00	23-Mar-2026 10:48:30
Max of Max	122.06 Label tc3	122.12 Label tc1	121.12 Label tc2	Duration: 00:27:43
Time	23-Mar-2026 10:32:00	23-Mar-2026 10:35:30	23-Mar-2026 10:47:30	Min ALeth 8.99
Max Range	46.67	0.52	0.65	Label tc11
Max Spread/Time	7.50 Time 23-Mar-2026 10:21:00	0.38 Time 23-Mar-2026 10:32:30	0.07 Time 23-Mar-2026 10:48:00	Max ALeth 18.51
Min ALeth	0.00 Label tc1	8.99 Label tc11	0.00 Label tc1	Label tc3
Max ALeth	0.00 Label tc1	18.51 Label tc3	0.00 Label tc1	Study Min 75.39
Avg of Avg	106.19	121.94	120.80	Label tc1
Max(Max-Min)	46.64 Label tc1	0.45 Label tc1	0.60 Label tc1	Study Max 122.12
				Label tc1

Study Name: 26-Feb-2026 15_47_48

Comments

System Messages

Time Stamp	Event Description	Event Type	User Name	Console ID
23-Mar-2026 10:20:46	NBT2_Start	Manual		
23-Mar-2026 10:20:47	***Study Start***	Manual	wusitan	6364071B54B2039A7382
23-Mar-2026 10:20:48	NBS5_Start	Manual		
23-Mar-2026 10:20:49	MDV3_Start	Manual		
23-Mar-2026 10:32:30	***Exposure Start***			6364071B54B2039A7382
23-Mar-2026 10:47:30	***Exposure End***			6364071B54B2039A7382
23-Mar-2026 10:48:00	MDV3_End	Manual		
23-Mar-2026 10:48:00	NBS5_End	Manual		
23-Mar-2026 10:48:30	NBT2_End	Manual		
23-Mar-2026 10:48:30	***Study End***	Manual	wusitan	6364071B54B2039A7382

Study Name: 26-Feb-2026 15_47_48

Cycles Selection Graph



Performed by: _____

Date: _____

Reviewed by: _____

Date: _____

Study Name: 26-Feb-2026 15_47_48

