



ValProbe RT Qualification Detailed Report

Printed on 23-Mar-2026 at 10:52:46 by wusitan

Study Name: 26-Feb-2026 15_47_48	Company: Hisun	ValProbe RT Version: 1.3.0.35
SOP / Protocol #: H3-VD-26684	Run#: 2	Common Reporting Tool Version: 1.4.0.8
Asset Name: Y10sterilizer-2	Asset ID: 152903-109-0008	BS Firmware Version: 0.11.34
Started By: wusitan	Date: 23-Mar-2026 10:20:47 (UTC+08:00)	Stopped By: wusitan
Setup Created By: qinpengyun	Setup Created Date: 26-Feb-2026 15:56:56 (UTC+08:00)	Study File Comments: yt20max
Setup Comments:	Setup Modified Date: 23-Mar-2026 10:20:42	Setup Description:

Comments

Calculation Criteria

Lethality has been selected with the following criteria : Base Temperature : 121.1 °C D Value : 1.00 Z Value : 10.0 TimeScale : Minutes
Lethality Calculated: During Exposure Cycle

Calculations in the report is based on 30 Seconds sampling rate. Report content is based on 30 Seconds printing rate.

Groups included in report

Temp - Include sensor inputs, statistical calculations, Lethality

Base Station Details

Name: VPRT_Basestation	Location: Anywhere	Hardware Version: 2.4.0	IP Address: 192.168.99.50
Serial No: T220806	Model No: VPRT_BS_1	Firmware Version: 0.11.34	

Performed by: _____

Date: _____

Reviewed by: _____

Date: _____

Study Name: 26-Feb-2026 15_47_48

Loggers Included in Study

Serial No	Logger Type	Model No	MFG Cal Date	FW Version	Battery Life (%) at Start
MDV3	5 Ch Temp-Flexible	XC5FVP40-RT	30-Sep-2024	2.3.5	92
NBS5	5 Ch Temp-Flexible	XC5FVP40-RT	17-Dec-2024	2.3.5	90
NBT2	5 Ch Temp-Flexible	XC5FVP40-RT	17-Dec-2024	2.3.5	91

Sensor/Group Definitions

Group Name	Sensor Label	Sensor Type	Sensor SN-Probe ID	Description
Temp	tc1	Temperature	MDV3-A	
	tc2	Temperature	MDV3-B	
	tc3	Temperature	MDV3-C	
	tc4	Temperature	MDV3-D	
	tc5	Temperature	MDV3-E	
	tc6	Temperature	NBS5-A	
	tc7	Temperature	NBS5-B	
	tc8	Temperature	NBS5-C	
	tc9	Temperature	NBS5-D	
	tc10	Temperature	NBS5-E	
	tc11	Temperature	NBT2-A	
	tc12	Temperature	NBT2-B	

Study Name: 26-Feb-2026 15_47_48

Temp												
Asset ID: 152903-109-0008						SOP / Protocol #: H3-VD-26684						
Comments:						Description:						

Temperature Data(°C)												
Time Stamp	tc1 MDV3-A	tc2 MDV3-B	tc3 MDV3-C	tc4 MDV3-D	tc5 MDV3-E	tc6 NBS5-A	tc7 NBS5-B	tc8 NBS5-C	tc9 NBS5-D	tc10 NBS5-E	tc11 NBT2-A	tc12 NBT2-B
23-Mar-2026												
10:20:47	***** ***** Study Start ***** *****											
10:20:47	*****1*****											
10:20:47											80.71	76.41
10:20:49						81.75	81.68	80.29	79.27	79.80		
10:20:50	75.39	81.87	82.00	82.66	78.15							
10:21:00	80.78	87.71	88.03	87.52	86.79	88.28	88.24	87.75	86.48	87.04	87.48	84.98
10:21:30	92.36	96.32	96.28	95.19	96.28	96.21	96.14	96.23	96.07	95.90	95.82	94.37
10:22:00	98.74	100.23	100.15	99.98	100.12	100.05	100.03	100.26	100.21	100.05	100.33	99.42
10:22:30	102.71	102.85	102.70	103.54	102.66	102.69	102.67	103.09	103.08	102.87	102.80	102.21
10:23:00	105.73	104.27	104.84	105.82	104.57	103.91	104.15	103.89	103.97	104.06	103.92	104.27
10:23:30	107.17	105.88	106.42	107.08	106.32	105.62	105.80	105.48	105.41	105.67	105.59	105.93
10:24:00	107.94	107.80	107.50	107.63	107.45	107.20	107.34	107.07	107.06	107.57	107.08	107.36
10:24:30	108.40	108.22	108.19	108.28	108.18	108.14	108.15	108.08	108.14	108.15	108.04	108.28
10:25:00	108.48	108.46	108.43	108.45	108.39	108.35	108.36	108.24	108.41	108.37	108.35	108.48
10:25:30	108.32	108.34	108.32	108.33	108.28	108.25	108.28	108.12	108.28	108.25	108.23	108.32
10:26:00	107.99	108.08	108.05	108.05	108.02	107.95	108.01	107.85	108.02	107.98	108.00	108.03
10:26:30	107.71	107.80	107.79	107.76	107.75	107.67	107.72	107.58	107.74	107.71	107.73	107.75
10:27:00	107.53	107.61	107.59	107.58	107.57	107.49	107.53	107.39	107.55	107.52	107.55	107.56
10:27:30	107.41	107.48	107.47	107.46	107.45	107.37	107.41	107.28	107.42	107.39	107.41	107.44
10:28:00	107.33	107.41	107.40	107.38	107.38	107.28	107.32	107.21	107.33	107.30	107.34	107.35
10:28:30	107.24	107.32	107.31	107.29	107.28	107.21	107.25	107.14	107.25	107.23	107.27	107.27
10:29:00	107.24	107.29	107.29	107.28	107.27	107.18	107.21	107.12	107.20	107.20	107.22	107.26
10:29:30	107.36	107.41	107.41	107.39	107.39	107.31	107.34	107.27	107.32	107.31	107.34	107.38
10:30:00	107.45	107.50	107.50	107.48	107.48	107.39	107.43	107.36	107.42	107.42	107.44	107.46
10:30:30	112.42	112.34	112.40	112.29	112.53	112.16	111.92	112.07	111.42	112.00	111.82	112.10
10:31:00	117.78	117.74	117.79	117.69	117.91	117.56	117.34	117.45	116.92	117.45	117.24	117.52

Study Name: 26-Feb-2026 15_47_48

Temp												
Asset ID: 152903-109-0008						SOP / Protocol #: H3-VD-26684						
Comments:						Description:						

Temperature Data(°C)												
Time Stamp	tc1 MDV3-A	tc2 MDV3-B	tc3 MDV3-C	tc4 MDV3-D	tc5 MDV3-E	tc6 NBS5-A	tc7 NBS5-B	tc8 NBS5-C	tc9 NBS5-D	tc10 NBS5-E	tc11 NBT2-A	tc12 NBT2-B
23-Mar-2026												
10:31:30	121.86	121.77	121.84	121.80	121.91	121.79	121.62	121.65	121.30	121.68	121.51	121.79
10:32:00	122.03	121.95	122.06	122.05	122.04	122.01	122.00	121.86	122.03	121.96	121.95	122.06
10:32:30	*****Exposure Start*****											
10:32:30	*****Exposure*****											
10:32:30	121.91	121.86	121.98	121.93	121.95	121.71	121.77	121.60	121.83	121.70	121.71	121.76
10:33:00	122.06	122.00	122.10	122.03	122.04	121.98	121.96	121.88	121.92	121.94	121.90	122.00
10:33:30	121.89	121.92	121.95	121.92	121.88	121.93	121.95	121.79	121.93	121.88	121.95	121.99
10:34:00	121.83	121.90	121.91	121.89	121.87	121.82	121.86	121.70	121.88	121.76	121.87	121.87
10:34:30	121.85	121.93	121.93	121.91	121.90	121.80	121.84	121.72	121.87	121.76	121.87	121.85
10:35:00	122.07	122.05	122.10	122.03	122.07	121.86	121.89	121.85	121.89	121.87	121.83	121.90
10:35:30	122.12	122.09	122.12	122.07	122.11	121.90	121.93	121.90	121.91	121.91	121.88	121.97
10:36:00	122.10	122.08	122.10	122.06	122.09	121.95	121.97	121.95	121.93	121.95	121.92	121.99
10:36:30	122.10	122.07	122.11	122.06	122.10	121.88	121.92	121.91	121.90	121.90	121.89	121.92
10:37:00	122.11	122.07	122.10	122.06	122.10	121.92	121.94	121.93	121.91	121.93	121.91	121.97
10:37:30	122.11	122.08	122.11	122.07	122.11	121.90	121.93	121.93	121.91	121.92	121.88	121.93
10:38:00	122.08	122.06	122.09	122.05	122.09	121.85	121.89	121.90	121.89	121.87	121.86	121.87
10:38:30	122.05	122.05	122.07	122.03	122.08	121.82	121.85	121.85	121.88	121.82	121.86	121.83
10:39:00	121.97	121.99	122.00	121.99	122.02	121.81	121.83	121.84	121.87	121.81	121.87	121.84
10:39:30	121.88	121.93	121.91	121.90	121.87	121.88	121.90	121.90	121.92	121.87	121.96	121.93
10:40:00	122.00	122.02	122.01	121.99	121.97	121.97	122.01	121.98	121.98	121.97	122.05	122.07
10:40:30	122.10	122.08	122.10	122.06	122.08	122.01	122.02	122.00	121.96	122.01	INVALID	INVALID
10:41:00	122.11	122.07	122.11	122.06	122.10	121.91	121.94	121.94	121.92	121.92	INVALID	INVALID
10:41:30	122.05	122.05	122.07	122.04	122.08	121.83	121.85	121.86	121.87	121.84	INVALID	INVALID
10:42:00	121.93	121.97	121.97	121.96	121.99	121.84	121.85	121.86	121.88	121.83	INVALID	INVALID
10:42:30	121.92	121.95	121.93	121.93	121.88	121.91	121.95	121.94	121.95	121.92	INVALID	INVALID
10:43:00	122.05	122.05	122.04	122.03	122.02	121.99	122.03	122.00	121.99	122.00	INVALID	INVALID

Study Name: 26-Feb-2026 15_47_48

Temp												
Asset ID: 152903-109-0008						SOP / Protocol #: H3-VD-26684						
Comments:						Description:						

Temperature Data(°C)												
Time Stamp	tc1 MDV3-A	tc2 MDV3-B	tc3 MDV3-C	tc4 MDV3-D	tc5 MDV3-E	tc6 NBS5-A	tc7 NBS5-B	tc8 NBS5-C	tc9 NBS5-D	tc10 NBS5-E	tc11 NBT2-A	tc12 NBT2-B
23-Mar-2026												
10:43:30	122.11	122.08	122.09	122.06	122.10	121.98	122.00	121.99	121.96	121.99	INVALID	INVALID
10:44:00	122.07	122.06	122.07	122.05	122.10	121.84	121.87	121.89	121.89	121.87	INVALID	INVALID
10:44:30	121.95	121.97	121.98	121.98	121.99	121.83	121.85	121.85	121.88	121.84	INVALID	INVALID
10:45:00	121.92	121.95	121.94	121.93	121.89	121.92	121.94	121.94	121.95	121.93	INVALID	INVALID
10:45:30	122.07	122.06	122.07	122.05	122.05	122.01	122.03	122.00	121.99	122.01	INVALID	INVALID
10:46:00	122.12	122.09	122.10	122.08	122.12	121.98	121.99	121.98	121.95	121.99	INVALID	INVALID
10:46:30	121.97	122.00	122.01	122.00	122.03	121.83	121.84	121.85	121.88	121.84	INVALID	INVALID
10:47:00	121.67	121.73	121.71	121.70	121.67	121.68	121.68	121.69	121.73	121.69	INVALID	INVALID
10:47:30	*****Exposure End*****											
10:47:30	*****3*****											
10:47:30	121.09	121.12	121.12	121.11	121.07	121.07	121.08	121.09	121.12	121.08	INVALID	INVALID
10:48:00	120.49	120.54	120.53	120.51	120.49	120.47	120.48	120.50	120.53	120.50	INVALID	INVALID
10:48:30											INVALID	INVALID
10:48:30	***** ***** Study End ***** *****											

Study Name: 26-Feb-2026 15_47_48

Temp												
AssetID: 152903-109-0008						SOP / Protocol #: H3-VD-26684						
Comments:						Description:						

Temperature Data(°C) - Statistical Calculations

Time Stamp	MIN			MAX			Max-Min	Avg	Std Dev	Max-Avg	Avg-Min	CycleTime
	Value	Label	S/N	Value	Label	S/N						
23-Mar-2026												
10:20:47	***** ***** Study Start ***** *****											
10:20:47	*****1*****											
10:20:47	76.41	tc12	NBT2-B	80.71	tc11	NBT2-A	4.30	78.56	3.04	2.15	2.15	00:00:00
10:20:49	79.27	tc9	NBS5-D	81.75	tc6	NBS5-A	2.48	80.56	1.12	1.19	1.29	00:00:02
10:20:50	75.39	tc1	MDV3-A	82.66	tc4	MDV3-D	7.27	80.01	3.13	2.65	4.62	00:00:03
10:21:00	80.78	tc1	MDV3-A	88.28	tc6	NBS5-A	7.50	86.76	2.09	1.52	5.98	00:00:13
10:21:30	92.36	tc1	MDV3-A	96.32	tc2	MDV3-B	3.96	95.60	1.17	0.72	3.24	00:00:43
10:22:00	98.74	tc1	MDV3-A	100.33	tc11	NBT2-A	1.59	99.96	0.45	0.37	1.22	00:01:13
10:22:30	102.21	tc12	NBT2-B	103.54	tc4	MDV3-D	1.33	102.82	0.32	0.72	0.61	00:01:43
10:23:00	103.89	tc8	NBS5-C	105.82	tc4	MDV3-D	1.93	104.45	0.68	1.37	0.56	00:02:13
10:23:30	105.41	tc9	NBS5-D	107.17	tc1	MDV3-A	1.76	106.03	0.60	1.14	0.62	00:02:43
10:24:00	107.06	tc9	NBS5-D	107.94	tc1	MDV3-A	0.88	107.42	0.29	0.52	0.36	00:03:13
10:24:30	108.04	tc11	NBT2-A	108.40	tc1	MDV3-A	0.36	108.19	0.10	0.21	0.15	00:03:43
10:25:00	108.24	tc8	NBS5-C	108.48	tc1	MDV3-A	0.24	108.40	0.07	0.08	0.16	00:04:13
10:25:30	108.12	tc8	NBS5-C	108.34	tc2	MDV3-B	0.22	108.28	0.06	0.06	0.16	00:04:43
10:26:00	107.85	tc8	NBS5-C	108.08	tc2	MDV3-B	0.23	108.00	0.06	0.08	0.15	00:05:13
10:26:30	107.58	tc8	NBS5-C	107.80	tc2	MDV3-B	0.22	107.73	0.06	0.07	0.15	00:05:43
10:27:00	107.39	tc8	NBS5-C	107.61	tc2	MDV3-B	0.22	107.54	0.06	0.07	0.15	00:06:13
10:27:30	107.28	tc8	NBS5-C	107.48	tc2	MDV3-B	0.20	107.42	0.05	0.06	0.14	00:06:43
10:28:00	107.21	tc8	NBS5-C	107.41	tc2	MDV3-B	0.20	107.34	0.06	0.07	0.13	00:07:13
10:28:30	107.14	tc8	NBS5-C	107.32	tc2	MDV3-B	0.18	107.25	0.05	0.07	0.11	00:07:43
10:29:00	107.12	tc8	NBS5-C	107.29	tc2	MDV3-B	0.17	107.23	0.05	0.06	0.11	00:08:13
10:29:30	107.27	tc8	NBS5-C	107.41	tc2	MDV3-B	0.14	107.35	0.04	0.06	0.08	00:08:43
10:30:00	107.36	tc8	NBS5-C	107.50	tc2	MDV3-B	0.14	107.44	0.04	0.06	0.08	00:09:13
10:30:30	111.42	tc9	NBS5-D	112.53	tc5	MDV3-E	1.11	112.12	0.31	0.41	0.70	00:09:43

Study Name: 26-Feb-2026 15_47_48

Temp												
AssetID: 152903-109-0008						SOP / Protocol #: H3-VD-26684						
Comments:						Description:						

Temperature Data(°C) - Statistical Calculations

Time Stamp	MIN			MAX			Max-Min	Avg	Std Dev	Max-Avg	Avg-Min	CycleTime
	Value	Label	S/N	Value	Label	S/N						
23-Mar-2026												
10:31:00	116.92	tc9	NBS5-D	117.91	tc5	MDV3-E	0.99	117.53	0.28	0.38	0.61	00:10:13
10:31:30	121.30	tc9	NBS5-D	121.91	tc5	MDV3-E	0.61	121.71	0.17	0.20	0.41	00:10:43
10:32:00	121.86	tc8	NBS5-C	122.06	tc3	MDV3-C	0.20	122.00	0.06	0.06	0.14	00:11:13
10:32:30	*****Exposure Start*****											
10:32:30	*****Exposure*****											
10:32:30	121.60	tc8	NBS5-C	121.98	tc3	MDV3-C	0.38	121.81	0.12	0.17	0.21	00:00:00
10:33:00	121.88	tc8	NBS5-C	122.10	tc3	MDV3-C	0.22	121.98	0.07	0.12	0.10	00:00:30
10:33:30	121.79	tc8	NBS5-C	121.99	tc12	NBT2-B	0.20	121.92	0.05	0.07	0.13	00:01:00
10:34:00	121.70	tc8	NBS5-C	121.91	tc3	MDV3-C	0.21	121.85	0.06	0.06	0.15	00:01:30
10:34:30	121.72	tc8	NBS5-C	121.93	tc2	MDV3-B	0.21	121.85	0.07	0.08	0.13	00:02:00
10:35:00	121.83	tc11	NBT2-A	122.10	tc3	MDV3-C	0.27	121.95	0.10	0.15	0.12	00:02:30
10:35:30	121.88	tc11	NBT2-A	122.12	tc1	MDV3-A	0.24	121.99	0.10	0.13	0.11	00:03:00
10:36:00	121.92	tc11	NBT2-A	122.10	tc1	MDV3-A	0.18	122.01	0.07	0.09	0.09	00:03:30
10:36:30	121.88	tc6	NBS5-A	122.11	tc3	MDV3-C	0.23	121.98	0.10	0.13	0.10	00:04:00
10:37:00	121.91	tc9	NBS5-D	122.11	tc1	MDV3-A	0.20	122.00	0.08	0.11	0.09	00:04:30
10:37:30	121.88	tc11	NBT2-A	122.11	tc1	MDV3-A	0.23	121.99	0.10	0.12	0.11	00:05:00
10:38:00	121.85	tc6	NBS5-A	122.09	tc3	MDV3-C	0.24	121.96	0.10	0.13	0.11	00:05:30
10:38:30	121.82	tc6	NBS5-A	122.08	tc5	MDV3-E	0.26	121.93	0.11	0.15	0.11	00:06:00
10:39:00	121.81	tc6	NBS5-A	122.02	tc5	MDV3-E	0.21	121.90	0.08	0.12	0.09	00:06:30
10:39:30	121.87	tc5	MDV3-E	121.96	tc11	NBT2-A	0.09	121.90	0.03	0.06	0.03	00:07:00
10:40:00	121.97	tc5	MDV3-E	122.07	tc12	NBT2-B	0.10	122.00	0.03	0.07	0.03	00:07:30
10:40:30	121.96	tc9	NBS5-D	122.10	tc1	MDV3-A	0.14	122.04	0.05	0.06	0.08	00:08:00
10:41:00	121.91	tc6	NBS5-A	122.11	tc1	MDV3-A	0.20	122.01	0.09	0.10	0.10	00:08:30
10:41:30	121.83	tc6	NBS5-A	122.08	tc5	MDV3-E	0.25	121.95	0.11	0.13	0.12	00:09:00
10:42:00	121.83	tc10	NBS5-E	121.99	tc5	MDV3-E	0.16	121.91	0.06	0.08	0.08	00:09:30

Study Name: 26-Feb-2026 15_47_48

Temp												
AssetID: 152903-109-0008						SOP / Protocol #: H3-VD-26684						
Comments:						Description:						

Temperature Data(°C) - Statistical Calculations

Time Stamp	MIN			MAX			Max-Min	Avg	Std Dev	Max-Avg	Avg-Min	CycleTime
	Value	Label	S/N	Value	Label	S/N						
23-Mar-2026												
10:42:30	121.88	tc5	MDV3-E	121.95	tc2	MDV3-B	0.07	121.93	0.02	0.02	0.05	00:10:00
10:43:00	121.99	tc6	NBS5-A	122.05	tc1	MDV3-A	0.06	122.02	0.02	0.03	0.03	00:10:30
10:43:30	121.96	tc9	NBS5-D	122.11	tc1	MDV3-A	0.15	122.04	0.06	0.07	0.08	00:11:00
10:44:00	121.84	tc6	NBS5-A	122.10	tc5	MDV3-E	0.26	121.97	0.11	0.13	0.13	00:11:30
10:44:30	121.83	tc6	NBS5-A	121.99	tc5	MDV3-E	0.16	121.91	0.07	0.08	0.08	00:12:00
10:45:00	121.89	tc5	MDV3-E	121.95	tc2	MDV3-B	0.06	121.93	0.02	0.02	0.04	00:12:30
10:45:30	121.99	tc9	NBS5-D	122.07	tc1	MDV3-A	0.08	122.03	0.03	0.04	0.04	00:13:00
10:46:00	121.95	tc9	NBS5-D	122.12	tc1	MDV3-A	0.17	122.04	0.07	0.08	0.09	00:13:30
10:46:30	121.83	tc6	NBS5-A	122.03	tc5	MDV3-E	0.20	121.93	0.08	0.10	0.10	00:14:00
10:47:00	121.67	tc1	MDV3-A	121.73	tc2	MDV3-B	0.06	121.69	0.02	0.04	0.02	00:14:30
10:47:30	*****Exposure End*****											
10:47:30	*****3*****											
10:47:30	121.07	tc5	MDV3-E	121.12	tc2	MDV3-B	0.05	121.10	0.02	0.02	0.03	00:00:00
10:48:00	120.47	tc6	NBS5-A	120.54	tc2	MDV3-B	0.07	120.50	0.02	0.04	0.03	00:00:30
10:48:30												00:01:00
10:48:30	***** ***** Study End ***** *****											

Study Name: 26-Feb-2026 15_47_48

Temp																
Asset ID: 152903-109-0008								SOP / Protocol #: H3-VD-26684								
Comments:								Description:								

Lethality Data

Time Stamp	tc1 MDV3-A	tc2 MDV3-B	tc3 MDV3-C	tc4 MDV3-D	tc5 MDV3-E	tc6 NBS5-A	tc7 NBS5-B	tc8 NBS5-C	tc9 NBS5-D	tc10 NBS5-E	tc11 NBT2-A	tc12 NBT2-B	Min	S/N Min	Max	S/N Max	CycleTime
23-Mar-2026																	
10:20:47	***** Study Start *****																
10:20:47	*****1*****																
10:20:47	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	tc1	0.00	tc1	00:00:00
10:20:49	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	tc1	0.00	tc1	00:00:02
10:20:50	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	tc1	0.00	tc1	00:00:03
10:21:00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	tc1	0.00	tc1	00:00:13
10:21:30	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	tc1	0.00	tc1	00:00:43
10:22:00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	tc1	0.00	tc1	00:01:13
10:22:30	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	tc1	0.00	tc1	00:01:43
10:23:00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	tc1	0.00	tc1	00:02:13
10:23:30	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	tc1	0.00	tc1	00:02:43
10:24:00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	tc1	0.00	tc1	00:03:13
10:24:30	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	tc1	0.00	tc1	00:03:43
10:25:00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	tc1	0.00	tc1	00:04:13
10:25:30	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	tc1	0.00	tc1	00:04:43
10:26:00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	tc1	0.00	tc1	00:05:13
10:26:30	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	tc1	0.00	tc1	00:05:43
10:27:00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	tc1	0.00	tc1	00:06:13
10:27:30	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	tc1	0.00	tc1	00:06:43
10:28:00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	tc1	0.00	tc1	00:07:13
10:28:30	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	tc1	0.00	tc1	00:07:43
10:29:00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	tc1	0.00	tc1	00:08:13
10:29:30	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	tc1	0.00	tc1	00:08:43
10:30:00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	tc1	0.00	tc1	00:09:13
10:30:30	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	tc1	0.00	tc1	00:09:43
10:31:00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	tc1	0.00	tc1	00:10:13

Study Name: 26-Feb-2026 15_47_48

Temp																
Asset ID: 152903-109-0008								SOP / Protocol #: H3-VD-26684								
Comments:								Description:								

Lethality Data																	
Time Stamp	tc1 MDV3-A	tc2 MDV3-B	tc3 MDV3-C	tc4 MDV3-D	tc5 MDV3-E	tc6 NBS5-A	tc7 NBS5-B	tc8 NBS5-C	tc9 NBS5-D	tc10 NBS5-E	tc11 NBT2-A	tc12 NBT2-B	Min	S/N Min	Max	S/N Max	CycleTime
23-Mar-2026																	
10:31:30	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	tc1	0.00	tc1	00:10:43
10:32:00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	tc1	0.00	tc1	00:11:13
10:32:30	*****Exposure Start*****																
10:32:30	*****Exposure*****																
10:32:30	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	tc1	0.00	tc1	00:00:00
10:33:00	0.61	0.61	0.62	0.61	0.61	0.59	0.60	0.58	0.60	0.59	0.59	0.60	0.58	tc8	0.62	tc3	00:00:30
10:33:30	1.22	1.21	1.24	1.22	1.22	1.20	1.21	1.17	1.20	1.19	1.19	1.21	1.17	tc8	1.24	tc3	00:01:00
10:34:00	1.82	1.82	1.84	1.83	1.82	1.80	1.81	1.75	1.80	1.78	1.80	1.82	1.75	tc8	1.84	tc3	00:01:30
10:34:30	2.41	2.42	2.45	2.43	2.42	2.39	2.40	2.33	2.40	2.37	2.39	2.41	2.33	tc8	2.45	tc3	00:02:00
10:35:00	3.02	3.03	3.07	3.04	3.03	2.98	3.00	2.91	3.00	2.95	2.99	3.01	2.91	tc8	3.07	tc3	00:02:30
10:35:30	3.65	3.66	3.70	3.66	3.66	3.58	3.60	3.51	3.60	3.55	3.58	3.62	3.51	tc8	3.70	tc3	00:03:00
10:36:00	4.28	4.29	4.33	4.28	4.29	4.18	4.21	4.12	4.21	4.16	4.18	4.23	4.12	tc8	4.33	tc3	00:03:30
10:36:30	4.91	4.91	4.96	4.91	4.92	4.79	4.82	4.72	4.81	4.76	4.78	4.84	4.72	tc8	4.96	tc3	00:04:00
10:37:00	5.54	5.54	5.59	5.53	5.55	5.39	5.42	5.32	5.41	5.37	5.39	5.45	5.32	tc8	5.59	tc3	00:04:30
10:37:30	6.17	6.16	6.22	6.16	6.18	5.99	6.03	5.93	6.01	5.97	5.99	6.05	5.93	tc8	6.22	tc3	00:05:00
10:38:00	6.80	6.79	6.85	6.78	6.81	6.59	6.63	6.53	6.61	6.57	6.58	6.66	6.53	tc8	6.85	tc3	00:05:30
10:38:30	7.43	7.41	7.47	7.40	7.44	7.18	7.23	7.13	7.21	7.17	7.18	7.25	7.13	tc8	7.47	tc3	00:06:00
10:39:00	8.04	8.03	8.09	8.02	8.06	7.77	7.82	7.72	7.81	7.76	7.78	7.84	7.72	tc8	8.09	tc3	00:06:30
10:39:30	8.65	8.64	8.70	8.62	8.67	8.36	8.42	8.32	8.41	8.35	8.38	8.44	8.32	tc8	8.70	tc3	00:07:00
10:40:00	9.25	9.25	9.31	9.23	9.27	8.97	9.02	8.93	9.02	8.95	8.99	9.06	8.93	tc8	9.31	tc3	00:07:30
10:40:30	9.88	9.87	9.94	9.85	9.89	9.58	9.64	9.54	9.63	9.57	8.99	9.06	8.99	tc11	9.94	tc3	00:08:00
10:41:00	10.51	10.50	10.57	10.47	10.52	10.19	10.25	10.15	10.24	10.18	8.99	9.06	8.99	tc11	10.57	tc3	00:08:30
10:41:30	11.13	11.12	11.19	11.10	11.15	10.79	10.85	10.75	10.84	10.77	8.99	9.06	8.99	tc11	11.19	tc3	00:09:00
10:42:00	11.75	11.74	11.81	11.71	11.77	11.38	11.45	11.35	11.43	11.37	8.99	9.06	8.99	tc11	11.81	tc3	00:09:30
10:42:30	12.35	12.35	12.42	12.32	12.37	11.98	12.05	11.95	12.04	11.96	8.99	9.06	8.99	tc11	12.42	tc3	00:10:00
10:43:00	12.96	12.96	13.03	12.93	12.98	12.59	12.66	12.56	12.65	12.57	8.99	9.06	8.99	tc11	13.03	tc3	00:10:30

Study Name: 26-Feb-2026 15_47_48

Temp																
Asset ID: 152903-109-0008								SOP / Protocol #: H3-VD-26684								
Comments:								Description:								

Lethality Data																	
Time Stamp	tc1 MDV3-A	tc2 MDV3-B	tc3 MDV3-C	tc4 MDV3-D	tc5 MDV3-E	tc6 NBS5-A	tc7 NBS5-B	tc8 NBS5-C	tc9 NBS5-D	tc10 NBS5-E	tc11 NBT2-A	tc12 NBT2-B	Min	S/N Min	Max	S/N Max	CycleTime
23-Mar-2026																	
10:43:30	13.59	13.59	13.66	13.55	13.60	13.20	13.28	13.18	13.26	13.19	8.99	9.06	8.99	tc11	13.66	tc3	00:11:00
10:44:00	14.22	14.21	14.28	14.18	14.23	13.80	13.89	13.78	13.87	13.79	8.99	9.06	8.99	tc11	14.28	tc3	00:11:30
10:44:30	14.84	14.83	14.90	14.79	14.85	14.39	14.48	14.38	14.46	14.39	8.99	9.06	8.99	tc11	14.90	tc3	00:12:00
10:45:00	15.44	15.44	15.51	15.40	15.46	14.99	15.08	14.98	15.07	14.99	8.99	9.06	8.99	tc11	15.51	tc3	00:12:30
10:45:30	16.06	16.06	16.13	16.02	16.07	15.60	15.70	15.59	15.68	15.60	8.99	9.06	8.99	tc11	16.13	tc3	00:13:00
10:46:00	16.69	16.68	16.76	16.64	16.70	16.22	16.31	16.20	16.29	16.21	8.99	9.06	8.99	tc11	16.76	tc3	00:13:30
10:46:30	17.31	17.30	17.38	17.26	17.33	16.82	16.92	16.81	16.89	16.82	8.99	9.06	8.99	tc11	17.38	tc3	00:14:00
10:47:00	17.90	17.90	17.97	17.86	17.92	17.40	17.50	17.39	17.48	17.40	8.99	9.06	8.99	tc11	17.97	tc3	00:14:30
10:47:30	*****Exposure End*****																
10:47:30	*****3*****																
10:47:30	18.43	18.44	18.51	18.39	18.45	17.93	18.03	17.93	18.02	17.94	8.99	9.06	8.99	tc11	18.51	tc3	00:00:00
10:48:00	18.43	18.44	18.51	18.39	18.45	17.93	18.03	17.93	18.02	17.94	8.99	9.06	8.99	tc11	18.51	tc3	00:00:30
10:48:30	18.43	18.44	18.51	18.39	18.45	17.93	18.03	17.93	18.02	17.94	8.99	9.06	8.99	tc11	18.51	tc3	00:01:00
10:48:30	***** ***** Study End ***** *****																

Study Name: 26-Feb-2026 15_47_48

Comments

System Messages

Time Stamp	Event Description	Event Type	User Name	Console ID
23-Mar-2026 10:20:46	NBT2_Start	Manual		
23-Mar-2026 10:20:47	***Study Start***	Manual	wusitan	6364071B54B2039A7382
23-Mar-2026 10:20:48	NBS5_Start	Manual		
23-Mar-2026 10:20:49	MDV3_Start	Manual		
23-Mar-2026 10:32:30	***Exposure Start***			6364071B54B2039A7382
23-Mar-2026 10:47:30	***Exposure End***			6364071B54B2039A7382
23-Mar-2026 10:48:00	MDV3_End	Manual		
23-Mar-2026 10:48:00	NBS5_End	Manual		
23-Mar-2026 10:48:30	NBT2_End	Manual		
23-Mar-2026 10:48:30	***Study End***	Manual	wusitan	6364071B54B2039A7382

Performed by: _____

Date: _____

Reviewed by: _____

Date: _____