



[_ASSETS](#) · [SAI-LIFE-SCIENCES](#)

Equipment Calibration Certificate

VERSION

0.1 DRAFT

STATUS

DRAFT

OWNER

Hawkeye Platform

RENDERED

2026-06-16

CONTENTS

- 1. Sai Life Sciences Pvt. Ltd. · Equipment Identifier EQ-RX-204
- 2. Calibration results
 - 1. Temperature loop (RTD-Type Pt100, jacket-fluid sensor)
 - 2. Pressure transmitter (jacket pressure, 0-10 bar)
 - 3. Agitation RPM (tachometer verification)
- 3. Conclusion
- 4. References

Sai Life Sciences Pvt. Ltd. · Equipment Identifier EQ-RX-204

FIELD	VALUE
Equipment Description	Glass-lined Reactor 2,000 L · Block 2 · Reactor R-204
Asset Tag	EQ-RX-204
Manufacturer	De Dietrich Process Systems
Serial Number	DD-GLR-2KL-2019-0148
Location	API Manufacturing Block 2 · Hall A · Bay 4
Calibration Type	Routine annual · temperature + pressure + agitation RPM
Performed by	Sai Metrology Cell (in-house) + Bureau Veritas (witness)
Calibration Date	2026-03-18
Next Due	2027-03-18
Certificate No.	SAI-CAL-2026-RX-204-018

Calibration results

Temperature loop (RTD-Type Pt100, jacket-fluid sensor)

TEST POINT (°C)	STANDARD READING	MEASURED	DEVIATION	PASS / FAIL
0.0	0.00	-0.02	-0.02	Pass (within ±0.5 °C)
25.0	25.00	24.97	-0.03	Pass
60.0	60.00	59.94	-0.06	Pass
100.0	100.00	99.91	-0.09	Pass
130.0	130.00	129.88	-0.12	Pass

Pressure transmitter (jacket pressure, 0-10 bar)

TEST POINT (BAR)	STANDARD	MEASURED	DEVIATION	PASS / FAIL
0.0	0.00	0.01	+0.01	Pass (within ±0.05 bar)
2.5	2.50	2.49	-0.01	Pass
5.0	5.00	4.98	-0.02	Pass
7.5	7.50	7.47	-0.03	Pass
10.0	10.00	9.96	-0.04	Pass

Agitation RPM (tachometer verification)

SET POINT (RPM)	MEASURED	DEVIATION	PASS / FAIL
50	50.2	+0.4%	Pass (within ±2%)
100	100.1	+0.1%	Pass
150	149.6	-0.3%	Pass
200	199.4	-0.3%	Pass

Conclusion

Equipment EQ-RX-204 meets calibration acceptance criteria across temperature, pressure, and agitation RPM channels. Release for continued production use approved.

SIGNATORY	ROLE	DATE
K. Iyer	Metrology Engineer	2026-03-18
R. Sharma	QA Verification	2026-03-19
Bureau Veritas Auditor	Witness (third party)	2026-03-18

References

- Validated calibration method: VM-RX-CAL-007 v3
- Reference standards traceable to NABL-accredited PTB / NIST chain
- Equipment qualification: IQ/OQ/PQ documents EQQ-RX-204 v2 (2019); annual re-qualification not yet triggered