

## Table of contents

	0.1	T · · · ·		
		Sample Holding Time:		
	0.3	Materials Required for Titration	2	
	0.4	Glassware Preparation:	3	
	0.5	Personal Protective Equipment / Waste Disposal:	3	
	0.6	Quality Assurance/Quality Control:	3	
	0.7	Waste Disposal:	3	
	0.8	Consumables Ordering:	4	
1	Preparing for Analysis			
2	Ana	lysis of Standard	ndard 6	
Б	. 1	C W:11: M 1 2022		

Revised: Grace Wilkinson, March 2023

#### 0.1 Purpose:

This procedure describes the steps to potentiometrically titrate water samples with standard-ized hydrochloric acid to calculate alkalinity according to Andersen, 2002 and USGS National Field Manual for the Collection of Water-Quality Data. The units of alkalinity for this analysis are microequivalents of carbonate per liter.

### 0.2 Sample Holding Time:

14 days @ 4° C unpreserved

#### 0.3 Materials Required for Titration

(see materials for regents in step 5):

- MilliQ water
- 0.05N Hydrochloric Acid

- 1000 ueg/L Sodium Carbonate standard
- pH meter and probe
- pH buffers
- Analytical balance
- P200 manual, adjustable pipette
- P100 electronic, adjustable pipette
- Pipet tips
- Small, graduated cups for titration
- Micro-stirbar
- MilliQ squirt bottle
- Kimwipes

#### 0.4 Glassware Preparation:

The glass jar to hold waste rinses and the graduated cups used for titration should be rinsed with DI water from the tap at the sink and set upside down to dry.

#### 0.5 Personal Protective Equipment / Waste Disposal:

Nitrile gloves and eye protection should be worn while titrating. Always use chemical resistant gloves (not latex), safety glasses, lab coat, and a fume hood while using concentrated acids to prepare the 0.05N HCl. This is not only for your protection, but also to prevent contamination of samples. Proper personal protective equipment is always required for safety and contamination prevention.

#### 0.6 Quality Assurance/Quality Control:

- Blind samples for analysis (i.e., field duplicates)
- Triplicate analysis of standard solution

#### 0.7 Waste Disposal:

Most of the reagent solutions used in this procedure can go down the drain; however the pH should be near neutral (pH 5-8). Flush during and after disposal by running tap water. Excess dry reagents from preparing the stock can go in the trash.

Consumables Ordering: Item Catalog # Item Catalog # Na2CO3 salt Fisher AA3648522 Buffer 3.557 (Ricca) Fisher 149816 Optima HCl (500 mL) Fisher A466500 Buffer 6.87 (Ricca) Fisher 154016

## 0.8 Consumables Ordering:

Item	Catalog #	Item	Catalog #
$Na_2CO_3$ salt	Fisher $AA3648522$	Buffer 3.557 (Ricca)	Fisher 149816
Optima HCl (500	Fisher A466500	Buffer 6.87 (Ricca)	Fisher 154016
mL)			

# 1 Preparing for Analysis

Remove samples for analysis from the fridge to allow them to warm up to room temperature prior to analysis.

Turn on the pH probe by pressing any key. Make sure it is reading in mV; if not, press 'MODE' until mV is being read. NOTE: put the meter on "Standby" after analysis

Prepare the pH probe for analysis. NOTE: follow these directions in reverse to store the pH probe after analysis

- Remove the storage solution and parafilm from the probe
- Check that there is enough liquid in the probe
- Rinse the probe with MilliQ water from the squirt bottle and dab with a Kimwipe. NOTE: Do not rub the probe with a Kimwipe as this creates static.

Using the pH probe, measure the standard buffers (pH = 3.557 and 6.87) to create a calibration curve.

- Place the pH probe in the buffer solution and allow the reading to stabilize. Record the millivolts (mV) on the sample data sheet.
- Rinse off the pH probe with MilliQ water and dab dry with a Kimwipe
- If the millivolts are not close to the values recorded below, the pH probe may be faulty or need recalibration. Consult with the lab manager before proceeding.

Buffer pH	millivolts
3.557	168
6.87	-19

Make sure that all the sample cups and stir bars are clean and dry.

Turn on the electronic micropipette.

# 2 Analysis of Standard