



Production Test Report

Model: SR50AH

Serial #: 10933

Temperature Rating: -45°C to +50°C

Manufacture date: 12/5/2018

Room Temperature Test	Minimum	Maximum	Standard	Tolerance	Pass	Test Duration
	4989.0 mm	4999.0 mm	1.7 mm	±20 mm	YES	2Hrs
Cold Temperature Test	Minimum	Maximum	Standard	Tolerance	Pass	Test Duration
	860.0 mm	862.0 mm	0.4 mm	±10 mm	YES	4Hrs

Test Details:

- 1) Room temperature test with a target at a fixed distance of 5.00 Metres. 300 measurements are taken over a 2 hour period. The minimum, maximum and standard
- 2) Cold test is conducted at a minimum of 5°C below the cold temperature rating with a target at a distance of ≈1.00 Metre. 600 measurements are taken over a 4 hour

Technician: Emir Muratovic
Edmonton, AB
Canada