

PRODUCT SPECIFICATION

제품 사양서

Product Name	DIGITAL THERMOMETER	Model	MT-D100
Brand	MEDITEMP	Manufacturer	TAIPEI MEDICAL DEVICE CO., LTD
Suggested HS Code	9025.19-1000	Country of Origin	TAIWAN

Technical Specification

Item	Specification
Measurement Method	Digital electronic thermometer, oral/axillary use
Measurement Range	32.0°C ~ 42.9°C
Accuracy	±0.1°C in standard range
Display	LCD display
Power	Button cell battery included
Composition	Electronic circuit, LCD, sensor probe, plastic housing, battery
Packaging	Individual blister pack with instruction leaflet

본 사양서는 수입신고 심사 참고용으로 구성된 샘플입니다.