TREND REPORT

Review recommended

Patient ID: Wheeler **Period:** April 1 to April 30, 2026 **Report Date:** May 1, 2026 **Coverage:** 240/720h (33%)

Elevated Breathing (sustained, 14h, peak 33), during a reduced coverage window. Review recommended.

Last 24 Hours Episodic Events

Heart rate

No measured windows

Spread: 51 to 80 (avg: 64) bpm

Breathing

Two measured windows

Elevated Breathing — sustained, 13h
Apr 30, 3 AM to 4 PM, 26 brpm (10 to 34)


Elevated Breathing — brief, 1h
Apr 30, 6 PM to 7 PM, 24 brpm (10 to 30)

Spread: 8 to 34 (avg: 24) brpm

Last 30 days Episodic events from Apr 01 to Apr 30, 2026

HR Breathing

Episodic events per day · Apr 01 to Apr 30, 2026 (days 1 to 30)



Bars: count of episodic events per day (breathing blue, heart rate red). Daily counts sum to the 26 measured windows for the period.

NOTABLE DAYS

- ***Apr 03 (day 3)*** *— busiest day, 4 windows; Elevated Breathing (sustained, 7h, peak 39)*

Measured windows


Elevated Breathing sustained, 14h	Apr 01, 1 AM to 3 PM · 26 brpm (12 to 33)	*Recurring pattern, 26 windows this period.*
Elevated Breathing sustained, 13h	Apr 30, 3 AM to 4 PM · 26 brpm (10 to 34)	
Elevated Breathing sustained, 10h	Apr 27, 4 AM to 2 PM · 27 brpm (14 to 47)	
Elevated Breathing sustained, 9h	Apr 28, 4 AM to 1 PM · 27 brpm (9 to 43)	

plus 22 more windows


Elevated Breathing Apr 01 to Apr 03 (3d)	Elevated Breathing Apr 15 to Apr 30 (9d)		
54 to 73 *Heart rate P5 to P95 (bpm)*	**18 to 28** *Breathing P5 to P95 (brpm)*	**33%** *Coverage*	**26** *Measured windows*


Measurement data only. Not a diagnosis. Values reflect selected windows where heart rate or breathing was outside the stated ranges, from radar based vital sign measurement. Signal coverage 33%; interpret windows during reduced coverage with caution.

Daily trend




Distribution





Daily monitoring coverage



Measurement thresholds

Low Heart Rate *< 45 bpm*
Low Breathing *< 10 brpm*
High Heart Rate *> 95 bpm*
Elevated Breathing *> 24 brpm*
Very High Heart Rate *> 110 bpm*
High Breathing *> 40 brpm*

Measured window: a value outside the stated range continuously for ≥ 1 hour. Windows separated by ≤ 1 hour of in range values are counted as the same window. Brief crossings that do not sustain are not counted.