TREND REPORT

Review recommended

Patient ID: Thornburg **Period:** April 1 to April 30, 2026 **Report Date:** May 1, 2026 **Coverage:** 293/720h (41%)

Elevated Breathing (sustained, 6h, peak 32). Closer observation suggested.

Last 24 Hours Episodic Events

Heart rate

No measured windows

Spread: 52 to 73 (avg: 61) bpm

Breathing

Two measured windows

Elevated Breathing — brief, 1h
Apr 30, 12 AM to 1 AM, 24 brpm (22 to 32)


Elevated Breathing — brief, 1h
Apr 30, 11 PM to 12 AM, 24 brpm (20 to 27)

Spread: 14 to 32 (avg: 22) brpm

Last 30 days Episodic events from Apr 01 to Apr 30, 2026

HR Breathing

Episodic events per day · Apr 01 to Apr 30, 2026 (days 1 to 30)



NOTABLE DAYS

- ***Apr 25 (day 25)*** *— busiest day, 3 windows; Elevated Breathing (sustained, 6h, peak 32)*

Bars: count of episodic events per day (breathing blue, heart rate red). Daily counts sum to the 10 measured windows for the period.

Measured windows


Elevated Breathing sustained, 6h	Apr 25, 9 AM to 3 PM · 25 brpm (16 to 32)	*Recurring pattern, 10 windows this period.*
Elevated Breathing sustained, 4h	Apr 25, 4 PM to 8 PM · 26 brpm (18 to 32)	
Elevated Breathing sustained, 4h	Apr 28, 6 PM to 10 PM · 26 brpm (15 to 33)	
Elevated Breathing sustained, 4h	Apr 29, 9 PM to 1 AM · 25 brpm (17 to 32)	

plus 6 more windows


Elevated Breathing Apr 25 to Apr 25 (1d)	Elevated Breathing Apr 28 to Apr 30 (3d)		
54 to 70 *Heart rate P5 to P95 (bpm)*	**16 to 25** *Breathing P5 to P95 (brpm)*	**41%** *Coverage*	**10** *Measured windows*


Measurement data only. Not a diagnosis. Values reflect selected windows where heart rate or breathing was outside the stated ranges, from radar based vital sign measurement. Signal coverage 41%; interpret windows during reduced coverage with caution.

Daily trend




Distribution





Daily monitoring coverage



Measurement thresholds

Low Heart Rate *< 45 bpm*
Low Breathing *< 10 brpm*
High Heart Rate *> 95 bpm*
Elevated Breathing *> 24 brpm*
Very High Heart Rate *> 110 bpm*
High Breathing *> 40 brpm*

Measured window: a value outside the stated range continuously for ≥ 1 hour. Windows separated by ≤ 1 hour of in range values are counted as the same window. Brief crossings that do not sustain are not counted.