JD-Photo Data Photo Data House, Knowl Piece, Hitchin, Herts. SG4 0TY

COMPANY. **TTK University of Applied Sciences**

ADDRESS. **Pärnu rd 62**
Tallinn, 10135
Estonia

CERT No: **008/194821-001** **DATE.** **22/05/2026**

CERTIFICATE OF CONFORMITY.

ADVICE No:	**194821**	**CUSTOMERS REF:**	**SL12 Dot Grid Distortion**
ORDER No:	**n/a**	**PHOTO DATA REF:**	**JD194821a.gbx**

DESCRIPTION: **Required measurements on a LowReflective Chrome Mask**
Measured at a temperature of 21 Deg C and 50% humidity

Imaged at Class A
Size 125mm x 125mm
Thickness 1.5mm
SL12-4

We hereby certify that the information detailed above has been produced, inspected and unless otherwise stated conforms in all respects to the specifications, drawings and orders relative there to.

Equipment used for calibration have been calibrated traceable to the National Institute of Standards and Technology (N.I.S.T.), in accordance with procedures set forth in: ISO/IEC 17025-2005 and ANSI/NCSL Z540-1-1994

Please see successive pages for the measurement results

Signed

Fil Router
Quality Assurance.

CERT No: 008/194821-001

	Nominal mm	Measured mm	Tol +/- um	Error um	Polarity of Feature
A1	98.0000	97.9994	3.00	-0.60	Chrome
A2	98.0000	97.9994	3.00	-0.60	Chrome
A3	98.0000	97.9992	3.00	-0.80	Chrome
A4	98.0000	97.9993	3.00	-0.70	Chrome
A5	138.5930	138.5915	3.00	-1.50	Chrome
CD1	1.0000	1.0002	0.50	0.20	Chrome

