weighing instrument services (PTY) LTD.

Reg. No. 1982/002054/07

63 EARP STREET, OPHIRTON
P.O. BOX 38122, BOOYSENS 2016
TEL: +27 - 11 - 493-6075
FAX: +27 - 11 - 493-2706
E-Mail: wis@wissa.co.za

sanas Calibration Laboratory
No. 1407

WOW Scales
C/O Vlei & John Mitten Street Douglas Valley Small Holdings Dealsgift
Bloemfontein

CALIBRATION CERTIFICATE No. 317/25

Calibration of:	A set of weights
Calibrated At:	**Weighing Instrument Services**
Environmental Conditions:	**Suitable for calibration**
Calibration procedure:	The instrument has been calibrated with standard weights of known value, traceable to international standards, in accordance with our procedure No. **WIS (1)**
Traceability of standards:	**Certificate No: NMISA-MAS-2024-6787** **Dated: May 2024**

Results:

Identification Number	Nominal Value (g)	Actual Value (g)	Uncertainty of Calibration (g) + or -
SET WOW S1	2 000	2 000,01	0,05
	2 000	2 000,00	0,05
	1 000	1 000,00	0,03
	500	500,002	0,010
	200	200,001	0,005
	200	200,001	0,005
	100	100,000	0,003
	50	50,000	0,002
	20	20,000	0,001
	20	20,000	0,001
	10	10,000 1	0,000 5
	5	5,000 1	0,000 5
	2	2,000 0	0,000 5
	2	2,000 1	0,000 5
	1	1,000 0	0,000 5

The results only relate to the calibrated item (s) specified above.

Uncertainty: The uncertainties of calibration were calculated and expressed in accordance with the BIPM, IEC, ISO, IUPAP, OIML document entitled "Guide to the Expression of Uncertainties in Measurement" (International Organisation for Standardisation, Geneva, Switzerland, 1993) and are based on a standard uncertainty multiplied by a coverage factor of k=2, which provides a level of confidence of approximately 95%

Validity: The values given in this certificates are correct at the time of calibration. Subsequently the accuracy will depend on such factors as the care exercised in handling and use of the instruments as well as the frequency of use. Recalibration should be performed after such a period which is chosen to ensure that the accuracy remains within the desired limits.

Date Calibrated: **31 March 2025**

Calibrated By: **Chezrae Taylor**

Graeme Brolly
TECHNICAL SIGNATORY

Date Issued: **01 April 2025**

weighing instrument services (PTY) LTD.

Reg. No. 1982/002054/07

63 EARP STREET, OPHIRTON
P.O. BOX 38122, BOOYSENS 2016
TEL: +27 - 11 - 493-6075
FAX: +27 - 11 - 493-2706
E-Mail: wis@wissa.co.za

sanas Calibration Laboratory
No. 1407

WOW Scales
C/O Vlei & John Mitten Street Douglas Valley Small Holdings Dealsgift
Bloemfontein

CALIBRATION CERTIFICATE No. 318/25

Calibration of:	A set of weights
Calibrated At:	Weighing Instrument Services
Environmental Conditions:	Suitable for calibration
Calibration procedure:	The instrument has been calibrated with standard weights of known value, traceable to international standards, in accordance with our procedure No. WIS (1)
Traceability of standards:	Certificate No: NMISA-MAS-2024-6787 Dated: May 2024

Results:

Identification Number	Nominal Value (g)	Actual Value (g)	Uncertainty of Calibration (g) + or -
SET WOW S2	2 000	2 000,00	0,05
	2 000	2 000,00	0,05
	1 000	1 000,00	0,03
	500	500,001	0,010
	200	200,001	0,005
	200	200,001	0,005
	100	100,000	0,003
	50	50,000	0,002
	20	20,000	0,001
	20	20,000	0,001
	10	10,000 0	0,000 5
	5	5,000 1	0,000 5
	2	2,000 0	0,000 5
	2	2,000 0	0,000 5
	1	1,000 1	0,000 5

The results only relate to the calibrated item (s) specified above.

Uncertainty: The uncertainties of calibration were calculated and expressed in accordance with the BIPM, IEC, ISO, IUPAP, OIML document entitled "Guide to the Expression of Uncertainties in Measurement" (International Organisation for Standardisation, Geneva, Switzerland, 1993) and are based on a standard uncertainty multiplied by a coverage factor of k=2, which provides a level of confidence of approximately 95%

Validity: The values given in this certificates are correct at the time of calibration. Subsequently the accuracy will depend on such factors as the care exercised in handling and use of the instruments as well as the frequency of use. Recalibration should be performed after such a period which is chosen to ensure that the accuracy remains within the desired limits.

Date Calibrated: 1 April 2025

Calibrated By: Chezrae Taylor

Graeme Brolly
TECHNICAL SIGNATORY

Date Issued: 01 April 2025