



WOW Calibration Pty Ltd  
Reg no. 2021/697529/07  
Van Dyk Business park Unit 71 c/o Van Dyk rd and  
Brakpan rd, Boksburg  
TEL: 011 749 3102



## Mass Piece calibration

### Calibration certificate number: **W25/009**

This certificate is a correct record of the measurements made by standards traceable to National Measurement Institute, South Africa (NMISA) and International Standards. Copyright of this certificate is owned by the issuing laboratory. This certificate may not be reproduced except in full, without written approval of WOW Calibration (Pty) Ltd.

Company: WOW Scales	Procedure followed: WCP001
Contact Person: Mr Rhyno Page	Standard used: 20kg;
Physical Address: John Mitten street, Vlei Ave, Dealesgift, Bloemfontein	Standard Traceability: 254/25
Location of calibration: wow laboratory	Date of calibration: 03/04/2025-04/04/2025
Calibration of: weights	Calibrated by: S.NKOSI

The accuracies of the standards used are, by way of regular inter-comparison, traceable to National Standards, with the limitations of the services available. The measurement results in the certificate are correct at the time of calibration. Subsequently the accuracy will depend on such factors as the care exercised in handling as well as use of instruments, equipment and the frequency of use. No long-term stability checks were carried out. This certificate automatically expires if the instrument is repaired or if the seal is broken. It is recommended that re-calibration be undertaken at an interval that will ensure that the instrument remains within the desired limits. The measurement results relate only to this item calibrated.

unit: g

No.	Serial Number	Nominal value	Actual value	Uncertainty of calibration $\pm$
1.	WOW1	20 000	20 000.6	1.0
2.	WOW2	20 000	20 000.9	1.0
3.	WOW3	20 000	20 000.1	1.0
4.	WOW4	20 000	20 000.3	1.0
5.	WOW5	20 000	20 000.7	1.0
6.	WOW6	20 000	20 000.0	1.0
7.	WOW7	20 000	19 999.7	1.0
8.	WOW8	20 000	19 999.9	1.0
9.	WOW9	20 000	20 000.0	1.0
10.	WOW10	20 000	19 999.9	1.0
11.	WOW11	20 000	20 000.0	1.0
12.	WOW12	20 000	20 000.5	1.0
13.	WOW13	20 000	20 000.2	1.0
14.	WOW14	20 000	19 999.9	1.0
15.	WOW15	20 000	19 999.8	1.0
16.	WOW16	20 000	19 999.7	1.0
17.	WOW17	20 000	19 999.8	1.0
18.	WOW18	20 000	19 999.7	1.0
19.	WOW19	20 000	19 999.6	1.0
20.	WOW20	20 000	19 999.5	1.0
21.				
22.				
23.				
24.				
25.				

Checked by:

H. McGee

Date of issue:

15/04/2025

Technical Signatory

S.Nkosi

*[Signature]*

Certificate end



WOW Calibration Pty Ltd  
Reg no. 2021/697529/07  
Van Dyk Business park Unit 71 c/o Van Dyk rd and  
Brakpan rd, Boksburg  
TEL: 011 749 3102



## Mass Piece calibration

Calibration certificate number: **W25/009B**

This certificate is a correct record of the measurements made by standards traceable to National Measurement Institute, South Africa (NMISA) and International Standards. Copyright of this certificate is owned by the issuing laboratory. This certificate may not be reproduced except in full, without written approval of WOW Calibration (Pty) Ltd.

Company: WOW Scales	Procedure followed: WCP001
Contact Person: Mr Rhyno Page	Standard used: 20kg;
Physical Address: John Mitten street, Vlei Ave, Dealesgift, Bloemfontein	Standard Traceability: 254/25
Location of calibration: wow laboratory	Date of calibration: 03/04/2025-04/04/2025
Calibration of: weights	Calibrated by: S.NKOSI

The accuracies of the standards used are, by way of regular inter-comparison, traceable to National Standards, with the limitations of the services available. The measurement results in the certificate are correct at the time of calibration. Subsequently the accuracy will depend on such factors as the care exercised in handling as well as use of instruments, equipment and the frequency of use. No long-term stability checks were carried out. This certificate automatically expires if the instrument is repaired or if the seal is broken. It is recommended that re-calibration be undertaken at an interval that will ensure that the instrument remains within the desired limits. The measurement results relate only to this item calibrated.

unit: g

No.	Serial Number	Nominal value	Actual value	Uncertainty of calibration $\pm$
1.	WOW21	20 000	20 000.0	1.0
2.	WOW22	20 000	20 000.5	1.0
3.	WOW23	20 000	20 000.8	1.0
4.	WOW24	20 000	20 000.6	1.0
5.	WOW25	20 000	20 000.2	1.0
6.	WOW26	20 000	20 000.0	1.0
7.	WOW27	20 000	20 000.3	1.0
8.	WOW28	20 000	20 000.7	1.0
9.	WOW29	20 000	20 000.5	1.0
10.	WOW30	20 000	20 000.7	1.0
11.	WOW31	20 000	20 000.1	1.0
12.	WOW32	20 000	20 000.0	1.0
13.	WOW33	20 000	20 000.3	1.0
14.	WOW34	20 000	20 000.5	1.0
15.	WOW35	20 000	20 000.2	1.0
16.	WOW66	20 000	20 000.7	1.0
17.	WOW37	20 000	20 000.5	1.0
18.	WOW38	20 000	20 000.7	1.0
19.	WOW39	20 000	20 000.5	1.0
20.	WOW40	20 000	20 000.1	1.0
21.				
22.				
23.				
24.				
25.				

Checked by:

H. McGeev

Date of issue:

15/04/2025

Technical Signatory

S.Nkosi

Certificate end