


REVISION			
REV.	CHANGE ORDER	APPROVED BY	DATE
A	ECO-R165612	Refer to Agile	Refer to Agile





ITEM	SUPPLIER	PART NO.	QTY.	DESCRIPTION
1	RS	488-438	1	A/C MAINS PLUG - DURAPLUG BS1363A AMP
2	FELLER	110-1120	1	TWO CORE MAINS LEAD - AIROX DEL PWR CORD AIROX P/N: 2966200
3	FARNELL	119-0173	1	MOULDED USB CABLE - USB TYPE A FREE END B CABLE ASSEMBLY

ASSEMBLY INSTRUCTIONS:

- Trim back and isolate the power and data lines of the USB cable.
- Connect the USB cable shielding to the Earth Pin to AC Plug.
- Overall cable Length 0.5 to 0.75 meters.
- Equivalent parts may be used with test Engineering approval.

DBASE FILENAME: 10021744 Approved Revision A.DWG

MODEL FILENAME:

UNLESS OTHERWISE SPECIFIED		UNLESS OTHERWISE STATED, THIS PRINT AND SUBJECT MATTER SHOWN HEREIN ARE PROPRIETARY ITEMS TO WHICH NELLCOR PURITAN BENNETT RETAINS THE EXCLUSIVE RIGHT OF REPRODUCTION MANUFACTURE AND SALE. THIS PRINT IS SUBMITTED IN CONFIDENCE FOR THE USE OF THE RECIPIENT ALONE UNLESS PERMISSION FOR FURTHER DISCLOSURE IS EXPRESSLY GRANTED IN WRITING.		Puritan-Bennett Ireland Ltd. Mervue, Galway, Ireland	
XX → ±0.5 XXX → ±0.25 ANGLES → ±1/2° XXX → ±0.125 DRAFT → ±1/2"		MATERIAL:		TITLE	
<div><div>NO</div><div></div></div> SURFACE ROUGHNESS → BREAK ALL SHARP EDGES → FILLET AND CORNER RADIUS → DIMENSIONS IN MILLIMETERS DO NOT SCALE DRAWING DIMENSIONING AND TOLERANCING PER BS 308 OR ANSI Y14.5 ALL DIMENSIONS AFTER FINISH		FINISH:		PB540 SYSTEM HYPOT TEST POWER CABLE	
DRAWN Mark Sweetman		DATE 28 December 07		SIZE A4	
ENGINEER Refer to Agile		DATE Refer to Agile		DWG. NO. 10021744	
CHECKED Refer to Agile		DATE Refer to Agile		REV. A	
		SCALE: NTS		SHEET 1 OF 1	