

NATIONAL LEGAL METROLOGY LABORATORY

OFFICIAL VERIFICATION CERTIFICATE & TEST REPORT • OIML R 76-1:2006

Test Report ID: TEST-2026-9001	Date of Test: 2026-09-29	Overall Verdict: PASS
Laboratory: National Standards Metrology Lab	Tester: Lab Tester One	Reviewer: Pending Review

1. INSTRUMENT UNDER TEST (EUT)

Instrument ID: INST-P45-001	Manufacturer: Mettler Toledo	Model: Precision Pro 15k	Serial No.: SN-P45-9988
Accuracy Class: Class III	Max Capacity: 15000.0 kg	Min Capacity: 100.0 kg	Interval (e): 5.0 kg
Interval (d): 5.0 kg	Tare Capacity: 0.0 kg	Verification Intervals (n): 3000	Classification: OIML R 76 NAWI

2. ENVIRONMENTAL & LABORATORY CONDITIONS

Temperature: 21.5 °C	Relative Humidity: 48.0 %	Pressure: 101.3 kPa
Test Location: Lab Room 1	Reference Standards: OIML Class F1 Weights Set #900	Standard Range: 10°C – 30°C (Class III/IIII)

3. OIML R-76 COMPLIANCE EVALUATION SUMMARY

Test Procedure	OIML Clause	Points	Max Deviation	Permissible Limit	Verdict
	Clause A.4.7	0	N/A	Formula	PASS

4. OBSERVATIONS & ERROR CALCULATIONS BREAKDOWN

4.1 Eccentricity Test (Clause A.4.7) — PASS

Pos	Load (kg)	Reading I (kg)	ΔL (kg)	Calc P (kg)	Error E (kg)	Ec (kg)	MPE	Status
Center	5000.000	5000.000	2.5000	5000.0000	+0.0000	+0.0000	±5.000	PASS
Front-Left	5000.000	5001.000	2.5000	5001.0000	+1.0000	+1.0000	±5.000	PASS
Back-Right	5000.000	5000.000	2.5000	5000.0000	+0.0000	+0.0000	±5.000	PASS

5. SUMMARY REMARKS & METROLOGICAL STATEMENT

Annual legal metrology recalibration verification. Met all criteria.

Tamper-Evident SHA-256 Checksum: f659aa4c1dc65b7d939ab02952f66a204a99e8d93355c69dcb1b41abb7de513c

TEST CONDUCTED BY	VERIFIED & APPROVED BY	OFFICIAL LABORATORY STAMP
Lab Tester One Authorized Metrology Tester	Pending Review Lead Metrologist / Reviewer	[ACCREDITED LAB SEAL] ISO/IEC 17025 Compliant