

NATIONAL LEGAL METROLOGY LABORATORY

OFFICIAL VERIFICATION CERTIFICATE & TEST REPORT • OIML R 76-1:2006

Test Report ID: TEST-2026-0016	Date of Test: 2026-09-29	Overall Verdict: PASS
Laboratory: National Metrology Institute (NMI) Calibration Lab	Tester: Marcus Vance	Reviewer: Elena Rostova

1. INSTRUMENT UNDER TEST (EUT)

Instrument ID: E2E-9A6366	Manufacturer: Kern & Sohn GmbH	Model: EG 2200-2NM	Serial No.: SN-E2E-9A6366
Accuracy Class: Class II	Max Capacity: 2200.0 g	Min Capacity: 0.5 g	Interval (e): 0.01 g
Interval (d): 0.01 g	Tare Capacity: 0.0 g	Verification Intervals (n): 220000	Classification: OIML R 76 NAWI

2. ENVIRONMENTAL & LABORATORY CONDITIONS

Temperature: 20.4 °C	Relative Humidity: 48.5 %	Pressure: 101.32 kPa
Test Location: Lab Room 1	Reference Standards: OIML Class E2 Stainless Steel Mass Standards (Kit #E2-992)	Standard Range: 10°C – 30°C (Class III/IIII)

3. OIML R-76 COMPLIANCE EVALUATION SUMMARY

Test Procedure	OIML Clause	Points	Max Deviation	Permissible Limit	Verdict
	OIML R 76-1 Clause A.4.4	0	N/A	Formula	PASS
	OIML R 76-1 Clause A.4.4.1	0	N/A	Formula	PASS

4. OBSERVATIONS & ERROR CALCULATIONS BREAKDOWN

4.1 Repeatability Test (OIML R 76-1 Clause A.4.4) — PASS

Run #	Load (g)	Zero I ₀	Indication I (g)	ΔL (g)	Calc P (g)	Error E (g)
1	1000.000	0.000	1000.000	0.0050	1000.0000	+0.0000
2	1000.000	0.000	1000.000	0.0050	1000.0000	+0.0000
3	1000.000	0.000	1000.000	0.0050	1000.0000	+0.0000

4.2 Weighing Performance Test (OIML R 76-1 Clause A.4.4.1) — PASS

#	Load (g)	Reading I (g)	ΔL (g)	Calc P (g)	Error E (g)	Ec (g)	MPE
1	0.000	0.000	0.0050	0.0000	+0.0000	+0.0000	±0.010
2	100.000	100.000	0.0050	100.0000	+0.0000	+0.0000	±0.010
3	500.000	500.000	0.0050	500.0000	+0.0000	+0.0000	±0.010
4	1000.000	1000.000	0.0050	1000.0000	+0.0000	+0.0000	±0.010
5	2000.000	2000.000	0.0050	2000.0000	+0.0000	+0.0000	±0.010

5. SUMMARY REMARKS & METROLOGICAL STATEMENT

TEST CONDUCTED BY	VERIFIED & APPROVED BY	OFFICIAL LABORATORY STAMP
Marcus Vance Authorized Metrology Tester	Elena Rostova Lead Metrologist / Reviewer	[ACCREDITED LAB SEAL] ISO/IEC 17025 Compliant