

NATIONAL LEGAL METROLOGY LABORATORY

OFFICIAL VERIFICATION CERTIFICATE & TEST REPORT • OIML R 76-1:2006

Test Report ID: TEST-2026-0007	Date of Test: 2026-09-29	Overall Verdict: PASS
Laboratory: Federal Metrology Testing Center	Tester: Marcus Vance	Reviewer: Elena Rostova

1. INSTRUMENT UNDER TEST (EUT)

Instrument ID: NAWI-2026-001	Manufacturer: Mettler Toledo	Model: XPR205	Serial No.: MT-8849201
Accuracy Class: Class I	Max Capacity: 220.0 g	Min Capacity: 0.01 g	Interval (e): 0.001 g
Interval (d): 0.001 g	Tare Capacity: 0.0 g	Verification Intervals (n): 220000	Classification: OIML R 76 NAWI

2. ENVIRONMENTAL & LABORATORY CONDITIONS

Temperature: 20.4 °C	Relative Humidity: 48.5 %	Pressure: 101.32 kPa
Test Location: Lab Room 1	Reference Standards: OIML Class E2 Standard Weights Set 01	Standard Range: 10°C – 30°C (Class III/IIII)

3. OIML R-76 COMPLIANCE EVALUATION SUMMARY

Test Procedure	OIML Clause	Points	Max Deviation	Permissible Limit	Verdict
	OIML R 76-1 Clause A.4.4	0	N/A	Formula	PASS

4. OBSERVATIONS & ERROR CALCULATIONS BREAKDOWN

4.1 Repeatability Test (OIML R 76-1 Clause A.4.4) — PASS

Run #	Load (g)	Zero I_0	Indication I (g)	ΔL (g)	Calc P (g)	Error E (g)
1	200.000	0.000	200.000	0.0004	200.0006	+0.0006
2	200.000	0.000	200.001	0.0005	200.0010	+0.0010
3	200.000	0.000	200.000	0.0003	200.0002	+0.0002

5. SUMMARY REMARKS & METROLOGICAL STATEMENT

Metrological verification test conducted in accordance with OIML R 76-1:2006. Overall test outcome: PASS.

Tamper-Evident SHA-256 Checksum: 0713d55f8b0a49808d52306af614db98ce5ad0486665df9b8d54350084d28db3

TEST CONDUCTED BY	VERIFIED & APPROVED BY	OFFICIAL LABORATORY STAMP
Marcus Vance Authorized Metrology Tester	Elena Rostova Lead Metrologist / Reviewer	[ACCREDITED LAB SEAL] ISO/IEC 17025 Compliant