

FDA Clearances

Date	Product	Clearance	Details
December 28, 2018	EyeBOX®	De Novo DEN170091	First FDA authorization for an objective, eye-tracking aid in diagnosing concussion that does not require a pre-injury baseline. Cleared with the broadest age range for use from 5-67 years for concussion from any cause of injury.
July 31, 2019	EyeBOX®	510(k) K191183	Updated indications to clarify that concussion is also known as mild traumatic brain injury (mTBI); moved EyeBOX from a wheeled chassis to a tabletop, WiFi-enabled design.
September 6, 2020	EyeBOX®	510(k) K201841	No change to indications. Upgraded the eye-tracking camera and supporting hardware to equivalent, validated components.
December 22, 2021	EyeBOX®	510(k) K212310	Next-generation, portable EyeBOX.
April 4, 2025	EyeBOX®	510(k) K242116	Label expansion enabling comparison of a patient's results to a normative population. Supports recovery monitoring and return-to-activity decisions over time.
June 5, 2026	SNAP™	510(k) K254086	90-second, multi-modal concussion test combining eye-tracking with symptom severity 90% sensitivity/94% specificity. SNAP also does not require a pre-injury baseline.