

Date Range:  
 15-Aug-2012 01:50:05  
 15-Aug-2012 03:58:04

Number of Traces:  
 17

Length of Fiber:  
 1524.48

#### Calibration Reference Location and Temperature:

|        | Index          |                | Distance (m)   |                | Tref   |
|--------|----------------|----------------|----------------|----------------|--------|
|        | Z <sub>1</sub> | Z <sub>2</sub> | Z <sub>1</sub> | Z <sub>2</sub> |        |
| Ref. 1 | 66             | 219            | 116.2          | 271.5          | tref 1 |
| Ref. 2 | 1351           | 1375           | 1420           | 1444.3         | tref 2 |
| Ref. 3 | 1380           | 1397           | 1449.4         | 1466.6         | tref 3 |
| Val.   | 220            | 230            | 272.5          | 282.6          | tref 1 |

#### Calibration Step Loss Data:

Step Loss:

Index Location: 728

Fiber Distance: 788.897

Equal Distance Temperature:

Index Location: 5

Fiber Distance: 55.344

#### Calibration Results Summary:

Calibration RMSE: 0.396

Calibration Bias: 0.001

Validation RMSE: 0.829

Validation Bias: -0.824

|           | Mean  | Min   | Max   |
|-----------|-------|-------|-------|
| Ref. 1    | 0.445 | 0.347 | 0.484 |
| Ref. 2    | 0.088 | 0.075 | 0.103 |
| Ref. 3    | 0.098 | 0.077 | 0.121 |
| Ref. Val. | 0.829 | 0.674 | 0.894 |

