



Certificate No. 34141/DBBPAM
Date: September 30, 2019



Komite Akreditasi Nasional
Laboratorium Kalibrasi
LK - 023 - IDN



Issuing Office:
Jl. Arteri Tol Cibitung No. 1, Cibitung Bekasi 17520, Indonesia
Phone/Fax: +62 21 88321176/88321166
Email: jum.cbt@sucofindo.co.id

CERTIFICATE OF CALIBRATION

CLIENT	:	SENTRAL TEHNOLOGY MANAGEMEN, PT
ADDRESS	:	CIKARANG SQUARE BLOK B NO. 11 JL. RAYA CIBARUSAH CIKARANG SELATAN CIKARANG - JAWA BARAT
NAME OF EQUIPMENT	:	ANAK TIMBANGAN
SERIAL NO./CODE	:	- / CSM 019
MATERIALS	:	STAINLESS STEEL
WEIGHT/CLASS	:	5000 g / F1
MANUFACTURER	:	EXCELLENT

The Attachment available is an integral part of this certificate.

This calibration result (s) related to the sample (s) submitted only and the certificate cannot be reproduced in any way, except in full context and with the prior approval in writing from Sucofindo Laboratory.

This Certificate/report is issued under our General Terms and Conditions, copy of which is available upon request or may be accessed at www.sucofindo.co.id

SBU Laboratorium

Husen Muhajir

CBT20295119-006



3328103

SCI-2007A



Amendment No. AM 139/DBBPAM
Date: October 14, 2019

Issuing Office:
Jl. Arteri Tol Cibitung No. 1, Cibitung Bekasi 17520, Indonesia
Phone/Fax: +62 21 88321176/88321166
Email: jum.cbt@sucofindo.co.id

**AMENDMENT
TO CERTIFICATE NO. 34141/DBBPAM DATE: SEPTEMBER 30, 2019**

WRITTEN	SHOULD READ
CLIENT : SENTRAL TEHNOLOGY MANAGEMEN, PT	CLIENT : SENTRAL TEHNOLOGI MANAGEMEN, PT

All other matters in the certificate remain unchanged.

This calibration result (s) related to the sample (s) submitted only and the certificate cannot be reproduced in any way, except in full context and with the prior approval in writing from Sucofindo Laboratory.

This Certificate/report is issued under our General Terms and Conditions, copy of which is available upon request or may be accessed at www.sucofindo.co.id

SBU Laboratorium



Husen Muhajir

CBT20295119-006



3346110

SCI - 2007 A

Attachment
To Certificate No. 34141/DBBPAM
Date: September 30, 2019

Page 2 of 2



Phone/Fax: +62 21 88321176/88321166
Jl. Arteri Tol Cibitung No. 1, Cibitung Bekasi 17520, Indonesia
Email: jum.cbt@sucofindo.co.id

CERTIFICATE OF CALIBRATION

TESTING CONDITION,

- TEMPERATURE
- HUMIDITY

STANDARD & EQUIPMENT

: 21.7 ± 0.3 °C
: 56 ± 1 %
: CALIBRATED MASSES CLASS E2
Serial No. 15850-2
Certificate of Calibration S. 052098
: THERMOHYGROMETER
Serial No. MS-2
Certificate of Calibration Cal. 3292019

*The measurement is traceable to International System of Units (SI)
through P2M-LIPI*

CALIBRATION METHOD

RECEIVED ON

DATE OF CALIBRATION

: PO / MS / 05
: September 19, 2019
: September 20, 2019

RESULTS :

ANAK TIMBANGAN Serial No / Code : -
WEIGHT / CLASS : 5000 g / F1

NOMINAL	CONVENTIONAL MASS	UNCERTAINTY OF MEASUREMENT at 95% Confidence Level	k =
g	g	mg	
5000	5000.007	8.3	1.97



1537711

G.C.L. 0.0007 D