

Certificate No. 34134/DBBPAM  
Date: September 30, 2019



Komite Akreditasi Nasional

Laboratorium Kalibrasi

LK - 023 - IDN

Issuing Office:

Jl. Arteri Tol Cibitung No. 1, Cibitung Bekasi 17520, Indonesia

Phone/Fax: +62 21 88321176/88321166

Email: jum.cbt@sucofindo.co.id

## CERTIFICATE OF CALIBRATION

CLIENT	:	SENTRAL TEHNOLOGY MANAGEMEN, PT
ADDRESS	:	CIKARANG SQUARE BLOK B NO. 11 JL. RAYA CIBARUSAH CIKARANG SELATAN CIKARANG - JAWA BARAT
NAME OF EQUIPMENT	:	ANAK TIMBANGAN
SERIAL NO./CODE	:	- / CSM-017
MATERIALS	:	STAINLESS STEEL
WEIGHT/CLASS	:	1000 g / F1
MANUFACTURER	:	EXCELLENT

The Attachment available is an integral part of this certificate.

This calibration result (s) related to the sample (s) submitted only and the certificate cannot be reproduced in any way, except in full context and with the prior approval in writing from Sucofindo Laboratory.

This Certificate/report is issued under our General Terms and Conditions, copy of which is available upon request or may be accessed at [www.sucofindo.co.id](http://www.sucofindo.co.id)

**SBU Laboratorium**



Husen Muhamir

CBT20295119-004



3328101

SCI - 2007 A



Amendment No. AM 137/DBBPAM  
Date: October 14, 2019

Issuing Office:  
Jl. Arteri Tol Cibitung No. 1, Cibitung Bekasi 17520, Indonesia  
Phone/Fax: +62 21 88321176/88321166  
Email: jum.cbt@sucofindo.co.id


**AMENDMENT  
TO CERTIFICATE NO. 34134/DBBPAM DATE: SEPTEMBER 30, 2019**

WRITTEN	SHOULD READ
CLIENT : SENTRAL TEHNOLOGY MANAGEMEN, PT	CLIENT : SENTRAL TEHNOLOGI MANAGEMEN, PT

All other matters in the certificate remain unchanged.

This calibration result (s) related to the sample (s) submitted only and the certificate cannot be reproduced in any way, except in full context and with the prior approval in writing from Sucofindo Laboratory.

This Certificate/report is issued under our General Terms and Conditions, copy of which is available upon request or may be accessed at [www.sucofindo.co.id](http://www.sucofindo.co.id)



CBT20295119-004



3346108

SCI - 2007A

Attachment

To Certificate No. 34134/DBBPAM

Date: September 30, 2019

Page 2 of 2



SUCOFINDO  
Issuing Office:

Phone/Fax: +62 21 88321176/88321166

Jl. Arteri Tol Cibitung No. 1, Cibitung Bekasi 17520, Indonesia

Email: jum.cbt@sucofindo.co.id

## CERTIFICATE OF CALIBRATION

### TESTING CONDITION,

- TEMPERATURE
- HUMIDITY

### STANDARD & EQUIPMENT

: 21.4 ± 0.3 °C  
: 51 ± 1 %  
: CALIBRATED MASSES CLASS E2  
Serial No. 35329151  
Certificate of Calibration No. 108353

: THERMOHYGROMETER  
Serial No. MS-2

Certificate of Calibration Cal. 3292019

*The measurement is traceable to International System of Units (SI) through P2M-LIPI & Sartorius*

### CALIBRATION METHOD

: PO / MS / 05

### RECEIVED ON

: September 19, 2019

### DATE OF CALIBRATION

: September 23, 2019

### RESULTS :

ANAK TIMBANGAN Serial No / Code : -

WEIGHT / CLASS : 1000 g / F1

NOMINAL	CONVENTIONAL MASS	UNCERTAINTY OF MEASUREMENT at 95% Confidence Level	k =
g	g	mg	
1000	1000.0020	1.7	2.00



1537709

S.C.I.-2007 P