Cambridge Instruments
& Engg. Co.

EST. 1963 · HOWRAH, INDIA

INSULATION TESTER · HAND-DRIVEN, BAKELITE BODY

CIE/555

250 V / 0–20 MΩ

Hand-driven insulation tester · clutch-type generator · bakelite body

01 KEY FEATURES

- Silent clutch-type mechanism
- High-sensitivity cross-coil indicator
- Moulded bakelite body, rubber shoes
- Collapsible driving handle

Supplied as base instrument only - carrying case and test leads available at extra cost.



CIE/555 · INSULATION TESTER · ILLUSTRATIVE IMAGE

02 OPTIONAL ACCESSORIES (extra cost)

+ Carrying case (bakelite body)
+ Testing leads, 1m

03 TECHNICAL SPECIFICATIONS

PARAMETER	SPECIFICATION
CONFIGURATION	
Rated Test Voltage	250 V D.C.
Insulation Resistance Range	0–20 MΩ
MEASUREMENT	
Indicator	High-sensitivity cross-coil
Scale Length	75 mm (approx.)
Accuracy	Conforms to IS: 2992-1987
Response Time	3 seconds
GENERATOR	
Generator Type	Hand-driven, clutch-type generator
Rotation Speed	160 r.p.m.
GENERAL	
Body	Moulded high-grade bakelite
Dimensions (L×W×H)	155 × 104 × 70 mm (approx.)
Weight	1.4 kg (approx.)
Warranty	12 months against manufacturing defects only

Dual-voltage, dual-range options and custom ranges are available against specific requirements - enquire.

All specifications are subject to change without prior notice. © Cambridge Instruments & Engg. Co. - CIE/555 250 V, 0-20 MΩ Product Catalogue.